

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2177

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Datafile : VN086838.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0603WBSD01	Vinyl chloride	20	16.1	ug/L	81	4		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	17.9	ug/L	90	3		70 (74)	130 (110)	20 (20)
	2-Butanone	100	94.6	ug/L	95	1		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	18.5	ug/L	93	1		70 (77)	130 (113)	20 (20)
	Chloroform	20	18.2	ug/L	91	1		70 (79)	130 (113)	20 (20)
	Benzene	20	18.8	ug/L	94	2		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	18.3	ug/L	92	0		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	17.9	ug/L	90	3		70 (77)	130 (113)	20 (20)
	Tetrachloroethene	20	15.4	ug/L	77	3		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.4	ug/L	97	1		70 (82)	130 (109)	20 (20)