

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177  
 Lab File ID: VN086835.D Date Analyzed: 06/03/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 11:30  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 307948        | 8.22  | 547716        | 9.09  | 465701        | 11.87  |
| UPPER LIMIT    | 615896        | 8.718 | 1095430       | 9.594 | 931402        | 12.365 |
| LOWER LIMIT    | 153974        | 7.718 | 273858        | 8.594 | 232851        | 11.365 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| B-187-SB01     | 220507        | 8.22  | 428516        | 9.10  | 389120        | 11.87  |
| B-187-SB02     | 221135        | 8.22  | 428494        | 9.10  | 404884        | 11.87  |
| B-202-SB01     | 214739        | 8.22  | 418395        | 9.10  | 402557        | 11.87  |
| VN0603WBL01    | 257056        | 8.22  | 493879        | 9.09  | 460882        | 11.87  |
| VN0603WBS01    | 311874        | 8.22  | 552803        | 9.09  | 467842        | 11.87  |
| VN0603WBSD01   | 283988        | 8.22  | 510275        | 9.09  | 430553        | 11.87  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.