

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177
 Lab File ID: VN086835.D Date Analyzed: 06/03/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:30
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	208517	13.788				
UPPER LIMIT	417034	14.288				
LOWER LIMIT	104259	13.288				
EPA SAMPLE NO.						
B-187-SB01	163382	13.79				
B-187-SB02	181011	13.79				
B-202-SB01	183854	13.79				
VN0603WBL01	202290	13.79				
VN0603WBS01	208414	13.79				
VN0603WBSD01	192303	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.