

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133811
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP133838
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060225

Review By	John Carlone	Review On	6/3/2025 9:49:55 AM
Supervise By	Maresh Dadoda	Supervise On	6/3/2025 2:37:54 PM
SubDirectory	VX060225	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134094
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134095,VP134096

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046432.D	02 Jun 2025 09:42		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046433.D	02 Jun 2025 10:21	pH#Lot#V12668	JC/MD	Ok,M
3	VX0602MBL01	VX0602MBL01	VX046434.D	02 Jun 2025 10:51		JC/MD	Ok
4	VX0602WBL01	VX0602WBL01	VX046435.D	02 Jun 2025 11:14		JC/MD	Ok
5	VX0602WBS01	VX0602WBS01	VX046436.D	02 Jun 2025 11:37		JC/MD	Ok,M
6	VX0602WBSD01	VX0602WBSD01	VX046437.D	02 Jun 2025 12:04		JC/MD	Ok,M
7	IBLK	IBLK	VX046438.D	02 Jun 2025 12:27		JC/MD	Ok
8	Q2164-01	STORAGE-BLANK-SO	VX046439.D	02 Jun 2025 12:50	vial A pH<2	JC/MD	Ok
9	Q2164-02	STORAGE-BLANK-WA	VX046440.D	02 Jun 2025 13:14	vial A pH<2	JC/MD	Ok
10	Q2164-03	STORAGE-BLANK-WA	VX046441.D	02 Jun 2025 13:37	vial A pH<2	JC/MD	Ok
11	Q2164-04	STORAGE-BLANK-SAI	VX046442.D	02 Jun 2025 14:00	vial A pH<2	JC/MD	Ok
12	Q2177-08	EB05312025	VX046443.D	02 Jun 2025 14:23	vial A pH<2 EB;Hit of comp.#16	JC/MD	ReRun
13	Q2177-09	TB05312025	VX046444.D	02 Jun 2025 14:47	vial A pH<2 TB	JC/MD	Ok
14	Q2157-01	291	VX046445.D	02 Jun 2025 15:36	cloth material	JC/MD	Ok
15	VX0602MBS01	VX0602MBS01	VX046446.D	02 Jun 2025 15:59		JC/MD	Ok,M
16	VX0602MBSD01	VX0602MBSD01	VX046447.D	02 Jun 2025 16:22		JC/MD	Ok,M
17	Q2155-01	446	VX046448.D	02 Jun 2025 16:46	Need Straight Run	JC/MD	Not Ok
18	Q2155-02	447-SOLID	VX046449.D	02 Jun 2025 17:09	Need Soil Run	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060225

Review By	John Carlone	Review On	6/3/2025 9:49:55 AM
Supervise By	Maresh Dadoda	Supervise On	6/3/2025 2:37:54 PM
SubDirectory	VX060225	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134094		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134095,VP134096		

19	Q2161-01	B27-SOIL-SAMPLE	VX046450.D	02 Jun 2025 17:32		JC/MD	Ok
20	Q2161-02	B28-SOIL-SAMPLE	VX046451.D	02 Jun 2025 17:56	Need Soil Run	JC/MD	Ok
21	IBLK	IBLK	VX046452.D	02 Jun 2025 18:19		JC/MD	Ok
22	IBLK	IBLK	VX046453.D	02 Jun 2025 18:42		JC/MD	Ok
23	Q2177-08	EB05312025	VX046454.D	02 Jun 2025 19:06	vial B pH<2	JC/MD	Ok
24	Q2169-01	303-PPR-1	VX046455.D	02 Jun 2025 19:29	Need Straight Run	JC/MD	Not Ok
25	Q2169-03	303-PPR-2	VX046456.D	02 Jun 2025 19:52	Need Straight Run	JC/MD	Not Ok
26	Q2155-01	446	VX046457.D	02 Jun 2025 20:39	Need Soil Run	JC/MD	Not Ok
27	VSTDCCC050	VSTDCCC050EC	VX046458.D	02 Jun 2025 21:03		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM		
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y060225S.M

STD. NAME	STD REF.#
Tune/Reschk	VP134084
Initial Calibration Stds	VP134085,VP134086,VP134087,VP134088,VP134089,VP134090
CCC	VP134092,VP134093
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134091
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022488.D	02 Jun 2025 08:31		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022489.D	02 Jun 2025 09:08	Need ICAL	SY/MD	Not Ok
3	VY0602SBL01	VY0602SBL01	VY022490.D	02 Jun 2025 10:38		SY/MD	Not Ok
4	VSTDICC005	VSTDICC005	VY022491.D	02 Jun 2025 11:46		SY/MD	Ok,M
5	VSTDICC010	VSTDICC010	VY022492.D	02 Jun 2025 12:09		SY/MD	Ok,M
6	VSTDICC020	VSTDICC020	VY022493.D	02 Jun 2025 12:32		SY/MD	Ok,M
7	VSTDICCC050	VSTDICCC050	VY022494.D	02 Jun 2025 12:54		SY/MD	Ok,M
8	VSTDICC100	VSTDICC100	VY022495.D	02 Jun 2025 13:17		SY/MD	Ok,M
9	VSTDICC150	VSTDICC150	VY022496.D	02 Jun 2025 13:39		SY/MD	Ok,M
10	VIBLK	VIBLK	VY022497.D	02 Jun 2025 14:03		SY/MD	Ok
11	VSTDICV050	ICVVY060225	VY022498.D	02 Jun 2025 14:59	Icv Fail For DOD For Com.# 18	SY/MD	Ok,M
12	VY0602SBL02	VY0602SBL02	VY022499.D	02 Jun 2025 16:00		SY/MD	Ok
13	VY0602SBS01	VY0602SBS01	VY022500.D	02 Jun 2025 16:24		SY/MD	Ok,M
14	VY0602SBSD01	VY0602SBSD01	VY022501.D	02 Jun 2025 16:47		SY/MD	Ok,M
15	Q2160-02	TP04-MHG-VOC	VY022502.D	02 Jun 2025 17:10	vial A	SY/MD	Ok
16	Q2152-01	OK-02-05292025	VY022503.D	02 Jun 2025 17:34	Not Purged,vial B	SY/MD	Not Ok
17	Q2153-01RE	TR-04-0592025RE	VY022504.D	02 Jun 2025 17:57	Surrogate Fail;Internal Standard Fail, vial B	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM		
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y060225S.M
STD. NAME		STD REF.#			
Tune/Reschk		VP134084			
Initial Calibration Stds		VP134085,VP134086,VP134087,VP134088,VP134089,VP134090			
CCC		VP134092,VP134093			
Internal Standard/PEM		VP133934			
ICV/I.BLK		VP134091			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2147-02	VOC	VY022505.D	02 Jun 2025 18:21	Not purge,vial B	SY/MD	Not Ok
19	Q2150-02	TP-42	VY022506.D	02 Jun 2025 18:44	vial B	SY/MD	Ok
20	Q2150-05RE	TP-47RE	VY022507.D	02 Jun 2025 19:08	Internal Standard Fail,vial B	SY/MD	Confirms
21	Q2161-01	B27-SOIL-SAMPLE	VY022508.D	02 Jun 2025 19:31	vial-B Not purge	SY/MD	Not Ok
22	VSTDCCC050	VSTDCCC050EC	VY022509.D	02 Jun 2025 19:54		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060325

Review By	Maresh Dadoda	Review On	6/4/2025 4:21:05 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/4/2025 4:23:38 PM
SubDirectory	VY060325	HP Acquire Method	MSVOA_Y HP Processing Method 82y060225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134097
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134098,VP134099 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022510.D	03 Jun 2025 09:03		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022511.D	03 Jun 2025 10:03		SY/MD	Ok,M
3	VY0603SBL01	VY0603SBL01	VY022512.D	03 Jun 2025 10:40		SY/MD	Ok
4	VY0603SBS01	VY0603SBS01	VY022513.D	03 Jun 2025 11:15		SY/MD	Ok,M
5	VY0603SBSD01	VY0603SBSD01	VY022514.D	03 Jun 2025 11:38		SY/MD	Ok,M
6	Q2155-01	446	VY022515.D	03 Jun 2025 12:17	vial-A	SY/MD	Ok
7	Q2155-02	447-SOLID	VY022516.D	03 Jun 2025 12:40	vial-A	SY/MD	Ok,M
8	Q2185-02	TP02-MHB-VOC	VY022517.D	03 Jun 2025 13:04	vial-A	SY/MD	Ok
9	Q2185-06	TP01-MHA-VOC	VY022518.D	03 Jun 2025 13:27	vial-A	SY/MD	Ok
10	Q2172-02	TP06-MHQ-VOC	VY022519.D	03 Jun 2025 13:51	vial-A	SY/MD	Ok
11	Q2168-03	A3	VY022520.D	03 Jun 2025 14:14	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
12	Q2168-07	B3	VY022521.D	03 Jun 2025 14:38	vial-A Need MeOH;Internal Standard Fail	SY/MD	Dilution
13	Q2168-11	C2	VY022522.D	03 Jun 2025 15:01	vial-A Need MeOH;Internal Standard Fail	SY/MD	Dilution
14	Q2182-01	OR-03-06022025	VY022523.D	03 Jun 2025 15:24	vial-A Internal Standard Fail	SY/MD	ReRun
15	Q2162-01	BP-VPB-182-GW-580-5	VY022524.D	03 Jun 2025 15:48	vial-A	SY/MD	Ok
16	Q2161-02	B28-SOIL-SAMPLE	VY022525.D	03 Jun 2025 16:11	vial-B Not purge	SY/MD	Not Ok
17	Q2176-01	TP-46	VY022526.D	03 Jun 2025 16:35	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060325

Review By	Mahesh Dadoda	Review On	6/4/2025 4:21:05 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/4/2025 4:23:38 PM		
SubDirectory	VY060325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y060225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134097			
CCC		VP134098,VP134099			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2176-02	TP-56	VY022527.D	03 Jun 2025 16:58	vial-A	SY/MD	Ok
19	Q2176-03	TP-25	VY022528.D	03 Jun 2025 17:22	vial-A	SY/MD	Ok
20	Q2176-04	TP-26	VY022529.D	03 Jun 2025 17:45	vial-A	SY/MD	Ok
21	Q2176-05	TP-28	VY022530.D	03 Jun 2025 18:09	vial-A	SY/MD	Ok
22	Q2176-06	TP-27	VY022531.D	03 Jun 2025 18:32	vial-A	SY/MD	Ok
23	Q2176-07	TP-31	VY022532.D	03 Jun 2025 18:56	vial-A	SY/MD	Ok
24	Q2176-08	TP-65	VY022533.D	03 Jun 2025 19:19	vial-A	SY/MD	Ok
25	Q2177-01	B-187-SB00	VY022534.D	03 Jun 2025 19:42	vial-A	SY/MD	Ok
26	VIBLK	VIBLK	VY022535.D	03 Jun 2025 20:06		SY/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VY022536.D	03 Jun 2025 20:51		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060425

Review By	Maresh Dadoda	Review On	6/5/2025 4:43:35 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:47:03 PM
SubDirectory	VY060425	HP Acquire Method	MSVOA_Y HP Processing Method 82y060225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134109
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134110,VP134111,MDL VP134112,VP134113 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022537.D	04 Jun 2025 08:48		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022538.D	04 Jun 2025 09:54		SY/MD	Ok,M
3	VY0604SBL01	VY0604SBL01	VY022539.D	04 Jun 2025 10:24		SY/MD	Ok
4	VY0604SBS01	VY0604SBS01	VY022540.D	04 Jun 2025 10:54		SY/MD	Ok,M
5	VY0604SBSD01	VY0604SBSD01	VY022541.D	04 Jun 2025 11:16		SY/MD	Ok,M
6	Q2126-03	MDL-SOIL-03-QT2-202	VY022542.D	04 Jun 2025 11:53		SY/MD	Ok,M
7	Q2126-03	MDL-SOIL-03-QT2-202	VY022543.D	04 Jun 2025 12:16		SY/MD	Ok,M
8	Q2168-03RE	A3RE	VY022544.D	04 Jun 2025 12:39	vial-B Internal Standard Fail; Surrogate Fail	SY/MD	Confirms
9	Q2182-01RE	OR-03-06022025RE	VY022545.D	04 Jun 2025 13:03	vial-B Internal Standard Fail	SY/MD	Confirms
10	Q2195-01	OK-01-060325	VY022546.D	04 Jun 2025 13:26	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q2194-01	COMP-12	VY022547.D	04 Jun 2025 13:50	vial-A	SY/MD	Ok
12	Q2194-03	COMP-13	VY022548.D	04 Jun 2025 14:13	vial-A Internal Standard Fail	SY/MD	ReRun
13	Q2199-02	ETGI-343	VY022549.D	04 Jun 2025 14:37	vial-A Internal Standard Fail	SY/MD	ReRun
14	Q2199-04	VNJ-231	VY022550.D	04 Jun 2025 15:00	vial-A Not purge	SY/MD	Not Ok
15	Q2199-06	72-11978	VY022551.D	04 Jun 2025 15:24	vial-A	SY/MD	Ok
16	Q2198-01	B-202-SB02	VY022552.D	04 Jun 2025 15:47	vial-A	SY/MD	Ok
17	Q2198-03	B-207-SB02	VY022553.D	04 Jun 2025 16:11	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060425

Review By	Mahesh Dadoda	Review On	6/5/2025 4:43:35 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/5/2025 4:47:03 PM		
SubDirectory	VY060425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y060225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134109			
CCC		VP134110,VP134111,MDL VP134112,VP134113			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2177-02	B-187-SB01	VY022554.D	04 Jun 2025 16:34	vial-A	SY/MD	Ok
19	Q2177-04	B-187-SB02	VY022555.D	04 Jun 2025 16:58	vial-A	SY/MD	Ok
20	Q2177-06	B-202-SB01	VY022556.D	04 Jun 2025 17:21	vial-A Internal Standard Fail	SY/MD	ReRun
21	VIBLK	VIBLK	VY022557.D	04 Jun 2025 17:45		SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VY022558.D	04 Jun 2025 18:30		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060525

Review By	Maresh Dadoda	Review On	6/6/2025 5:55:27 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/6/2025 5:56:08 PM
SubDirectory	VY060525	HP Acquire Method	MSVOA_Y HP Processing Method 82y060225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134127
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134128,VP134129 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022559.D	05 Jun 2025 08:08		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022560.D	05 Jun 2025 08:40		SY/MD	Ok,M
3	VY0605SBL01	VY0605SBL01	VY022561.D	05 Jun 2025 09:11		SY/MD	Ok
4	VY0605SBS01	VY0605SBS01	VY022562.D	05 Jun 2025 09:41		SY/MD	Ok,M
5	VY0605SBSD01	VY0605SBSD01	VY022563.D	05 Jun 2025 10:31		SY/MD	Ok,M
6	Q2199-02RE	ETGI-343RE	VY022564.D	05 Jun 2025 11:12	vial-B Internal Standard Fail; Surrogate Fail	SY/MD	Confirms
7	Q2199-06	72-11978	VY022565.D	05 Jun 2025 11:35	vial-B Already analyzed	SY/MD	Not Ok
8	Q2194-03	COMP-13	VY022566.D	05 Jun 2025 11:59	vial-B	SY/MD	Ok
9	Q2195-01RE	OK-01-060325RE	VY022567.D	05 Jun 2025 12:22	Internal Standard Fail	SY/MD	Confirms
10	Q2177-06	B-202-SB01	VY022568.D	05 Jun 2025 12:45	vial-B	SY/MD	Ok
11	Q2219-01	72-11995	VY022569.D	05 Jun 2025 13:09	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
12	Q2214-01	RBR251425	VY022570.D	05 Jun 2025 13:33	vial-A NOT purged	SY/MD	Not Ok
13	Q2224-01	EO-01-06042025	VY022571.D	05 Jun 2025 13:56	vial-A Internal Standard Fail	SY/MD	ReRun
14	Q2208-02	BU-703-VOC-06	VY022572.D	05 Jun 2025 14:20	vial-A Internal Standard Fail	SY/MD	ReRun
15	Q2208-11	BU-703-VOC-07	VY022573.D	05 Jun 2025 14:44	vial-A Internal Standard Fail	SY/MD	ReRun
16	Q2208-20	BU-703-VOC-08	VY022574.D	05 Jun 2025 15:07	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q2208-29	BU-703-VOC-09	VY022575.D	05 Jun 2025 15:31	vial-A Internal Standard Fail	SY/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060525

Review By	Mahesh Dadoda	Review On	6/6/2025 5:55:27 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/6/2025 5:56:08 PM		
SubDirectory	VY060525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y060225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134127			
CCC		VP134128,VP134129			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2207-02	BU-703-VOC-01	VY022576.D	05 Jun 2025 15:55	vial-A Internal Standard Fail	SY/MD	ReRun
19	Q2207-11	BU-703-VOC-02	VY022577.D	05 Jun 2025 16:19	vial-A Internal Standard Fail	SY/MD	ReRun
20	Q2207-20	BU-703-VOC-03	VY022578.D	05 Jun 2025 16:42	vial-A Internal Standard Fail	SY/MD	ReRun
21	Q2207-29	BU-703-VOC-04	VY022579.D	05 Jun 2025 17:06	vial-A Internal Standard Fail	SY/MD	ReRun
22	Q2207-38	BU-703-VOC-05	VY022580.D	05 Jun 2025 17:30	vial-A Internal Standard Fail	SY/MD	ReRun
23	Q2225-01	SU-03-06042025	VY022581.D	05 Jun 2025 17:54	vial-A Internal Standard Fail	SY/MD	ReRun
24	Q2226-02	TP06-MHI-VOC	VY022582.D	05 Jun 2025 18:17	vial-A Internal Standard Fail	SY/MD	ReRun
25	Q2227-02	TP07-MHH-VOC	VY022583.D	05 Jun 2025 18:41	vial-A Internal Standard Fail	SY/MD	ReRun
26	VIBLK	VIBLK	VY022584.D	05 Jun 2025 19:05		SY/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VY022585.D	05 Jun 2025 19:51		SY/MD	Ok,M

M : Manual Integration