

Surrogate Summary

SDG No.: Q2177

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2177-01	B-187-SB00	1,2-Dichloroethane-d4	50	48.3	97	70 (63)	130 (155)
		Dibromofluoromethane	50	51.7	103	70 (70)	130 (134)
		Toluene-d8	50	48.7	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	30.8	62 *	70 (17)	130 (146)
Q2177-02	B-187-SB01	1,2-Dichloroethane-d4	50	52.2	104	70 (63)	130 (155)
		Dibromofluoromethane	50	51.2	102	70 (70)	130 (134)
		Toluene-d8	50	49.8	100	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.8	86	70 (17)	130 (146)
Q2177-04	B-187-SB02	1,2-Dichloroethane-d4	50	52.0	104	70 (63)	130 (155)
		Dibromofluoromethane	50	52.1	104	70 (70)	130 (134)
		Toluene-d8	50	49.7	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	33.1	66 *	70 (17)	130 (146)
Q2177-06	B-202-SB01	1,2-Dichloroethane-d4	50	49.9	100	70 (63)	130 (155)
		Dibromofluoromethane	50	49.5	99	70 (70)	130 (134)
		Toluene-d8	50	49.8	100	70 (74)	130 (123)
		4-Bromofluorobenzene	50	41.7	83	70 (17)	130 (146)
VY0603SBL01	VY0603SBL01	1,2-Dichloroethane-d4	50	53.5	107	70 (63)	130 (155)
		Dibromofluoromethane	50	50.9	102	70 (70)	130 (134)
		Toluene-d8	50	49.9	100	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.0	84	70 (17)	130 (146)
VY0603SBS01	VY0603SBS01	1,2-Dichloroethane-d4	50	50.0	100	70 (63)	130 (155)
		Dibromofluoromethane	50	51.0	102	70 (70)	130 (134)
		Toluene-d8	50	51.5	103	70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.0	98	70 (17)	130 (146)
VY0603SBSD01	VY0603SBSD01	1,2-Dichloroethane-d4	50	50.7	101	70 (63)	130 (155)
		Dibromofluoromethane	50	51.5	103	70 (70)	130 (134)
		Toluene-d8	50	52.6	105	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.1	100	70 (17)	130 (146)
VY0604SBL01	VY0604SBL01	1,2-Dichloroethane-d4	50	52.7	105	70 (63)	130 (155)
		Dibromofluoromethane	50	50.7	101	70 (70)	130 (134)
		Toluene-d8	50	49.3	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	41.7	83	70 (17)	130 (146)
VY0604SBS01	VY0604SBS01	1,2-Dichloroethane-d4	50	53.9	108	70 (63)	130 (155)
		Dibromofluoromethane	50	53.3	107	70 (70)	130 (134)
		Toluene-d8	50	53.4	107	70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.0	104	70 (17)	130 (146)
VY0605SBL01	VY0605SBL01	1,2-Dichloroethane-d4	50	51.7	103	70 (63)	130 (155)
		Dibromofluoromethane	50	50.3	101	70 (70)	130 (134)
		Toluene-d8	50	49.6	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	40.7	81	70 (17)	130 (146)
VY0605SBS01	VY0605SBS01	1,2-Dichloroethane-d4	50	49.6	99	70 (63)	130 (155)
		Dibromofluoromethane	50	50.3	101	70 (70)	130 (134)
		Toluene-d8	50	50.5	101	70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.3	97	70 (17)	130 (146)

Surrogate Summary

SDG No.: Q2177

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2177-08	EB05312025	1,2-Dichloroethane-d4	50	51.9	104	70 (74)	130 (125)
		Dibromofluoromethane	50	49.1	98	70 (75)	130 (124)
		Toluene-d8	50	50.3	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.1	106	70 (77)	130 (121)
Q2177-09	TB05312025	1,2-Dichloroethane-d4	50	55.7	111	70 (74)	130 (125)
		Dibromofluoromethane	50	51.7	103	70 (75)	130 (124)
		Toluene-d8	50	50.1	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.4	103	70 (77)	130 (121)
VX0602WBL01	VX0602WBL01	1,2-Dichloroethane-d4	50	53.9	108	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	50.9	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.8	102	70 (77)	130 (121)
VX0602WBS01	VX0602WBS01	1,2-Dichloroethane-d4	50	49.5	99	70 (74)	130 (125)
		Dibromofluoromethane	50	51.6	103	70 (75)	130 (124)
		Toluene-d8	50	49.2	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.8	100	70 (77)	130 (121)
VX0602WBSD0	VX0602WBSD01	1,2-Dichloroethane-d4	50	50.1	100	70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101	70 (75)	130 (124)
		Toluene-d8	50	48.3	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)

