

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177
Lab File ID: VX046433.D Date Analyzed: 06/02/2025
Instrument ID: MSVOA_X Time Analyzed: 10:21
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	91247	5.54	155497	6.75	135003	10.05
UPPER LIMIT	182494	6.038	310994	7.251	270006	10.549
LOWER LIMIT	45623.5	5.038	77748.5	6.251	67501.5	9.549
EPA SAMPLE NO.						
EB05312025	65892	5.54	131320	6.76	124794	10.06
TB05312025	60024	5.54	122332	6.76	117285	10.05
VX0602WBL01	66983	5.54	131566	6.76	125222	10.05
VX0602WBS01	88536	5.54	152496	6.76	133074	10.05
VX0602WBSD01	82810	5.54	146139	6.76	127160	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.