

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177  
 Lab File ID: VY022511.D Date Analyzed: 06/03/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 10:03  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|-------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 187797        | 7.71  | 325302        | 8.62  | 276655        | 11.42 |
| UPPER LIMIT    | 375594        | 8.213 | 650604        | 9.115 | 553310        | 11.92 |
| LOWER LIMIT    | 93898.5       | 7.213 | 162651        | 8.115 | 138328        | 10.92 |
| EPA SAMPLE NO. |               |       |               |       |               |       |
| B-187-SB00     | 249503        | 7.71  | 446768        | 8.62  | 313918        | 11.41 |
| VY0603SBL01    | 287398        | 7.71  | 534399        | 8.62  | 428748        | 11.42 |
| VY0603SBS01    | 197612        | 7.71  | 337632        | 8.62  | 284268        | 11.42 |
| VY0603SBSD01   | 196295        | 7.71  | 333757        | 8.62  | 277257        | 11.41 |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.