

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177
 Lab File ID: VY022538.D Date Analyzed: 06/04/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	186956	7.71	316261	8.62	266067	11.42
UPPER LIMIT	373912	8.213	632522	9.116	532134	11.92
LOWER LIMIT	93478	7.213	158131	8.116	133034	10.92
EPA SAMPLE NO.						
B-187-SB01	261314	7.71	474753	8.61	386745	11.41
B-187-SB02	280753	7.71	497958	8.61	360175	11.41
VY0604SBL01	271205	7.71	498185	8.62	390593	11.41
VY0604SBS01	177182	7.71	307142	8.62	260223	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.