

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177
 Lab File ID: VY022560.D Date Analyzed: 06/05/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	179733	7.71	304462	8.62	260955	11.41
UPPER LIMIT	359466	8.207	608924	9.116	521910	11.914
LOWER LIMIT	89866.5	7.207	152231	8.116	130478	10.914
EPA SAMPLE NO.						
B-202-SB01	270828	7.71	498576	8.61	399252	11.41
VY0605SBL01	260317	7.71	481555	8.61	383268	11.41
VY0605SBS01	185960	7.71	319893	8.62	265176	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.