

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177
 Lab File ID: VX046433.D Date Analyzed: 06/02/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:21
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	65264	12.018				
UPPER LIMIT	130528	12.518				
LOWER LIMIT	32632	11.518				
EPA SAMPLE NO.						
EB05312025	55855	12.02				
TB05312025	49224	12.02				
VX0602WBL01	51823	12.02				
VX0602WBS01	62985	12.02				
VX0602WBSD01	61222	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.