

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q2177 SAS No.: Q2177 SDG NO.: Q2177
 Lab File ID: VY022560.D Date Analyzed: 06/05/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	122317	13.346				
UPPER LIMIT	244634	13.846				
LOWER LIMIT	61158.5	12.846				
EPA SAMPLE NO.						
B-202-SB01	144715	13.34				
VY0605SBL01	134444	13.35				
VY0605SBS01	121174	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.