



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2180

Project: Linden Generating Station - 402

RunNo.: LB136072

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.01	7	100	90-110	06/02/2025
Sample ID:	CCV1						
Field pH		pH	7.03	7	100	90-110	06/02/2025
Sample ID:	CCV2						
Field pH		pH	7.02	7	100	90-110	06/02/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2180

Project: Linden Generating Station - 402

RunNo.: LB136072

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------