ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2181 MATRIX:		IATRIX: Water			
METH	OD: 8270E/3510				
				NO	T/EG
1.	Chromatograms Labeled/Compounds Identified. (Field samples an	d Method Blanks)	NA	NO	YES ✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY A CLP AND NJ)	SP CLP,			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 seri 8000 Series.	es and 12 hours for			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days be analysis and continuing calibration performed within 24 hours of sa 600 series and 12 hours for 8000 series.				✓
5.	GC/MS Calibration Requirements.				\checkmark
	The Initial Calibration met the requirements .				
	The Continuous Calibration met the requirements .				
6.	Blank Contamination - If yes, list compounds and concentrations in	each blank:		\checkmark	
7.	Surrogate Recoveries Meet Criteria				\checkmark
	If not met, list those compounds and their recoveries which fall out ranges.	side the acceptable			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria				\checkmark
	If not met, list those compounds and their recoveries which fall out range.	side the acceptable			
	The Blank Spike met requirements for all samples .				
	The Blank Spike Duplicate met requirements for all samples .				
9.	Internal Standard Area/Retention Time Shift Meet Criteria				\checkmark
	Comments:				
10.	Extraction Holding Time Met				\checkmark
	If not met, list number of days exceeded for each sample:				

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

<u>GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY</u> (CONTINUED)

NA NO YES

 \checkmark

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points. Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the % RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date