

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF052025

Review By	Rahul	Review On	5/21/2025 2:52:20 PM		
Supervise By	Jagrut	Supervise On	5/21/2025 4:44:06 PM		
SubDirectory	BF052025	HP Acquire Method	BNA_F	HP Processing Method	bf052025
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791			
CCC		SP6787			
Internal Standard/PEM		S12665,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142465.D	20 May 2025 11:13	RC/JU	Ok
2	SSTDCCC040	BF142466.D	20 May 2025 11:41	RC/JU	Not Ok
3	SSTDICC2.5	BF142467.D	20 May 2025 12:10	RC/JU	Ok
4	SSTDICC005	BF142468.D	20 May 2025 12:38	RC/JU	Ok
5	SSTDICC010	BF142469.D	20 May 2025 13:07	RC/JU	Ok,M
6	SSTDICC020	BF142470.D	20 May 2025 13:36	RC/JU	Ok
7	SSTDICCC040	BF142471.D	20 May 2025 14:05	RC/JU	Ok
8	SSTDICC050	BF142472.D	20 May 2025 14:34	RC/JU	Ok
9	SSTDICC060	BF142473.D	20 May 2025 15:03	RC/JU	Ok
10	SSTDICC080	BF142474.D	20 May 2025 15:31	RC/JU	Ok
11	SSTDICV040	BF142475.D	20 May 2025 16:31	RC/JU	Ok
12	PB168067TB	BF142476.D	20 May 2025 17:29	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF060525

Review By	Rahul	Review On	6/6/2025 12:18:39 PM
Supervise By	Jagrut	Supervise On	6/6/2025 1:17:23 PM
SubDirectory	BF060525	HP Acquire Method	BNA_F
		HP Processing Method	bf052025
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791		
CCC	SP6787		
Internal Standard/PEM	S12667,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142624.D	05 Jun 2025 09:07	RC/JU	Ok
2	SSTDCCC040	BF142625.D	05 Jun 2025 09:37	RC/JU	Ok
3	PB168285BL	BF142626.D	05 Jun 2025 11:04	RC/JU	Ok
4	Q2181-02	BF142627.D	05 Jun 2025 11:35	RC/JU	Ok
5	Q2194-03	BF142628.D	05 Jun 2025 12:08	RC/JU	Ok
6	Q2194-03MS	BF142629.D	05 Jun 2025 12:38	RC/JU	Ok,M
7	Q2194-03MSD	BF142630.D	05 Jun 2025 13:07	RC/JU	Ok,M
8	Q2185-05	BF142631.D	05 Jun 2025 13:37	RC/JU	Ok
9	Q2194-01	BF142632.D	05 Jun 2025 14:07	RC/JU	Ok
10	Q2199-01	BF142633.D	05 Jun 2025 14:37	RC/JU	Ok,M
11	Q2199-03	BF142634.D	05 Jun 2025 15:06	RC/JU	Ok,M
12	Q2199-05	BF142635.D	05 Jun 2025 15:36	RC/JU	Ok,M
13	Q2178-01DL	BF142636.D	05 Jun 2025 16:07	RC/JU	Dilution
14	Q2178-01DL2	BF142637.D	05 Jun 2025 16:38	RC/JU	Ok,M
15	SSTDCCC040	BF142638.D	05 Jun 2025 17:18	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF061125

Review By	Rahul	Review On	6/11/2025 9:54:45 AM
Supervise By	Jagrut	Supervise On	6/11/2025 9:58:22 AM
SubDirectory	BF061125	HP Acquire Method	BNA_F
		HP Processing Method	BF061125
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791		
CCC	SP6787		
Internal Standard/PEM	S12668,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142710.D	10 Jun 2025 15:42	RC/JU	Ok
2	SSTDCCC040	BF142711.D	10 Jun 2025 16:11	RC/JU	Not Ok
3	SSTDICC2.5	BF142712.D	10 Jun 2025 16:54	RC/JU	Ok
4	SSTDICC005	BF142713.D	10 Jun 2025 17:24	RC/JU	Ok
5	SSTDICC010	BF142714.D	10 Jun 2025 17:53	RC/JU	Ok,M
6	SSTDICC020	BF142715.D	10 Jun 2025 18:22	RC/JU	Ok
7	SSTDICCC040	BF142716.D	10 Jun 2025 18:52	RC/JU	Ok
8	SSTDICC050	BF142717.D	10 Jun 2025 19:21	RC/JU	Ok
9	SSTDICC060	BF142718.D	10 Jun 2025 19:50	RC/JU	Ok
10	SSTDICC080	BF142719.D	10 Jun 2025 20:19	RC/JU	Ok,M
11	SSTDICV040	BF142720.D	10 Jun 2025 20:49	RC/JU	Ok
12	PB168323BL	BF142721.D	10 Jun 2025 21:47	RC/JU	Ok
13	DFTPP	BF142722.D	11 Jun 2025 08:56	RC/JU	Ok
14	SSTDCCC040	BF142723.D	11 Jun 2025 09:24	RC/JU	Ok
15	PB168376BL	BF142724.D	11 Jun 2025 09:53	RC/JU	Ok
16	PB168376BS	BF142725.D	11 Jun 2025 10:22	RC/JU	Ok,M
17	PB168234BS	BF142726.D	11 Jun 2025 10:51	RC/JU	Ok,M
18	PB168285BS	BF142727.D	11 Jun 2025 11:21	RC/JU	Ok,M
19	PB168285BSD	BF142728.D	11 Jun 2025 11:50	RC/JU	Ok,M
20	PB168378BS	BF142729.D	11 Jun 2025 12:19	RC/JU	Ok,M
21	PB168378BSD	BF142730.D	11 Jun 2025 12:49	RC/JU	Ok,M

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF061125

Review By	Rahul	Review On	6/11/2025 9:54:45 AM
Supervise By	Jagrut	Supervise On	6/11/2025 9:58:22 AM
SubDirectory	BF061125	HP Acquire Method	BNA_F
		HP Processing Method	BF061125
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791		
CCC	SP6787		
Internal Standard/PEM	S12668,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB168378BL	BF142731.D	11 Jun 2025 13:18	RC/JU	Ok
23	Q2264-04	BF142732.D	11 Jun 2025 13:52	RC/JU	Ok
24	Q2268-10	BF142733.D	11 Jun 2025 14:21	RC/JU	Ok
25	Q2268-03	BF142734.D	11 Jun 2025 14:50	RC/JU	Dilution
26	Q2268-04MS	BF142735.D	11 Jun 2025 15:20	RC/JU	Ok,M
27	Q2268-05MSD	BF142736.D	11 Jun 2025 15:50	RC/JU	Ok
28	Q2268-06	BF142737.D	11 Jun 2025 16:19	RC/JU	Dilution
29	Q2268-07	BF142738.D	11 Jun 2025 16:49	RC/JU	Dilution
30	Q2268-08	BF142739.D	11 Jun 2025 17:19	RC/JU	Dilution
31	Q2273-01	BF142740.D	11 Jun 2025 17:49	RC/JU	Ok
32	Q2273-05	BF142741.D	11 Jun 2025 18:18	RC/JU	Ok
33	Q2273-05MS	BF142742.D	11 Jun 2025 18:48	RC/JU	Ok
34	Q2273-05MSD	BF142743.D	11 Jun 2025 19:18	RC/JU	Ok
35	Q2268-03DL	BF142744.D	11 Jun 2025 19:48	RC/JU	Ok
36	Q2268-03	BF142745.D	11 Jun 2025 20:17	RC/JU	Not Ok
37	Q2280-01	BF142746.D	11 Jun 2025 20:47	RC/JU	ReRun

M : Manual Integration