

Surrogate Summary

SW-846

SDG No.: Q2181

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168285BL	PB168285BL	2-Fluorophenol	150	125	83		23	138
		Phenol-d6	150	124	83		10	134
		2,4,6-Tribromophenol	150	122	82		44	137
PB168285BS	PB168285BS	2-Fluorophenol	150	126	84		23	138
		Phenol-d6	150	127	85		10	134
		2,4,6-Tribromophenol	150	129	86		44	137
PB168285BSD	PB168285BSD	2-Fluorophenol	150	125	83		23	138
		Phenol-d6	150	127	84		10	134
		2,4,6-Tribromophenol	150	129	86		44	137
Q2181-02	38073-032124	2-Fluorophenol	150	136	91		23	138
		Phenol-d6	150	140	93		10	134
		2,4,6-Tribromophenol	150	146	98		44	137