

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2181

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E **DataFile:** BF142728.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168285BSD	2,4-Dimethylphenol	50	46.3	ug/L	93	2			42	142	20