

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2181 SAS No.: Q2181 SDG NO.: Q2181

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/03/2025

Lab File ID: BN037145.D Time Analyzed: 12:51

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2099	7.59	5221	10.37	2886	14.24
UPPER LIMIT	4198	8.09	10442	10.872	5772	14.735
LOWER LIMIT	1049.5	7.09	2610.5	9.872	1443	13.735
EPA SAMPLE NO.						
01 38072-062624	1726	7.60	4746	10.37	2693	14.25
02 38072-062624DL	2324	7.60	6120	10.37	3333	14.23

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.