

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2181 SAS No.: Q2181 SDG NO.: Q2181

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/04/2025

Lab File ID: BN037166.D Time Analyzed: 10:27

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2757	7.589	7191	10.37	3834	14.23
UPPER LIMIT	5514	8.089	14382	10.872	7668	14.734
LOWER LIMIT	1378.5	7.089	3595.5	9.872	1917	13.734
EPA SAMPLE NO.						
01 PB168238BL	2941	7.59	7107	10.37	3690	14.25
02 PB168238BS	2103	7.59	5280	10.37	2700	14.24
03 PB168238BSD	2147	7.59	5350	10.37	2661	14.24

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.