

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2181 SAS No.: Q2181 SDG NO.: Q2181

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/03/2025

Lab File ID: BN037145.D Time Analyzed: 12:51

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5446	16.984	3769	21.189	3386	23.377
UPPER LIMIT	10892	17.484	7538	21.689	6772	23.877
LOWER LIMIT	2723	16.484	1884.5	20.689	1693	22.877
EPA SAMPLE NO.						
01 38072-062624	4905	16.98	3958	21.19	3402	23.37
02 38072-062624DL	6288	16.98	4597	21.18	3669	23.38

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.