

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2186

SDG NO.: Q2186

Lab File ID: BN037142.D

DFTPP Injection Date: 06/03/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	69.8
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	58.7
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	53.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	4.5
441	Present, but less than mass 443	10.3
442	Greater than 50% of mass 198	61.1
443	15.0 - 24.0% of mass 442	12.1 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037143.D	06/03/2025	11:39
SSTDICC0.2	SSTDICC0.2	BN037144.D	06/03/2025	12:15
SSTDICCC0.4	SSTDICCC0.4	BN037145.D	06/03/2025	12:51
SSTDICC0.8	SSTDICC0.8	BN037146.D	06/03/2025	13:26
SSTDICC1.6	SSTDICC1.6	BN037147.D	06/03/2025	14:02
SSTDICC3.2	SSTDICC3.2	BN037148.D	06/03/2025	14:38
SSTDICC5.0	SSTDICC5.0	BN037149.D	06/03/2025	15:14