

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2186

SDG NO.: Q2186

Lab File ID: BN037171.D

DFTPP Injection Date: 06/05/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:03

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	74.9
68	Less than 2.0% of mass 69	0.4 (0.7) 1
69	Mass 69 relative abundance	60.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	54.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	8.7
442	Greater than 50% of mass 198	52.4
443	15.0 - 24.0% of mass 442	10.3 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037172.D	06/05/2025	09:42
PB168286BL	PB168286BL	BN037173.D	06/05/2025	10:19
BP-VPB-182-GW-680-682	Q2186-03	BN037174.D	06/05/2025	10:55
BP-VPB-182-GW-720-722	Q2186-07	BN037175.D	06/05/2025	11:31
VPB182-HYD-20250530	Q2186-09	BN037176.D	06/05/2025	12:08
PB168286BS	PB168286BS	BN037182.D	06/05/2025	16:21
PB168286BSD	PB168286BSD	BN037186.D	06/05/2025	18:46
SSTDCCC0.4EC	SSTDCCC0.4	BN037187.D	06/05/2025	19:22