

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060325

Review By	John Carlone	Review On	6/4/2025 10:07:03 AM
Supervise By	Maresh Dadoda	Supervise On	6/4/2025 4:22:37 PM
SubDirectory	VX060325	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134100
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134101,VP134102

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046459.D	03 Jun 2025 09:40		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046460.D	03 Jun 2025 10:10	pH#Lot#V12668	JC/MD	Ok,M
3	VX0603MBL01	VX0603MBL01	VX046461.D	03 Jun 2025 10:38		JC/MD	Ok
4	VX0603WBL01	VX0603WBL01	VX046462.D	03 Jun 2025 11:01		JC/MD	Ok
5	VX0603WBS01	VX0603WBS01	VX046463.D	03 Jun 2025 11:24	BSD failed high for comp. #18	JC/MD	Ok,M
6	VX0603WBSD01	VX0603WBSD01	VX046464.D	03 Jun 2025 11:50	BSD failed high for comp. #18	JC/MD	Ok,M
7	Q2186-09	VPB182-HYD-2025053	VX046465.D	03 Jun 2025 12:13	vial A pH<2	JC/MD	Ok
8	Q2186-01	BP-VPB-182-TB-20250	VX046466.D	03 Jun 2025 12:36	vial A pH<2 TB	JC/MD	Ok
9	Q2186-02	BP-VPB-182-GW-670-6	VX046467.D	03 Jun 2025 12:59	vial A pH<2	JC/MD	Ok
10	Q2186-03	BP-VPB-182-GW-680-6	VX046468.D	03 Jun 2025 13:23	vial A pH<2	JC/MD	Ok
11	Q2186-06	BP-VPB-182-GW-705-7	VX046469.D	03 Jun 2025 13:46	vial A pH<2	JC/MD	Ok
12	Q2186-07	BP-VPB-182-GW-720-7	VX046470.D	03 Jun 2025 14:09	vial A pH<2	JC/MD	Ok
13	Q2186-08	BP-VPB-182-DUP-2025	VX046471.D	03 Jun 2025 14:33	vial A pH<2	JC/MD	Ok
14	Q2159-04	TP05-MHO-WC	VX046472.D	03 Jun 2025 14:56	vial A pH#5.0	JC/MD	Ok
15	Q2160-04	TP04-MHG-WC	VX046473.D	03 Jun 2025 15:20	vial A pH#5.0	JC/MD	Ok,M
16	Q2160-08	TP05-MHH-WC	VX046474.D	03 Jun 2025 15:43	vial A pH#5.0	JC/MD	Ok,M
17	Q2173-06	OR-400-CF-402B-COM	VX046475.D	03 Jun 2025 16:06	vial A pH#5.0	JC/MD	Ok,M
18	Q2173-12	OR-400-CF-402B-COM	VX046476.D	03 Jun 2025 16:30	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060325

Review By	John Carlone	Review On	6/4/2025 10:07:03 AM
Supervise By	Maresh Dadoda	Supervise On	6/4/2025 4:22:37 PM
SubDirectory	VX060325	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134100		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134101,VP134102		

19	Q2173-18	OR-400-CF-402B-COM	VX046477.D	03 Jun 2025 16:53	vial A pH#5.0	JC/MD	Ok
20	Q2189-01	MW-7-20250602	VX046478.D	03 Jun 2025 17:17	vial A pH<2 CCAL,BS and BSD failed high	JC/MD	ReRun
21	Q2189-02	MW-8-20250602	VX046479.D	03 Jun 2025 17:40	vial A pH<2	JC/MD	Ok
22	Q2172-04	TP06-MHQ	VX046480.D	03 Jun 2025 18:04	vial A pH#5.0	JC/MD	Ok
23	Q2168-04	A3	VX046481.D	03 Jun 2025 18:28	vial A pH#5.0	JC/MD	Ok,M
24	Q2168-08	B3	VX046482.D	03 Jun 2025 18:51	vial A pH#5.0 Need 5X	JC/MD	Dilution
25	Q2168-12	C2	VX046483.D	03 Jun 2025 19:15	vial A pH#5.0 Need 10X	JC/MD	Dilution
26	Q2169-01	303-PPR-1	VX046484.D	03 Jun 2025 19:38	vial A pH<2 E flag in previous sample	JC/MD	ReRun
27	Q2169-03	303-PPR-2	VX046485.D	03 Jun 2025 20:02	vial A pH<2 Confirm concentration; Need 5X	JC/MD	Dilution
28	VSTDCCC050	VSTDCCC050EC	VX046486.D	03 Jun 2025 20:25		JC/MD	Ok,M

M : Manual Integration