

Cover Page

Order ID : Q2186

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q2186-01
Q2186-02
Q2186-03
Q2186-06
Q2186-07
Q2186-08
Q2186-09

Client Sample Number

BP-VPB-182-TB-20250530
BP-VPB-182-GW-670-672
BP-VPB-182-GW-680-682
BP-VPB-182-GW-705-707
BP-VPB-182-GW-720-722
BP-VPB-182-DUP-20250602
VPB182-HYD-20250530

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/5/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager# Ernie Wu

Order ID # Q2186

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

7 Water samples were received on 06/02/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data Summary forms (i.e. Form Is)."

The not QT review data is reported in the Miscellaneous.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2186

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

summary forms (i.e. Form Is).”

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2186

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/05/2025

LAB CHRONICLE

OrderID:	Q2186	OrderDate:	6/2/2025 4:04:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	N13,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2186-01	BP-VPB-182-TB-2025 0530	Water			05/30/25			06/02/25
			VOCMS Group1	8260-Low			06/03/25	
Q2186-02	BP-VPB-182-GW-670- 672	Water			05/30/25			06/02/25
			VOCMS Group1	8260-Low			06/03/25	
Q2186-03	BP-VPB-182-GW-680- 682	Water			05/30/25			06/02/25
			VOCMS Group1	8260-Low			06/03/25	
Q2186-06	BP-VPB-182-GW-705- 707	Water			06/02/25			06/02/25
			VOCMS Group1	8260-Low			06/03/25	
Q2186-07	BP-VPB-182-GW-720- 722	Water			06/02/25			06/02/25
			VOCMS Group1	8260-Low			06/03/25	
Q2186-08	BP-VPB-182-DUP-202 50602	Water			06/02/25			06/02/25
			VOCMS Group1	8260-Low			06/03/25	
Q2186-09	VPB182-HYD-202505 30	Water			05/30/25			06/02/25
			VOCMS Group1	8260-Low			06/03/25	

Hit Summary Sheet
SW-846

SDG No.: Q2186

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	BP-VPB-182-TB-20250530								
Q2186-01	BP-VPB-182-TB-20 Water	Acetone		1.70	J	1.50	3.80	5.00	ug/L
		Total Voc :		1.70					
		Total Concentration:		1.70					
Client ID:	BP-VPB-182-GW-670-672								
Q2186-02	BP-VPB-182-GW-6 Water	Acetone		15.2		1.50	3.80	5.00	ug/L
Q2186-02	BP-VPB-182-GW-6 Water	Carbon Disulfide		1.00		0.21	0.75	1.00	ug/L
		Total Voc :		16.2					
		Total Concentration:		16.2					
Client ID:	BP-VPB-182-GW-680-682								
Q2186-03	BP-VPB-182-GW-6 Water	Acetone		7.30		1.50	3.80	5.00	ug/L
		Total Voc :		7.30					
		Total Concentration:		7.30					
Client ID:	BP-VPB-182-GW-705-707								
Q2186-06	BP-VPB-182-GW-7 Water	Acetone		7.00		1.50	3.80	5.00	ug/L
		Total Voc :		7.00					
		Total Concentration:		7.00					
Client ID:	BP-VPB-182-GW-720-722								
Q2186-07	BP-VPB-182-GW-7 Water	Acetone		6.30		1.50	3.80	5.00	ug/L
		Total Voc :		6.30					
		Total Concentration:		6.30					
Client ID:	BP-VPB-182-DUP-20250602								
Q2186-08	BP-VPB-182-DUP- Water	Acetone		7.20		1.50	3.80	5.00	ug/L
		Total Voc :		7.20					
		Total Concentration:		7.20					
Client ID:	VPB182-HYD-20250530								
Q2186-09	VPB182-HYD-2025 Water	Acetone		2.50	J	1.50	3.80	5.00	ug/L
Q2186-09	VPB182-HYD-2025 Water	Dibromochloromethane		1.70		0.18	0.50	1.00	ug/L
		Total Voc :		4.20					
		Total Concentration:		4.20					



QC SUMMARY

Surrogate Summary

SDG No.: **Q2186**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2186-01	BP-VPB-182-TB-20250530	1,2-Dichloroethane-d4	50	51.0	102	81	118
		Dibromofluoromethane	50	48.6	97	80	119
		Toluene-d8	50	49.4	99	89	112
		4-Bromofluorobenzene	50	51.4	103	85	114
Q2186-02	BP-VPB-182-GW-670-672	1,2-Dichloroethane-d4	50	51.7	103	81	118
		Dibromofluoromethane	50	49.8	100	80	119
		Toluene-d8	50	49.8	100	89	112
		4-Bromofluorobenzene	50	51.0	102	85	114
Q2186-03	BP-VPB-182-GW-680-682	1,2-Dichloroethane-d4	50	51.3	103	81	118
		Dibromofluoromethane	50	50.0	100	80	119
		Toluene-d8	50	49.9	100	89	112
		4-Bromofluorobenzene	50	51.8	104	85	114
Q2186-06	BP-VPB-182-GW-705-707	1,2-Dichloroethane-d4	50	52.1	104	81	118
		Dibromofluoromethane	50	50.0	100	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	52.1	104	85	114
Q2186-07	BP-VPB-182-GW-720-722	1,2-Dichloroethane-d4	50	52.2	104	81	118
		Dibromofluoromethane	50	48.7	97	80	119
		Toluene-d8	50	49.1	98	89	112
		4-Bromofluorobenzene	50	49.7	99	85	114
Q2186-08	BP-VPB-182-DUP-20250602	1,2-Dichloroethane-d4	50	52.5	105	81	118
		Dibromofluoromethane	50	49.0	98	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	51.0	102	85	114
Q2186-09	VPB182-HYD-20250530	1,2-Dichloroethane-d4	50	51.2	102	81	118
		Dibromofluoromethane	50	49.3	98	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	53.0	106	85	114
VX0603WBL01	VX0603WBL01	1,2-Dichloroethane-d4	50	51.5	103	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	49.8	100	89	112
		4-Bromofluorobenzene	50	51.4	103	85	114
VX0603WBS01	VX0603WBS01	1,2-Dichloroethane-d4	50	48.8	98	81	118
		Dibromofluoromethane	50	50.4	101	80	119
		Toluene-d8	50	47.8	96	89	112
		4-Bromofluorobenzene	50	50.6	101	85	114
VX0603WBSD0	VX0603WBSD01	1,2-Dichloroethane-d4	50	50.4	101	81	118
		Dibromofluoromethane	50	51.6	103	80	119
		Toluene-d8	50	48.2	96	89	112
		4-Bromofluorobenzene	50	50.5	101	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2186

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046463.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0603WBS01	Chloromethane	20	17.0	ug/L	85			50	139	
	Vinyl chloride	20	16.6	ug/L	83			58	137	
	Bromomethane	20	16.9	ug/L	85			53	141	
	Chloroethane	20	19.1	ug/L	96			60	138	
	Trichlorofluoromethane	20	18.8	ug/L	94			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.7	ug/L	94			70	136	
	1,1-Dichloroethene	20	17.9	ug/L	90			71	131	
	Acetone	100	100	ug/L	100			39	160	
	Carbon disulfide	20	13.3	ug/L	67			64	133	
	Methyl tert-butyl Ether	20	20.3	ug/L	102			71	124	
	Methylene Chloride	20	18.4	ug/L	92			74	124	
	trans-1,2-Dichloroethene	20	18.0	ug/L	90			75	124	
	1,1-Dichloroethane	20	20.5	ug/L	103			77	125	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	18.5	ug/L	93			72	136	
	cis-1,2-Dichloroethene	20	20.4	ug/L	102			78	123	
	Chloroform	20	20.9	ug/L	104			79	124	
	1,1,1-Trichloroethane	20	20.1	ug/L	101			74	131	
	Methylcyclohexane	20	17.7	ug/L	89			72	132	
	Benzene	20	19.5	ug/L	98			79	120	
	1,2-Dichloroethane	20	20.2	ug/L	101			73	128	
	Trichloroethene	20	18.9	ug/L	95			79	123	
	1,2-Dichloropropane	20	20.9	ug/L	104			78	122	
	Bromodichloromethane	20	20.0	ug/L	100			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	19.9	ug/L	100			80	121	
	t-1,3-Dichloropropene	20	19.2	ug/L	96			73	127	
	cis-1,3-Dichloropropene	20	19.9	ug/L	100			75	124	
	1,1,2-Trichloroethane	20	21.5	ug/L	108			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	20.5	ug/L	103			74	126	
	Tetrachloroethene	20	19.8	ug/L	99			74	129	
	Chlorobenzene	20	20.2	ug/L	101			82	118	
	Ethyl Benzene	20	20.4	ug/L	102			79	121	
	m/p-Xylenes	40	40.4	ug/L	101			80	121	
	o-Xylene	20	21.0	ug/L	105			78	122	
	Styrene	20	21.2	ug/L	106			78	123	
	Bromoform	20	19.9	ug/L	100			66	130	
	Isopropylbenzene	20	21.1	ug/L	106			72	131	
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107			71	121	
	1,3-Dichlorobenzene	20	20.1	ug/L	101			80	119	
	1,4-Dichlorobenzene	20	19.6	ug/L	98			79	118	
	1,2-Dichlorobenzene	20	21.3	ug/L	106			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2186

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046464.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0603WBSD01	Chloromethane	20	16.4	ug/L	82	4		50	139	20
	Vinyl chloride	20	16.0	ug/L	80	4		58	137	20
	Bromomethane	20	16.8	ug/L	84	1		53	141	20
	Chloroethane	20	17.8	ug/L	89	8		60	138	20
	Trichlorofluoromethane	20	18.1	ug/L	91	3		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.1	ug/L	96	2		70	136	20
	1,1-Dichloroethene	20	17.8	ug/L	89	1		71	131	20
	Acetone	100	110	ug/L	110	10		39	160	20
	Carbon disulfide	20	13.3	ug/L	67	0		64	133	20
	Methyl tert-butyl Ether	20	21.5	ug/L	108	6		71	124	20
	Methylene Chloride	20	18.3	ug/L	92	0		74	124	20
	trans-1,2-Dichloroethene	20	18.0	ug/L	90	0		75	124	20
	1,1-Dichloroethane	20	20.1	ug/L	101	2		77	125	20
	2-Butanone	100	110	ug/L	110	0		56	143	20
	Carbon Tetrachloride	20	18.6	ug/L	93	0		72	136	20
	cis-1,2-Dichloroethene	20	20.1	ug/L	101	1		78	123	20
	Chloroform	20	20.7	ug/L	104	0		79	124	20
	1,1,1-Trichloroethane	20	19.1	ug/L	96	5		74	131	20
	Methylcyclohexane	20	17.5	ug/L	88	1		72	132	20
	Benzene	20	19.4	ug/L	97	1		79	120	20
	1,2-Dichloroethane	20	20.9	ug/L	104	3		73	128	20
	Trichloroethene	20	19.2	ug/L	96	1		79	123	20
	1,2-Dichloropropane	20	21.1	ug/L	106	2		78	122	20
	Bromodichloromethane	20	20.9	ug/L	104	4		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		67	130	20
	Toluene	20	19.7	ug/L	99	1		80	121	20
	t-1,3-Dichloropropene	20	19.7	ug/L	99	3		73	127	20
	cis-1,3-Dichloropropene	20	20.5	ug/L	103	3		75	124	20
	1,1,2-Trichloroethane	20	22.2	ug/L	111	3		80	119	20
	2-Hexanone	100	110	ug/L	110	0		57	139	20
	Dibromochloromethane	20	21.0	ug/L	105	2		74	126	20
	Tetrachloroethene	20	19.4	ug/L	97	2		74	129	20
	Chlorobenzene	20	20.0	ug/L	100	1		82	118	20
	Ethyl Benzene	20	20.4	ug/L	102	0		79	121	20
	m/p-Xylenes	40	40.2	ug/L	101	0		80	121	20
	o-Xylene	20	21.4	ug/L	107	2		78	122	20
	Styrene	20	21.2	ug/L	106	0		78	123	20
	Bromoform	20	20.6	ug/L	103	3		66	130	20
	Isopropylbenzene	20	21.0	ug/L	105	1		72	131	20
	1,1,2,2-Tetrachloroethane	20	21.3	ug/L	106	1		71	121	20
	1,3-Dichlorobenzene	20	20.2	ug/L	101	0		80	119	20
	1,4-Dichlorobenzene	20	20.2	ug/L	101	3		79	118	20
	1,2-Dichlorobenzene	20	21.2	ug/L	106	0		80	119	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0603WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q2186

SAS No.: Q2186 SDG NO.: Q2186

Lab File ID: VX046462.D

Lab Sample ID: VX0603WBL01

Date Analyzed: 06/03/2025

Time Analyzed: 11:01

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0603WBS01	VX0603WBS01	VX046463.D	06/03/2025
VX0603WBSD01	VX0603WBSD01	VX046464.D	06/03/2025
VPB182-HYD-20250530	Q2186-09	VX046465.D	06/03/2025
BP-VPB-182-TB-20250530	Q2186-01	VX046466.D	06/03/2025
BP-VPB-182-GW-670-672	Q2186-02	VX046467.D	06/03/2025
BP-VPB-182-GW-680-682	Q2186-03	VX046468.D	06/03/2025
BP-VPB-182-GW-705-707	Q2186-06	VX046469.D	06/03/2025
BP-VPB-182-GW-720-722	Q2186-07	VX046470.D	06/03/2025
BP-VPB-182-DUP-20250602	Q2186-08	VX046471.D	06/03/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG NO.: Q2186
Lab File ID: VX046038.D BFB Injection Date: 05/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG NO.: Q2186
Lab File ID: VX046459.D BFB Injection Date: 06/03/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:40
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	56.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1.1) 1
174	50.0 - 100.0% of mass 95	69.6
175	5.0 - 9.0% of mass 174	5.1 (7.3) 1
176	95.0 - 101.0% of mass 174	66.3 (95.2) 1
177	5.0 - 9.0% of mass 176	4.1 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046460.D	06/03/2025	10:10
VX0603WBL01	VX0603WBL01	VX046462.D	06/03/2025	11:01
VX0603WBS01	VX0603WBS01	VX046463.D	06/03/2025	11:24
VX0603WBSD01	VX0603WBSD01	VX046464.D	06/03/2025	11:50
VPB182-HYD-20250530	Q2186-09	VX046465.D	06/03/2025	12:13
BP-VPB-182-TB-20250530	Q2186-01	VX046466.D	06/03/2025	12:36
BP-VPB-182-GW-670-672	Q2186-02	VX046467.D	06/03/2025	12:59
BP-VPB-182-GW-680-682	Q2186-03	VX046468.D	06/03/2025	13:23
BP-VPB-182-GW-705-707	Q2186-06	VX046469.D	06/03/2025	13:46
BP-VPB-182-GW-720-722	Q2186-07	VX046470.D	06/03/2025	14:09
BP-VPB-182-DUP-20250602	Q2186-08	VX046471.D	06/03/2025	14:33
VSTDCCC050EC	VSTDCCC050	VX046486.D	06/03/2025	20:25

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG NO.: Q2186
Lab File ID: VX046460.D Date Analyzed: 06/03/2025
Instrument ID: MSVOA_X Time Analyzed: 10:10
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97221	5.54	166757	6.75	142113	10.05
UPPER LIMIT	194442	6.038	333514	7.251	284226	10.549
LOWER LIMIT	48610.5	5.038	83378.5	6.251	71056.5	9.549
EPA SAMPLE NO.						
BP-VPB-182-TB-20250530	73085	5.54	148251	6.76	138870	10.06
BP-VPB-182-GW-670-672	70404	5.54	140445	6.76	131337	10.05
BP-VPB-182-GW-680-682	70004	5.55	138414	6.76	130629	10.05
BP-VPB-182-GW-705-707	68425	5.54	137561	6.76	128221	10.05
BP-VPB-182-GW-720-722	65938	5.55	133475	6.76	122673	10.06
BP-VPB-182-DUP-20250602	68690	5.55	138832	6.76	129871	10.05
VPB182-HYD-20250530	72286	5.54	144078	6.76	137138	10.05
VX0603WBL01	73151	5.55	147671	6.76	138701	10.05
VX0603WBS01	89858	5.54	159016	6.76	137079	10.05
VX0603WBSD01	86032	5.54	150958	6.76	130475	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG NO.: Q2186
Lab File ID: VX046460.D Date Analyzed: 06/03/2025
Instrument ID: MSVOA_X Time Analyzed: 10:10
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	69603	12.018				
UPPER LIMIT	139206	12.518				
LOWER LIMIT	34801.5	11.518				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250530	59352	12.02				
BP-VPB-182-GW-670-672	57166	12.02				
BP-VPB-182-GW-680-682	57376	12.02				
BP-VPB-182-GW-705-707	56111	12.02				
BP-VPB-182-GW-720-722	50694	12.02				
BP-VPB-182-DUP-20250602	55355	12.02				
VPB182-HYD-20250530	61107	12.02				
VX0603WBL01	60850	12.02				
VX0603WBS01	65270	12.02				
VX0603WBSD01	62365	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/30/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	06/02/25	
Client Sample ID:	BP-VPB-182-TB-20250530			SDG No.:	Q2186	
Lab Sample ID:	Q2186-01			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046466.D	1		06/03/25 12:36	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.70	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	06/02/25	
Client Sample ID:	BP-VPB-182-TB-20250530		SDG No.:	Q2186	
Lab Sample ID:	Q2186-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046466.D	1		06/03/25 12:36	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.0		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.6		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	49.4		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.4		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	73100	5.543				
540-36-3	1,4-Difluorobenzene	148000	6.757				
3114-55-4	Chlorobenzene-d5	139000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	59400	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/30/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/02/25
Client Sample ID:	BP-VPB-182-TB-20250530	SDG No.:	Q2186
Lab Sample ID:	Q2186-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046466.D	1		06/03/25 12:36	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046466.D
 Acq On : 03 Jun 2025 12:36
 Operator : JC/MD
 Sample : Q2186-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-TB-20250530

Quant Time: Jun 04 01:25:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.543	168	73085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	148251	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59352	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69528	51.028	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.060%
35) Dibromofluoromethane	5.379	113	51892	48.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.220%
50) Toluene-d8	8.647	98	182654	49.433	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.079	95	72809	51.370	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.740%
Target Compounds						Qvalue
16) Acetone	2.386	43	941	1.722	ug/l	# 79


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046466.D
Acq On : 03 Jun 2025 12:36
Operator : JC/MD
Sample : Q2186-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-TB-20250530


Quant Time: Jun 04 01:25:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 731
Delta R.T. -0.001 min
Lab File: VX046466.D
Acq: 03 Jun 2025 12:36

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-TB-20250530







Tgt Ion:168 Resp: 73085
Ion Ratio Lower Upper
168 100
99 70.1 54.9 82.3



#16
Acetone
Concen: 1.722 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046466.D
Acq: 03 Jun 2025 12:36

Tgt Ion: 43 Resp: 941
Ion Ratio Lower Upper
43 100
58 15.5 21.2 31.8#





#33

1,2-Dichloroethane-d4

Concen: 51.028 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046466.D

Acq: 03 Jun 2025 12:36

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-TB-20250530

Tgt Ion: 65 Resp: 69528

Ion Ratio Lower Upper

65 100

67 49.8 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046466.D

Acq: 03 Jun 2025 12:36




Tgt Ion: 114 Resp: 148251


Ion Ratio Lower Upper

114 100

63 25.9 0.0 49.2

88 17.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 48.608 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046466.D


Acq: 03 Jun 2025 12:36

Instrument :

MSVOA_X




ClientSampleId :

BP-VPB-182-TB-20250530



Tgt Ion:113 Resp: 51892

Ion	Ratio	Lower	Upper
113	100		
111	101.8	83.1	124.7
192	16.5	13.3	19.9



#50

Toluene-d8

Concen: 49.433 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. -0.000 min


Lab File: VX046466.D

Acq: 03 Jun 2025 12:36

Tgt Ion: 98 Resp: 182654

Ion	Ratio	Lower	Upper
98	100		
100	65.3	53.5	80.3





#62

4-Bromofluorobenzene

Concen: 51.370 ug/l

RT: 11.079 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046466.D

Acq: 03 Jun 2025 12:36

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-TB-20250530





Tgt Ion: 95 Resp: 72809

Ion Ratio Lower Upper

95 100

174 67.0 0.0 135.8

176 64.6 0.0 131.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.006 min

Lab File: VX046466.D

Acq: 03 Jun 2025 12:36




Tgt Ion: 117 Resp: 138870


Ion Ratio Lower Upper

117 100

82 59.8 50.6 76.0

119 32.5 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046466.D


Acq: 03 Jun 2025 12:36

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-TB-20250530





Tgt Ion:152 Resp: 59352

Ion Ratio Lower Upper

152 100

115 65.1 46.9 140.7

150 162.2 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046466.D
Acq On : 03 Jun 2025 12:36
Operator : JC/MD
Sample : Q2186-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-TB-20250530

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046466.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.569	71	79	80	rBV3	4264	7819	1.53%	0.286%
2	5.379	692	704	722	rBV	58109	174571	34.09%	6.389%
3	5.550	722	732	749	rVB2	85972	245881	48.01%	8.999%
4	5.952	786	798	813	rBV	69678	183251	35.78%	6.707%
5	6.757	919	930	944	rBV	155054	372838	72.80%	13.646%
6	8.647	1234	1240	1255	rBV	324097	512109	100.00%	18.743%
7	10.049	1465	1470	1484	rBV	331015	471421	92.05%	17.254%
8	11.079	1634	1639	1654	rBV	275382	356596	69.63%	13.052%
9	12.018	1788	1793	1809	rBV	330313	407729	79.62%	14.923%




Sum of corrected areas: 2732215

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046466.D
Acq On : 03 Jun 2025 12:36
Operator : JC/MD
Sample : Q2186-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-TB-20250530

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046466.D
Acq On : 03 Jun 2025 12:36
Operator : JC/MD
Sample : Q2186-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-TB-20250530

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046466.D
Acq On : 03 Jun 2025 12:36
Operator : JC/MD
Sample : Q2186-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-TB-20250530

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/30/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	06/02/25		
Client Sample ID:	BP-VPB-182-GW-670-672			SDG No.:	Q2186		
Lab Sample ID:	Q2186-02			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046467.D	1		06/03/25 12:59	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	15.2		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	1.00		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	06/02/25	
Client Sample ID:	BP-VPB-182-GW-670-672		SDG No.:	Q2186	
Lab Sample ID:	Q2186-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046467.D	1		06/03/25 12:59	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.7		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.8		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.0		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70400	5.544				
540-36-3	1,4-Difluorobenzene	140000	6.757				
3114-55-4	Chlorobenzene-d5	131000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	57200	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/30/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/02/25
Client Sample ID:	BP-VPB-182-GW-670-672	SDG No.:	Q2186
Lab Sample ID:	Q2186-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046467.D	1		06/03/25 12:59	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046467.D
 Acq On : 03 Jun 2025 12:59
 Operator : JC/MD
 Sample : Q2186-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-670-672

Quant Time: Jun 04 01:26:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	70404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	140445	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	131337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67836	51.682	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.360%			
35) Dibromofluoromethane	5.385	113	50330	49.765	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.540%			
50) Toluene-d8	8.647	98	174190	49.763	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.520%			
62) 4-Bromofluorobenzene	11.079	95	68437	50.969	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.940%			
Target Compounds						
16) Acetone	2.380	43	7996	15.194	ug/l	98
17) Carbon Disulfide	2.508	76	2010	1.016	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046467.D
Acq On : 03 Jun 2025 12:59
Operator : JC/MD
Sample : Q2186-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-670-672

Quant Time: Jun 04 01:26:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046467.D

Acq: 03 Jun 2025 12:59

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-GW-670-672

Tgt Ion:168 Resp: 70404

Ion Ratio Lower Upper

168 100

99 72.2 54.9 82.3



#16

Acetone

Concen: 15.194 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046467.D




Acq: 03 Jun 2025 12:59


Tgt Ion: 43 Resp: 7996

Ion Ratio Lower Upper

43 100

58 25.6 21.2 31.8





#17

Carbon Disulfide

Concen: 1.016 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046467.D

Acq: 03 Jun 2025 12:59

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-GW-670-672

Tgt Ion: 76 Resp: 2010

Ion Ratio Lower Upper

76 100

78 9.2 7.0 10.4



#33

1,2-Dichloroethane-d4

Concen: 51.682 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046467.D




Acq: 03 Jun 2025 12:59


Tgt Ion: 65 Resp: 67836

Ion Ratio Lower Upper

65 100

67 50.0 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. 0.000 min
Lab File: VX046467.D
Acq: 03 Jun 2025 12:59




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-670-672


Tgt Ion	Ratio	Lower	Upper
114	100		
63	24.3	0.0	49.2
88	17.1	0.0	33.6



#35
Dibromofluoromethane
Concen: 49.765 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX046467.D
Acq: 03 Jun 2025 12:59

Tgt Ion	Ratio	Lower	Upper
113	100		
111	102.0	83.1	124.7
192	16.3	13.3	19.9





#50

Toluene-d8

Concen: 49.763 ug/l

RT: 8.647 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046467.D

Acq: 03 Jun 2025 12:59

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-GW-670-672

Tgt Ion: 98 Resp: 174190

Ion Ratio Lower Upper

98 100

100 64.9 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 50.969 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX046467.D

Acq: 03 Jun 2025 12:59




Tgt Ion: 95 Resp: 68437


Ion Ratio Lower Upper

95 100

174 67.6 0.0 135.8

176 64.6 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX046467.D

Acq: 03 Jun 2025 12:59

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-670-672





Tgt Ion:117 Resp: 131337

Ion Ratio Lower Upper

117 100

82 61.2 50.6 76.0

119 31.9 25.8 38.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046467.D

Acq: 03 Jun 2025 12:59




Tgt Ion:152 Resp: 57166

Ion Ratio Lower Upper

152 100

115 65.5 46.9 140.7

150 156.1 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046467.D
 Acq On : 03 Jun 2025 12:59
 Operator : JC/MD
 Sample : Q2186-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-670-672

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046467.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.142	6	9	19	rVB	17305	18144	3.72%	0.685%
2	1.368	41	46	50	rBV2	6130	7301	1.50%	0.275%
3	1.453	56	60	67	rVV	8564	10296	2.11%	0.388%
4	1.569	71	79	83	rVB4	3603	7852	1.61%	0.296%
5	2.380	207	212	221	rVB	7418	12439	2.55%	0.469%
6	5.385	693	705	721	rBV	57066	169362	34.69%	6.390%
7	5.544	721	731	748	rVB	81458	233599	47.84%	8.813%
8	5.952	788	798	812	rBV	65759	177536	36.36%	6.698%
9	6.757	922	930	947	rBV	147896	355933	72.90%	13.429%
10	8.647	1234	1240	1253	rBV	314252	488267	100.00%	18.421%
11	10.049	1465	1470	1493	rBV	322861	447236	91.60%	16.873%
12	11.079	1634	1639	1654	rBV	260893	336247	68.87%	12.686%
13	12.018	1788	1793	1804	rBV	307717	386347	79.13%	14.576%




Sum of corrected areas: 2650559

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046467.D
Acq On : 03 Jun 2025 12:59
Operator : JC/MD
Sample : Q2186-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-670-672

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046467.D
Acq On : 03 Jun 2025 12:59
Operator : JC/MD
Sample : Q2186-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-670-672

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046467.D
Acq On : 03 Jun 2025 12:59
Operator : JC/MD
Sample : Q2186-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-670-672

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/30/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	06/02/25		
Client Sample ID:	BP-VPB-182-GW-680-682			SDG No.:	Q2186		
Lab Sample ID:	Q2186-03			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046468.D	1		06/03/25 13:23	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	7.30		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	06/02/25	
Client Sample ID:	BP-VPB-182-GW-680-682		SDG No.:	Q2186	
Lab Sample ID:	Q2186-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046468.D	1		06/03/25 13:23	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.3		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	50.0		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.9		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.8		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70000	5.55				
540-36-3	1,4-Difluorobenzene	138000	6.757				
3114-55-4	Chlorobenzene-d5	131000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	57400	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/30/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/02/25
Client Sample ID:	BP-VPB-182-GW-680-682	SDG No.:	Q2186
Lab Sample ID:	Q2186-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046468.D	1		06/03/25 13:23	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046468.D
 Acq On : 03 Jun 2025 13:23
 Operator : JC/MD
 Sample : Q2186-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-680-682

Quant Time: Jun 04 01:26:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.550	168	70004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138414	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66991	51.330	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.660%	
35) Dibromofluoromethane	5.385	113	49830	49.994	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.980%	
50) Toluene-d8	8.647	98	172065	49.877	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.760%	
62) 4-Bromofluorobenzene	11.079	95	68507	51.770	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.540%	
Target Compounds						
16) Acetone	2.380	43	3841	7.340	ug/l	Qvalue 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046468.D
Acq On : 03 Jun 2025 13:23
Operator : JC/MD
Sample : Q2186-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-680-682

Quant Time: Jun 04 01:26:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046468.D
Acq: 03 Jun 2025 13:23




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-680-682


Tgt Ion:168 Resp: 70004
Ion Ratio Lower Upper
168 100
99 69.5 54.9 82.3



#16
Acetone
Concen: 7.340 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046468.D
Acq: 03 Jun 2025 13:23

Tgt Ion: 43 Resp: 3841
Ion Ratio Lower Upper
43 100
58 30.5 21.2 31.8





#33

1,2-Dichloroethane-d4

Concen: 51.330 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046468.D

Acq: 03 Jun 2025 13:23

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-GW-680-682

Tgt Ion: 65 Resp: 66991

Ion Ratio Lower Upper

65 100

67 49.7 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046468.D

Acq: 03 Jun 2025 13:23




Tgt Ion: 114 Resp: 138414


Ion Ratio Lower Upper

114 100

63 24.9 0.0 49.2

88 17.9 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.994 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046468.D


Acq: 03 Jun 2025 13:23

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-680-682






Tgt Ion:113 Resp: 49830

Ion Ratio Lower Upper

113 100

111 100.5 83.1 124.7

192 16.3 13.3 19.9



#50

Toluene-d8

Concen: 49.877 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046468.D




Acq: 03 Jun 2025 13:23


Tgt Ion: 98 Resp: 172065

Ion Ratio Lower Upper

98 100

100 65.1 53.5 80.3





#62

4-Bromofluorobenzene

Concen: 51.770 ug/l

RT: 11.079 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046468.D

Acq: 03 Jun 2025 13:23

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-680-682





Tgt Ion: 95 Resp: 68507

Ion Ratio Lower Upper

95 100

174 67.1 0.0 135.8

176 65.2 0.0 131.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX046468.D

Acq: 03 Jun 2025 13:23




Tgt Ion: 117 Resp: 130629


Ion Ratio Lower Upper

117 100

82 62.3 50.6 76.0

119 31.7 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046468.D


Acq: 03 Jun 2025 13:23

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-680-682





Tgt Ion:152 Resp: 57376

Ion Ratio Lower Upper

152 100

115 64.2 46.9 140.7

150 158.6 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046468.D
 Acq On : 03 Jun 2025 13:23
 Operator : JC/MD
 Sample : Q2186-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-680-682

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046468.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.142	6	9	17	rVB	10484	10412	2.17%	0.399%
2	1.368	40	46	51	rVB3	3885	5552	1.15%	0.213%
3	1.447	56	59	65	rVB	4120	4986	1.04%	0.191%
4	1.575	69	80	84	rBV2	4382	10761	2.24%	0.412%
5	2.380	207	212	220	rVB	3889	6243	1.30%	0.239%
6	3.440	378	386	393	rVB3	2354	4876	1.01%	0.187%
7	5.385	694	705	722	rBV2	56104	167880	34.92%	6.434%
8	5.544	722	731	743	rBV2	80048	228622	47.56%	8.762%
9	5.952	789	798	812	rBV	63762	173526	36.10%	6.650%
10	6.757	921	930	948	rBV	144430	350440	72.90%	13.431%
11	8.647	1234	1240	1255	rBV	307003	480693	100.00%	18.423%
12	10.049	1465	1470	1483	rBV	311110	441786	91.91%	16.931%
13	11.079	1634	1639	1651	rBV	266988	333555	69.39%	12.784%
14	12.018	1788	1793	1807	rBV	313976	389923	81.12%	14.944%




Sum of corrected areas: 2609255

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046468.D
Acq On : 03 Jun 2025 13:23
Operator : JC/MD
Sample : Q2186-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-680-682

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046468.D
Acq On : 03 Jun 2025 13:23
Operator : JC/MD
Sample : Q2186-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-680-682

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046468.D
Acq On : 03 Jun 2025 13:23
Operator : JC/MD
Sample : Q2186-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-680-682

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	06/02/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	06/02/25		
Client Sample ID:	BP-VPB-182-GW-705-707			SDG No.:	Q2186		
Lab Sample ID:	Q2186-06			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046469.D	1		06/03/25 13:46	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	7.00		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	06/02/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	06/02/25	
Client Sample ID:	BP-VPB-182-GW-705-707		SDG No.:	Q2186	
Lab Sample ID:	Q2186-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046469.D	1		06/03/25 13:46	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.1		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	50.0		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.1		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	68400	5.544				
540-36-3	1,4-Difluorobenzene	138000	6.757				
3114-55-4	Chlorobenzene-d5	128000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	56100	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	06/02/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/02/25
Client Sample ID:	BP-VPB-182-GW-705-707	SDG No.:	Q2186
Lab Sample ID:	Q2186-06	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046469.D	1		06/03/25 13:46	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046469.D
 Acq On : 03 Jun 2025 13:46
 Operator : JC/MD
 Sample : Q2186-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-705-707

Quant Time: Jun 04 01:27:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	68425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137561	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56111	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66512	52.139	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.280%	
35) Dibromofluoromethane	5.385	113	49566	50.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.080%	
50) Toluene-d8	8.647	98	172025	50.174	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	11.079	95	68551	52.124	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.240%	
Target Compounds						
16) Acetone	2.380	43	3563	6.966	ug/l	Qvalue # 87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046469.D
Acq On : 03 Jun 2025 13:46
Operator : JC/MD
Sample : Q2186-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-705-707

Quant Time: Jun 04 01:27:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046469.D
Acq: 03 Jun 2025 13:46




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-705-707


Tgt Ion:168 Resp: 68425
Ion Ratio Lower Upper
168 100
99 72.3 54.9 82.3



#16
Acetone
Concen: 6.966 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046469.D
Acq: 03 Jun 2025 13:46

Tgt Ion: 43 Resp: 3563
Ion Ratio Lower Upper
43 100
58 33.2 21.2 31.8#





#33

1,2-Dichloroethane-d4

Concen: 52.139 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046469.D

Acq: 03 Jun 2025 13:46

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-GW-705-707

Tgt Ion: 65 Resp: 66512

Ion Ratio Lower Upper

65 100

67 49.6 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046469.D

Acq: 03 Jun 2025 13:46




Tgt Ion: 114 Resp: 137561


Ion Ratio Lower Upper

114 100

63 25.9 0.0 49.2

88 17.5 0.0 33.6





#35

Dibromofluoromethane

Concen: 50.037 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046469.D


Acq: 03 Jun 2025 13:46

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-705-707






Tgt Ion:113 Resp: 49566

Ion Ratio Lower Upper

113 100

111 102.4 83.1 124.7

192 16.8 13.3 19.9



#50

Toluene-d8

Concen: 50.174 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046469.D




Acq: 03 Jun 2025 13:46


Tgt Ion: 98 Resp: 172025

Ion Ratio Lower Upper

98 100

100 64.3 53.5 80.3





#62

4-Bromofluorobenzene

Concen: 52.124 ug/l

RT: 11.079 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046469.D

Acq: 03 Jun 2025 13:46

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-705-707





Tgt Ion: 95 Resp: 68551

Ion Ratio Lower Upper

95 100

174 66.5 0.0 135.8

176 62.2 0.0 131.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX046469.D

Acq: 03 Jun 2025 13:46




Tgt Ion: 117 Resp: 128221


Ion Ratio Lower Upper

117 100

82 64.0 50.6 76.0

119 32.3 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046469.D


Acq: 03 Jun 2025 13:46

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-705-707





Tgt Ion:152 Resp: 56111

Ion Ratio Lower Upper

152 100

115 65.0 46.9 140.7

150 153.8 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046469.D
 Acq On : 03 Jun 2025 13:46
 Operator : JC/MD
 Sample : Q2186-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-705-707

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046469.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.142	6	9	15	rVB	7159	6707	1.40%	0.260%
2	1.453	56	60	66	rVB	3800	4889	1.02%	0.190%
3	1.581	70	81	84	rBV4	4430	10478	2.19%	0.407%
4	2.380	208	212	218	rBV	3652	5552	1.16%	0.216%
5	3.447	377	387	396	rBV3	2223	5134	1.07%	0.199%
6	5.385	695	705	721	rBV	55818	165629	34.55%	6.432%
7	5.544	721	731	748	rVB	80007	228619	47.68%	8.878%
8	5.952	788	798	814	rBV2	64734	176123	36.73%	6.839%
9	6.757	920	930	944	rBV	142177	348439	72.67%	13.531%
10	8.647	1233	1240	1251	rBV	302309	479452	100.00%	18.618%
11	10.049	1465	1470	1479	rBV	311262	435455	90.82%	16.910%
12	11.079	1634	1639	1654	rBV	258261	329746	68.78%	12.805%
13	12.018	1788	1793	1809	rBV	302194	378952	79.04%	14.716%




Sum of corrected areas: 2575175

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046469.D
Acq On : 03 Jun 2025 13:46
Operator : JC/MD
Sample : Q2186-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-705-707

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046469.D
Acq On : 03 Jun 2025 13:46
Operator : JC/MD
Sample : Q2186-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-705-707

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046469.D
Acq On : 03 Jun 2025 13:46
Operator : JC/MD
Sample : Q2186-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-705-707

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	06/02/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	06/02/25		
Client Sample ID:	BP-VPB-182-GW-720-722			SDG No.:	Q2186		
Lab Sample ID:	Q2186-07			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046470.D	1		06/03/25 14:09	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	6.30		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	06/02/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	06/02/25	
Client Sample ID:	BP-VPB-182-GW-720-722		SDG No.:	Q2186	
Lab Sample ID:	Q2186-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046470.D	1		06/03/25 14:09	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.2		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	48.7		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	49.1		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.7		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	65900	5.55				
540-36-3	1,4-Difluorobenzene	133000	6.757				
3114-55-4	Chlorobenzene-d5	123000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	50700	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	06/02/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/02/25
Client Sample ID:	BP-VPB-182-GW-720-722	SDG No.:	Q2186
Lab Sample ID:	Q2186-07	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046470.D	1		06/03/25 14:09	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046470.D
 Acq On : 03 Jun 2025 14:09
 Operator : JC/MD
 Sample : Q2186-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-720-722

Quant Time: Jun 04 01:27:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.550	168	65938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133475	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50694	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64202	52.227	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.460%
35) Dibromofluoromethane	5.379	113	46770	48.660	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.320%
50) Toluene-d8	8.647	98	163222	49.064	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.120%
62) 4-Bromofluorobenzene	11.079	95	63368	49.658	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.320%
Target Compounds						
16) Acetone	2.380	43	3123	6.336	ug/l	Qvalue 96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046470.D
Acq On : 03 Jun 2025 14:09
Operator : JC/MD
Sample : Q2186-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-720-722

Quant Time: Jun 04 01:27:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.550 min Scan# 71

Delta R.T. 0.006 min

Lab File: VX046470.D

Acq: 03 Jun 2025 14:09

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-GW-720-722

Tgt Ion:168 Resp: 65938

Ion Ratio Lower Upper

168 100

99 68.1 54.9 82.3



#16

Acetone

Concen: 6.336 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046470.D




Acq: 03 Jun 2025 14:09


Tgt Ion: 43 Resp: 3123

Ion Ratio Lower Upper

43 100

58 28.4 21.2 31.8





#33

1,2-Dichloroethane-d4

Concen: 52.227 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046470.D

Acq: 03 Jun 2025 14:09

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-GW-720-722

Tgt Ion: 65 Resp: 64202

Ion Ratio Lower Upper

65 100

67 49.5 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046470.D

Acq: 03 Jun 2025 14:09




Tgt Ion: 114 Resp: 133475


Ion Ratio Lower Upper

114 100

63 22.8 0.0 49.2

88 16.9 0.0 33.6





#35

Dibromofluoromethane

Concen: 48.660 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046470.D


Acq: 03 Jun 2025 14:09

Instrument :

MSVOA_X




ClientSampleId :

BP-VPB-182-GW-720-722



Tgt Ion:113 Resp: 46770

Ion	Ratio	Lower	Upper
113	100		
111	105.2	83.1	124.7
192	17.5	13.3	19.9



#50

Toluene-d8

Concen: 49.064 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. 0.000 min


Lab File: VX046470.D

Acq: 03 Jun 2025 14:09

Tgt Ion: 98 Resp: 163222

Ion	Ratio	Lower	Upper
98	100		
100	65.9	53.5	80.3





#62

4-Bromofluorobenzene

Concen: 49.658 ug/l

RT: 11.079 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046470.D

Acq: 03 Jun 2025 14:09

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-720-722





Tgt Ion: 95 Resp: 63368

Ion Ratio Lower Upper

95 100

174 66.8 0.0 135.8

176 64.1 0.0 131.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.006 min

Lab File: VX046470.D

Acq: 03 Jun 2025 14:09




Tgt Ion: 117 Resp: 122673


Ion Ratio Lower Upper

117 100

82 58.5 50.6 76.0

119 32.4 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046470.D


Acq: 03 Jun 2025 14:09

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-GW-720-722





Tgt Ion:152 Resp: 50694

Ion Ratio Lower Upper

152 100

115 64.3 46.9 140.7

150 157.5 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046470.D
 Acq On : 03 Jun 2025 14:09
 Operator : JC/MD
 Sample : Q2186-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-720-722

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046470.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.142	6	9	19	rVB	12500	12463	2.69%	0.508%
2	1.368	40	46	49	rBV3	4320	6629	1.43%	0.270%
3	1.575	74	80	83	rVB3	3767	6848	1.48%	0.279%
4	2.380	207	212	222	rBV	2642	5583	1.20%	0.227%
5	3.440	380	386	393	rBV3	2322	5065	1.09%	0.206%
6	5.385	692	705	720	rBV	53114	160821	34.70%	6.553%
7	5.544	722	731	746	rVB	77628	220609	47.60%	8.989%
8	5.952	789	798	811	rBV	65050	168360	36.33%	6.860%
9	6.757	919	930	949	rBV	138294	337386	72.79%	13.748%
10	8.647	1234	1240	1257	rBV	294546	463477	100.00%	18.886%
11	10.049	1465	1470	1482	rBV	292174	413979	89.32%	16.869%
12	11.079	1634	1639	1651	rBV	241265	308006	66.46%	12.551%
13	12.018	1788	1793	1805	rBV	273960	344900	74.42%	14.054%




Sum of corrected areas: 2454126

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046470.D
Acq On : 03 Jun 2025 14:09
Operator : JC/MD
Sample : Q2186-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-720-722

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046470.D
Acq On : 03 Jun 2025 14:09
Operator : JC/MD
Sample : Q2186-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-720-722

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046470.D
Acq On : 03 Jun 2025 14:09
Operator : JC/MD
Sample : Q2186-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-720-722

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	06/02/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	06/02/25		
Client Sample ID:	BP-VPB-182-DUP-20250602			SDG No.:	Q2186		
Lab Sample ID:	Q2186-08			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046471.D	1		06/03/25 14:33	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	7.20		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	06/02/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	06/02/25	
Client Sample ID:	BP-VPB-182-DUP-20250602		SDG No.:	Q2186	
Lab Sample ID:	Q2186-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046471.D	1		06/03/25 14:33	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.5		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	49.7		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.0		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	68700	5.55				
540-36-3	1,4-Difluorobenzene	139000	6.757				
3114-55-4	Chlorobenzene-d5	130000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	55400	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	06/02/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/02/25
Client Sample ID:	BP-VPB-182-DUP-20250602	SDG No.:	Q2186
Lab Sample ID:	Q2186-08	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046471.D	1		06/03/25 14:33	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046471.D
 Acq On : 03 Jun 2025 14:33
 Operator : JC/MD
 Sample : Q2186-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-DUP-20250602

Quant Time: Jun 04 01:27:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.550	168	68690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138832	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129871	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67222	52.493	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.980%
35) Dibromofluoromethane	5.379	113	48986	48.999	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	8.647	98	172017	49.713	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.079	95	67628	50.952	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.900%
Target Compounds						Qvalue
16) Acetone	2.386	43	3717	7.239	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046471.D
Acq On : 03 Jun 2025 14:33
Operator : JC/MD
Sample : Q2186-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250602

Quant Time: Jun 04 01:27:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046471.D
Acq: 03 Jun 2025 14:33




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250602


Tgt Ion:168 Resp: 68690
Ion Ratio Lower Upper
168 100
99 73.3 54.9 82.3



#16
Acetone
Concen: 7.239 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX046471.D
Acq: 03 Jun 2025 14:33

Tgt Ion: 43 Resp: 3717
Ion Ratio Lower Upper
43 100
58 29.4 21.2 31.8





#33

1,2-Dichloroethane-d4

Concen: 52.493 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046471.D

Acq: 03 Jun 2025 14:33

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-182-DUP-20250602

Tgt Ion: 65 Resp: 67222

Ion Ratio Lower Upper

65 100

67 50.2 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046471.D

Acq: 03 Jun 2025 14:33




Tgt Ion: 114 Resp: 138832


Ion Ratio Lower Upper

114 100

63 24.0 0.0 49.2

88 17.9 0.0 33.6





#35

Dibromofluoromethane

Concen: 48.999 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046471.D


Acq: 03 Jun 2025 14:33

Instrument :

MSVOA_X




ClientSampleId :

BP-VPB-182-DUP-20250602



Tgt Ion:113 Resp: 48986

Ion	Ratio	Lower	Upper
113	100		
111	103.1	83.1	124.7
192	16.9	13.3	19.9



#50

Toluene-d8

Concen: 49.713 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. 0.000 min


Lab File: VX046471.D

Acq: 03 Jun 2025 14:33

Tgt Ion: 98 Resp: 172017

Ion	Ratio	Lower	Upper
98	100		
100	65.2	53.5	80.3









#62
4-Bromofluorobenzene
Concen: 50.952 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046471.D
Acq: 03 Jun 2025 14:33




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250602


Tgt Ion: 95 Resp: 67628
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.0 0.0 131.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046471.D
Acq: 03 Jun 2025 14:33

Tgt Ion: 117 Resp: 129871
Ion Ratio Lower Upper
117 100
82 61.6 50.6 76.0
119 31.8 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046471.D


Acq: 03 Jun 2025 14:33

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-182-DUP-20250602





Tgt Ion:152 Resp: 55355

Ion Ratio Lower Upper

152 100

115 65.5 46.9 140.7

150 158.4 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046471.D
 Acq On : 03 Jun 2025 14:33
 Operator : JC/MD
 Sample : Q2186-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-DUP-20250602

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046471.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.142	6	9	15	rVB	6312	6454	1.34%	0.250%
2	1.581	70	81	86	rBB2	4347	9829	2.04%	0.381%
3	2.380	205	212	219	rBV	3302	5662	1.17%	0.220%
4	3.434	380	385	392	rVB4	2169	4859	1.01%	0.189%
5	5.379	694	704	719	rBV	56058	165295	34.29%	6.413%
6	5.550	722	732	746	rVB	79720	228164	47.33%	8.853%
7	5.952	787	798	814	rBV	65346	176617	36.64%	6.853%
8	6.757	921	930	945	rBV	143421	349694	72.54%	13.568%
9	8.647	1233	1240	1256	rBV	298981	482081	100.00%	18.705%
10	10.049	1465	1470	1481	rBV	317759	438314	90.92%	17.006%
11	11.079	1634	1639	1652	rBV	257631	331776	68.82%	12.873%
12	12.018	1788	1793	1806	rBV	307293	378591	78.53%	14.689%




Sum of corrected areas: 2577336

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046471.D
Acq On : 03 Jun 2025 14:33
Operator : JC/MD
Sample : Q2186-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250602

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046471.D
Acq On : 03 Jun 2025 14:33
Operator : JC/MD
Sample : Q2186-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250602

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046471.D
Acq On : 03 Jun 2025 14:33
Operator : JC/MD
Sample : Q2186-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250602

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/30/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	06/02/25		
Client Sample ID:	VPB182-HYD-20250530			SDG No.:	Q2186		
Lab Sample ID:	Q2186-09			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046465.D	1		06/03/25 12:13	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.50	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	06/02/25	
Client Sample ID:	VPB182-HYD-20250530		SDG No.:	Q2186	
Lab Sample ID:	Q2186-09		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046465.D	1		06/03/25 12:13	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	1.70		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.2		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	49.2		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	50.0		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.0		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	72300	5.544				
540-36-3	1,4-Difluorobenzene	144000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	61100	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/30/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/02/25
Client Sample ID:	VPB182-HYD-20250530	SDG No.:	Q2186
Lab Sample ID:	Q2186-09	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046465.D	1		06/03/25 12:13	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046465.D
 Acq On : 03 Jun 2025 12:13
 Operator : JC/MD
 Sample : Q2186-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB182-HYD-20250530

Quant Time: Jun 04 01:25:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	72286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144078	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69043	51.232	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.460%			
35) Dibromofluoromethane	5.385	113	51094	49.247	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.500%			
50) Toluene-d8	8.647	98	179500	49.986	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.980%			
62) 4-Bromofluorobenzene	11.079	95	72967	52.973	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.940%			
Target Compounds						
16) Acetone	2.380	43	1333	2.467	ug/l #	74
60) Dibromochloromethane	9.519	129	1867	1.722	ug/l	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046465.D
Acq On : 03 Jun 2025 12:13
Operator : JC/MD
Sample : Q2186-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB182-HYD-20250530

Quant Time: Jun 04 01:25:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046465.D
Acq: 03 Jun 2025 12:13




Instrument :
MSVOA_X
ClientSampleId :
VPB182-HYD-20250530


Tgt Ion:168 Resp: 72286
Ion Ratio Lower Upper
168 100
99 71.1 54.9 82.3



#16
Acetone
Concen: 2.467 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046465.D
Acq: 03 Jun 2025 12:13

Tgt Ion: 43 Resp: 1333
Ion Ratio Lower Upper
43 100
58 40.0 21.2 31.8#





#33

1,2-Dichloroethane-d4

Concen: 51.232 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046465.D

Acq: 03 Jun 2025 12:13

Instrument :

MSVOA_X

ClientSampleId :





VPB182-HYD-20250530

Tgt Ion: 65 Resp: 69043

Ion Ratio Lower Upper

65 100

67 49.4 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046465.D

Acq: 03 Jun 2025 12:13




Tgt Ion: 114 Resp: 144078


Ion Ratio Lower Upper

114 100

63 25.1 0.0 49.2

88 17.8 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.247 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046465.D


Acq: 03 Jun 2025 12:13

Instrument :

MSVOA_X

ClientSampleId :

VPB182-HYD-20250530






Tgt Ion:113 Resp: 51094

Ion Ratio Lower Upper

113 100

111 103.4 83.1 124.7

192 16.7 13.3 19.9



#50

Toluene-d8

Concen: 49.986 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046465.D




Acq: 03 Jun 2025 12:13


Tgt Ion: 98 Resp: 179500

Ion Ratio Lower Upper

98 100

100 65.1 53.5 80.3





#60

Dibromochloromethane

Concen: 1.722 ug/l

RT: 9.519 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046465.D

Acq: 03 Jun 2025 12:13

Instrument :

MSVOA_X

ClientSampleId :





VPB182-HYD-20250530

Tgt Ion:129 Resp: 1867

Ion Ratio Lower Upper

129 100

127 72.7 39.3 117.8



#62

4-Bromofluorobenzene

Concen: 52.973 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX046465.D

Acq: 03 Jun 2025 12:13




Tgt Ion: 95 Resp: 72967


Ion Ratio Lower Upper

95 100

174 66.2 0.0 135.8

176 64.5 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX046465.D

Acq: 03 Jun 2025 12:13

Instrument :

MSVOA_X

ClientSampleId :

VPB182-HYD-20250530





Tgt Ion:117 Resp: 137138

Ion Ratio Lower Upper

117 100

82 62.4 50.6 76.0

119 32.5 25.8 38.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046465.D

Acq: 03 Jun 2025 12:13




Tgt Ion:152 Resp: 61107

Ion Ratio Lower Upper

152 100

115 64.6 46.9 140.7

150 156.0 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046465.D
Acq On : 03 Jun 2025 12:13
Operator : JC/MD
Sample : Q2186-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB182-HYD-20250530

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046465.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.587	70	82	87	rBV5	5140	14775	2.95%	0.545%
2	5.379	694	704	721	rBV	58212	172939	34.49%	6.374%
3	5.544	721	731	751	rBV2	82219	240797	48.02%	8.875%
4	5.952	788	798	816	rBV	68330	181015	36.10%	6.671%
5	6.757	921	930	945	rBV	150870	364200	72.63%	13.423%
6	8.647	1233	1240	1254	rBV	316026	501447	100.00%	18.481%
7	9.525	1379	1384	1390	rBV	3666	5800	1.16%	0.214%
8	10.049	1465	1470	1483	rBV	335638	462933	92.32%	17.062%
9	11.079	1634	1639	1657	rBV	283035	355360	70.87%	13.097%
10	12.018	1788	1793	1806	rBV	338840	414023	82.57%	15.259%


Sum of corrected areas: 2713289

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046465.D
Acq On : 03 Jun 2025 12:13
Operator : JC/MD
Sample : Q2186-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB182-HYD-20250530

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046465.D
Acq On : 03 Jun 2025 12:13
Operator : JC/MD
Sample : Q2186-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB182-HYD-20250530

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046465.D
Acq On : 03 Jun 2025 12:13
Operator : JC/MD
Sample : Q2186-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB182-HYD-20250530

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc



CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG No.: Q2186
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D	
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD				
Chloromethane	0.727	0.775	0.787	0.791	0.679	0.694	0.742	6.6				
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2				
Bromomethane	0.296	0.326	0.340	0.334	0.305		0.320	5.8				
Chloroethane	0.354	0.378	0.329	0.317	0.368	0.467	0.369	14.4				
Trichlorofluoromethane	1.035	1.068	0.983	0.985	0.990	1.064	1.021	3.9				
1,1,2-Trichlorotrifluoroethane	0.628	0.641	0.629	0.648	0.610	0.633	0.632	2.1				
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9				
Acetone	0.361	0.362	0.361	0.370	0.408	0.380	0.374	4.9				
Carbon Disulfide	1.295	1.455	1.522	1.597	1.141	1.423	1.406	11.7				
Methyl tert-butyl Ether	2.044	2.160	2.172	2.239	1.908	1.949	2.079	6.4				
Methylene Chloride	0.689	0.684	0.691	0.691	0.689	0.853	0.716	9.4				
trans-1,2-Dichloroethene	0.573	0.610	0.612	0.622	0.557	0.604	0.596	4.3				
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6				
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8				
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6				
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5				
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3				
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6				
Methylcyclohexane	0.596	0.641	0.627	0.658	0.587	0.627	0.623	4.3				
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3				
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5				
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3				
1,2-Dichloropropane	0.356	0.371	0.368	0.378	0.324	0.317	0.352	7.4				
Bromodichloromethane	0.557	0.577	0.573	0.594	0.498	0.485	0.547	8.2				
4-Methyl-2-Pentanone	0.620	0.634	0.630	0.631	0.555	0.561	0.605	6				
Toluene	0.884	0.898	0.885	0.904	0.838	0.803	0.869	4.5				
t-1,3-Dichloropropene	0.468	0.528	0.555	0.591	0.406	0.371	0.487	17.9				
cis-1,3-Dichloropropene	0.531	0.578	0.602	0.623	0.469	0.423	0.538	14.6				
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3				
2-Hexanone	0.466	0.473	0.477	0.473	0.414	0.385	0.448	8.7				

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG No.: Q2186
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.378	0.400	0.415	0.431	0.326	0.306	0.376	13.3	
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8	
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7	
Ethyl Benzene	1.919	2.022	1.979	2.036	1.816	1.803	1.929	5.2	
m/p-Xylenes	0.706	0.740	0.721	0.740	0.678	0.648	0.706	5.2	
o-Xylene	0.688	0.727	0.706	0.726	0.639	0.642	0.688	5.7	
Styrene	1.135	1.219	1.214	1.230	1.012	0.951	1.127	10.6	
Bromoform	0.270	0.304	0.312	0.327	0.236	0.234	0.281	14.2	
Isopropylbenzene	3.843	4.130	3.876	4.156	3.562	3.789	3.893	5.7	
1,1,2,2-Tetrachloroethane	1.315	1.338	1.284	1.345	1.350	1.552	1.364	7	
1,3-Dichlorobenzene	1.633	1.701	1.656	1.730	1.558	1.619	1.649	3.7	
1,4-Dichlorobenzene	1.629	1.693	1.639	1.722	1.606	1.817	1.684	4.6	
1,2-Dichlorobenzene	1.613	1.696	1.634	1.702	1.577	1.710	1.655	3.3	
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6	
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7	
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2	
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M
 Title : SW846 8260
 Last Update : Tue May 06 07:12:22 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX046047.D 5 =VX046046.D 20 =VX046041.D 50 =VX046042.D 100 =VX046043.D 150 =VX046044.D

	Compound	1	5	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.658	0.639	0.697	0.864	0.859	0.875	0.765	14.62
3) P	Chloromethane	0.694	0.679	0.727	0.775	0.787	0.791	0.742	6.58
4) C	Vinyl Chloride	0.673	0.619	0.660	0.710	0.727	0.755	0.691	7.17#
5) T	Bromomethane		0.305	0.296	0.326	0.340	0.334	0.320	5.84
6) T	Chloroethane	0.467	0.368	0.354	0.378	0.329	0.317	0.369	14.44
7) T	Trichlorofluor...	1.064	0.990	1.035	1.068	0.983	0.985	1.021	3.92
8) T	Diethyl Ether	0.403	0.311	0.340	0.337	0.338	0.355	0.347	8.83
9) T	1,1,2-Trichlor...	0.633	0.610	0.628	0.641	0.629	0.648	0.632	2.10
10) T	Methyl Iodide		0.608	0.767	0.806	0.793	0.763	0.747	10.68
11) T	Tert butyl alc...		0.114	0.122	0.129	0.144	0.146	0.131	10.45
12) CM	1,1-Dichloroet...	0.594	0.567	0.565	0.601	0.607	0.625	0.593	3.94#
13) T	Acrolein		0.117	0.158	0.152	0.154	0.163	0.149	12.33
14) T	Allyl chloride	1.052	1.058	1.127	1.179	1.187	1.196	1.133	5.75
15) T	Acrylonitrile	0.363	0.345	0.378	0.388	0.381	0.390	0.374	4.52
16) T	Acetone	0.380	0.408	0.361	0.362	0.361	0.370	0.374	4.90
17) T	Carbon Disulfide	1.423	1.141	1.295	1.455	1.522	1.597	1.406	11.68
18) T	Methyl Acetate	1.006	0.816	0.814	0.848	0.845	0.875	0.867	8.27
19) T	Methyl tert-bu...	1.949	1.908	2.044	2.160	2.172	2.239	2.079	6.39
20) T	Methylene Chlo...	0.853	0.689	0.689	0.684	0.691	0.691	0.716	9.39
21) T	trans-1,2-Dich...	0.604	0.557	0.573	0.610	0.612	0.622	0.596	4.27
22) T	Diisopropyl ether	2.095	1.924	2.219	2.278	2.295	2.321	2.189	6.97
23) T	Vinyl Acetate	1.660	1.698	1.928	2.048	2.082	2.134	1.925	10.52
24) P	1,1-Dichloroet...	1.116	1.154	1.233	1.263	1.263	1.286	1.219	5.60
25) T	2-Butanone	0.495	0.539	0.540	0.555	0.558	0.569	0.543	4.77
26) T	2,2-Dichloropr...	0.965	0.850	0.910	0.957	1.003	1.039	0.954	7.03
27) T	cis-1,2-Dichlo...	0.719	0.642	0.716	0.737	0.738	0.755	0.718	5.55
28) T	Bromochloromet...	0.576	0.553	0.628	0.578	0.595	0.590	0.587	4.28
29) T	Tetrahydrofuran	0.318	0.318	0.340	0.350	0.351	0.362	0.340	5.36
30) C	Chloroform	1.265	1.199	1.287	1.296	1.277	1.300	1.271	2.95#
31) T	Cyclohexane		1.059	1.090	1.128	1.128	1.150	1.111	3.26
32) T	1,1,1-Trichlor...	1.015	1.013	1.106	1.131	1.155	1.188	1.101	6.60
33) S	1,2-Dichloroet...		0.935	0.953	0.910	0.930	0.932	0.932	1.65
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.354	0.359	0.355	0.364	0.368	0.360	1.70
36) T	1,1-Dichloropr...	0.493	0.462	0.463	0.495	0.483	0.505	0.484	3.68
37) T	Ethyl Acetate	0.586	0.582	0.569	0.611	0.609	0.631	0.598	3.83
38) T	Carbon Tetrach...	0.541	0.505	0.528	0.558	0.552	0.577	0.544	4.61
39) T	Methylcyclohexane	0.627	0.587	0.596	0.641	0.627	0.658	0.623	4.34
40) TM	Benzene	1.348	1.337	1.426	1.474	1.441	1.477	1.417	4.30
41) T	Methacrylonitrile	0.233	0.288	0.318	0.346	0.343	0.348	0.313	14.50
42) TM	1,2-Dichloroet...	0.579	0.594	0.632	0.627	0.611	0.625	0.612	3.45
43) T	Isopropyl Acetate	0.764	0.826	0.905	0.963	0.982	1.030	0.912	11.05
44) TM	Trichloroethene	0.324	0.315	0.344	0.355	0.345	0.362	0.341	5.28
45) C	1,2-Dichloropr...	0.317	0.324	0.356	0.371	0.368	0.378	0.352	7.37#
46) T	Dibromomethane	0.263	0.262	0.285	0.287	0.280	0.289	0.278	4.35
47) T	Bromodichlorom...	0.485	0.498	0.557	0.577	0.573	0.594	0.547	8.22
48) T	Methyl methacr...	0.370	0.426	0.465	0.502	0.500	0.531	0.466	12.75
49) T	1,4-Dioxane	0.007	0.009	0.009	0.009	0.009	0.010	0.009	8.45
50) S	Toluene-d8		1.221	1.246	1.223	1.266	1.275	1.246	1.95
51) T	4-Methyl-2-Pen...	0.561	0.555	0.620	0.634	0.630	0.631	0.605	6.04
52) CM	Toluene	0.803	0.838	0.884	0.898	0.885	0.904	0.869	4.54#
53) T	t-1,3-Dichloro...	0.371	0.406	0.468	0.528	0.555	0.591	0.487	17.86
54) T	cis-1,3-Dichlo...	0.423	0.469	0.531	0.578	0.602	0.623	0.538	14.61
55) T	1,1,2-Trichlor...	0.308	0.337	0.349	0.354	0.351	0.356	0.343	5.30
56) T	Ethyl methacry...	0.377	0.508	0.540	0.595	0.617	0.639	0.546	17.58

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M

57)	T	1,3-Dichloropr...	0.610	0.601	0.618	0.623	0.613	0.627	0.615	1.53
58)	T	2-Chloroethyl ...	0.230	0.247	0.270	0.307	0.303	0.313	0.278	12.41
59)	T	2-Hexanone	0.385	0.414	0.466	0.473	0.477	0.473	0.448	8.69
60)	T	Dibromochlorom...	0.306	0.326	0.378	0.400	0.415	0.431	0.376	13.28
61)	T	1,2-Dibromoethane	0.322	0.333	0.359	0.373	0.368	0.381	0.356	6.54
62)	S	4-Bromofluorob...		0.464	0.455	0.470	0.500	0.500	0.478	4.41
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.347	0.323	0.390	0.375	0.345	0.344	0.354	6.84
65)	PM	Chlorobenzene	1.131	1.046	1.093	1.098	1.085	1.114	1.094	2.65
66)	T	1,1,1,2-Tetrac...	0.369	0.341	0.365	0.390	0.382	0.395	0.374	5.22
67)	C	Ethyl Benzene	1.803	1.816	1.919	2.022	1.979	2.036	1.929	5.24#
68)	T	m/p-Xylenes	0.648	0.678	0.706	0.740	0.721	0.740	0.706	5.21
69)	T	o-Xylene	0.642	0.639	0.688	0.727	0.706	0.726	0.688	5.75
70)	T	Styrene	0.951	1.012	1.135	1.219	1.214	1.230	1.127	10.56
71)	P	Bromoform	0.234	0.236	0.270	0.304	0.312	0.327	0.281	14.17
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.789	3.562	3.843	4.130	3.876	4.156	3.893	5.72
74)	T	N-amyl acetate	1.715	1.652	1.846	2.067	2.068	2.192	1.924	11.32
75)	P	1,1,2,2-Tetrac...	1.552	1.350	1.315	1.338	1.284	1.345	1.364	6.97
76)	T	1,2,3-Trichlor...	1.405	1.167	1.151	1.187	1.131	1.181	1.204	8.36
77)	T	Bromobenzene	0.926	0.862	0.896	0.928	0.883	0.928	0.904	3.08
78)	T	n-propylbenzene	4.272	4.186	4.394	4.854	4.583	4.868	4.526	6.45
79)	T	2-Chlorotoluene	3.184	2.748	2.832	2.994	2.805	2.953	2.919	5.45
80)	T	1,3,5-Trimethy...	3.036	3.053	3.275	3.487	3.255	3.405	3.252	5.60
81)	T	trans-1,4-Dich...		0.269	0.335	0.385	0.410	0.449	0.370	18.91
82)	T	4-Chlorotoluene	3.226	2.939	3.196	3.430	3.255	3.379	3.238	5.32
83)	T	tert-Butylbenzene	3.341	3.098	3.115	3.435	3.255	3.411	3.276	4.44
84)	T	1,2,4-Trimethy...	3.150	3.034	3.274	3.522	3.335	3.444	3.293	5.52
85)	T	sec-Butylbenzene	3.708	3.767	3.937	4.282	4.095	4.343	4.022	6.55
86)	T	p-Isopropyltol...	3.025	3.084	3.206	3.555	3.450	3.599	3.320	7.44
87)	T	1,3-Dichlorobe...	1.619	1.558	1.633	1.701	1.656	1.729	1.649	3.71
88)	T	1,4-Dichlorobe...	1.817	1.606	1.629	1.693	1.639	1.722	1.684	4.64
89)	T	n-Butylbenzene	2.443	2.650	2.748	3.147	3.139	3.346	2.912	12.00
90)	T	Hexachloroethane	0.511	0.523	0.551	0.622	0.622	0.680	0.585	11.44
91)	T	1,2-Dichlorobe...	1.710	1.577	1.613	1.696	1.634	1.702	1.655	3.34
92)	T	1,2-Dibromo-3-...	0.259	0.248	0.299	0.322	0.329	0.356	0.302	13.89
93)	T	1,2,4-Trichlor...	0.862	0.842	0.861	0.981	1.035	1.123	0.951	12.03
94)	T	Hexachlorobuta...	0.393	0.414	0.394	0.427	0.418	0.445	0.415	4.79
95)	T	Naphthalene	3.499	2.929	3.204	3.613	3.690	3.984	3.487	10.69
96)	T	1,2,3-Trichlor...	0.941	0.846	0.921	1.019	1.051	1.107	0.981	9.74

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	83671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	31909	12.820	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	25.640%#		
35) Dibromofluoromethane	5.385	113	21112	12.804	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	25.600%#		
50) Toluene-d8	8.646	98	73333	13.200	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	26.400%#		
62) 4-Bromofluorobenzene	11.079	95	26760	13.236	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	26.480%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	23341	12.638	ug/l	96
3) Chloromethane	1.306	50	24316	13.371	ug/l	98
4) Vinyl Chloride	1.373	62	22096	13.868	ug/l	94
5) Bromomethane	1.593	94	9920	12.279	ug/l	93
6) Chloroethane	1.666	64	11832	14.414	ug/l	95
7) Trichlorofluoromethane	1.873	101	34636	14.271	ug/l	100
8) Diethyl Ether	2.129	74	11392	14.278	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.324	101	21027	14.499	ug/l	98
10) Methyl Iodide	2.446	142	25661	15.070	ug/l	99
11) Tert butyl alcohol	2.971	59	20370	68.662	ug/l	97
12) 1,1-Dichloroethene	2.312	96	18899	13.385	ug/l	98
13) Acrolein	2.233	56	26427	76.689	ug/l	95
14) Allyl chloride	2.660	41	37708	14.133	ug/l	99
15) Acrylonitrile	3.062	53	63290	71.312	ug/l	98
16) Acetone	2.379	43	60363	70.884	ug/l	99
17) Carbon Disulfide	2.501	76	43348	13.251	ug/l	100
18) Methyl Acetate	2.702	43	27234	13.397	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	68402	13.890	ug/l	100
20) Methylene Chloride	2.782	84	23050	13.362	ug/l	96
21) trans-1,2-Dichloroethene	3.086	96	19161	13.294	ug/l	96
22) Diisopropyl ether	3.763	45	74257	14.896	ug/l	95
23) Vinyl Acetate	3.720	43	322598	73.230	ug/l	100
24) 1,1-Dichloroethane	3.605	63	41257	14.082	ug/l	98
25) 2-Butanone	4.556	43	90331	73.489	ug/l	98
26) 2,2-Dichloropropane	4.470	77	30472	13.729	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23970	13.791	ug/l	98
28) Bromochloromethane	4.897	49	21028	13.497	ug/l	99
29) Tetrahydrofuran	5.007	42	56819	71.175	ug/l	100
30) Chloroform	5.086	83	43065	14.144	ug/l	95
31) Cyclohexane	5.458	56	36470	14.769	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	37003	14.092	ug/l	98
36) 1,1-Dichloropropene	5.684	75	27242	14.231	ug/l	98
37) Ethyl Acetate	4.720	43	33458	13.931	ug/l	98
38) Carbon Tetrachloride	5.671	117	31091	14.137	ug/l	98
39) Methylcyclohexane	7.372	83	35077	14.623	ug/l	99
40) Benzene	6.031	78	83898	14.140	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	18732	14.038	ug/l	95
42) 1,2-Dichloroethane	6.080	62	37203	15.165	ug/l	100
43) Isopropyl Acetate	6.342	43	53275	14.512	ug/l	99
44) Trichloroethene	7.116	130	20239	14.429	ug/l	99
45) 1,2-Dichloropropane	7.427	63	20940	14.204	ug/l	98
46) Dibromomethane	7.580	93	16781	14.439	ug/l	99
47) Bromodichloromethane	7.817	83	32762	14.545	ug/l	94
48) Methyl methacrylate	7.689	41	27374	14.521	ug/l	100
49) 1,4-Dioxane	7.659	88	10585	282.442	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	182455	75.775	ug/l	98
52) Toluene	8.713	92	52030	14.788	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27549	14.566	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	31226	14.154	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	20536	14.442	ug/l	96
56) Ethyl methacrylate	9.116	69	31787	14.425	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36341	14.417	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	79531	80.312	ug/l	100
59) 2-Hexanone	9.427	43	137054	74.997	ug/l	99
60) Dibromochloromethane	9.518	129	22267	14.387	ug/l	99
61) 1,2-Dibromoethane	9.610	107	21113	14.458	ug/l	96
64) Tetrachloroethene	9.268	164	19924	15.263	ug/l	97
65) Chlorobenzene	10.079	112	55863	14.202	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.158	131	18671	14.231	ug/l	98
67) Ethyl Benzene	10.189	91	98113	14.631	ug/l	100
68) m/p-Xylenes	10.299	106	72202	29.806	ug/l	99
69) o-Xylene	10.640	106	35178	14.408	ug/l	96
70) Styrene	10.652	104	58034	14.853	ug/l	98
71) Bromoform	10.799	173	13825	14.030	ug/l #	100
73) Isopropylbenzene	10.957	105	93013	14.096	ug/l	99
74) N-amyl acetate	10.841	43	44684	13.581	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	31830	13.560	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	27845m	11.039	ug/l	
77) Bromobenzene	11.195	156	21695	14.212	ug/l	96
78) n-propylbenzene	11.298	91	106337	14.376	ug/l	99
79) 2-Chlorotoluene	11.359	91	68534	13.748	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79261	14.468	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8104	13.371	ug/l	91
82) 4-Chlorotoluene	11.451	91	77338	14.116	ug/l	100
83) tert-Butylbenzene	11.713	119	75382	14.011	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	79232	14.480	ug/l	99
85) sec-Butylbenzene	11.890	105	95271	14.263	ug/l	100
86) p-Isopropyltoluene	12.006	119	77599	14.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	39520	13.898	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	39419	14.146	ug/l	99
89) n-Butylbenzene	12.329	91	66506	14.184	ug/l	99
90) Hexachloroethane	12.536	117	13332	13.287	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	39031	14.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7232	13.535	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	20846	13.745	ug/l	99
94) Hexachlorobutadiene	13.725	225	9524	13.426	ug/l	97
95) Naphthalene	13.774	128	77542	14.040	ug/l	99
96) 1,2,3-Trichlorobenzene	13.956	180	22301	14.023	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:168 Resp: 8367
Ion Ratio Lower Upper
168 100
99 72.7 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 12.638 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 85 Resp: 23341
Ion Ratio Lower Upper
85 100
87 33.5 15.7 47.1









#3
Chloromethane
Concen: 13.371 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 24310
Ion Ratio Lower Upper
50 100
52 33.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 13.868 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 62 Resp: 22096
Ion Ratio Lower Upper
62 100
64 34.9 25.2 37.8





#5
Bromomethane
Concen: 12.279 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 94 Resp: 9920
Ion Ratio Lower Upper
94 100
96 98.1 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 14.414 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 64 Resp: 11832
Ion Ratio Lower Upper
64 100
66 32.9 24.3 36.5





#7
Trichlorofluoromethane
Concen: 14.271 ug/l
RT: 1.873 min Scan# 110
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:101 Resp: 34630
Ion Ratio Lower Upper
101 100
103 63.5 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 14.278 ug/l
RT: 2.129 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 74 Resp: 11392
Ion Ratio Lower Upper
74 100
45 103.7 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 14.499 ug/l

RT: 2.324 min Scan# 202

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	101	Resp:	2102
Ion	Ratio	Lower	Upper
101	100		
85	46.6	38.6	58.0
151	70.4	55.2	82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 15.070 ug/l




RT: 2.446 min Scan# 223


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	142	Resp:	25661
Ion	Ratio	Lower	Upper
142	100		
127	53.0	41.7	62.5
141	14.7	11.5	17.3





#11

Tert butyl alcohol

Concen: 68.662 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 59 Resp: 20370

Ion Ratio Lower Upper

59 100





57 11.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 13.385 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 96 Resp: 18899


Ion Ratio Lower Upper

96 100

61 179.8 146.2 219.2

98 61.8 51.0 76.6





#13

Acrolein

Concen: 76.689 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 2642

Ion Ratio Lower Upper

56 100





55 74.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 14.133 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 41 Resp: 37708


Ion Ratio Lower Upper

41 100

39 75.0 60.6 90.8

76 30.6 24.9 37.3





#15

Acrylonitrile

Concen: 71.312 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	53	Resp:	63290
Ion	Ratio	Lower	Upper
53	100		
52	83.5	65.3	97.9
51	38.4	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 70.884 ug/l




RT: 2.379 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	43	Resp:	60363
Ion	Ratio	Lower	Upper
43	100		
58	26.0	21.2	31.8









#17
Carbon Disulfide
Concen: 13.251 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 76 Resp: 43348
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#18
Methyl Acetate
Concen: 13.397 ug/l
RT: 2.702 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 27234
Ion Ratio Lower Upper
43 100
74 20.7 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 13.890 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 73 Resp: 6840
Ion Ratio Lower Upper
73 100
57 22.2 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 13.362 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 84 Resp: 23050
Ion Ratio Lower Upper
84 100
49 146.0 113.9 170.9
51 43.8 33.5 50.3
86 62.2 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 13.294 ug/l
RT: 3.086 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 96 Resp: 1916
Ion Ratio Lower Upper
96 100
61 155.6 119.5 179.3
98 63.2 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 14.896 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 45 Resp: 74257
Ion Ratio Lower Upper
45 100
43 78.2 66.6 100.0
87 24.0 19.8 29.6
59 11.7 8.6 12.8









#23
Vinyl Acetate
Concen: 73.230 ug/l
RT: 3.720 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 322598
Ion Ratio Lower Upper
43 100
86 9.3 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 14.082 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 63 Resp: 41257
Ion Ratio Lower Upper
63 100
98 7.9 3.6 10.8
100 5.0 2.1 6.3





#25

2-Butanone

Concen: 73.489 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 9033

Ion Ratio Lower Upper

43 100





72 23.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 13.729 ug/l

RT: 4.470 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 77 Resp: 30472

Ion Ratio Lower Upper

77 100

97 20.9 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 13.791 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 96 Resp: 23970

Ion Ratio Lower Upper

96 100

61 158.3 0.0 322.8





98 63.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 13.497 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 49 Resp: 21028


Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.0

130 70.9 56.2 84.2





#29

Tetrahydrofuran

Concen: 71.175 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	42	Resp:	56819
Ion Ratio	Lower	Upper	
42	100		
72	42.3	33.7	50.5
71	39.2	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 14.144 ug/l




RT: 5.086 min Scan# 656


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	83	Resp:	43065
Ion Ratio	Lower	Upper	
83	100		
85	65.7	49.3	73.9









#31
Cyclohexane
Concen: 14.769 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 56 Resp: 36470
Ion Ratio Lower Upper
56 100
69 28.8 24.4 36.6
84 78.5 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 14.092 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 97 Resp: 37003
Ion Ratio Lower Upper
97 100
99 62.8 51.8 77.6
61 49.5 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 12.820 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 31909

Ion Ratio Lower Upper

65 100





67 49.9 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:114 Resp: 147096


Ion Ratio Lower Upper

114 100

63 25.2 0.0 49.2

88 18.0 0.0 33.6





#35

Dibromofluoromethane

Concen: 12.804 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	113	Resp:	2111
Ion	Ratio	Lower	Upper
113	100		
111	103.0	83.1	124.7
192	17.0	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 14.231 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	75	Resp:	27242
Ion	Ratio	Lower	Upper
75	100		
110	31.6	16.3	48.9
77	31.9	24.3	36.5





#37

Ethyl Acetate

Concen: 13.931 ug/l

RT: 4.720 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	43	Resp:	3345
Ion	Ratio	Lower	Upper
43	100		
61	12.6	10.3	15.5
70	10.9	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 14.137 ug/l




RT: 5.671 min Scan# 752


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	117	Resp:	31091
Ion	Ratio	Lower	Upper
117	100		
119	95.6	75.2	112.8
121	32.1	24.2	36.4





#39

Methylcyclohexane

Concen: 14.623 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 83	Resp: 3507
Ion Ratio	Lower Upper
83	100
55	80.7 64.7 97.1
98	44.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 14.140 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35


Tgt Ion: 78	Resp: 83898
Ion Ratio	Lower Upper
78	100
77	22.3 19.0 28.4





#41
Methacrylonitrile
Concen: 14.038 ug/l
RT: 4.915 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 41 Resp: 18732
Ion Ratio Lower Upper
41 100
39 62.6 47.2 70.8
67 68.3 50.7 76.1
52 32.5 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 15.165 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 62 Resp: 37203
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 14.512 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 43	Resp: 5327
Ion Ratio	Lower Upper
43	100
61	17.3 14.3 21.5
87	11.6 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 14.429 ug/l




RT: 7.116 min Scan# 989


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:130	Resp: 20239
Ion Ratio	Lower Upper
130	100
95	102.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 14.204 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 20940

Ion Ratio Lower Upper

63 100





65 29.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 14.439 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 93 Resp: 16781


Ion Ratio Lower Upper

93 100

95 81.5 65.6 98.4

174 83.1 68.2 102.2





#47

Bromodichloromethane

Concen: 14.545 ug/l

RT: 7.817 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	83	Resp:	32762
Ion	Ratio	Lower	Upper
83	100		
85	68.4	50.5	75.7
127	8.3	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 14.521 ug/l




RT: 7.689 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	41	Resp:	27374
Ion	Ratio	Lower	Upper
41	100		
69	73.5	58.5	87.7
39	64.9	51.7	77.5





#49

1,4-Dioxane

Concen: 282.442 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	88	Resp:	1058
Ion	Ratio	Lower	Upper
88	100		
43	45.5	33.4	50.2
58	73.6	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 13.200 ug/l




RT: 8.646 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35


Tgt Ion:	98	Resp:	73333
Ion	Ratio	Lower	Upper
98	100		
100	64.8	53.5	80.3





#51
4-Methyl-2-Pentanone
Concen: 75.775 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 18245
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 14.788 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 92 Resp: 52030
Ion Ratio Lower Upper
92 100
91 172.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 14.566 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 75 Resp: 27549

Ion Ratio Lower Upper

75 100





77 31.4 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 14.154 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 75 Resp: 31226


Ion Ratio Lower Upper

75 100

77 30.8 25.4 38.0

39 64.0 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 14.442 ug/l

RT: 9.152 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion: 97 Resp: 20530

Ion Ratio Lower Upper

97 100

83 87.0 69.0 103.6

85 58.7 44.3 66.5

99 66.8 47.8 71.6

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 14.425 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 69 Resp: 31787


Ion Ratio Lower Upper

69 100

41 78.2 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 14.417 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 76 Resp: 3634

Ion Ratio Lower Upper

76 100





78 31.8 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 80.312 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 63 Resp: 79531

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 74.997 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 137054

Ion Ratio Lower Upper

43 100





58 48.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 14.387 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:129 Resp: 22267

Ion Ratio Lower Upper

129 100

127 78.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 14.458 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:107 Resp: 2111
Ion Ratio Lower Upper
107 100
109 95.3 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 13.236 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 95 Resp: 26760
Ion Ratio Lower Upper
95 100
174 67.4 0.0 135.8
176 67.0 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:117 Resp: 127829
Ion Ratio Lower Upper
117 100
82 57.7 50.6 76.0
119 32.7 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 15.263 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:164 Resp: 19924
Ion Ratio Lower Upper
164 100
166 127.2 105.0 157.6
129 107.6 83.5 125.3
131 97.7 76.5 114.7









#65
Chlorobenzene
Concen: 14.202 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:112 Resp: 5586
Ion Ratio Lower Upper
112 100
114 32.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 14.231 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:131 Resp: 18671
Ion Ratio Lower Upper
131 100
133 94.5 47.3 141.9
119 66.7 31.6 95.0





#67

Ethyl Benzene

Concen: 14.631 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 91 Resp: 9811

Ion Ratio Lower Upper

91 100





106 29.3 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 29.806 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:106 Resp: 72202

Ion Ratio Lower Upper

106 100

91 216.1 171.2 256.8





#69

o-Xylene

Concen: 14.408 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 35178

Ion Ratio Lower Upper

106 100





91 231.7 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 14.853 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:104 Resp: 58034


Ion Ratio Lower Upper

104 100

78 58.7 45.7 68.5


103 55.6 43.7 65.5





#71
Bromoform
Concen: 14.030 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:173 Resp: 13825
Ion Ratio Lower Upper
173 100
175 46.4 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:152 Resp: 60503
Ion Ratio Lower Upper
152 100
115 76.6 46.9 140.7
150 163.6 0.0 351.0









#73
Isopropylbenzene
Concen: 14.096 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:105 Resp: 9301
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 13.581 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 44684
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 24.1 18.7 28.1
61 22.3 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 13.560 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 83 Resp: 31830
Ion Ratio Lower Upper
83 100
131 10.1 5.0 14.9
85 63.3 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 11.039 ug/l m
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 75 Resp: 27845
Ion Ratio Lower Upper
75 100
77 43.0 20.5 61.5





#77
Bromobenzene
Concen: 14.212 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020




Tgt Ion:156 Resp: 2169
Ion Ratio Lower Upper
156 100
77 179.9 92.8 278.3
158 93.1 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#78
n-propylbenzene
Concen: 14.376 ug/l
RT: 11.298 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35



Tgt Ion: 91 Resp: 106337
Ion Ratio Lower Upper
91 100
120 22.0 10.8 32.4





#79

2-Chlorotoluene

Concen: 13.748 ug/l

RT: 11.359 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 68534

Ion Ratio Lower Upper

91 100





126 31.4 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 14.468 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:105 Resp: 79261

Ion Ratio Lower Upper

105 100

120 45.9 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 13.371 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 75 Resp: 8104
Ion Ratio Lower Upper
75 100
53 138.3 100.5 150.7
89 44.6 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 14.116 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 91 Resp: 77338
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 14.011 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:119 Resp: 75382
Ion Ratio Lower Upper
119 100
91 66.8 32.9 98.7
134 23.7 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 14.480 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:105 Resp: 79232
Ion Ratio Lower Upper
105 100
120 42.0 21.2 63.6





#85

sec-Butylbenzene

Concen: 14.263 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:105 Resp: 9527

Ion Ratio Lower Upper

105 100





134 19.3 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 14.387 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:119 Resp: 77599


Ion Ratio Lower Upper

119 100

134 25.9 12.5 37.5

91 28.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 13.898 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 39520

Ion Ratio Lower Upper

146 100

111 42.7 22.1 66.3





148 62.9 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 14.146 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:146 Resp: 39419


Ion Ratio Lower Upper

146 100

111 42.8 21.3 63.9

148 64.2 31.9 95.5





#89

n-Butylbenzene

Concen: 14.184 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 66500

Ion Ratio Lower Upper

91 100

92 52.9 26.9 80.7





134 23.1 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 13.287 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:117 Resp: 13332

Ion Ratio Lower Upper

117 100


201 62.9 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 14.092 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:146 Resp: 3903
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 13.535 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 75 Resp: 7232
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 88.4 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 13.745 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:180 Resp: 20840
Ion Ratio Lower Upper
180 100
182 96.6 47.9 143.8
145 34.1 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 13.426 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion:225 Resp: 9524
Ion Ratio Lower Upper
225 100
223 65.6 30.8 92.4
227 61.4 30.9 92.7





#95
Naphthalene
Concen: 14.040 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:128 Resp: 7754
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 11.1 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 14.023 ug/l
RT: 13.956 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:180 Resp: 22301
Ion Ratio Lower Upper
180 100
182 92.1 47.8 143.3
145 36.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	80678m	28.468	ug/l	
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:168 Resp: 97070
Ion Ratio Lower Upper
168 100
99 68.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 39.120 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 85 Resp: 83826
Ion Ratio Lower Upper
85 100
87 31.4 15.7 47.1









#3
Chloromethane
Concen: 35.658 ug/l
RT: 1.307 min Scan# 36
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 50 Resp: 75238
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 37.273 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 62 Resp: 68901
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8





#5
Bromomethane
Concen: 33.765 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 31648
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 38.558 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 64 Resp: 36722
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5





#7
Trichlorofluoromethane
Concen: 36.824 ug/l
RT: 1.874 min Scan# 129
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:101 Resp: 10369
Ion Ratio Lower Upper
101 100
103 63.7 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 35.350 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 74 Resp: 32724
Ion Ratio Lower Upper
74 100
45 109.9 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 36.997 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:101 Resp: 6225

Ion Ratio Lower Upper

101 100

85 48.3 38.6 58.0





151 69.0 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 39.622 ug/l

RT: 2.447 min Scan# 223

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:142 Resp: 78277


Ion Ratio Lower Upper

142 100

127 52.1 41.7 62.5

141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 181.745 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 6255

Ion Ratio Lower Upper

59 100





57 10.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 35.610 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 96 Resp: 58334


Ion Ratio Lower Upper

96 100

61 182.7 146.2 219.2

98 63.8 51.0 76.6





#13

Acrolein

Concen: 184.907 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 7392

Ion Ratio Lower Upper

56 100





55 70.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 36.971 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 41 Resp: 114445


Ion Ratio Lower Upper

41 100

39 75.7 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 182.874 ug/l

RT: 3.062 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	53	Resp:	188304
Ion	Ratio	Lower	Upper
53	100		
52	81.6	65.3	97.9
51	37.3	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 177.910 ug/l




RT: 2.386 min Scan# 213


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	43	Resp:	175777
Ion	Ratio	Lower	Upper
43	100		
58	26.5	21.2	31.8





#17

Carbon Disulfide

Concen: 37.223 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 14128

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 34.915 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 43 Resp: 82347

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 36.695 ug/l

RT: 3.111 min Scan# 311

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 20966

Ion Ratio Lower Upper

73 100





57 22.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 33.182 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion: 84 Resp: 66412




Ion Ratio Lower Upper


84 100

49 142.4 113.9 170.9

51 41.9 33.5 50.3


86 67.3 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 35.421 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 59234
Ion Ratio Lower Upper
96 100
61 149.4 119.5 179.3
98 62.5 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 38.241 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 45 Resp: 221172
Ion Ratio Lower Upper
45 100
43 83.3 66.6 100.0
87 24.7 19.8 29.6
59 10.7 8.6 12.8





#23

Vinyl Acetate

Concen: 194.528 ug/l

RT: 3.721 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 994230

Ion Ratio Lower Upper

43 100





86 9.4 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#24

1,1-Dichloroethane

Concen: 36.067 ug/l

RT: 3.605 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 63 Resp: 122601


Ion Ratio Lower Upper

63 100

98 7.2 3.6 10.8

100 4.2 2.1 6.3





#25

2-Butanone

Concen: 188.783 ug/l

RT: 4.556 min Scan# 569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 269224

Ion Ratio Lower Upper

43 100





72 23.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 36.059 ug/l

RT: 4.465 min Scan# 554

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 77 Resp: 92857

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 35.466 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 71518

Ion Ratio Lower Upper

96 100

61 161.4 0.0 322.8





98 64.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 31.031 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 49 Resp: 56091


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.0

130 70.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 183.679 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	42	Resp:	170124
Ion	Ratio	Lower	Upper
42	100		
72	42.1	33.7	50.5
71	39.3	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 35.626 ug/l




RT: 5.087 min Scan# 656


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58


Tgt Ion:	83	Resp:	125850
Ion	Ratio	Lower	Upper
83	100		
85	61.6	49.3	73.9





#31
Cyclohexane
Concen: 38.205 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 109459
Ion Ratio Lower Upper
56 100
69 30.5 24.4 36.6
84 83.6 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 36.034 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 97 Resp: 109781
Ion Ratio Lower Upper
97 100
99 64.7 51.8 77.6
61 50.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 30.601 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 8837

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:114 Resp: 167947


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 16.8 0.0 33.6





#35

Dibromofluoromethane

Concen: 31.633 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	113	Resp:	59550
Ion Ratio	Lower	Upper	
113	100		
111	103.9	83.1	124.7
192	16.6	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 38.073 ug/l




RT: 5.684 min Scan# 754


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	75	Resp:	83215
Ion Ratio	Lower	Upper	
75	100		
110	32.6	16.3	48.9
77	30.4	24.3	36.5





#37

Ethyl Acetate

Concen: 37.394 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 10253

Ion Ratio Lower Upper

43 100

61 12.9 10.3 15.5





70 9.9 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 37.320 ug/l

RT: 5.672 min Scan# 752

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:117 Resp: 93712


Ion Ratio Lower Upper

117 100

119 94.0 75.2 112.8

121 30.3 24.2 36.4





#39

Methylcyclohexane

Concen: 39.307 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 10765

Ion Ratio Lower Upper

83 100

55 80.9 64.7 97.1





98 46.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 36.530 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 78 Resp: 247476

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 38.123 ug/l
RT: 4.916 min Scan# 612
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 58082
Ion Ratio Lower Upper
41 100
39 59.0 47.2 70.8
67 63.4 50.7 76.1
52 30.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 37.605 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 62 Resp: 105332
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 38.598 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 16178

Ion Ratio Lower Upper

43 100

61 17.9 14.3 21.5





87 11.9 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 37.230 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:130 Resp: 59623

Ion Ratio Lower Upper

130 100

95 102.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 37.064 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 6238

Ion Ratio Lower Upper

63 100





65 30.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 36.326 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 93 Resp: 48201


Ion Ratio Lower Upper

93 100

95 82.0 65.6 98.4

174 85.2 68.2 102.2





#47

Bromodichloromethane

Concen: 37.685 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	83	Resp:	96910
Ion	Ratio	Lower	Upper
83	100		
85	63.1	50.5	75.7
127	8.1	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 39.157 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	41	Resp:	84277
Ion	Ratio	Lower	Upper
41	100		
69	73.1	58.5	87.7
39	64.6	51.7	77.5









#49
1,4-Dioxane
Concen: 713.477 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 88 Resp: 30529
Ion Ratio Lower Upper
88 100
43 41.8 33.4 50.2
58 73.3 58.6 88.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#50
Toluene-d8
Concen: 32.395 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 98 Resp: 205479
Ion Ratio Lower Upper
98 100
100 66.9 53.5 80.3





#51

4-Methyl-2-Pentanone

Concen: 193.653 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 53238

Ion Ratio Lower Upper

43 100





58 36.1 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#52

Toluene

Concen: 37.530 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 92 Resp: 150757

Ion Ratio Lower Upper

92 100

91 170.8 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 41.056 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 8865

Ion Ratio Lower Upper

75 100

77 31.3 25.0 37.6

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 38.521 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 75 Resp: 97031


Ion Ratio Lower Upper

75 100

77 31.7 25.4 38.0

39 65.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 36.648 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 5949

Ion Ratio Lower Upper

97 100

83 86.3 69.0 103.6

85 55.4 44.3 66.5





99 59.7 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 39.726 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 69 Resp: 99947


Ion Ratio Lower Upper

69 100

41 76.0 60.8 91.2

39 48.8 39.0 58.6





#57

1,3-Dichloropropane

Concen: 36.347 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 104600

Ion Ratio Lower Upper

76 100





78 32.9 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 227.902 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 63 Resp: 257678

Ion Ratio Lower Upper

63 100

106 26.9 21.5 32.3





#59

2-Hexanone

Concen: 190.168 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 396784

Ion Ratio Lower Upper

43 100





58 49.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 38.027 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:129 Resp: 67198

Ion Ratio Lower Upper

129 100

127 78.5 39.3 117.8









#61
1,2-Dibromoethane
Concen: 37.567 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 6263
Ion Ratio Lower Upper
107 100
109 91.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 34.229 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 95 Resp: 79012
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.7 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 14625
Ion Ratio Lower Upper
117 100
82 63.3 50.6 76.0
119 32.2 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 36.726 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:164 Resp: 54853
Ion Ratio Lower Upper
164 100
166 131.3 105.0 157.6
129 104.4 83.5 125.3
131 95.6 76.5 114.7









#65
Chlorobenzene
Concen: 35.686 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 160600
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 38.017 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:131 Resp: 57066
Ion Ratio Lower Upper
131 100
133 94.6 47.3 141.9
119 63.3 31.6 95.0









#67
Ethyl Benzene
Concen: 38.543 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 295712
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 78.142 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:106 Resp: 216578
Ion Ratio Lower Upper
106 100
91 214.0 171.2 256.8





#69

o-Xylene

Concen: 38.064 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:106 Resp: 10633

Ion Ratio Lower Upper

106 100





91 225.4 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 39.888 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:104 Resp: 178313


Ion Ratio Lower Upper

104 100

78 57.1 45.7 68.5


103 54.6 43.7 65.5





#71
Bromoform
Concen: 39.457 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:173 Resp: 44480
Ion Ratio Lower Upper
173 100
175 46.7 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:152 Resp: 67976
Ion Ratio Lower Upper
152 100
115 93.8 46.9 140.7
150 175.5 0.0 351.0









#73
Isopropylbenzene
Concen: 37.865 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 280719
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 38.012 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 140509
Ion Ratio Lower Upper
43 100
70 38.6 30.9 46.3
55 23.4 18.7 28.1
61 21.4 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 34.495 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 83 Resp: 90975
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 63.8 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 28.468 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 75 Resp: 80678
Ion Ratio Lower Upper
75 100
77 41.0 20.5 61.5





#77
Bromobenzene
Concen: 36.763 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:156 Resp: 6305
Ion Ratio Lower Upper
156 100
77 185.5 92.8 278.3
158 97.5 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 39.707 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 329981
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4





#79

2-Chlorotoluene

Concen: 36.336 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 203509

Ion Ratio Lower Upper

91 100





126 31.1 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 38.516 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:105 Resp: 237066

Ion Ratio Lower Upper

105 100

120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 38.445 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 26179
Ion Ratio Lower Upper
75 100
53 125.6 100.5 150.7
89 46.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 37.880 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 91 Resp: 233172
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 38.622 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:119 Resp: 233468
Ion Ratio Lower Upper
119 100
91 65.8 32.9 98.7
134 22.7 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 38.946 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:105 Resp: 239433
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6









#85
sec-Butylbenzene
Concen: 38.786 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 291074
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 39.879 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:119 Resp: 241658
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 36.199 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 11565

Ion Ratio Lower Upper

146 100

111 44.2 22.1 66.3





148 64.3 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 36.763 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:146 Resp: 115097


Ion Ratio Lower Upper

146 100

111 42.6 21.3 63.9

148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 40.603 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 21389

Ion Ratio Lower Upper

91 100

92 53.8 26.9 80.7





134 23.5 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 37.534 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:117 Resp: 42314

Ion Ratio Lower Upper

117 100


201 63.1 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 37.047 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:146 Resp: 115280
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.3
148 63.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 36.496 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 21909
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 87.6 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 39.117 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:180 Resp: 6665
Ion Ratio Lower Upper
180 100
182 95.9 47.9 143.8
145 34.1 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 36.395 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion:225 Resp: 29007
Ion Ratio Lower Upper
225 100
223 61.6 30.8 92.4
227 61.8 30.9 92.7





#95
Naphthalene
Concen: 39.575 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:128 Resp: 245568
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.8 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 38.765 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:180 Resp: 69262
Ion Ratio Lower Upper
180 100
182 95.5 47.8 143.3
145 36.2 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161877	62.526	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 125.060%#			
35) Dibromofluoromethane	5.379	113	111455	65.006	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 130.020%#			
50) Toluene-d8	8.647	98	387230	67.031	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 134.060%#			
62) 4-Bromofluorobenzene	11.079	95	153085	72.818	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 145.640%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	149508	77.828	ug/l	99
3) Chloromethane	1.307	50	136937	72.392	ug/l	98
4) Vinyl Chloride	1.374	62	126609	76.398	ug/l	98
5) Bromomethane	1.593	94	59246	70.507	ug/l	99
6) Chloroethane	1.660	64	57332	67.148	ug/l	99
7) Trichlorofluoromethane	1.867	101	171042	67.754	ug/l	99
8) Diethyl Ether	2.136	74	58762	70.807	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	109520	72.604	ug/l	99
10) Methyl Iodide	2.440	142	138045	77.943	ug/l	100
11) Tert butyl alcohol	2.995	59	125094	405.394	ug/l	100
12) 1,1-Dichloroethene	2.306	96	105566	71.884	ug/l	96
13) Acrolein	2.239	56	134456	375.132	ug/l	99
14) Allyl chloride	2.660	41	206519	74.419	ug/l	98
15) Acrylonitrile	3.068	53	331142	358.723	ug/l	97
16) Acetone	2.386	43	314383	354.936	ug/l	98
17) Carbon Disulfide	2.501	76	264956	77.867	ug/l	100
18) Methyl Acetate	2.703	43	147014	69.531	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	378113	73.818	ug/l	100
20) Methylene Chloride	2.782	84	120270	67.030	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	106606	71.109	ug/l	99
22) Diisopropyl ether	3.763	45	399494	77.048	ug/l	90
23) Vinyl Acetate	3.721	43	1811550	395.363	ug/l	100
24) 1,1-Dichloroethane	3.605	63	219759	72.114	ug/l	99
25) 2-Butanone	4.556	43	485448	379.704	ug/l	100
26) 2,2-Dichloropropane	4.464	77	174579	75.621	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	128454	71.056	ug/l	99
28) Bromochloromethane	4.891	49	103602	63.932	ug/l	100
29) Tetrahydrofuran	5.001	42	305757	368.234	ug/l	99
30) Chloroform	5.086	83	222253	70.180	ug/l	95
31) Cyclohexane	5.458	56	196302	76.428	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	201112	73.635	ug/l	98
36) 1,1-Dichloropropene	5.690	75	147873	74.285	ug/l	99
37) Ethyl Acetate	4.714	43	186215	74.565	ug/l	99
38) Carbon Tetrachloride	5.672	117	168946	73.875	ug/l	99
39) Methylcyclohexane	7.372	83	191941	76.952	ug/l	96
40) Benzene	6.031	78	440885	71.457	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	104863	75.574	ug/l	98
42) 1,2-Dichloroethane	6.080	62	187036	73.318	ug/l	99
43) Isopropyl Acetate	6.342	43	300491	78.714	ug/l	100
44) Trichloroethene	7.123	130	105628	72.420	ug/l	98
45) 1,2-Dichloropropane	7.427	63	112554	73.420	ug/l	100
46) Dibromomethane	7.574	93	85728	70.938	ug/l	100
47) Bromodichloromethane	7.818	83	175301	74.845	ug/l	99
48) Methyl methacrylate	7.689	41	153093	78.101	ug/l	99
49) 1,4-Dioxane	7.659	88	57313	1470.687	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	963319	384.740	ug/l	100
52) Toluene	8.714	92	270763	74.010	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	169863	86.372	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	184134	80.264	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	107264	72.543	ug/l	99
56) Ethyl methacrylate	9.116	69	188717	82.359	ug/l	99
57) 1,3-Dichloropropane	9.305	76	187649	71.590	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	463169	449.791	ug/l	100
59) 2-Hexanone	9.433	43	729602	383.945	ug/l	100
60) Dibromochloromethane	9.518	129	127012	78.918	ug/l	98
61) 1,2-Dibromoethane	9.610	107	112669	74.200	ug/l	98
64) Tetrachloroethene	9.268	164	93195	67.493	ug/l	96
65) Chlorobenzene	10.079	112	293284	70.490	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	103177	74.349	ug/l	99
67) Ethyl Benzene	10.189	91	535122	75.444	ug/l	99
68) m/p-Xylenes	10.299	106	389935	152.181	ug/l	99
69) o-Xylene	10.640	106	190833	73.890	ug/l	97
70) Styrene	10.652	104	328195	79.411	ug/l	100
71) Bromoform	10.799	173	84353	80.927	ug/l #	98
73) Isopropylbenzene	10.957	105	514528	71.083	ug/l	100
74) N-amyl acetate	10.841	43	274553	76.073	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	170481	66.207	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	150182m	54.277	ug/l	
77) Bromobenzene	11.195	156	117151	69.959	ug/l	99
78) n-propylbenzene	11.305	91	608332	74.974	ug/l	100
79) 2-Chlorotoluene	11.360	91	372392	68.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	432096	71.903	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	54398	81.820	ug/l	93
82) 4-Chlorotoluene	11.451	91	432053	71.890	ug/l	100
83) tert-Butylbenzene	11.713	119	432011	73.197	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	442728	73.758	ug/l	100
85) sec-Butylbenzene	11.890	105	543597	74.188	ug/l	100
86) p-Isopropyltoluene	12.006	119	457898	77.393	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	219753	70.449	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	217613	71.190	ug/l	99
89) n-Butylbenzene	12.329	91	416602	80.996	ug/l	100
90) Hexachloroethane	12.536	117	82528	74.978	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	216834	71.367	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43673	74.512	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	137430	82.605	ug/l	98
94) Hexachlorobutadiene	13.725	225	55494	71.314	ug/l	98
95) Naphthalene	13.774	128	489778	80.842	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	139525	79.981	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion:168 Resp: 87028
Ion Ratio Lower Upper
168 100
99 72.1 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 77.828 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 85 Resp: 149508
Ion Ratio Lower Upper
85 100
87 32.0 15.7 47.1





#3
Chloromethane
Concen: 72.392 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 50 Resp: 13693
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 76.398 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 62 Resp: 126609
Ion Ratio Lower Upper
62 100
64 30.6 25.2 37.8





#5
Bromomethane
Concen: 70.507 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 94 Resp: 59240
Ion Ratio Lower Upper
94 100
96 91.7 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 67.148 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 64 Resp: 57332
Ion Ratio Lower Upper
64 100
66 30.8 24.3 36.5





#7
Trichlorofluoromethane
Concen: 67.754 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:101 Resp: 17104
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 70.807 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 74 Resp: 58762
Ion Ratio Lower Upper
74 100
45 110.1 54.9 164.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 72.604 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:101 Resp: 109520
Ion Ratio Lower Upper
101 100
85 48.1 38.6 58.0
151 69.9 55.2 82.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#10
Methyl Iodide
Concen: 77.943 ug/l
RT: 2.440 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion:142 Resp: 138045
Ion Ratio Lower Upper
142 100
127 52.0 41.7 62.5
141 14.5 11.5 17.3





#11

Tert butyl alcohol

Concen: 405.394 ug/l

RT: 2.995 min Scan# 310

Delta R.T. 0.018 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 125094

Ion Ratio Lower Upper

59 100





57 10.9 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 71.884 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 96 Resp: 105566


Ion Ratio Lower Upper

96 100

61 189.9 146.2 219.2

98 63.3 51.0 76.6





#13

Acrolein

Concen: 375.132 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 56 Resp: 134450

Ion Ratio Lower Upper

56 100





55 71.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 74.419 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 206519


Ion Ratio Lower Upper

41 100

39 73.8 60.6 90.8


76 31.4 24.9 37.3





#15
Acrylonitrile
Concen: 358.723 ug/l
RT: 3.068 min Scan# 3114
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 53 Resp: 33114
Ion Ratio Lower Upper
53 100
52 84.7 65.3 97.9
51 37.9 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16
Acetone
Concen: 354.936 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 314383
Ion Ratio Lower Upper
43 100
58 27.6 21.2 31.8





#17

Carbon Disulfide

Concen: 77.867 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 264950

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 69.531 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 43 Resp: 147014

Ion Ratio Lower Upper

43 100

74 22.6 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 73.818 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 73 Resp: 37811
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 67.030 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 84 Resp: 120270
Ion Ratio Lower Upper
84 100
49 136.4 113.9 170.9
51 41.1 33.5 50.3
86 64.0 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 71.109 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 106600
Ion Ratio Lower Upper
96 100
61 149.7 119.5 179.3
98 64.4 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 77.048 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 45 Resp: 399494
Ion Ratio Lower Upper
45 100
43 71.2 66.6 100.0
87 25.7 19.8 29.6
59 11.0 8.6 12.8









#23
Vinyl Acetate
Concen: 395.363 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1811550
Ion Ratio Lower Upper
43 100
86 9.5 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 72.114 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 63 Resp: 219759
Ion Ratio Lower Upper
63 100
98 6.8 3.6 10.8
100 4.5 2.1 6.3





#25

2-Butanone

Concen: 379.704 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 485448

Ion Ratio Lower Upper

43 100





72 23.1 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 75.621 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 77 Resp: 174579

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 71.056 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 96 Resp: 128454

Ion Ratio Lower Upper

96 100

61 160.1 0.0 322.8





98 64.9 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 63.932 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 49 Resp: 103602


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.1 56.2 84.2





#29

Tetrahydrofuran

Concen: 368.234 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 42 Resp: 30575

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 70.180 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 83 Resp: 222253

Ion Ratio Lower Upper

83 100


85 65.1 49.3 73.9





#31
Cyclohexane
Concen: 76.428 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 196302
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.3 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 73.635 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 97 Resp: 201112
Ion Ratio Lower Upper
97 100
99 63.8 51.8 77.6
61 48.7 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 62.526 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 65 Resp: 16187

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:114 Resp: 152958


Ion Ratio Lower Upper

114 100

63 24.3 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 65.006 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 11145

Ion Ratio Lower Upper

113 100

111 101.7 83.1 124.7





192 16.6 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 74.285 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 147873


Ion Ratio Lower Upper

75 100

110 33.5 16.3 48.9

77 31.1 24.3 36.5





#37

Ethyl Acetate

Concen: 74.565 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 186215

Ion Ratio Lower Upper

43 100

61 12.4 10.3 15.5





70 9.9 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 73.875 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:117 Resp: 168946


Ion Ratio Lower Upper

117 100

119 95.0 75.2 112.8

121 29.6 24.2 36.4





#39

Methylcyclohexane

Concen: 76.952 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 19194

Ion Ratio Lower Upper

83 100

55 84.2 64.7 97.1





98 49.0 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 71.457 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 78 Resp: 440885

Ion Ratio Lower Upper

78 100


77 24.6 19.0 28.4





#41
Methacrylonitrile
Concen: 75.574 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 41 Resp: 10486
Ion Ratio Lower Upper
41 100
39 57.0 47.2 70.8
67 64.6 50.7 76.1
52 29.7 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 73.318 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 187036
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 78.714 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046043.D


Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100






Tgt Ion:	43	Resp:	30049
Ion	Ratio	Lower	Upper
43	100		
61	18.0	14.3	21.5
87	11.6	9.5	14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 72.420 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Tgt Ion:	130	Resp:	105628
Ion	Ratio	Lower	Upper
130	100		
95	103.8	0.0	204.2





#45

1,2-Dichloropropane

Concen: 73.420 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 63 Resp: 112554

Ion Ratio Lower Upper

63 100





65 30.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 70.938 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 93 Resp: 85728


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2





#47

Bromodichloromethane

Concen: 74.845 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 83 Resp: 17530

Ion Ratio Lower Upper

83 100

85 63.6 50.5 75.7





127 7.8 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 78.101 ug/l

RT: 7.689 min Scan# 1083

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 153093


Ion Ratio Lower Upper

41 100

69 74.1 58.5 87.7

39 65.5 51.7 77.5





#49

1,4-Dioxane

Concen: 1470.687 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 88 Resp: 5731

Ion Ratio Lower Upper

88 100

43 40.1 33.4 50.2





58 73.3 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 67.031 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 98 Resp: 387230

Ion Ratio Lower Upper

98 100


100 65.2 53.5 80.3





#51
4-Methyl-2-Pentanone
Concen: 384.740 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 963319
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 74.010 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 92 Resp: 270763
Ion Ratio Lower Upper
92 100
91 171.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 86.372 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 75 Resp: 16986

Ion Ratio Lower Upper

75 100





77 31.5 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 80.264 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 184134


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 59.6 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 72.543 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 107264

Ion Ratio Lower Upper

97 100

83 86.8 69.0 103.6

85 56.4 44.3 66.5





99 61.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 82.359 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 69 Resp: 188717


Ion Ratio Lower Upper

69 100

41 75.7 60.8 91.2

39 47.7 39.0 58.6





#57

1,3-Dichloropropane

Concen: 71.590 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 187649

Ion Ratio Lower Upper

76 100





78 32.1 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 449.791 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 63 Resp: 463169

Ion Ratio Lower Upper

63 100


106 26.7 21.5 32.3





#59
2-Hexanone
Concen: 383.945 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 729602
Ion Ratio Lower Upper
43 100
58 49.9 24.9 74.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#60
Dibromochloromethane
Concen: 78.918 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:129 Resp: 127012
Ion Ratio Lower Upper
129 100
127 77.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 74.200 ug/l
RT: 9.610 min Scan# 11266
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:107 Resp: 11266
Ion Ratio Lower Upper
107 100
109 93.9 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 72.818 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 95 Resp: 153085
Ion Ratio Lower Upper
95 100
174 67.1 0.0 135.8
176 65.3 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:117 Resp: 135214
Ion Ratio Lower Upper
117 100
82 61.7 50.6 76.0
119 30.6 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 67.493 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:164 Resp: 93195
Ion Ratio Lower Upper
164 100
166 123.4 105.0 157.6
129 101.4 83.5 125.3
131 98.0 76.5 114.7





#65
Chlorobenzene
Concen: 70.490 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:112 Resp: 293284
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 74.349 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:131 Resp: 103177
Ion Ratio Lower Upper
131 100
133 95.3 47.3 141.9
119 64.5 31.6 95.0





#67
Ethyl Benzene
Concen: 75.444 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 91 Resp: 53512
Ion Ratio Lower Upper
91 100
106 29.6 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 152.181 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:106 Resp: 389935
Ion Ratio Lower Upper
106 100
91 215.8 171.2 256.8









#69
o-Xylene
Concen: 73.890 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:106 Resp: 190833
Ion Ratio Lower Upper
106 100
91 230.3 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 79.411 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:104 Resp: 328195
Ion Ratio Lower Upper
104 100
78 57.4 45.7 68.5
103 54.8 43.7 65.5





#71
Bromoform
Concen: 80.927 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:173 Resp: 8435
Ion Ratio Lower Upper
173 100
175 48.3 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:152 Resp: 66369
Ion Ratio Lower Upper
152 100
115 120.4 46.9 140.7
150 190.4 0.0 351.0





#73
Isopropylbenzene
Concen: 71.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion:105 Resp: 514528
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 76.073 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 274553
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.5 18.7 28.1
61 22.2 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 66.207 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 17048

Ion Ratio Lower Upper

83 100

131 9.7 5.0 14.9





85 64.9 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 54.277 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 75 Resp: 150182

Ion Ratio Lower Upper

75 100


77 42.2 20.5 61.5





#77
Bromobenzene
Concen: 69.959 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:156 Resp: 11715
Ion Ratio Lower Upper
156 100
77 187.4 92.8 278.3
158 96.1 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#78
n-propylbenzene
Concen: 74.974 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion: 91 Resp: 608332
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 68.100 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 372392
Ion Ratio Lower Upper
91 100
126 30.9 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 71.903 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 432096
Ion Ratio Lower Upper
105 100
120 46.3 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 81.820 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 54398
Ion Ratio Lower Upper
75 100
53 115.8 100.5 150.7
89 47.0 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 71.890 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 91 Resp: 432053
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 73.197 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:119 Resp: 43201
Ion Ratio Lower Upper
119 100
91 64.4 32.9 98.7
134 22.8 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 73.758 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:105 Resp: 442728
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6









#85
sec-Butylbenzene
Concen: 74.188 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:105 Resp: 54359
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 77.393 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:119 Resp: 457898
Ion Ratio Lower Upper
119 100
134 24.6 12.5 37.5
91 26.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 70.449 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:146 Resp: 21975

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3





148 63.5 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 71.190 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:146 Resp: 217613


Ion Ratio Lower Upper

146 100

111 43.4 21.3 63.9

148 63.2 31.9 95.5





#89

n-Butylbenzene

Concen: 80.996 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 41660

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 74.978 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:117 Resp: 82528

Ion Ratio Lower Upper

117 100


201 63.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 71.367 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:146 Resp: 216834
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 64.0 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 74.512 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 43673
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 90.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 82.605 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:180 Resp: 137430

Ion Ratio Lower Upper

180 100

182 93.7 47.9 143.8





145 34.6 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carbone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 71.314 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:225 Resp: 55494


Ion Ratio Lower Upper

225 100

223 62.3 30.8 92.4


227 64.9 30.9 92.7





#95
Naphthalene
Concen: 80.842 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 489778
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 10.9 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 79.981 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:180 Resp: 139525
Ion Ratio Lower Upper
180 100
182 93.7 47.8 143.3
145 35.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60345	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	231952	93.983	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 187.960%#			
35) Dibromofluoromethane	5.385	113	160168	98.562	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 197.120%#			
50) Toluene-d8	8.647	98	554390	101.252	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 202.500%#			
62) 4-Bromofluorobenzene	11.079	95	217536	109.174	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 218.340%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	217793	118.929	ug/l	98
3) Chloromethane	1.307	50	196843	109.161	ug/l	99
4) Vinyl Chloride	1.374	62	187924	118.953	ug/l	100
5) Bromomethane	1.593	94	83008	103.625	ug/l	95
6) Chloroethane	1.660	64	78806	96.822	ug/l	98
7) Trichlorofluoromethane	1.868	101	245108	101.851	ug/l	97
8) Diethyl Ether	2.136	74	88441	111.791	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.313	101	161404	112.243	ug/l	99
10) Methyl Iodide	2.441	142	189843	112.441	ug/l	100
11) Tert butyl alcohol	3.002	59	181238	616.119	ug/l	100
12) 1,1-Dichloroethene	2.307	96	155451	111.039	ug/l	99
13) Acrolein	2.239	56	203257	594.872	ug/l	98
14) Allyl chloride	2.654	41	297776	112.561	ug/l	98
15) Acrylonitrile	3.069	53	484854	550.973	ug/l	98
16) Acetone	2.392	43	460820	545.754	ug/l	98
17) Carbon Disulfide	2.502	76	397353	122.499	ug/l	100
18) Methyl Acetate	2.709	43	217885	108.099	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	557210	114.112	ug/l	100
20) Methylene Chloride	2.782	84	172035	100.579	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	154694	108.241	ug/l	96
22) Diisopropyl ether	3.764	45	577610	116.858	ug/l	90
23) Vinyl Acetate	3.721	43	2655751	608.005	ug/l	100
24) 1,1-Dichloroethane	3.605	63	320083	110.182	ug/l	99
25) 2-Butanone	4.562	43	708463	581.292	ug/l	98
26) 2,2-Dichloropropane	4.471	77	258698	117.548	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	187835	108.994	ug/l	99
28) Bromochloromethane	4.898	49	146923	95.108	ug/l	100
29) Tetrahydrofuran	5.007	42	450456	569.081	ug/l	99
30) Chloroform	5.087	83	323659	107.208	ug/l	95
31) Cyclohexane	5.458	56	286312	116.934	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	295702	113.572	ug/l	98
36) 1,1-Dichloropropene	5.684	75	219731	116.462	ug/l	99
37) Ethyl Acetate	4.715	43	274312	115.889	ug/l	99
38) Carbon Tetrachloride	5.666	117	250916	115.760	ug/l	99
39) Methylcyclohexane	7.373	83	286311	121.107	ug/l	97
40) Benzene	6.031	78	642178	109.813	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	151424	115.139	ug/l	98
42) 1,2-Dichloroethane	6.080	62	271992	112.492	ug/l	99
43) Isopropyl Acetate	6.342	43	447871	123.780	ug/l	100
44) Trichloroethene	7.117	130	157641	114.032	ug/l	96
45) 1,2-Dichloropropane	7.428	63	164505	113.217	ug/l	99
46) Dibromomethane	7.574	93	125846	109.869	ug/l	100
47) Bromodichloromethane	7.818	83	258326	116.365	ug/l	98
48) Methyl methacrylate	7.690	41	231040	124.356	ug/l	99
49) 1,4-Dioxane	7.665	88	82800	2241.695	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	1371275	577.831	ug/l	100
52) Toluene	8.714	92	393323	113.430	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	257233	138.001	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	271054	124.658	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	154932	110.551	ug/l	97
56) Ethyl methacrylate	9.116	69	277921	127.968	ug/l	99
57) 1,3-Dichloropropane	9.305	76	272814	109.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	680608	697.344	ug/l	100
59) 2-Hexanone	9.433	43	1028597	571.094	ug/l	99
60) Dibromochloromethane	9.519	129	187497	122.915	ug/l	98
61) 1,2-Dibromoethane	9.604	107	165634	115.087	ug/l	97
64) Tetrachloroethene	9.269	164	132488	100.918	ug/l	96
65) Chlorobenzene	10.080	112	429656	108.615	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	152173	115.333	ug/l	99
67) Ethyl Benzene	10.189	91	785121	116.421	ug/l	99
68) m/p-Xylenes	10.299	106	570913	234.349	ug/l	100
69) o-Xylene	10.640	106	279906	113.990	ug/l	99
70) Styrene	10.653	104	474254	120.694	ug/l	99
71) Bromoform	10.799	173	125986	127.128	ug/l #	99
73) Isopropylbenzene	10.957	105	752337	114.312	ug/l	99
74) N-amyl acetate	10.842	43	396877	120.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	243532	104.018	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	213744m	84.960	ug/l	
77) Bromobenzene	11.195	156	167934	110.297	ug/l	99
78) n-propylbenzene	11.305	91	881218	119.448	ug/l	100
79) 2-Chlorotoluene	11.360	91	534560	107.514	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	616477	112.825	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	81372	134.609	ug/l	93
82) 4-Chlorotoluene	11.451	91	611736	111.948	ug/l	100
83) tert-Butylbenzene	11.713	119	617473	115.065	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	623456	114.236	ug/l	99
85) sec-Butylbenzene	11.890	105	786293	118.023	ug/l	99
86) p-Isopropyltoluene	12.006	119	651531	121.113	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	313100	110.394	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	311752	112.167	ug/l	100
89) n-Butylbenzene	12.329	91	605783	129.533	ug/l	100
90) Hexachloroethane	12.536	117	123188	123.091	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	308039	111.506	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64438	120.915	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	203255	134.366	ug/l	98
94) Hexachlorobutadiene	13.719	225	80558	113.857	ug/l	99
95) Naphthalene	13.774	128	721236	130.929	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	200379	126.331	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:168 Resp: 82961
Ion Ratio Lower Upper
168 100
99 67.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 118.929 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 85 Resp: 217793
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1





#3
Chloromethane
Concen: 109.161 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 50 Resp: 19684
Ion Ratio Lower Upper
50 100
52 32.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 118.953 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 187924
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8





#5

Bromomethane

Concen: 103.625 ug/l

RT: 1.593 min Scan# 81

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 94 Resp: 83008

Ion Ratio Lower Upper

94 100





96 95.8 72.8 109.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#6

Chloroethane

Concen: 96.822 ug/l

RT: 1.660 min Scan# 94

Delta R.T. -0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 64 Resp: 78806

Ion Ratio Lower Upper

64 100


66 31.3 24.3 36.5





#7
Trichlorofluoromethane
Concen: 101.851 ug/l
RT: 1.868 min Scan# 129
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:101 Resp: 245108
Ion Ratio Lower Upper
101 100
103 66.4 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 111.791 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 74 Resp: 88441
Ion Ratio Lower Upper
74 100
45 104.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 112.243 ug/l
RT: 2.313 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:101 Resp: 161404
Ion Ratio Lower Upper
101 100
85 46.8 38.6 58.0
151 69.8 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 112.441 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:142 Resp: 189843
Ion Ratio Lower Upper
142 100
127 52.4 41.7 62.5
141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 616.119 ug/l

RT: 3.002 min Scan# 310

Delta R.T. 0.024 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 59 Resp: 181238

Ion Ratio Lower Upper

59 100





57 10.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 111.039 ug/l

RT: 2.307 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 96 Resp: 155451


Ion Ratio Lower Upper

96 100

61 184.5 146.2 219.2

98 64.3 51.0 76.6





#13

Acrolein

Concen: 594.872 ug/l

RT: 2.239 min Scan# 188

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 56 Resp: 20325

Ion Ratio Lower Upper

56 100





55 72.1 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 112.561 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 41 Resp: 297776


Ion Ratio Lower Upper

41 100

39 73.7 60.6 90.8

76 31.8 24.9 37.3





#15

Acrylonitrile

Concen: 550.973 ug/l

RT: 3.069 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion:	53	Resp:	484854
Ion	Ratio	Lower	Upper
53	100		
52	84.2	65.3	97.9
51	37.5	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 545.754 ug/l




RT: 2.392 min Scan# 214


Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45


Tgt Ion:	43	Resp:	460820
Ion	Ratio	Lower	Upper
43	100		
58	27.3	21.2	31.8





#17
Carbon Disulfide
Concen: 122.499 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 76 Resp: 39735
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#18
Methyl Acetate
Concen: 108.099 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 217885
Ion Ratio Lower Upper
43 100
74 21.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 114.112 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 73 Resp: 557210
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 100.579 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 84 Resp: 172035
Ion Ratio Lower Upper
84 100
49 136.0 113.9 170.9
51 41.5 33.5 50.3
86 63.7 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 108.241 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 96 Resp: 154694
Ion Ratio Lower Upper
96 100
61 154.6 119.5 179.3
98 64.3 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 116.858 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 45 Resp: 577610
Ion Ratio Lower Upper
45 100
43 70.3 66.6 100.0
87 24.7 19.8 29.6
59 10.6 8.6 12.8





#23
Vinyl Acetate
Concen: 608.005 ug/l
RT: 3.721 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150




Tgt Ion: 43 Resp: 265575
Ion Ratio Lower Upper
43 100
86 9.6 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#24
1,1-Dichloroethane
Concen: 110.182 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45



Tgt Ion: 63 Resp: 320083
Ion Ratio Lower Upper
63 100
98 7.0 3.6 10.8
100 4.7 2.1 6.3





#25

2-Butanone

Concen: 581.292 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 70846

Ion Ratio Lower Upper

43 100





72 24.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 117.548 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 77 Resp: 258698

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 108.994 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 96 Resp: 18783

Ion Ratio Lower Upper

96 100

61 159.8 0.0 322.8





98 64.8 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 95.108 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 49 Resp: 146923


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 569.081 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 42 Resp: 450450

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 107.208 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 83 Resp: 323659

Ion Ratio Lower Upper

83 100


85 65.1 49.3 73.9





#31
Cyclohexane
Concen: 116.934 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 56 Resp: 286312
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.8 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 113.572 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 97 Resp: 295702
Ion Ratio Lower Upper
97 100
99 64.3 51.8 77.6
61 48.0 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 93.983 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 65 Resp: 231952

Ion Ratio Lower Upper

65 100





67 49.4 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:114 Resp: 144975


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 18.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 98.562 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046044.D


Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150



Tgt Ion: 113 Resp: 16016

Ion Ratio Lower Upper

113 100

111 103.0 83.1 124.7




192 16.9 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 116.462 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 219731


Ion Ratio Lower Upper

75 100

110 32.9 16.3 48.9

77 30.8 24.3 36.5





#37

Ethyl Acetate

Concen: 115.889 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 274312

Ion Ratio Lower Upper

43 100

61 12.5 10.3 15.5





70 10.2 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 115.760 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:117 Resp: 250916


Ion Ratio Lower Upper

117 100

119 94.3 75.2 112.8

121 29.7 24.2 36.4





#39

Methylcyclohexane

Concen: 121.107 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 28631

Ion Ratio Lower Upper

83 100

55 83.8 64.7 97.1





98 47.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 109.813 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 78 Resp: 642178

Ion Ratio Lower Upper

78 100


77 23.4 19.0 28.4





#41
Methacrylonitrile
Concen: 115.139 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 41 Resp: 151424
Ion Ratio Lower Upper
41 100
39 60.0 47.2 70.8
67 65.9 50.7 76.1
52 30.9 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 112.492 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 271992
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 123.780 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 44787

Ion Ratio Lower Upper

43 100

61 18.1 14.3 21.5





87 11.7 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 114.032 ug/l

RT: 7.117 min Scan# 989

Delta R.T. -0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:130 Resp: 157641

Ion Ratio Lower Upper

130 100

95 105.9 0.0 204.2





#45

1,2-Dichloropropane

Concen: 113.217 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 63 Resp: 16450

Ion Ratio Lower Upper

63 100





65 30.8 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 109.869 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 93 Resp: 125846


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4


174 85.0 68.2 102.2





#47
Bromodichloromethane
Concen: 116.365 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 83 Resp: 258320
Ion Ratio Lower Upper
83 100
85 64.8 50.5 75.7
127 8.0 6.5 9.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#48
Methyl methacrylate
Concen: 124.356 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 41 Resp: 231040
Ion Ratio Lower Upper
41 100
69 73.0 58.5 87.7
39 63.8 51.7 77.5





#49

1,4-Dioxane

Concen: 2241.695 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 88 Resp: 82800

Ion Ratio Lower Upper

88 100

43 40.0 33.4 50.2





58 74.8 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 101.252 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 98 Resp: 554390

Ion Ratio Lower Upper

98 100

100 65.4 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 577.831 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 43 Resp: 137127
Ion Ratio Lower Upper
43 100
58 36.2 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 113.430 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 92 Resp: 393323
Ion Ratio Lower Upper
92 100
91 171.0 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 138.001 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 75 Resp: 25723

Ion Ratio Lower Upper

75 100





77 31.7 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 124.658 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 271054


Ion Ratio Lower Upper

75 100

77 31.4 25.4 38.0

39 60.0 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 110.551 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 97 Resp: 154932
Ion Ratio Lower Upper
97 100
83 88.5 69.0 103.6
85 56.0 44.3 66.5
99 64.5 47.8 71.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#56
Ethyl methacrylate
Concen: 127.968 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 69 Resp: 277921
Ion Ratio Lower Upper
69 100
41 75.4 60.8 91.2
39 47.2 39.0 58.6





#57

1,3-Dichloropropane

Concen: 109.813 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 272814

Ion Ratio Lower Upper

76 100





78 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 697.344 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 63 Resp: 680608

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 571.094 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 102859

Ion Ratio Lower Upper

43 100





58 50.2 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 122.915 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:129 Resp: 187497

Ion Ratio Lower Upper

129 100

127 77.1 39.3 117.8









#61
1,2-Dibromoethane
Concen: 115.087 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:107 Resp: 165634
Ion Ratio Lower Upper
107 100
109 94.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 109.174 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 95 Resp: 217536
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 64.4 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:117 Resp: 12855
Ion Ratio Lower Upper
117 100
82 61.4 50.6 76.0
119 31.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 100.918 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:164 Resp: 132488
Ion Ratio Lower Upper
164 100
166 123.3 105.0 157.6
129 101.6 83.5 125.3
131 97.3 76.5 114.7









#65
Chlorobenzene
Concen: 108.615 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:112 Resp: 429650
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 115.333 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion:131 Resp: 152173
Ion Ratio Lower Upper
131 100
133 95.7 47.3 141.9
119 64.0 31.6 95.0





#67
Ethyl Benzene
Concen: 116.421 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 91 Resp: 78512
Ion Ratio Lower Upper
91 100
106 29.9 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 234.349 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:106 Resp: 570913
Ion Ratio Lower Upper
106 100
91 214.2 171.2 256.8









#69
o-Xylene
Concen: 113.990 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:106 Resp: 279900
Ion Ratio Lower Upper
106 100
91 226.9 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 120.694 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion:104 Resp: 474254
Ion Ratio Lower Upper
104 100
78 57.2 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 127.128 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:173 Resp: 125980
Ion Ratio Lower Upper
173 100
175 47.3 23.4 70.0
254 0.1 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion:152 Resp: 60345
Ion Ratio Lower Upper
152 100
115 150.7 46.9 140.7#
150 211.6 0.0 351.0





#73
Isopropylbenzene
Concen: 114.312 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:105 Resp: 75233
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 120.944 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 396877
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 22.5 18.7 28.1
61 22.5 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 104.018 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 83 Resp: 24353
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 64.4 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 84.960 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 75 Resp: 213744
Ion Ratio Lower Upper
75 100
77 42.4 20.5 61.5





#77
Bromobenzene
Concen: 110.297 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:156 Resp: 167934
Ion Ratio Lower Upper
156 100
77 187.1 92.8 278.3
158 97.3 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 119.448 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 881218
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 107.514 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 534560

Ion Ratio Lower Upper

91 100





126 31.0 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 112.825 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 616477

Ion Ratio Lower Upper

105 100

120 46.0 23.1 69.2





#81

trans-1,4-Dichloro-2-butene

Concen: 134.609 ug/l

RT: 11.018 min Scan# 1629

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 75 Resp: 81372

Ion Ratio Lower Upper

75 100

53 114.9 100.5 150.7





89 46.5 37.0 55.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#82

4-Chlorotoluene

Concen: 111.948 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 91 Resp: 611736

Ion Ratio Lower Upper

91 100

126 26.7 13.3 39.8





#83

tert-Butylbenzene

Concen: 115.065 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:119 Resp: 61747

Ion Ratio Lower Upper

119 100

91 65.6 32.9 98.7





134 22.6 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 114.236 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 623456

Ion Ratio Lower Upper

105 100

120 43.1 21.2 63.6









#85
sec-Butylbenzene
Concen: 118.023 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:105 Resp: 78629
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 121.113 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:119 Resp: 651531
Ion Ratio Lower Upper
119 100
134 25.4 12.5 37.5
91 27.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 110.394 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:146 Resp: 313100

Ion Ratio Lower Upper

146 100

111 44.5 22.1 66.3





148 64.0 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 112.167 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:146 Resp: 311752


Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.9

148 63.9 31.9 95.5





#89

n-Butylbenzene

Concen: 129.533 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 60578

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 123.091 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:117 Resp: 123188

Ion Ratio Lower Upper

117 100


201 63.2 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 111.506 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:146 Resp: 308039
Ion Ratio Lower Upper
146 100
111 46.3 22.4 67.3
148 63.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 120.915 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 64438
Ion Ratio Lower Upper
75 100
155 69.6 34.9 104.8
157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 134.366 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:180 Resp: 20325

Ion Ratio Lower Upper

180 100

182 94.1 47.9 143.8





145 34.5 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 113.857 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:225 Resp: 80558


Ion Ratio Lower Upper

225 100

223 62.5 30.8 92.4


227 63.1 30.9 92.7





#95
Naphthalene
Concen: 130.929 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:128 Resp: 721236
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 11.0 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 126.331 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:180 Resp: 200379
Ion Ratio Lower Upper
180 100
182 96.5 47.8 143.3
145 36.7 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	168484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	9067	3.143	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	6.280%#		
35) Dibromofluoromethane	5.373	113	5969	3.161	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	6.320%#		
50) Toluene-d8	8.647	98	20566	3.232	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	6.460%#		
62) 4-Bromofluorobenzene	11.079	95	7822	3.378	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	6.760%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	6195	2.894	ug/l	95
3) Chloromethane	1.307	50	6587	3.125	ug/l	99
4) Vinyl Chloride	1.374	62	5999	3.249	ug/l	96
5) Bromomethane	1.593	94	2962	3.164	ug/l	88
6) Chloroethane	1.666	64	3567	3.750	ug/l	94
7) Trichlorofluoromethane	1.874	101	9599	3.413	ug/l	88
8) Diethyl Ether	2.130	74	3020	3.266	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	5911	3.517	ug/l	99
10) Methyl Iodide	2.441	142	5900	2.990	ug/l	99
11) Tert butyl alcohol	2.965	59	5539	16.111	ug/l	98
12) 1,1-Dichloroethene	2.306	96	5496	3.359	ug/l	97
13) Acrolein	2.233	56	5673	14.206	ug/l	95
14) Allyl chloride	2.654	41	10262	3.319	ug/l	98
15) Acrylonitrile	3.062	53	16746	16.282	ug/l	97
16) Acetone	2.380	43	19773	20.036	ug/l	98
17) Carbon Disulfide	2.501	76	11068	2.919	ug/l	96
18) Methyl Acetate	2.703	43	7909	3.357	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	18497	3.241	ug/l	97
20) Methylene Chloride	2.782	84	6680	3.341	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	5402	3.234	ug/l	100
22) Diisopropyl ether	3.751	45	18657	3.230	ug/l #	65
23) Vinyl Acetate	3.715	43	82328	16.127	ug/l	99
24) 1,1-Dichloroethane	3.605	63	11193	3.297	ug/l	99
25) 2-Butanone	4.562	43	26125	18.340	ug/l	98
26) 2,2-Dichloropropane	4.465	77	8245	3.205	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	6222	3.089	ug/l	95
28) Bromochloromethane	4.891	49	5360	2.969	ug/l #	100
29) Tetrahydrofuran	5.001	42	15431	16.680	ug/l	98
30) Chloroform	5.080	83	11624	3.294	ug/l	89
31) Cyclohexane	5.458	56	10269	3.588	ug/l	96
32) 1,1,1-Trichloroethane	5.367	97	9827	3.229	ug/l	97
36) 1,1-Dichloropropene	5.678	75	7787	3.551	ug/l	98
37) Ethyl Acetate	4.721	43	9806m	3.565	ug/l	
38) Carbon Tetrachloride	5.659	117	8505	3.376	ug/l	97
39) Methylcyclohexane	7.373	83	9882	3.597	ug/l	93
40) Benzene	6.031	78	22528	3.315	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	4851	3.174	ug/l #	65
42) 1,2-Dichloroethane	6.080	62	10011	3.563	ug/l	100
43) Isopropyl Acetate	6.348	43	13914	3.309	ug/l	96
44) Trichloroethene	7.123	130	5313	3.307	ug/l	94
45) 1,2-Dichloropropane	7.427	63	5451	3.228	ug/l #	89
46) Dibromomethane	7.580	93	4417	3.318	ug/l	98
47) Bromodichloromethane	7.818	83	8394	3.254	ug/l	98
48) Methyl methacrylate	7.696	41	7171	3.321	ug/l	97
49) 1,4-Dioxane	7.659	88	2907	67.721	ug/l	96
51) 4-Methyl-2-Pentanone	8.567	43	46794	16.967	ug/l	98
52) Toluene	8.714	92	14126	3.505	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	6834	3.155	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	7908	3.129	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	5685	3.490	ug/l	98
56) Ethyl methacrylate	9.116	69	8555	3.390	ug/l	98
57) 1,3-Dichloropropane	9.305	76	10127	3.508	ug/l	94
58) 2-Chloroethyl Vinyl ether	8.238	63	20809	18.346	ug/l	99
59) 2-Hexanone	9.427	43	34853	16.651	ug/l	97
60) Dibromochloromethane	9.518	129	5500	3.102	ug/l	98
61) 1,2-Dibromoethane	9.604	107	5614	3.356	ug/l	99
64) Tetrachloroethene	9.269	164	4752	3.162	ug/l	96
65) Chlorobenzene	10.079	112	15391	3.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	5024	3.326	ug/l	99
67) Ethyl Benzene	10.195	91	26732	3.463	ug/l	99
68) m/p-Xylenes	10.299	106	19957	7.156	ug/l	98
69) o-Xylene	10.640	106	9398	3.343	ug/l	95
70) Styrene	10.652	104	14900	3.312	ug/l	99
71) Bromoform	10.799	173	3470	3.059	ug/l #	99
73) Isopropylbenzene	10.957	105	24201	3.266	ug/l	98
74) N-amyl acetate	10.841	43	11224	3.038	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	9170	3.479	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	7928m	2.799	ug/l	
77) Bromobenzene	11.195	156	5857	3.417	ug/l	96
78) n-propylbenzene	11.299	91	28436	3.424	ug/l	100
79) 2-Chlorotoluene	11.360	91	18672	3.336	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	20742	3.372	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	1828	2.686	ug/l	99
82) 4-Chlorotoluene	11.451	91	19966	3.245	ug/l	98
83) tert-Butylbenzene	11.713	119	21049	3.484	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	20613	3.355	ug/l	97
85) sec-Butylbenzene	11.890	105	25590	3.412	ug/l	99
86) p-Isopropyltoluene	12.006	119	20955	3.460	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10585	3.315	ug/l	97
88) 1,4-Dichlorobenzene	12.036	146	10908	3.486	ug/l	90
89) n-Butylbenzene	12.329	91	18004	3.419	ug/l	94
90) Hexachloroethane	12.536	117	3553	3.153	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	10712	3.444	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1687	2.812	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	5720	3.359	ug/l	96
94) Hexachlorobutadiene	13.719	225	2816	3.535	ug/l	96
95) Naphthalene	13.774	128	19902	3.209	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	5745	3.217	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On : 05 May 2025 16:04
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On    : 05 May 2025  16:04
Operator  : JC/MD
Sample    : VSTDIC005
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 10   Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:168 Resp: 9696
Ion Ratio Lower Upper
168 100
99 74.0 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 2.894 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 85 Resp: 6195
Ion Ratio Lower Upper
85 100
87 28.7 15.7 47.1









#3
Chloromethane
Concen: 3.125 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 50 Resp: 658
Ion Ratio Lower Upper
50 100
52 32.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 3.249 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 62 Resp: 5999
Ion Ratio Lower Upper
62 100
64 29.1 25.2 37.8





#5
Bromomethane
Concen: 3.164 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 94 Resp: 2962
Ion Ratio Lower Upper
94 100
96 102.2 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 3.750 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 64 Resp: 3567
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5





#7
Trichlorofluoromethane
Concen: 3.413 ug/l
RT: 1.874 min Scan# 110
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:101 Resp: 959
Ion Ratio Lower Upper
101 100
103 54.0 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 3.266 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 74 Resp: 3020
Ion Ratio Lower Upper
74 100
45 115.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 3.517 ug/l
RT: 2.325 min Scan# 202
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:101 Resp: 591
Ion Ratio Lower Upper
101 100
85 46.3 38.6 58.0
151 68.8 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 2.990 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:142 Resp: 5900
Ion Ratio Lower Upper
142 100
127 52.8 41.7 62.5
141 14.3 11.5 17.3





#11

Tert butyl alcohol

Concen: 16.111 ug/l

RT: 2.965 min Scan# 30

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 59 Resp: 553

Ion Ratio Lower Upper

59 100





57 10.0 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 3.359 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 96 Resp: 5496


Ion Ratio Lower Upper

96 100

61 178.7 146.2 219.2

98 59.5 51.0 76.6





#13

Acrolein

Concen: 14.206 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 567

Ion Ratio Lower Upper

56 100





55 74.5 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carbone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 3.319 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 41 Resp: 10262


Ion Ratio Lower Upper

41 100

39 74.1 60.6 90.8

76 29.1 24.9 37.3





#15

Acrylonitrile

Concen: 16.282 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	53	Resp:	16740
Ion	Ratio	Lower	Upper
53	100		
52	84.3	65.3	97.9
51	38.5	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 20.036 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	43	Resp:	19773
Ion	Ratio	Lower	Upper
43	100		
58	27.7	21.2	31.8





#17

Carbon Disulfide

Concen: 2.919 ug/l

RT: 2.501 min Scan# 211

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 11068

Ion Ratio Lower Upper

76 100





78 10.2 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 3.357 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 43 Resp: 7909

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 3.241 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04

Tgt Ion: 73 Resp: 1849

Ion Ratio Lower Upper

73 100

57 20.5 17.7 26.5



Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 3.341 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 84 Resp: 6680




Ion Ratio Lower Upper


84 100

49 137.0 113.9 170.9

51 43.8 33.5 50.3


86 67.6 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 3.234 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 96 Resp: 540
Ion Ratio Lower Upper
96 100
61 149.9 119.5 179.3
98 62.9 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 3.230 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 45 Resp: 18657
Ion Ratio Lower Upper
45 100
43 128.0 66.6 100.0#
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
Vinyl Acetate
Concen: 16.127 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 82328
Ion Ratio Lower Upper
43 100
86 9.1 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 3.297 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 63 Resp: 11193
Ion Ratio Lower Upper
63 100
98 6.9 3.6 10.8
100 4.7 2.1 6.3





#25

2-Butanone

Concen: 18.340 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 2612

Ion Ratio Lower Upper

43 100





72 21.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 3.205 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 77 Resp: 8245

Ion Ratio Lower Upper

77 100

97 23.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 3.089 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 96 Resp: 622

Ion Ratio Lower Upper

96 100

61 168.5 0.0 322.8





98 67.6 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 2.969 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 49 Resp: 5360


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 16.680 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 42 Resp: 1543

Ion Ratio Lower Upper

42 100

72 42.2 33.7 50.5





71 37.4 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 3.294 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 83 Resp: 11624

Ion Ratio Lower Upper

83 100

85 69.9 49.3 73.9





#31

Cyclohexane

Concen: 3.588 ug/l

RT: 5.458 min Scan# 718

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 10269

Ion Ratio Lower Upper

56 100

69 31.6 24.4 36.6





84 79.1 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#32

1,1,1-Trichloroethane

Concen: 3.229 ug/l

RT: 5.367 min Scan# 702

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 97 Resp: 9827


Ion Ratio Lower Upper

97 100

99 62.6 51.8 77.6

61 51.8 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 3.143 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 906

Ion Ratio Lower Upper

65 100





67 50.1 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:114 Resp: 168484


Ion Ratio Lower Upper

114 100

63 24.5 0.0 49.2

88 17.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 3.161 ug/l

RT: 5.373 min Scan# 704

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	113	Resp:	596
Ion	Ratio	Lower	Upper
113	100		
111	104.4	83.1	124.7
192	17.5	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 3.551 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	75	Resp:	7787
Ion	Ratio	Lower	Upper
75	100		
110	30.6	16.3	48.9
77	30.5	24.3	36.5





#37

Ethyl Acetate

Concen: 3.565 ug/l m

RT: 4.721 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 9800

Ion Ratio Lower Upper

43 100

61 11.0 10.3 15.5





70 8.4 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 3.376 ug/l

RT: 5.659 min Scan# 750

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:117 Resp: 8505


Ion Ratio Lower Upper

117 100

119 90.6 75.2 112.8

121 29.5 24.2 36.4





#39

Methylcyclohexane

Concen: 3.597 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 83	Resp: 988
Ion Ratio	Lower Upper
83	100
55	73.3 64.7 97.1
98	50.3 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 3.315 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion: 78	Resp: 22528
Ion Ratio	Lower Upper
78	100
77	23.9 19.0 28.4





#41
Methacrylonitrile
Concen: 3.174 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 41 Resp: 485
Ion Ratio Lower Upper
41 100
39 59.7 47.2 70.8
67 0.0 50.7 76.1
52 28.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 3.563 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 62 Resp: 10011
Ion Ratio Lower Upper
62 100
98 7.8 0.0 15.2





#43

Isopropyl Acetate

Concen: 3.309 ug/l

RT: 6.348 min Scan# 863

Delta R.T. 0.012 min

Lab File: VX046046.D


Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005






Tgt Ion:	43	Resp:	13914
Ion Ratio	100	Lower	Upper
61	16.2	14.3	21.5
87	10.6	9.5	14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 3.307 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	130	Resp:	5313
Ion Ratio	100	Lower	Upper
95	108.1	0.0	204.2





#45

1,2-Dichloropropane

Concen: 3.228 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 545

Ion Ratio Lower Upper

63 100





65 36.2 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 3.318 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 93 Resp: 4417


Ion Ratio Lower Upper

93 100

95 80.1 65.6 98.4

174 87.6 68.2 102.2





#47

Bromodichloromethane

Concen: 3.254 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	8394
Ion	Ratio	Lower	Upper
83	100		
85	61.9	50.5	75.7
127	9.0	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 3.321 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	41	Resp:	7171
Ion	Ratio	Lower	Upper
41	100		
69	68.4	58.5	87.7
39	65.1	51.7	77.5





#49

1,4-Dioxane

Concen: 67.721 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	88	Resp:	290
Ion	Ratio	Lower	Upper
88	100		
43	45.3	33.4	50.2
58	71.0	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 3.232 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	98	Resp:	20566
Ion	Ratio	Lower	Upper
98	100		
100	66.1	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 16.967 ug/l
RT: 8.567 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 46794
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 3.505 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 92 Resp: 14126
Ion Ratio Lower Upper
92 100
91 167.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 3.155 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 75 Resp: 6834

Ion Ratio Lower Upper

75 100





77 31.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 3.129 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 7908


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 55.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 3.490 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 568

Ion Ratio Lower Upper

97 100

83 88.0 69.0 103.6

85 54.4 44.3 66.5





99 62.2 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 3.390 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 69 Resp: 8555


Ion Ratio Lower Upper

69 100

41 73.7 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 3.508 ug/l

RT: 9.305 min Scan# 1134

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 1012

Ion Ratio Lower Upper

76 100





78 29.5 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 18.346 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 63 Resp: 20809

Ion Ratio Lower Upper

63 100

106 26.4 21.5 32.3





#59

2-Hexanone

Concen: 16.651 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005

Tgt Ion: 43 Resp: 3485

Ion Ratio Lower Upper

43 100





58 47.5 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 3.102 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:129 Resp: 5500

Ion Ratio Lower Upper

129 100

127 80.4 39.3 117.8









#61
1,2-Dibromoethane
Concen: 3.356 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:107 Resp: 5614
Ion Ratio Lower Upper
107 100
109 91.0 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 3.378 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 95 Resp: 7822
Ion Ratio Lower Upper
95 100
174 65.0 0.0 135.8
176 66.6 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:117 Resp: 14716
Ion Ratio Lower Upper
117 100
82 61.1 50.6 76.0
119 32.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 3.162 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:164 Resp: 4752
Ion Ratio Lower Upper
164 100
166 127.8 105.0 157.6
129 105.8 83.5 125.3
131 104.1 76.5 114.7









#65
Chlorobenzene
Concen: 3.399 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:112 Resp: 1539
Ion Ratio Lower Upper
112 100
114 32.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 3.326 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:131 Resp: 5024
Ion Ratio Lower Upper
131 100
133 95.0 47.3 141.9
119 64.8 31.6 95.0





#67

Ethyl Benzene

Concen: 3.463 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 2673

Ion Ratio Lower Upper

91 100





106 29.9 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 7.156 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:106 Resp: 19957

Ion Ratio Lower Upper

106 100

91 210.1 171.2 256.8









#69
o-Xylene
Concen: 3.343 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:106 Resp: 9398
Ion Ratio Lower Upper
106 100
91 233.6 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 3.312 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:104 Resp: 14900
Ion Ratio Lower Upper
104 100
78 57.9 45.7 68.5
103 55.6 43.7 65.5





#71
Bromoform
Concen: 3.059 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:173 Resp: 3470
Ion Ratio Lower Upper
173 100
175 45.8 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:152 Resp: 67939
Ion Ratio Lower Upper
152 100
115 66.8 46.9 140.7
150 157.5 0.0 351.0









#73
Isopropylbenzene
Concen: 3.266 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 105 Resp: 2420
Ion Ratio Lower Upper
105 100
120 26.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 3.038 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 43 Resp: 11224
Ion Ratio Lower Upper
43 100
70 35.4 30.9 46.3
55 25.0 18.7 28.1
61 19.6 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 3.479 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 9170
Ion Ratio Lower Upper
83 100
131 8.3 5.0 14.9
85 61.7 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 2.799 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 75 Resp: 7928
Ion Ratio Lower Upper
75 100
77 40.6 20.5 61.5





#77
Bromobenzene
Concen: 3.417 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:156 Resp: 585
Ion Ratio Lower Upper
156 100
77 177.0 92.8 278.3
158 97.7 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 3.424 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 28436
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 3.336 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 18672
Ion Ratio Lower Upper
91 100
126 29.7 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 3.372 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20742
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 2.686 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 75 Resp: 1828
Ion Ratio Lower Upper
75 100
53 125.8 100.5 150.7
89 47.9 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 3.245 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 19966
Ion Ratio Lower Upper
91 100
126 27.3 13.3 39.8





#83

tert-Butylbenzene

Concen: 3.484 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	119	Resp:	21049
Ion Ratio	Lower	Upper	
119	100		
91	67.0	32.9	98.7
134	23.8	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 3.355 ug/l




RT: 11.750 min Scan# 1749


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion:	105	Resp:	20613
Ion Ratio	Lower	Upper	
105	100		
120	44.5	21.2	63.6





#85
sec-Butylbenzene
Concen: 3.412 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005




Tgt Ion:105 Resp: 25590
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#86
p-Isopropyltoluene
Concen: 3.460 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Tgt Ion:119 Resp: 20955
Ion Ratio Lower Upper
119 100
134 25.6 12.5 37.5
91 28.0 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 3.315 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion:146 Resp: 1058
Ion Ratio Lower Upper
146 100
111 46.2 22.1 66.3
148 62.4 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#88
1,4-Dichlorobenzene
Concen: 3.486 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04



Tgt Ion:146 Resp: 10908
Ion Ratio Lower Upper
146 100
111 52.1 21.3 63.9
148 69.4 31.9 95.5





#89

n-Butylbenzene

Concen: 3.419 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	91	Resp:	1800
Ion	Ratio	Lower	Upper
91	100		
92	49.5	26.9	80.7
134	21.2	11.8	35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 3.153 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion:	117	Resp:	3553
Ion	Ratio	Lower	Upper
117	100		
201	58.5	31.6	94.7





#91
1,2-Dichlorobenzene
Concen: 3.444 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:146 Resp: 10712
Ion Ratio Lower Upper
146 100
111 44.7 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 2.812 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 75 Resp: 1687
Ion Ratio Lower Upper
75 100
155 71.7 34.9 104.8
157 87.0 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 3.359 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:180 Resp: 5720
Ion Ratio Lower Upper
180 100
182 91.6 47.9 143.8
145 35.7 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 3.535 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:225 Resp: 2816
Ion Ratio Lower Upper
225 100
223 65.1 30.8 92.4
227 58.8 30.9 92.7





#95
Naphthalene
Concen: 3.209 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:128 Resp: 1990
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 11.5 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 3.217 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:180 Resp: 5745
Ion Ratio Lower Upper
180 100
182 99.1 47.8 143.3
145 34.8 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	1061m	0.619	ug/l	
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1956m	0.717	ug/l	
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1235m	0.577	ug/l	
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1661m	0.674	ug/l	
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	2149m	0.789	ug/l	
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion:168 Resp: 92040
Ion Ratio Lower Upper
168 100
99 71.5 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 0.596 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 85 Resp: 1211
Ion Ratio Lower Upper
85 100
87 32.4 15.7 47.1









#3
Chloromethane
Concen: 0.639 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 50 Resp: 1278
Ion Ratio Lower Upper
50 100
52 30.3 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 0.707 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 62 Resp: 1239
Ion Ratio Lower Upper
62 100
64 41.7 25.2 37.8#









#6
Chloroethane
Concen: 0.951 ug/l
RT: 1.678 min Scan# 91
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 64 Resp: 859
Ion Ratio Lower Upper
64 100
66 35.2 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#7
Trichlorofluoromethane
Concen: 0.734 ug/l
RT: 1.880 min Scan# 130
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1959
Ion Ratio Lower Upper
101 100
103 67.8 51.0 76.4









#8
Diethyl Ether
Concen: 0.845 ug/l
RT: 2.130 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 74 Resp: 74
Ion Ratio Lower Upper
74 100
45 123.6 54.9 164.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.731 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1166
Ion Ratio Lower Upper
101 100
85 47.3 38.6 58.0
151 73.1 55.2 82.8





#12

1,1-Dichloroethene

Concen: 0.704 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	96	Resp:	109
Ion	Ratio	Lower	Upper
96	100		
61	199.6	146.2	219.2
98	49.6	51.0	76.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 0.660 ug/l




RT: 2.654 min Scan# 257


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	41	Resp:	1936
Ion	Ratio	Lower	Upper
41	100		
39	85.5	60.6	90.8
76	26.4	24.9	37.3





#15

Acrylonitrile

Concen: 3.424 ug/l

RT: 3.075 min Scan# 31

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 53	Resp: 334
Ion Ratio	Lower Upper
53 100	
52 82.9	65.3 97.9
51 34.4	29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 3.736 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 43	Resp: 3500
Ion Ratio	Lower Upper
43 100	
58 28.1	21.2 31.8





#17

Carbon Disulfide

Concen: 0.728 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 76 Resp: 2619

Ion Ratio Lower Upper

76 100





78 8.0 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 0.828 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 1852

Ion Ratio Lower Upper

43 100

74 22.0 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 0.662 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 73 Resp: 3588
Ion Ratio Lower Upper
73 100
57 20.3 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 0.828 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 84 Resp: 1571
Ion Ratio Lower Upper
84 100
49 140.0 113.9 170.9
51 51.3 33.5 50.3#
86 85.2 53.8 80.8#





#21
trans-1,2-Dichloroethene
Concen: 0.701 ug/l
RT: 3.081 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 96 Resp: 111
Ion Ratio Lower Upper
96 100
61 119.0 119.5 179.3
98 30.1 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 0.703 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 45 Resp: 3857
Ion Ratio Lower Upper
45 100
43 176.2 66.6 100.0#
87 28.9 19.8 29.6
59 9.1 8.6 12.8





#23
Vinyl Acetate
Concen: 3.154 ug/l
RT: 3.727 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 43 Resp: 1528
Ion Ratio Lower Upper
43 100
86 10.8 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 0.637 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 63 Resp: 2054
Ion Ratio Lower Upper
63 100
98 6.3 3.6 10.8
100 5.8 2.1 6.3





#25

2-Butanone

Concen: 3.371 ug/l

RT: 4.593 min Scan# 51

Delta R.T. 0.037 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 4558

Ion Ratio Lower Upper

43 100





72 19.2 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 0.728 ug/l

RT: 4.458 min Scan# 553

Delta R.T. -0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 77 Resp: 1777

Ion Ratio Lower Upper

77 100

97 13.7 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 0.693 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 96	Resp: 1324
Ion Ratio	Lower Upper
96 100	
61 145.5	0.0 322.8
98 63.1	0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 0.619 ug/l m




RT: 4.910 min Scan# 627


Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 49	Resp: 1061
Ion Ratio	Lower Upper
49 100	
129 0.0	0.0 4.0
130 65.3	56.2 84.2





#29

Tetrahydrofuran

Concen: 3.338 ug/l

RT: 5.019 min Scan# 64

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 42 Resp: 293

Ion Ratio Lower Upper

42 100

72 41.9 33.7 50.5





71 37.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 0.695 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 83 Resp: 2328

Ion Ratio Lower Upper

83 100

85 60.4 49.3 73.9





#32

1,1,1-Trichloroethane

Concen: 0.647 ug/l

RT: 5.379 min Scan# 70

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 97 Resp: 186

Ion Ratio Lower Upper

97 100

99 0.0 51.8 77.6





61 55.0 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 5.95 min

Lab File: VX046047.D


Acq: 05 May 2025 16:27


Tgt Ion: 65

Sig Exp Ratio

65 100

67 49.5





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion:114 Resp: 16702

Ion Ratio Lower Upper

114 100

63 24.2 0.0 49.2





88 17.1 0.0 33.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 5.38 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 113


Sig Exp Ratio

113 100

111 103.9

192 16.6





#36

1,1-Dichloropropene

Concen: 0.758 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	75	Resp:	164
Ion	Ratio	Lower	Upper
75	100		
110	28.2	16.3	48.9
77	15.8	24.3	36.5

Manual Integrations

APPROVED

Reviewed By :John Carbone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#37

Ethyl Acetate

Concen: 0.717 ug/l m




RT: 4.733 min Scan# 598


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	43	Resp:	1956
Ion	Ratio	Lower	Upper
43	100		
61	0.0	10.3	15.5#
70	1.0	7.9	11.9#





#38

Carbon Tetrachloride

Concen: 0.724 ug/l

RT: 5.678 min Scan# 752

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:117 Resp: 180

Ion Ratio Lower Upper

117 100

119 67.0 75.2 112.8





121 28.5 24.2 36.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#39

Methylcyclohexane

Concen: 0.770 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 83 Resp: 2096


Ion Ratio Lower Upper

83 100

55 80.1 64.7 97.1

98 41.9 37.4 56.2





#40

Benzene

Concen: 0.668 ug/l

RT: 6.044 min Scan# 811

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 450

Ion Ratio Lower Upper

78 100





77 29.1 19.0 28.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#41

Methacrylonitrile

Concen: 0.513 ug/l

RT: 4.946 min Scan# 633

Delta R.T. 0.030 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 41 Resp: 778




Ion Ratio Lower Upper


41 100

39 51.7 47.2 70.8

67 0.0 50.7 76.1#

52 13.6 24.6 37.0#





#42

1,2-Dichloroethane

Concen: 0.694 ug/l

RT: 6.092 min Scan# 819

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 62 Resp: 1934

Ion Ratio Lower Upper

62 100





98 1.1 0.0 15.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#43

Isopropyl Acetate

Concen: 0.612 ug/l

RT: 6.354 min Scan# 864

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 43 Resp: 2553


Ion Ratio Lower Upper

43 100

61 10.7 14.3 21.5#


87 3.2 9.5 14.3#





#44
Trichloroethene
Concen: 0.680 ug/l
RT: 7.135 min Scan# 990
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:130 Resp: 1083
Ion Ratio Lower Upper
130 100
95 117.3 0.0 204.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#45
1,2-Dichloropropane
Concen: 0.633 ug/l
RT: 7.440 min Scan# 1042
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion: 63 Resp: 1059
Ion Ratio Lower Upper
63 100
65 36.1 24.1 36.1





#46

Dibromomethane

Concen: 0.667 ug/l

RT: 7.586 min Scan# 1106

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 93	Resp: 880
Ion Ratio	Lower Upper
93	100
95	84.5 65.6 98.4
174	88.5 68.2 102.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#47

Bromodichloromethane

Concen: 0.633 ug/l




RT: 7.818 min Scan# 1104


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 83	Resp: 1620
Ion Ratio	Lower Upper
83	100
85	57.3 50.5 75.7
127	0.0 6.5 9.7





#48

Methyl methacrylate

Concen: 0.577 ug/l

RT: 7.714 min Scan# 1083

Delta R.T. 0.024 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	41	Resp:	123
Ion	Ratio	Lower	Upper
41	100		
69	70.6	58.5	87.7
39	84.8	51.7	77.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#49

1,4-Dioxane

Concen: 11.703 ug/l




RT: 7.677 min Scan# 1081


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	88	Resp:	498
Ion	Ratio	Lower	Upper
88	100		
43	44.4	33.4	50.2
58	30.3	58.6	88.0





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 8.65 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 98
Sig Exp Ratio
98 100
100 66.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#51
4-Methyl-2-Pentanone
Concen: 3.430 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 9378
Ion Ratio Lower Upper
43 100
58 34.4 28.9 43.3









#52
Toluene
Concen: 0.672 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 92 Resp: 2684
Ion Ratio Lower Upper
92 100
91 165.7 136.6 205.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#53
t-1,3-Dichloropropene
Concen: 0.576 ug/l
RT: 8.988 min Scan# 1296
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 75 Resp: 1238
Ion Ratio Lower Upper
75 100
77 42.7 25.0 37.6#





#54
cis-1,3-Dichloropropene
Concen: 0.564 ug/l
RT: 8.379 min Scan# 1196
Delta R.T. 0.018 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 75 Resp: 1414
Ion Ratio Lower Upper
75 100
77 44.2 25.4 38.0
39 59.0 52.2 78.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#55
1,1,2-Trichloroethane
Concen: 0.637 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 97 Resp: 1029
Ion Ratio Lower Upper
97 100
83 74.8 69.0 103.6
85 59.5 44.3 66.5
99 63.0 47.8 71.6





#56

Ethyl methacrylate

Concen: 0.504 ug/l

RT: 9.128 min Scan# 1319

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 69 Resp: 1260

Ion Ratio Lower Upper

69 100

41 86.6 60.8 91.2





39 78.0 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#57

1,3-Dichloropropane

Concen: 0.711 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 76 Resp: 2036

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5









#58
2-Chloroethyl Vinyl ether
Concen: 3.420 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 63 Resp: 3840
Ion Ratio Lower Upper
63 100
106 24.0 21.5 32.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#59
2-Hexanone
Concen: 3.096 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 6424
Ion Ratio Lower Upper
43 100
58 41.6 24.9 74.6





#60
Dibromochloromethane
Concen: 0.582 ug/l
RT: 9.525 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:129 Resp: 102
Ion Ratio Lower Upper
129 100
127 72.5 39.3 117.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#61
1,2-Dibromoethane
Concen: 0.650 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:107 Resp: 1077
Ion Ratio Lower Upper
107 100
109 94.0 73.3 109.9





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 95
Sig Exp Ratio
95 100
174 67.9
176 65.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion:117 Resp: 140108
Ion Ratio Lower Upper
117 100
82 61.0 50.6 76.0
119 32.1 25.8 38.6





#64
Tetrachloroethene
Concen: 0.679 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:164 Resp: 972
Ion Ratio Lower Upper
164 100
166 119.1 105.0 157.6
129 103.4 83.5 125.3
131 95.4 76.5 114.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#65
Chlorobenzene
Concen: 0.735 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion:112 Resp: 3170
Ion Ratio Lower Upper
112 100
114 30.6 25.4 38.2





#66
1,1,1,2-Tetrachloroethane
Concen: 0.720 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:131 Resp: 1035
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 0.0 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#67
Ethyl Benzene
Concen: 0.687 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 5052
Ion Ratio Lower Upper
91 100
106 32.5 23.4 35.2









#68
m/p-Xylenes
Concen: 1.367 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion:106 Resp: 3630
Ion Ratio Lower Upper
106 100
91 209.8 171.2 256.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#69
o-Xylene
Concen: 0.672 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion:106 Resp: 1799
Ion Ratio Lower Upper
106 100
91 214.9 112.7 338.1





#70
Styrene
Concen: 0.622 ug/l
RT: 10.659 min Scan# 1569
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:104 Resp: 2664
Ion Ratio Lower Upper
104 100
78 54.3 45.7 68.5
103 53.9 43.7 65.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#71
Bromoform
Concen: 0.608 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:173 Resp: 657
Ion Ratio Lower Upper
173 100
175 46.6 23.4 70.0
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:152 Resp: 5912

Ion Ratio Lower Upper

152 100

115 66.6 46.9 140.7





150 156.3 0.0 351.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#73

Isopropylbenzene

Concen: 0.695 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:105 Resp: 4480

Ion Ratio Lower Upper

105 100

120 23.7 12.8 38.4





#74

N-amy1 acetate

Concen: 0.631 ug/l

RT: 10.854 min Scan# 1602

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 43	Resp: 2028
Ion Ratio	Lower Upper
43 100	
70 31.5	30.9 46.3
55 24.5	18.7 28.1
61 15.4	17.1 25.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#75

1,1,2,2-Tetrachloroethane

Concen: 0.800 ug/l




RT: 11.213 min Scan# 1661


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 83	Resp: 1835
Ion Ratio	Lower Upper
83 100	
131 7.5	5.0 14.9
85 58.7	31.9 95.7









#76
1,2,3-Trichloropropane
Concen: 0.674 ug/l m
RT: 11.244 min Scan# 1665
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 75 Resp: 1665
Ion Ratio Lower Upper
75 100
77 29.7 20.5 61.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#77
Bromobenzene
Concen: 0.734 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:156 Resp: 1095
Ion Ratio Lower Upper
156 100
77 179.5 92.8 278.3
158 93.5 48.8 146.3





#78

n-propylbenzene

Concen: 0.699 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 5052

Ion Ratio Lower Upper

91 100





120 20.1 10.8 32.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#79

2-Chlorotoluene

Concen: 0.773 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3765

Ion Ratio Lower Upper

91 100

126 28.0 15.6 46.7





#80

1,3,5-Trimethylbenzene

Concen: 0.671 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 3590

Ion Ratio Lower Upper

105 100





120 43.4 23.1 69.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#82

4-Chlorotoluene

Concen: 0.713 ug/l

RT: 11.457 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3815

Ion Ratio Lower Upper

91 100

126 24.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 0.751 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	119	Resp:	395
Ion	Ratio	Lower	Upper
119	100		
91	63.9	32.9	98.7
134	25.2	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 0.697 ug/l




RT: 11.756 min Scan# 1750


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	105	Resp:	3725
Ion	Ratio	Lower	Upper
105	100		
120	40.3	21.2	63.6





#85

sec-Butylbenzene

Concen: 0.672 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 438

Ion Ratio Lower Upper

105 100





134 19.2 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 0.679 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:119 Resp: 3577


Ion Ratio Lower Upper

119 100

134 25.0 12.5 37.5

91 33.1 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 0.689 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:146	Resp:	1914
Ion Ratio	Lower	Upper
146	100	
111	43.9	22.1 66.3
148	65.3	32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 0.789 ug/l m




RT: 12.036 min Scan# 1796


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:146	Resp:	2149
Ion Ratio	Lower	Upper
146	100	
111	39.1	21.3 63.9
148	58.2	31.9 95.5





#89

n-Butylbenzene

Concen: 0.631 ug/l

RT: 12.335 min Scan# 1844

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001





Tgt Ion:	91	Resp:	288
Ion	Ratio	Lower	Upper
91	100		
92	48.1	26.9	80.7
134	20.1	11.8	35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 0.616 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion:	117	Resp:	604
Ion	Ratio	Lower	Upper
117	100		
201	64.2	31.6	94.7





#91
1,2-Dichlorobenzene
Concen: 0.747 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Tgt Ion:146 Resp: 202
Ion Ratio Lower Upper
146 100
111 48.8 22.4 67.3
148 59.4 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 0.586 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion: 75 Resp: 306
Ion Ratio Lower Upper
75 100
155 49.3 34.9 104.8
157 78.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 0.688 ug/l

RT: 13.591 min Scan# 2050

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:180	Resp:	1019
Ion Ratio	Lower	Upper
180	100	
182	100.6	47.9 143.8
145	29.5	17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 0.671 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion:225	Resp:	465
Ion Ratio	Lower	Upper
225	100	
223	69.7	30.8 92.4
227	70.5	30.9 92.7





#95
Naphthalene
Concen: 0.767 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:128 Resp: 4138
Ion Ratio Lower Upper
128 100
127 12.4 10.4 15.6
129 8.9 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 0.716 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion:180 Resp: 1113
Ion Ratio Lower Upper
180 100
182 92.5 47.8 143.3
145 32.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83733	48.509	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.020%		
35) Dibromofluoromethane	5.379	113	58192	49.683	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.360%		
50) Toluene-d8	8.647	98	198575	48.984	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.960%		
62) 4-Bromofluorobenzene	11.079	95	75568	48.596	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.200%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	78783	55.594	ug/l	99
3) Chloromethane	1.307	50	72275	52.592	ug/l	100
4) Vinyl Chloride	1.374	62	65994	51.598	ug/l	98
5) Bromomethane	1.593	94	30033	50.627	ug/l	97
6) Chloroethane	1.666	64	35690	52.271	ug/l	94
7) Trichlorofluoromethane	1.874	101	99352	52.560	ug/l	99
8) Diethyl Ether	2.130	74	31875	49.536	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59451	50.822	ug/l	99
10) Methyl Iodide	2.447	142	73940	53.419	ug/l	99
11) Tert butyl alcohol	2.977	59	61065	252.025	ug/l	99
12) 1,1-Dichloroethene	2.313	96	56097	51.097	ug/l	98
13) Acrolein	2.233	56	64157	232.505	ug/l	99
14) Allyl chloride	2.660	41	108757	51.833	ug/l	100
15) Acrylonitrile	3.063	53	181808	262.412	ug/l	99
16) Acetone	2.380	43	169923	245.518	ug/l	98
17) Carbon Disulfide	2.502	76	132695	50.981	ug/l	99
18) Methyl Acetate	2.703	43	80251	49.968	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	195442	50.777	ug/l	99
20) Methylene Chloride	2.782	84	63369	47.779	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	55770	50.513	ug/l	99
22) Diisopropyl ether	3.757	45	209946	51.800	ug/l	97
23) Vinyl Acetate	3.715	43	946746	265.585	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114733	50.825	ug/l	98
25) 2-Butanone	4.556	43	258339	257.110	ug/l	97
26) 2,2-Dichloropropane	4.471	77	88229	49.934	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	67727	50.956	ug/l	100
28) Bromochloromethane	4.891	49	55390	50.975	ug/l	98
29) Tetrahydrofuran	5.001	42	163675	259.958	ug/l	99
30) Chloroform	5.093	83	120782	51.332	ug/l	97
31) Cyclohexane	5.458	56	103158	50.147	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	104243	51.108	ug/l	100
36) 1,1-Dichloropropene	5.690	75	77931	49.521	ug/l	99
37) Ethyl Acetate	4.715	43	98604	50.714	ug/l	100
38) Carbon Tetrachloride	5.666	117	88100	49.825	ug/l	97
39) Methylcyclohexane	7.373	83	100808	49.757	ug/l	98
40) Benzene	6.031	78	233977	50.759	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	56017	55.076	ug/l	100
42) 1,2-Dichloroethane	6.080	62	101185	50.861	ug/l	98
43) Isopropyl Acetate	6.336	43	154818	52.195	ug/l	99
44) Trichloroethene	7.123	130	55491	50.018	ug/l	100
45) 1,2-Dichloropropane	7.428	63	59582	51.983	ug/l	98
46) Dibromomethane	7.580	93	45181	49.978	ug/l	99
47) Bromodichloromethane	7.818	83	92388	51.889	ug/l	100
48) Methyl methacrylate	7.690	41	82144	54.226	ug/l	98
49) 1,4-Dioxane	7.659	88	30477	1059.599	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	510206	259.133	ug/l	100
52) Toluene	8.714	92	144574	51.150	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	83751	52.920	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	92517	52.892	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	56713	50.887	ug/l	96
56) Ethyl methacrylate	9.116	69	95772	53.918	ug/l	99
57) 1,3-Dichloropropane	9.305	76	99663	49.793	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.238	63	240681	265.780	ug/l	99
59) 2-Hexanone	9.427	43	390701	268.218	ug/l	100
60) Dibromochloromethane	9.519	129	63738	52.075	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59134	51.050	ug/l	98
64) Tetrachloroethene	9.269	164	48034	46.795	ug/l	98
65) Chlorobenzene	10.073	112	154640	48.699	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	53735	49.557	ug/l	100
67) Ethyl Benzene	10.189	91	281239	50.245	ug/l	100
68) m/p-Xylenes	10.299	106	206813	101.022	ug/l	99
69) o-Xylene	10.640	106	101469	50.841	ug/l	99
70) Styrene	10.653	104	169703	51.907	ug/l	99
71) Bromoform	10.799	173	40364	49.583	ug/l #	97
73) Isopropylbenzene	10.957	105	269484	50.733	ug/l	100
74) N-amyl acetate	10.842	43	137559	52.407	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	89027	47.827	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	79763m	48.568	ug/l	
77) Bromobenzene	11.195	156	59871	48.549	ug/l	98
78) n-propylbenzene	11.299	91	312285	50.562	ug/l	100
79) 2-Chlorotoluene	11.360	91	195630	49.107	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	228538	51.500	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	24857	49.276	ug/l	98
82) 4-Chlorotoluene	11.451	91	223811	50.661	ug/l	100
83) tert-Butylbenzene	11.713	119	225783	50.511	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	231371	51.485	ug/l	100
85) sec-Butylbenzene	11.890	105	280578	51.122	ug/l	99
86) p-Isopropyltoluene	12.006	119	231942	51.198	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	113146	50.272	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	111173	48.368	ug/l	99
89) n-Butylbenzene	12.329	91	205858	51.804	ug/l	100
90) Hexachloroethane	12.536	117	39823	49.895	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111482	49.361	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21220	51.458	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	64421	49.662	ug/l	100
94) Hexachlorobutadiene	13.719	225	27425	48.408	ug/l	98
95) Naphthalene	13.774	128	239805	50.404	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	64950	48.525	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion:168 Resp: 9258

Ion Ratio Lower Upper

168 100





99 74.5 54.9 82.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#2

Dichlorodifluoromethane

Concen: 55.594 ug/l

RT: 1.167 min Scan# 13

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 85 Resp: 78783

Ion Ratio Lower Upper

85 100

87 32.2 15.7 47.1









#3
Chloromethane
Concen: 52.592 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 50 Resp: 72275
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 51.598 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 62 Resp: 65994
Ion Ratio Lower Upper
62 100
64 32.3 25.2 37.8





#5
Bromomethane
Concen: 50.627 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 94 Resp: 3003
Ion Ratio Lower Upper
94 100
96 93.5 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 52.271 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 64 Resp: 35690
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5





#7
Trichlorofluoromethane
Concen: 52.560 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVX050525



Tgt Ion:101 Resp: 99352
Ion Ratio Lower Upper
101 100
103 62.6 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 49.536 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 74 Resp: 31875
Ion Ratio Lower Upper
74 100
45 107.1 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 50.822 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	101	Resp:	5945
Ion	Ratio	Lower	Upper
101	100		
85	47.8	38.6	58.0
151	69.4	55.2	82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 53.419 ug/l




RT: 2.447 min Scan# 223


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	142	Resp:	73940
Ion	Ratio	Lower	Upper
142	100		
127	52.9	41.7	62.5
141	14.4	11.5	17.3





#11

Tert butyl alcohol

Concen: 252.025 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 59 Resp: 6106

Ion Ratio Lower Upper

59 100





57 11.1 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 51.097 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 96 Resp: 56097


Ion Ratio Lower Upper

96 100

61 179.7 146.2 219.2

98 62.1 51.0 76.6





#13

Acrolein

Concen: 232.505 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 56 Resp: 6415

Ion Ratio Lower Upper

56 100





55 70.8 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 51.833 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 41 Resp: 108757


Ion Ratio Lower Upper

41 100

39 75.6 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 262.412 ug/l

RT: 3.063 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 53 Resp: 181808

Ion Ratio Lower Upper

53 100

52 82.9 65.3 97.9





51 36.5 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 245.518 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 169923

Ion Ratio Lower Upper

43 100

58 27.3 21.2 31.8





#17

Carbon Disulfide

Concen: 50.981 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 76 Resp: 13269

Ion Ratio Lower Upper

76 100





78 9.1 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 49.968 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 80251

Ion Ratio Lower Upper

43 100

74 22.1 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 50.777 ug/l

RT: 3.111 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 73 Resp: 19544

Ion Ratio Lower Upper

73 100





57 21.8 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 47.779 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion: 84 Resp: 63369




Ion Ratio Lower Upper


84 100

49 142.7 113.9 170.9

51 43.3 33.5 50.3


86 66.8 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 50.513 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 96 Resp: 55770
Ion Ratio Lower Upper
96 100
61 149.3 119.5 179.3
98 64.6 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 51.800 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 45 Resp: 209946
Ion Ratio Lower Upper
45 100
43 86.9 66.6 100.0
87 24.8 19.8 29.6
59 10.8 8.6 12.8





#23

Vinyl Acetate

Concen: 265.585 ug/l

RT: 3.715 min Scan# 41

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 43 Resp: 946740

Ion Ratio Lower Upper

43 100





86 9.4 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#24

1,1-Dichloroethane

Concen: 50.825 ug/l

RT: 3.605 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 63 Resp: 114733


Ion Ratio Lower Upper

63 100

98 6.4 3.6 10.8

100 4.8 2.1 6.3





#25

2-Butanone

Concen: 257.110 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 43 Resp: 258339

Ion Ratio Lower Upper

43 100





72 24.3 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 49.934 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 77 Resp: 88229

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 50.956 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 96 Resp: 6772

Ion Ratio Lower Upper

96 100

61 160.7 0.0 322.8





98 64.4 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 50.975 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 49 Resp: 55390


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 68.7 56.2 84.2





#29

Tetrahydrofuran

Concen: 259.958 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 42 Resp: 163679

Ion Ratio Lower Upper

42 100

72 42.0 33.7 50.5





71 38.5 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 51.332 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 83 Resp: 120782

Ion Ratio Lower Upper

83 100


85 63.6 49.3 73.9





#31
Cyclohexane
Concen: 50.147 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 56 Resp: 103158
Ion Ratio Lower Upper
56 100
69 27.8 24.4 36.6
84 80.5 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 51.108 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 97 Resp: 104243
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.6
61 50.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.509 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 65 Resp: 8373

Ion Ratio Lower Upper

65 100





67 50.0 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:114 Resp: 162651


Ion Ratio Lower Upper

114 100

63 23.1 0.0 49.2

88 17.2 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.683 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	113	Resp:	5819
Ion	Ratio	Lower	Upper
113	100		
111	102.1	83.1	124.7
192	16.2	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 49.521 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	75	Resp:	77931
Ion	Ratio	Lower	Upper
75	100		
110	32.7	16.3	48.9
77	31.1	24.3	36.5





#37

Ethyl Acetate

Concen: 50.714 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	43	Resp:	98604
Ion	Ratio	Lower	Upper
43	100		
61	12.6	10.3	15.5
70	10.0	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 49.825 ug/l




RT: 5.666 min Scan# 751


Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	117	Resp:	88100
Ion	Ratio	Lower	Upper
117	100		
119	90.7	75.2	112.8
121	29.4	24.2	36.4





#39

Methylcyclohexane

Concen: 49.757 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 83 Resp: 100808

Ion Ratio Lower Upper

83 100

55 82.7 64.7 97.1





98 48.3 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 50.759 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 78 Resp: 233977

Ion Ratio Lower Upper

78 100

77 24.0 19.0 28.4









#41
Methacrylonitrile
Concen: 55.076 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVX050525

Tgt Ion:	41	Resp:	5601
Ion	Ratio	Lower	Upper
41	100		
39	58.7	47.2	70.8
67	63.7	50.7	76.1
52	30.5	24.6	37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 50.861 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:	62	Resp:	101185
Ion	Ratio	Lower	Upper
62	100		
98	7.1	0.0	15.2





#43

Isopropyl Acetate

Concen: 52.195 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 43 Resp: 154818

Ion Ratio Lower Upper

43 100

61 17.9 14.3 21.5





87 11.4 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 50.018 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:130 Resp: 55491

Ion Ratio Lower Upper

130 100

95 102.6 0.0 204.2





#45

1,2-Dichloropropane

Concen: 51.983 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 63 Resp: 5958

Ion Ratio Lower Upper

63 100





65 31.4 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 49.978 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 93 Resp: 45181


Ion Ratio Lower Upper

93 100

95 83.0 65.6 98.4

174 84.8 68.2 102.2





#47

Bromodichloromethane

Concen: 51.889 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	83	Resp:	9238
Ion	Ratio	Lower	Upper
83	100		
85	63.0	50.5	75.7
127	8.8	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 54.226 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	41	Resp:	82144
Ion	Ratio	Lower	Upper
41	100		
69	70.4	58.5	87.7
39	64.2	51.7	77.5





#49

1,4-Dioxane

Concen: 1059.599 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	88	Resp:	3047
Ion	Ratio	Lower	Upper
88	100		
43	39.9	33.4	50.2
58	76.2	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 48.984 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	98	Resp:	198575
Ion	Ratio	Lower	Upper
98	100		
100	65.6	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 259.133 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 43 Resp: 510200
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 51.150 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 92 Resp: 144574
Ion Ratio Lower Upper
92 100
91 171.4 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 52.920 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 75 Resp: 8375

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 52.892 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 75 Resp: 92517


Ion Ratio Lower Upper

75 100

77 31.5 25.4 38.0

39 63.2 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 50.887 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 97 Resp: 5671

Ion Ratio Lower Upper

97 100

83 89.8 69.0 103.6

85 58.4 44.3 66.5





99 63.9 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 53.918 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 69 Resp: 95772


Ion Ratio Lower Upper

69 100

41 76.1 60.8 91.2

39 47.4 39.0 58.6





#57

1,3-Dichloropropane

Concen: 49.793 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 76 Resp: 9966

Ion Ratio Lower Upper

76 100





78 31.4 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 265.780 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 63 Resp: 240681

Ion Ratio Lower Upper

63 100

106 26.6 21.5 32.3





#59

2-Hexanone

Concen: 268.218 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 43 Resp: 39070

Ion Ratio Lower Upper

43 100





58 49.6 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 52.075 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:129 Resp: 63738

Ion Ratio Lower Upper

129 100

127 77.9 39.3 117.8









#61
1,2-Dibromoethane
Concen: 51.050 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVX050525

Tgt Ion:107 Resp: 59134
Ion Ratio Lower Upper
107 100
109 93.8 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 48.596 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 95 Resp: 75568
Ion Ratio Lower Upper
95 100
174 68.2 0.0 135.8
176 66.5 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:117 Resp: 145078
Ion Ratio Lower Upper
117 100
82 62.1 50.6 76.0
119 32.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 46.795 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:164 Resp: 48034
Ion Ratio Lower Upper
164 100
166 129.9 105.0 157.6
129 105.3 83.5 125.3
131 98.8 76.5 114.7





#65
Chlorobenzene
Concen: 48.699 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:112 Resp: 154640
Ion Ratio Lower Upper
112 100
114 31.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 49.557 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:131 Resp: 53735
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 63.7 31.6 95.0





#67
Ethyl Benzene
Concen: 50.245 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 91 Resp: 281239
Ion Ratio Lower Upper
91 100
106 29.1 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 101.022 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:106 Resp: 206813
Ion Ratio Lower Upper
106 100
91 215.6 171.2 256.8









#69
o-Xylene
Concen: 50.841 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:106 Resp: 101469
Ion Ratio Lower Upper
106 100
91 227.7 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 51.907 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:104 Resp: 169703
Ion Ratio Lower Upper
104 100
78 57.6 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 49.583 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:173 Resp: 40364
Ion Ratio Lower Upper
173 100
175 48.8 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:152 Resp: 68229
Ion Ratio Lower Upper
152 100
115 93.3 46.9 140.7
150 175.3 0.0 351.0





#73
Isopropylbenzene
Concen: 50.733 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:105 Resp: 269484
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 52.407 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 137559
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 22.7 18.7 28.1
61 21.8 17.1 25.7





#75
1,1,2,2-Tetrachloroethane
Concen: 47.827 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 83 Resp: 8902
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 65.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 48.568 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 79763
Ion Ratio Lower Upper
75 100
77 40.8 20.5 61.5





#77

Bromobenzene

Concen: 48.549 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:	156	Resp:	5987
Ion	Ratio	Lower	Upper
156	100		
77	189.0	92.8	278.3
158	98.9	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#78

n-propylbenzene

Concen: 50.562 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	91	Resp:	312285
Ion	Ratio	Lower	Upper
91	100		
120	21.4	10.8	32.4









#79
2-Chlorotoluene
Concen: 49.107 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 91 Resp: 195630
Ion Ratio Lower Upper
91 100
126 30.8 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 51.500 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:105 Resp: 228538
Ion Ratio Lower Upper
105 100
120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 49.276 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVX050525

Tgt Ion: 75 Resp: 2485
Ion Ratio Lower Upper
75 100
53 122.3 100.5 150.7
89 46.2 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 50.661 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 91 Resp: 223811
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 50.511 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion:119 Resp: 22578

Ion Ratio Lower Upper

119 100

91 64.9 32.9 98.7





134 22.6 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 51.485 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 231371

Ion Ratio Lower Upper

105 100


120 42.3 21.2 63.6





#85
sec-Butylbenzene
Concen: 51.122 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:105 Resp: 280578
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 51.198 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:119 Resp: 231942
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.5 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 50.272 ug/l
RT: 11.963 min Scan# 1784
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:146 Resp: 113140
Ion Ratio Lower Upper
146 100
111 44.1 22.1 66.3
148 63.5 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#88
1,4-Dichlorobenzene
Concen: 48.368 ug/l
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:146 Resp: 111173
Ion Ratio Lower Upper
146 100
111 43.9 21.3 63.9
148 63.3 31.9 95.5





#89

n-Butylbenzene

Concen: 51.804 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 91 Resp: 20585

Ion Ratio Lower Upper

91 100

92 53.7 26.9 80.7





134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 49.895 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:117 Resp: 39823

Ion Ratio Lower Upper

117 100


201 62.4 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 49.361 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:146 Resp: 11148
Ion Ratio Lower Upper
146 100
111 45.8 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.458 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 75 Resp: 21220
Ion Ratio Lower Upper
75 100
155 68.6 34.9 104.8
157 86.7 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 49.662 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:180 Resp: 6442
Ion Ratio Lower Upper
180 100
182 95.5 47.9 143.8
145 34.4 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 48.408 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:225 Resp: 27425
Ion Ratio Lower Upper
225 100
223 63.1 30.8 92.4
227 63.7 30.9 92.7





#95
Naphthalene
Concen: 50.404 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:128 Resp: 239809
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.7 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 48.525 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:180 Resp: 64950
Ion Ratio Lower Upper
180 100
182 97.0 47.8 143.3
145 37.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.765	0.851	-11.2	94	0.00
3 P	Chloromethane	0.742	0.781	-5.3	96	0.00
4 C	Vinyl Chloride	0.691	0.713	-3.2#	96	0.00
5 T	Bromomethane	0.320	0.324	-1.3	95	0.00
6 T	Chloroethane	0.369	0.385	-4.3	97	0.00
7 T	Trichlorofluoromethane	1.021	1.073	-5.1	96	0.00
8 T	Diethyl Ether	0.347	0.344	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.642	-1.6	96	0.00
10 T	Methyl Iodide	0.747	0.799	-7.0	94	0.00
11 T	Tert butyl alcohol	0.131	0.132	-0.8	98	0.00
12 CM	1,1-Dichloroethene	0.593	0.606	-2.2#	96	0.00
13 T	Acrolein	0.149	0.139	6.7	87	0.00
14 T	Allyl chloride	1.133	1.175	-3.7	95	0.00
15 T	Acrylonitrile	0.374	0.393	-5.1	97	0.00
16 T	Acetone	0.374	0.367	1.9	97	0.00
17 T	Carbon Disulfide	1.406	1.433	-1.9	94	0.00
18 T	Methyl Acetate	0.867	0.867	0.0	97	0.00
19 T	Methyl tert-butyl Ether	2.079	2.111	-1.5	93	0.00
20 T	Methylene Chloride	0.716	0.684	4.5	95	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.602	-1.0	94	0.00
22 T	Diisopropyl ether	2.189	2.268	-3.6	95	0.00
23 T	Vinyl Acetate	1.925	2.045	-6.2	95	0.00
24 P	1,1-Dichloroethane	1.219	1.239	-1.6	94	0.00
25 T	2-Butanone	0.543	0.558	-2.8	96	0.00
26 T	2,2-Dichloropropane	0.954	0.953	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.731	-1.8	95	0.00
28 T	Bromochloromethane	0.587	0.598	-1.9	99	0.00
29 T	Tetrahydrofuran	0.340	0.354	-4.1	96	0.00
30 C	Chloroform	1.271	1.305	-2.7#	96	0.00
31 T	Cyclohexane	1.111	1.114	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	1.101	1.126	-2.3	95	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.904	3.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.360	0.358	0.6	98	0.00
36 T	1,1-Dichloropropene	0.484	0.479	1.0	94	0.00
37 T	Ethyl Acetate	0.598	0.606	-1.3	96	0.00
38 T	Carbon Tetrachloride	0.544	0.542	0.4	94	0.00
39 T	Methylcyclohexane	0.623	0.620	0.5	94	0.00
40 TM	Benzene	1.417	1.439	-1.6	95	0.00
41 T	Methacrylonitrile	0.313	0.344	-9.9	96	0.00
42 TM	1,2-Dichloroethane	0.612	0.622	-1.6	96	0.00
43 T	Isopropyl Acetate	0.912	0.952	-4.4	96	0.00
44 TM	Trichloroethene	0.341	0.341	0.0	93	0.00
45 C	1,2-Dichloropropane	0.352	0.366	-4.0#	96	0.00
46 T	Dibromomethane	0.278	0.278	0.0	94	0.00
47 T	Bromodichloromethane	0.547	0.568	-3.8	95	0.00
48 T	Methyl methacrylate	0.466	0.505	-8.4	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	100	0.00
50 S	Toluene-d8	1.246	1.221	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.627	-3.6	96	0.00
52 CM	Toluene	0.869	0.889	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	0.487	0.515	-5.7	94	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.569	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	0.343	0.349	-1.7	95	0.00
56 T	Ethyl methacrylate	0.546	0.589	-7.9	96	0.00
57 T	1,3-Dichloropropane	0.615	0.613	0.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.296	-6.5	93	0.00
59 T	2-Hexanone	0.448	0.480	-7.1	98	0.00
60 T	Dibromochloromethane	0.376	0.392	-4.3	95	0.00
61 T	1,2-Dibromoethane	0.356	0.364	-2.2	94	0.00
62 S	4-Bromofluorobenzene	0.478	0.465	2.7	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.354	0.331	6.5	88	0.00
65 PM	Chlorobenzene	1.094	1.066	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.370	1.1	94	0.00
67 C	Ethyl Benzene	1.929	1.939	-0.5#	95	0.00
68 T	m/p-Xylenes	0.706	0.713	-1.0	95	0.00
69 T	o-Xylene	0.688	0.699	-1.6	95	0.00
70 T	Styrene	1.127	1.170	-3.8	95	0.00
71 P	Bromoform	0.281	0.278	1.1	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.893	3.950	-1.5	96	0.00
74 T	N-amyl acetate	1.924	2.016	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.305	4.3	98	0.00
76 T	1,2,3-Trichloropropane	1.204	1.169	2.9	99	0.00
77 T	Bromobenzene	0.904	0.878	2.9	95	0.00
78 T	n-propylbenzene	4.526	4.577	-1.1	95	0.00
79 T	2-Chlorotoluene	2.919	2.867	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.350	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.364	1.6	95	0.00
82 T	4-Chlorotoluene	3.238	3.280	-1.3	96	0.00
83 T	tert-Butylbenzene	3.276	3.309	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.391	-3.0	97	0.00
85 T	sec-Butylbenzene	4.022	4.112	-2.2	96	0.00
86 T	p-Isopropyltoluene	3.320	3.399	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	1.649	1.658	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	1.684	1.629	3.3	97	0.00
89 T	n-Butylbenzene	2.912	3.017	-3.6	96	0.00
90 T	Hexachloroethane	0.585	0.584	0.2	94	0.00
91 T	1,2-Dichlorobenzene	1.655	1.634	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.311	-3.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.944	0.7	97	0.00
94 T	Hexachlorobutadiene	0.415	0.402	3.1	95	0.00
95 T	Naphthalene	3.487	3.515	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	0.952	3.0	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	55.594	-11.2	94	0.00
3 P	Chloromethane	50.000	52.592	-5.2	96	0.00
4 C	Vinyl Chloride	50.000	51.598	-3.2#	96	0.00
5 T	Bromomethane	50.000	50.627	-1.3	95	0.00
6 T	Chloroethane	50.000	52.271	-4.5	97	0.00
7 T	Trichlorofluoromethane	50.000	52.560	-5.1	96	0.00
8 T	Diethyl Ether	50.000	49.536	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.822	-1.6	96	0.00
10 T	Methyl Iodide	50.000	53.419	-6.8	94	0.00
11 T	Tert butyl alcohol	250.000	252.025	-0.8	98	0.00
12 CM	1,1-Dichloroethene	50.000	51.097	-2.2#	96	0.00
13 T	Acrolein	250.000	232.505	7.0	87	0.00
14 T	Allyl chloride	50.000	51.833	-3.7	95	0.00
15 T	Acrylonitrile	250.000	262.412	-5.0	97	0.00
16 T	Acetone	250.000	245.518	1.8	97	0.00
17 T	Carbon Disulfide	50.000	50.981	-2.0	94	0.00
18 T	Methyl Acetate	50.000	49.968	0.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	50.777	-1.6	93	0.00
20 T	Methylene Chloride	50.000	47.779	4.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.513	-1.0	94	0.00
22 T	Diisopropyl ether	50.000	51.800	-3.6	95	0.00
23 T	Vinyl Acetate	250.000	265.585	-6.2	95	0.00
24 P	1,1-Dichloroethane	50.000	50.825	-1.7	94	0.00
25 T	2-Butanone	250.000	257.110	-2.8	96	0.00
26 T	2,2-Dichloropropane	50.000	49.934	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.956	-1.9	95	0.00
28 T	Bromochloromethane	50.000	50.975	-2.0	99	0.00
29 T	Tetrahydrofuran	250.000	259.958	-4.0	96	0.00
30 C	Chloroform	50.000	51.332	-2.7#	96	0.00
31 T	Cyclohexane	50.000	50.147	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	50.000	51.108	-2.2	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.509	3.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.683	0.6	98	0.00
36 T	1,1-Dichloropropene	50.000	49.521	1.0	94	0.00
37 T	Ethyl Acetate	50.000	50.714	-1.4	96	0.00
38 T	Carbon Tetrachloride	50.000	49.825	0.3	94	0.00
39 T	Methylcyclohexane	50.000	49.757	0.5	94	0.00
40 TM	Benzene	50.000	50.759	-1.5	95	0.00
41 T	Methacrylonitrile	50.000	55.076	-10.2	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.861	-1.7	96	0.00
43 T	Isopropyl Acetate	50.000	52.195	-4.4	96	0.00
44 TM	Trichloroethene	50.000	50.018	-0.0	93	0.00
45 C	1,2-Dichloropropane	50.000	51.983	-4.0#	96	0.00
46 T	Dibromomethane	50.000	49.978	0.0	94	0.00
47 T	Bromodichloromethane	50.000	51.889	-3.8	95	0.00
48 T	Methyl methacrylate	50.000	54.226	-8.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1059.599	-6.0	100	0.00
50 S	Toluene-d8	50.000	48.984	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	259.133	-3.7	96	0.00
52 CM	Toluene	50.000	51.150	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	52.920	-5.8	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.892	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	50.000	50.887	-1.8	95	0.00
56 T	Ethyl methacrylate	50.000	53.918	-7.8	96	0.00
57 T	1,3-Dichloropropane	50.000	49.793	0.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.780	-6.3	93	0.00
59 T	2-Hexanone	250.000	268.218	-7.3	98	0.00
60 T	Dibromochloromethane	50.000	52.075	-4.2	95	0.00
61 T	1,2-Dibromoethane	50.000	51.050	-2.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.596	2.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	46.795	6.4	88	0.00
65 PM	Chlorobenzene	50.000	48.699	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.557	0.9	94	0.00
67 C	Ethyl Benzene	50.000	50.245	-0.5#	95	0.00
68 T	m/p-Xylenes	100.000	101.022	-1.0	95	0.00
69 T	o-Xylene	50.000	50.841	-1.7	95	0.00
70 T	Styrene	50.000	51.907	-3.8	95	0.00
71 P	Bromoform	50.000	49.583	0.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	50.733	-1.5	96	0.00
74 T	N-ethyl acetate	50.000	52.407	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.827	4.3	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.568	2.9	99	0.00
77 T	Bromobenzene	50.000	48.549	2.9	95	0.00
78 T	n-propylbenzene	50.000	50.562	-1.1	95	0.00
79 T	2-Chlorotoluene	50.000	49.107	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.500	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.276	1.4	95	0.00
82 T	4-Chlorotoluene	50.000	50.661	-1.3	96	0.00
83 T	tert-Butylbenzene	50.000	50.511	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.485	-3.0	97	0.00
85 T	sec-Butylbenzene	50.000	51.122	-2.2	96	0.00
86 T	p-Isopropyltoluene	50.000	51.198	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.272	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	50.000	48.368	3.3	97	0.00
89 T	n-Butylbenzene	50.000	51.804	-3.6	96	0.00
90 T	Hexachloroethane	50.000	49.895	0.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	49.361	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.458	-2.9	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.662	0.7	97	0.00
94 T	Hexachlorobutadiene	50.000	48.408	3.2	95	0.00
95 T	Naphthalene	50.000	50.404	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.525	3.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG No.: Q2186

Instrument ID: MSVOA_X Calibration Date/Time: 06/03/2025 10:10

Lab File ID: VX046460.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.736	0.1	-0.81	20
Vinyl Chloride	0.691	0.673		-2.61	20
Bromomethane	0.320	0.292		-8.75	20
Chloroethane	0.369	0.371		0.54	20
Trichlorofluoromethane	1.021	1.024		0.29	20
1,1,2-Trichlorotrifluoroethane	0.632	0.643		1.74	20
1,1-Dichloroethene	0.593	0.595		0.34	20
Acetone	0.374	0.443		18.45	20
Carbon Disulfide	1.406	1.272		-9.53	20
Methyl tert-butyl Ether	2.079	2.188		5.24	20
Methylene Chloride	0.716	0.695		-2.93	20
trans-1,2-Dichloroethene	0.596	0.602		1.01	20
1,1-Dichloroethane	1.219	1.266	0.1	3.86	20
2-Butanone	0.543	0.573		5.53	20
Carbon Tetrachloride	0.544	0.550		1.1	20
cis-1,2-Dichloroethene	0.718	0.743		3.48	20
Chloroform	1.271	1.297		2.05	20
1,1,1-Trichloroethane	1.101	1.123		2	20
Methylcyclohexane	0.623	0.640		2.73	20
Benzene	1.417	1.465		3.39	20
1,2-Dichloroethane	0.612	0.633		3.43	20
Trichloroethene	0.341	0.351		2.93	20
1,2-Dichloropropane	0.352	0.380		7.95	20
Bromodichloromethane	0.547	0.587		7.31	20
4-Methyl-2-Pentanone	0.605	0.632		4.46	20
Toluene	0.869	0.896		3.11	20
t-1,3-Dichloropropene	0.487	0.526		8.01	20
cis-1,3-Dichloropropene	0.538	0.583		8.36	20
1,1,2-Trichloroethane	0.343	0.357		4.08	20
2-Hexanone	0.448	0.479		6.92	20
Dibromochloromethane	0.376	0.403		7.18	20
Tetrachloroethene	0.354	0.378		6.78	20
Chlorobenzene	1.094	1.132	0.3	3.47	20
Ethyl Benzene	1.929	2.094		8.55	20
m/p-Xylenes	0.706	0.754		6.8	20
o-Xylene	0.688	0.747		8.58	20
Styrene	1.127	1.256		11.45	20
Bromoform	0.281	0.299	0.1	6.41	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG No.: Q2186

Instrument ID: MSVOA_X Calibration Date/Time: 06/03/2025 10:10

Lab File ID: VX046460.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	4.081		4.83	20
1,1,2,2-Tetrachloroethane	1.364	1.325	0.3	-2.93	20
1,3-Dichlorobenzene	1.649	1.704		3.34	20
1,4-Dichlorobenzene	1.684	1.703		1.13	20
1,2-Dichlorobenzene	1.655	1.679		1.45	20
1,2-Dichloroethane-d4	0.932	0.859		-7.83	20
Dibromofluoromethane	0.360	0.352		-2.22	20
Toluene-d8	1.246	1.151		-7.62	20
4-Bromofluorobenzene	0.478	0.460		-3.77	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/04/2025
 Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	97221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	166757	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	142113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	83472	46.053	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.100%		
35) Dibromofluoromethane	5.379	113	58765	48.937	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.880%		
50) Toluene-d8	8.641	98	191895	46.171	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.340%		
62) 4-Bromofluorobenzene	11.079	95	76697	48.108	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.220%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	71992	48.381	ug/l	96
3) Chloromethane	1.301	50	71578	49.603	ug/l	100
4) Vinyl Chloride	1.374	62	65400	48.697	ug/l	100
5) Bromomethane	1.593	94	28400	45.593	ug/l	98
6) Chloroethane	1.666	64	36039	50.267	ug/l	99
7) Trichlorofluoromethane	1.874	101	99535	50.148	ug/l	100
8) Diethyl Ether	2.130	74	33963	50.265	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	62523	50.901	ug/l	99
10) Methyl Iodide	2.447	142	74693	51.391	ug/l	99
11) Tert butyl alcohol	2.971	59	60143	236.391	ug/l	100
12) 1,1-Dichloroethene	2.313	96	57843	50.176	ug/l	96
13) Acrolein	2.233	56	48798	168.416	ug/l	98
14) Allyl chloride	2.654	41	116985	53.097	ug/l	98
15) Acrylonitrile	3.056	53	186943	256.965	ug/l	99
16) Acetone	2.380	43	215550	296.602	ug/l	99
17) Carbon Disulfide	2.502	76	123668	45.249	ug/l	99
18) Methyl Acetate	2.703	43	106195	62.971	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	212708	52.629	ug/l	100
20) Methylene Chloride	2.782	84	67564	48.515	ug/l	94
21) trans-1,2-Dichloroethene	3.087	96	58500	50.461	ug/l	99
22) Diisopropyl ether	3.758	45	227076	53.357	ug/l	94
23) Vinyl Acetate	3.715	43	941542	251.538	ug/l	99
24) 1,1-Dichloroethane	3.599	63	123119	51.940	ug/l	99
25) 2-Butanone	4.550	43	278637	264.096	ug/l	99
26) 2,2-Dichloropropane	4.465	77	96338	51.925	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	72243	51.764	ug/l	99
28) Bromochloromethane	4.885	49	54439	47.712	ug/l	99
29) Tetrahydrofuran	4.995	42	166146	251.308	ug/l	100
30) Chloroform	5.080	83	126054	51.020	ug/l	95
31) Cyclohexane	5.458	56	106868	49.475	ug/l	95
32) 1,1,1-Trichloroethane	5.373	97	109132	50.955	ug/l	99
36) 1,1-Dichloropropene	5.678	75	81293	50.385	ug/l	98
37) Ethyl Acetate	4.709	43	98903	49.615	ug/l	100
38) Carbon Tetrachloride	5.666	117	91791	50.634	ug/l	99
39) Methylcyclohexane	7.373	83	106672	51.355	ug/l	98
40) Benzene	6.025	78	244281	51.690	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	58928	56.511	ug/l	99
42) 1,2-Dichloroethane	6.080	62	105576	51.762	ug/l	100
43) Isopropyl Acetate	6.330	43	160957	52.928	ug/l	100
44) Trichloroethene	7.117	130	58531	51.459	ug/l	94
45) 1,2-Dichloropropane	7.422	63	63306	53.872	ug/l	98
46) Dibromomethane	7.574	93	47187	50.912	ug/l	99
47) Bromodichloromethane	7.812	83	97879	53.619	ug/l	97
48) Methyl methacrylate	7.690	41	83841	53.984	ug/l	99
49) 1,4-Dioxane	7.653	88	27329	926.756	ug/l	98
51) 4-Methyl-2-Pentanone	8.568	43	526838	260.992	ug/l	99
52) Toluene	8.714	92	149333	51.533	ug/l	98
53) t-1,3-Dichloropropene	8.970	75	87793	54.108	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	97218	54.211	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	59472	52.048	ug/l	98
56) Ethyl methacrylate	9.110	69	99973	54.897	ug/l	98
57) 1,3-Dichloropropane	9.305	76	106135	51.721	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	241312	259.915	ug/l	99
59) 2-Hexanone	9.427	43	398970	267.150	ug/l	100
60) Dibromochloromethane	9.519	129	67274	53.611	ug/l	100
61) 1,2-Dibromoethane	9.604	107	61399	51.700	ug/l	98
64) Tetrachloroethene	9.269	164	53715	53.421	ug/l	97
65) Chlorobenzene	10.073	112	160933	51.739	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	57122	53.780	ug/l	99
67) Ethyl Benzene	10.189	91	297648	54.286	ug/l	100
68) m/p-Xylenes	10.299	106	214306	106.866	ug/l	99
69) o-Xylene	10.640	106	106129	54.285	ug/l	97
70) Styrene	10.653	104	178562	55.756	ug/l	99
71) Bromoform	10.799	173	42440	53.221	ug/l #	97
73) Isopropylbenzene	10.957	105	284040	52.417	ug/l	99
74) N-amyl acetate	10.842	43	141380	52.800	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	92189	48.548	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	83887m	50.070	ug/l	
77) Bromobenzene	11.195	156	63843	50.748	ug/l	99
78) n-propylbenzene	11.299	91	332789	52.818	ug/l	100
79) 2-Chlorotoluene	11.360	91	201843	49.667	ug/l	99
80) 1,3,5-Trimethylbenzene	11.445	105	238098	52.595	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.012	75	26522	51.539	ug/l	99
82) 4-Chlorotoluene	11.451	91	234790	52.097	ug/l	100
83) tert-Butylbenzene	11.713	119	235173	51.573	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	240382	52.434	ug/l	100
85) sec-Butylbenzene	11.884	105	298046	53.233	ug/l	100
86) p-Isopropyltoluene	12.006	119	249289	53.941	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	118632	51.669	ug/l	100
88) 1,4-Dichlorobenzene	12.037	146	118528	50.549	ug/l	99
89) n-Butylbenzene	12.329	91	227595	56.143	ug/l	99
90) Hexachloroethane	12.536	117	42405	52.082	ug/l	100
91) 1,2-Dichlorobenzene	12.329	146	116846	50.715	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21419	50.916	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	70545	53.309	ug/l	100
94) Hexachlorobutadiene	13.719	225	31033	53.695	ug/l	99
95) Naphthalene	13.774	128	248647	51.231	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	71443	52.322	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046460.D
Acq On : 03 Jun 2025 10:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:21:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046460.D
Acq On : 03 Jun 2025 10:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:21:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:168 Resp: 9722
Ion Ratio Lower Upper
168 100
99 70.0 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#2
Dichlorodifluoromethane
Concen: 48.381 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 85 Resp: 71992
Ion Ratio Lower Upper
85 100
87 33.7 15.7 47.1









#3
Chloromethane
Concen: 49.603 ug/l
RT: 1.301 min Scan# 35
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 71578
Ion Ratio Lower Upper
50 100
52 31.7 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#4
Vinyl Chloride
Concen: 48.697 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 62 Resp: 65400
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8









#5
Bromomethane
Concen: 45.593 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 28400
Ion Ratio Lower Upper
94 100
96 92.6 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#6
Chloroethane
Concen: 50.267 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion: 64 Resp: 36039
Ion Ratio Lower Upper
64 100
66 30.9 24.3 36.5





#7
Trichlorofluoromethane
Concen: 50.148 ug/l
RT: 1.874 min Scan# 110
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:101 Resp: 9953
Ion Ratio Lower Upper
101 100
103 63.7 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025




#8
Diethyl Ether
Concen: 50.265 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10



Tgt Ion: 74 Resp: 33963
Ion Ratio Lower Upper
74 100
45 106.2 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.901 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 6252
Ion Ratio Lower Upper
101 100
85 48.2 38.6 58.0
151 70.4 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#10
Methyl Iodide
Concen: 51.391 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:142 Resp: 74693
Ion Ratio Lower Upper
142 100
127 52.5 41.7 62.5
141 14.8 11.5 17.3





#11

Tert butyl alcohol

Concen: 236.391 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 6014

Ion Ratio Lower Upper

59 100





57 10.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#12

1,1-Dichloroethene

Concen: 50.176 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 96 Resp: 57843


Ion Ratio Lower Upper

96 100

61 187.2 146.2 219.2


98 68.1 51.0 76.6





#13
Acrolein
Concen: 168.416 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 48798
Ion Ratio Lower Upper
56 100
55 71.6 56.2 84.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#14
Allyl chloride
Concen: 53.097 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 41 Resp: 116985
Ion Ratio Lower Upper
41 100
39 73.9 60.6 90.8
76 30.2 24.9 37.3









#15
Acrylonitrile
Concen: 256.965 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 53 Resp: 18694
Ion Ratio Lower Upper
53 100
52 82.9 65.3 97.9
51 37.4 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#16
Acetone
Concen: 296.602 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion: 43 Resp: 215550
Ion Ratio Lower Upper
43 100
58 26.8 21.2 31.8





#17
Carbon Disulfide
Concen: 45.249 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 12366
Ion Ratio Lower Upper
76 100
78 8.9 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#18
Methyl Acetate
Concen: 62.971 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion: 43 Resp: 106195
Ion Ratio Lower Upper
43 100
74 22.1 16.7 25.1





#19
Methyl tert-butyl Ether
Concen: 52.629 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 73 Resp: 212708
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#20
Methylene Chloride
Concen: 48.515 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion: 84 Resp: 67564
Ion Ratio Lower Upper
84 100
49 135.4 113.9 170.9
51 40.7 33.5 50.3
86 61.3 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 50.461 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 58500
Ion Ratio Lower Upper
96 100
61 148.7 119.5 179.3
98 60.7 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#22
Diisopropyl ether
Concen: 53.357 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 45 Resp: 227076
Ion Ratio Lower Upper
45 100
43 76.3 66.6 100.0
87 24.2 19.8 29.6
59 10.4 8.6 12.8





#23

Vinyl Acetate

Concen: 251.538 ug/l

RT: 3.715 min Scan# 41

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 94154

Ion Ratio Lower Upper

43 100





86 9.1 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#24

1,1-Dichloroethane

Concen: 51.940 ug/l

RT: 3.599 min Scan# 412

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 63 Resp: 123119


Ion Ratio Lower Upper

63 100

98 6.6 3.6 10.8

100 4.3 2.1 6.3





#25

2-Butanone

Concen: 264.096 ug/l

RT: 4.550 min Scan# 50

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 27863

Ion Ratio Lower Upper

43 100





72 23.5 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#26

2,2-Dichloropropane

Concen: 51.925 ug/l

RT: 4.465 min Scan# 554

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 77 Resp: 96338

Ion Ratio Lower Upper

77 100

97 21.4 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 51.764 ug/l

RT: 4.477 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 7224

Ion Ratio Lower Upper

96 100

61 160.5 0.0 322.8





98 65.6 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#28

Bromochloromethane

Concen: 47.712 ug/l

RT: 4.885 min Scan# 623

Delta R.T. -0.012 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 49 Resp: 54439


Ion Ratio Lower Upper

49 100

129 1.7 0.0 4.0

130 69.7 56.2 84.2





#29

Tetrahydrofuran

Concen: 251.308 ug/l

RT: 4.995 min Scan# 641

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 166140

Ion Ratio Lower Upper

42 100

72 42.0 33.7 50.5





71 38.9 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#30

Chloroform

Concen: 51.020 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 83 Resp: 126054

Ion Ratio Lower Upper

83 100


85 65.3 49.3 73.9





#31
Cyclohexane
Concen: 49.475 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 10686
Ion Ratio Lower Upper
56 100
69 28.8 24.4 36.6
84 78.6 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#32
1,1,1-Trichloroethane
Concen: 50.955 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 97 Resp: 109132
Ion Ratio Lower Upper
97 100
99 63.5 51.8 77.6
61 50.5 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 46.053 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 83472

Ion Ratio Lower Upper

65 100





67 48.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion:114 Resp: 166757


Ion Ratio Lower Upper

114 100

63 23.8 0.0 49.2

88 17.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 48.937 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 5876

Ion Ratio Lower Upper

113 100

111 102.9 83.1 124.7





192 16.9 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#36

1,1-Dichloropropene

Concen: 50.385 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 75 Resp: 81293


Ion Ratio Lower Upper

75 100

110 33.5 16.3 48.9

77 31.6 24.3 36.5





#37

Ethyl Acetate

Concen: 49.615 ug/l

RT: 4.709 min Scan# 594

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 9890

Ion Ratio Lower Upper

43 100

61 12.8 10.3 15.5





70 9.8 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#38

Carbon Tetrachloride

Concen: 50.634 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion:117 Resp: 91791


Ion Ratio Lower Upper

117 100

119 92.5 75.2 112.8

121 29.8 24.2 36.4





#39

Methylcyclohexane

Concen: 51.355 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 106672

Ion Ratio Lower Upper

83 100

55 82.0 64.7 97.1





98 48.6 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#40

Benzene

Concen: 51.690 ug/l

RT: 6.025 min Scan# 810

Delta R.T. -0.006 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 78 Resp: 244281

Ion Ratio Lower Upper

78 100


77 24.2 19.0 28.4





#41
Methacrylonitrile
Concen: 56.511 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 58928
Ion Ratio Lower Upper
41 100
39 58.5 47.2 70.8
67 63.9 50.7 76.1
52 30.2 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#42
1,2-Dichloroethane
Concen: 51.762 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 62 Resp: 105576
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 52.928 ug/l

RT: 6.330 min Scan# 80

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 16095

Ion Ratio Lower Upper

43 100

61 18.0 14.3 21.5





87 11.7 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#44

Trichloroethene

Concen: 51.459 ug/l

RT: 7.117 min Scan# 989

Delta R.T. -0.006 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion:130 Resp: 58531

Ion Ratio Lower Upper

130 100

95 108.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 53.872 ug/l

RT: 7.422 min Scan# 1039

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 63300

Ion Ratio Lower Upper

63 100





65 29.2 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#46

Dibromomethane

Concen: 50.912 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 93 Resp: 47187


Ion Ratio Lower Upper

93 100

95 83.2 65.6 98.4

174 85.1 68.2 102.2





#47

Bromodichloromethane

Concen: 53.619 ug/l

RT: 7.812 min Scan# 1104

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 97879

Ion Ratio Lower Upper

83 100

85 60.6 50.5 75.7





127 7.3 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#48

Methyl methacrylate

Concen: 53.984 ug/l

RT: 7.690 min Scan# 1083

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 41 Resp: 83841


Ion Ratio Lower Upper

41 100

69 72.3 58.5 87.7

39 64.7 51.7 77.5





#49

1,4-Dioxane

Concen: 926.756 ug/l

RT: 7.653 min Scan# 1077

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 27329

Ion Ratio Lower Upper

88 100

43 42.5 33.4 50.2





58 75.3 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#50

Toluene-d8

Concen: 46.171 ug/l

RT: 8.641 min Scan# 1239

Delta R.T. -0.006 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 98 Resp: 191895

Ion Ratio Lower Upper

98 100

100 64.3 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 260.992 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 526838
Ion Ratio Lower Upper
43 100
58 35.7 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#52
Toluene
Concen: 51.533 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 92 Resp: 149333
Ion Ratio Lower Upper
92 100
91 172.9 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 54.108 ug/l

RT: 8.970 min Scan# 1193

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 8779

Ion Ratio Lower Upper

75 100





77 32.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#54

cis-1,3-Dichloropropene

Concen: 54.211 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 75 Resp: 97218


Ion Ratio Lower Upper

75 100

77 32.0 25.4 38.0

39 62.9 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 52.048 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 97 Resp: 59472

Ion Ratio Lower Upper

97 100

83 85.8 69.0 103.6

85 55.4 44.3 66.5





99 63.0 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#56

Ethyl methacrylate

Concen: 54.897 ug/l

RT: 9.110 min Scan# 1316

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 69 Resp: 99973


Ion Ratio Lower Upper

69 100

41 77.8 60.8 91.2

39 47.9 39.0 58.6





#57

1,3-Dichloropropane

Concen: 51.721 ug/l

RT: 9.305 min Scan# 1134

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 10613

Ion Ratio Lower Upper

76 100





78 32.5 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#58

2-Chloroethyl Vinyl ether

Concen: 259.915 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 63 Resp: 241312

Ion Ratio Lower Upper

63 100

106 26.4 21.5 32.3





#59

2-Hexanone

Concen: 267.150 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 398970

Ion Ratio Lower Upper

43 100





58 49.6 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#60

Dibromochloromethane

Concen: 53.611 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion:129 Resp: 67274

Ion Ratio Lower Upper

129 100


127 78.8 39.3 117.8





#61
1,2-Dibromoethane
Concen: 51.700 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:107 Resp: 61399
Ion Ratio Lower Upper
107 100
109 93.9 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#62
4-Bromofluorobenzene
Concen: 48.108 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion: 95 Resp: 76697
Ion Ratio Lower Upper
95 100
174 67.0 0.0 135.8
176 65.1 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 142115
Ion Ratio Lower Upper
117 100
82 62.9 50.6 76.0
119 31.7 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#64
Tetrachloroethene
Concen: 53.421 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:164 Resp: 53715
Ion Ratio Lower Upper
164 100
166 131.0 105.0 157.6
129 105.9 83.5 125.3
131 103.2 76.5 114.7









#65
Chlorobenzene
Concen: 51.739 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:112 Resp: 16093
Ion Ratio Lower Upper
112 100
114 31.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 53.780 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion:131 Resp: 57122
Ion Ratio Lower Upper
131 100
133 93.1 47.3 141.9
119 63.0 31.6 95.0





#67
Ethyl Benzene
Concen: 54.286 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 297648
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#68
m/p-Xylenes
Concen: 106.866 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:106 Resp: 214306
Ion Ratio Lower Upper
106 100
91 216.3 171.2 256.8









#69
o-Xylene
Concen: 54.285 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 106129
Ion Ratio Lower Upper
106 100
91 229.8 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#70
Styrene
Concen: 55.756 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion:104 Resp: 178562
Ion Ratio Lower Upper
104 100
78 57.7 45.7 68.5
103 54.8 43.7 65.5





#71
Bromoform
Concen: 53.221 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:173 Resp: 42440
Ion Ratio Lower Upper
173 100
175 48.7 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10


Tgt Ion:152 Resp: 69603
Ion Ratio Lower Upper
152 100
115 93.8 46.9 140.7
150 172.1 0.0 351.0





#73
Isopropylbenzene
Concen: 52.417 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:105 Resp: 284040
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#74
N-ethyl acetate
Concen: 52.800 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 43 Resp: 141380
Ion Ratio Lower Upper
43 100
70 37.7 30.9 46.3
55 22.6 18.7 28.1
61 22.0 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 48.548 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 92189
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 64.2 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#76
1,2,3-Trichloropropane
Concen: 50.070 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 75 Resp: 83887
Ion Ratio Lower Upper
75 100
77 39.5 20.5 61.5





#77

Bromobenzene

Concen: 50.748 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:156 Resp: 6384

Ion Ratio Lower Upper

156 100

77 186.5 92.8 278.3





158 96.5 48.8 146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#78

n-propylbenzene

Concen: 52.818 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 91 Resp: 332789

Ion Ratio Lower Upper

91 100

120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 49.667 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 20184
Ion Ratio Lower Upper
91 100
126 31.6 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#80
1,3,5-Trimethylbenzene
Concen: 52.595 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:105 Resp: 238098
Ion Ratio Lower Upper
105 100
120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 51.539 ug/l
RT: 11.012 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 2652
Ion Ratio Lower Upper
75 100
53 123.8 100.5 150.7
89 45.9 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#82
4-Chlorotoluene
Concen: 52.097 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 91 Resp: 234790
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 51.573 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 23517

Ion Ratio Lower Upper

119 100

91 65.9 32.9 98.7





134 22.7 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#84

1,2,4-Trimethylbenzene

Concen: 52.434 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion:105 Resp: 240382

Ion Ratio Lower Upper

105 100

120 42.1 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 53.233 ug/l
 RT: 11.884 min Scan# 1771
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tgt Ion:105 Resp: 298040
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.7 29.1




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/04/2025
 Supervised By :Mahesh Dadoda 06/04/2025



#86
 p-Isopropyltoluene
 Concen: 53.941 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10


Tgt Ion:119 Resp: 249289
 Ion Ratio Lower Upper
 119 100
 134 25.3 12.5 37.5
 91 27.7 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 51.669 ug/l
RT: 11.963 min Scan# 1784
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 118632
Ion Ratio Lower Upper
146 100
111 44.2 22.1 66.3
148 64.4 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#88
1,4-Dichlorobenzene
Concen: 50.549 ug/l
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:146 Resp: 118528
Ion Ratio Lower Upper
146 100
111 43.5 21.3 63.9
148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 56.143 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 22759

Ion Ratio Lower Upper

91 100

92 53.1 26.9 80.7





134 23.9 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#90

Hexachloroethane

Concen: 52.082 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion:117 Resp: 42405

Ion Ratio Lower Upper

117 100


201 63.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 50.715 ug/l
RT: 12.329 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 116840
Ion Ratio Lower Upper
146 100
111 46.7 22.4 67.3
148 63.5 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.916 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 75 Resp: 21419
Ion Ratio Lower Upper
75 100
155 68.3 34.9 104.8
157 89.7 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 53.309 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:180	Resp:	7054
Ion Ratio	Lower	Upper
180	100	
182	96.0	47.9 143.8
145	34.0	17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#94

Hexachlorobutadiene

Concen: 53.695 ug/l




RT: 13.719 min Scan# 2072


Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10


Tgt Ion:225	Resp:	31033
Ion Ratio	Lower	Upper
225	100	
223	61.2	30.8 92.4
227	61.3	30.9 92.7





#95
Naphthalene
Concen: 51.231 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 24864
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 11.1 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#96
1,2,3-Trichlorobenzene
Concen: 52.322 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:180 Resp: 71443
Ion Ratio Lower Upper
180 100
182 95.1 47.8 143.3
145 36.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.765	0.740	3.3	86	0.00
3 P	Chloromethane	0.742	0.736	0.8	95	0.00
4 C	Vinyl Chloride	0.691	0.673	2.6#	95	0.00
5 T	Bromomethane	0.320	0.292	8.8	90	0.00
6 T	Chloroethane	0.369	0.371	-0.5	98	0.00
7 T	Trichlorofluoromethane	1.021	1.024	-0.3	96	0.00
8 T	Diethyl Ether	0.347	0.349	-0.6	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.643	-1.7	100	0.00
10 T	Methyl Iodide	0.747	0.768	-2.8	95	0.00
11 T	Tert butyl alcohol	0.131	0.124	5.3	96	0.00
12 CM	1,1-Dichloroethene	0.593	0.595	-0.3#	99	0.00
13 T	Acrolein	0.149	0.100	32.9#	66	0.00
14 T	Allyl chloride	1.133	1.203	-6.2	102	0.00
15 T	Acrylonitrile	0.374	0.385	-2.9	99	0.00
16 T	Acetone	0.374	0.443	-18.4	123	0.00
17 T	Carbon Disulfide	1.406	1.272	9.5	88	0.00
18 T	Methyl Acetate	0.867	1.092	-26.0#	129	0.00
19 T	Methyl tert-butyl Ether	2.079	2.188	-5.2	101	0.00
20 T	Methylene Chloride	0.716	0.695	2.9	102	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.602	-1.0	99	0.00
22 T	Diisopropyl ether	2.189	2.336	-6.7	103	0.00
23 T	Vinyl Acetate	1.925	1.937	-0.6	95	0.00
24 P	1,1-Dichloroethane	1.219	1.266	-3.9	100	0.00
25 T	2-Butanone	0.543	0.573	-5.5	103	0.00
26 T	2,2-Dichloropropane	0.954	0.991	-3.9	104	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.743	-3.5	101	0.00
28 T	Bromochloromethane	0.587	0.560	4.6	97	-0.01
29 T	Tetrahydrofuran	0.340	0.342	-0.6	98	0.00
30 C	Chloroform	1.271	1.297	-2.0#	100	0.00
31 T	Cyclohexane	1.111	1.099	1.1	98	0.00
32 T	1,1,1-Trichloroethane	1.101	1.123	-2.0	99	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.859	7.8	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.360	0.352	2.2	99	0.00
36 T	1,1-Dichloropropene	0.484	0.487	-0.6	98	0.00
37 T	Ethyl Acetate	0.598	0.593	0.8	96	0.00
38 T	Carbon Tetrachloride	0.544	0.550	-1.1	98	0.00
39 T	Methylcyclohexane	0.623	0.640	-2.7	99	0.00
40 TM	Benzene	1.417	1.465	-3.4	99	0.00
41 T	Methacrylonitrile	0.313	0.353	-12.8	101	0.00
42 TM	1,2-Dichloroethane	0.612	0.633	-3.4	100	0.00
43 T	Isopropyl Acetate	0.912	0.965	-5.8	99	0.00
44 TM	Trichloroethene	0.341	0.351	-2.9	98	0.00
45 C	1,2-Dichloropropane	0.352	0.380	-8.0#	101	0.00
46 T	Dibromomethane	0.278	0.283	-1.8	98	0.00
47 T	Bromodichloromethane	0.547	0.587	-7.3	101	0.00
48 T	Methyl methacrylate	0.466	0.503	-7.9	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	90	0.00
50 S	Toluene-d8	1.246	1.151	7.6	93	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.632	-4.5	99	0.00
52 CM	Toluene	0.869	0.896	-3.1#	99	0.00
53 T	t-1,3-Dichloropropene	0.487	0.526	-8.0	99	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.583	-8.4	100	0.00
55 T	1,1,2-Trichloroethane	0.343	0.357	-4.1	100	0.00
56 T	Ethyl methacrylate	0.546	0.600	-9.9	100	0.00
57 T	1,3-Dichloropropane	0.615	0.636	-3.4	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.289	-4.0	94	0.00
59 T	2-Hexanone	0.448	0.479	-6.9	101	0.00
60 T	Dibromochloromethane	0.376	0.403	-7.2	100	0.00
61 T	1,2-Dibromoethane	0.356	0.368	-3.4	98	0.00
62 S	4-Bromofluorobenzene	0.478	0.460	3.8	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.354	0.378	-6.8	98	0.00
65 PM	Chlorobenzene	1.094	1.132	-3.5	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.402	-7.5	100	0.00
67 C	Ethyl Benzene	1.929	2.094	-8.6#	101	0.00
68 T	m/p-Xylenes	0.706	0.754	-6.8	99	0.00
69 T	o-Xylene	0.688	0.747	-8.6	100	0.00
70 T	Styrene	1.127	1.256	-11.4	100	0.00
71 P	Bromoform	0.281	0.299	-6.4	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.893	4.081	-4.8	101	0.00
74 T	N-amyl acetate	1.924	2.031	-5.6	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.324	2.9	101	0.00
76 T	1,2,3-Trichloropropane	1.204	1.205	-0.1	104	0.00
77 T	Bromobenzene	0.904	0.917	-1.4	101	0.00
78 T	n-propylbenzene	4.526	4.781	-5.6	101	0.00
79 T	2-Chlorotoluene	2.919	2.900	0.7	99	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.421	-5.2	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.381	-3.0	101	0.00
82 T	4-Chlorotoluene	3.238	3.373	-4.2	101	0.00
83 T	tert-Butylbenzene	3.276	3.379	-3.1	101	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.454	-4.9	100	0.00
85 T	sec-Butylbenzene	4.022	4.282	-6.5	102	0.00
86 T	p-Isopropyltoluene	3.320	3.582	-7.9	103	0.00
87 T	1,3-Dichlorobenzene	1.649	1.704	-3.3	103	0.00
88 T	1,4-Dichlorobenzene	1.684	1.703	-1.1	103	0.00
89 T	n-Butylbenzene	2.912	3.270	-12.3	106	0.00
90 T	Hexachloroethane	0.585	0.609	-4.1	100	0.00
91 T	1,2-Dichlorobenzene	1.655	1.679	-1.5	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.308	-2.0	98	0.00
93 T	1,2,4-Trichlorobenzene	0.951	1.014	-6.6	106	0.00
94 T	Hexachlorobutadiene	0.415	0.446	-7.5	107	0.00
95 T	Naphthalene	3.487	3.572	-2.4	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046460.D
Acq On : 03 Jun 2025 10:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	1.026	-4.6	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	50.000	48.381	3.2	86	0.00
3 P	Chloromethane	50.000	49.603	0.8	95	0.00
4 C	Vinyl Chloride	50.000	48.697	2.6#	95	0.00
5 T	Bromomethane	50.000	45.593	8.8	90	0.00
6 T	Chloroethane	50.000	50.267	-0.5	98	0.00
7 T	Trichlorofluoromethane	50.000	50.148	-0.3	96	0.00
8 T	Diethyl Ether	50.000	50.265	-0.5	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.901	-1.8	100	0.00
10 T	Methyl Iodide	50.000	51.391	-2.8	95	0.00
11 T	Tert butyl alcohol	250.000	236.391	5.4	96	0.00
12 CM	1,1-Dichloroethene	50.000	50.176	-0.4#	99	0.00
13 T	Acrolein	250.000	168.416	32.6#	66	0.00
14 T	Allyl chloride	50.000	53.097	-6.2	102	0.00
15 T	Acrylonitrile	250.000	256.965	-2.8	99	0.00
16 T	Acetone	250.000	296.602	-18.6	123	0.00
17 T	Carbon Disulfide	50.000	45.249	9.5	88	0.00
18 T	Methyl Acetate	50.000	62.971	-25.9#	129	0.00
19 T	Methyl tert-butyl Ether	50.000	52.629	-5.3	101	0.00
20 T	Methylene Chloride	50.000	48.515	3.0	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.461	-0.9	99	0.00
22 T	Diisopropyl ether	50.000	53.357	-6.7	103	0.00
23 T	Vinyl Acetate	250.000	251.538	-0.6	95	0.00
24 P	1,1-Dichloroethane	50.000	51.940	-3.9	100	0.00
25 T	2-Butanone	250.000	264.096	-5.6	103	0.00
26 T	2,2-Dichloropropane	50.000	51.925	-3.8	104	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.764	-3.5	101	0.00
28 T	Bromochloromethane	50.000	47.712	4.6	97	-0.01
29 T	Tetrahydrofuran	250.000	251.308	-0.5	98	0.00
30 C	Chloroform	50.000	51.020	-2.0#	100	0.00
31 T	Cyclohexane	50.000	49.475	1.0	98	0.00
32 T	1,1,1-Trichloroethane	50.000	50.955	-1.9	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.053	7.9	94	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	48.937	2.1	99	0.00
36 T	1,1-Dichloropropene	50.000	50.385	-0.8	98	0.00
37 T	Ethyl Acetate	50.000	49.615	0.8	96	0.00
38 T	Carbon Tetrachloride	50.000	50.634	-1.3	98	0.00
39 T	Methylcyclohexane	50.000	51.355	-2.7	99	0.00
40 TM	Benzene	50.000	51.690	-3.4	99	0.00
41 T	Methacrylonitrile	50.000	56.511	-13.0	101	0.00
42 TM	1,2-Dichloroethane	50.000	51.762	-3.5	100	0.00
43 T	Isopropyl Acetate	50.000	52.928	-5.9	99	0.00
44 TM	Trichloroethene	50.000	51.459	-2.9	98	0.00
45 C	1,2-Dichloropropane	50.000	53.872	-7.7#	101	0.00
46 T	Dibromomethane	50.000	50.912	-1.8	98	0.00
47 T	Bromodichloromethane	50.000	53.619	-7.2	101	0.00
48 T	Methyl methacrylate	50.000	53.984	-8.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	926.756	7.3	90	0.00
50 S	Toluene-d8	50.000	46.171	7.7	93	0.00
51 T	4-Methyl-2-Pentanone	250.000	260.992	-4.4	99	0.00
52 CM	Toluene	50.000	51.533	-3.1#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	54.108	-8.2	99	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.211	-8.4	100	0.00
55 T	1,1,2-Trichloroethane	50.000	52.048	-4.1	100	0.00
56 T	Ethyl methacrylate	50.000	54.897	-9.8	100	0.00
57 T	1,3-Dichloropropane	50.000	51.721	-3.4	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	259.915	-4.0	94	0.00
59 T	2-Hexanone	250.000	267.150	-6.9	101	0.00
60 T	Dibromochloromethane	50.000	53.611	-7.2	100	0.00
61 T	1,2-Dibromoethane	50.000	51.700	-3.4	98	0.00
62 S	4-Bromofluorobenzene	50.000	48.108	3.8	97	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	53.421	-6.8	98	0.00
65 PM	Chlorobenzene	50.000	51.739	-3.5	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.780	-7.6	100	0.00
67 C	Ethyl Benzene	50.000	54.286	-8.6#	101	0.00
68 T	m/p-Xylenes	100.000	106.866	-6.9	99	0.00
69 T	o-Xylene	50.000	54.285	-8.6	100	0.00
70 T	Styrene	50.000	55.756	-11.5	100	0.00
71 P	Bromoform	50.000	53.221	-6.4	95	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	52.417	-4.8	101	0.00
74 T	N-ethyl acetate	50.000	52.800	-5.6	101	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.548	2.9	101	0.00
76 T	1,2,3-Trichloropropane	50.000	50.070	-0.1	104	0.00
77 T	Bromobenzene	50.000	50.748	-1.5	101	0.00
78 T	n-propylbenzene	50.000	52.818	-5.6	101	0.00
79 T	2-Chlorotoluene	50.000	49.667	0.7	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.595	-5.2	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.539	-3.1	101	0.00
82 T	4-Chlorotoluene	50.000	52.097	-4.2	101	0.00
83 T	tert-Butylbenzene	50.000	51.573	-3.1	101	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.434	-4.9	100	0.00
85 T	sec-Butylbenzene	50.000	53.233	-6.5	102	0.00
86 T	p-Isopropyltoluene	50.000	53.941	-7.9	103	0.00
87 T	1,3-Dichlorobenzene	50.000	51.669	-3.3	103	0.00
88 T	1,4-Dichlorobenzene	50.000	50.549	-1.1	103	0.00
89 T	n-Butylbenzene	50.000	56.143	-12.3	106	0.00
90 T	Hexachloroethane	50.000	52.082	-4.2	100	0.00
91 T	1,2-Dichlorobenzene	50.000	50.715	-1.4	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.916	-1.8	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.309	-6.6	106	0.00
94 T	Hexachlorobutadiene	50.000	53.695	-7.4	107	0.00
95 T	Naphthalene	50.000	51.231	-2.5	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046460.D
Acq On : 03 Jun 2025 10:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.322	-4.6	103	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG No.: Q2186

Instrument ID: MSVOA_X Calibration Date/Time: 06/03/2025 20:25

Lab File ID: VX046486.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.594	0.1	-19.95	50
Vinyl Chloride	0.691	0.576		-16.64	50
Bromomethane	0.320	0.228		-28.75	50
Chloroethane	0.369	0.340		-7.86	50
Trichlorofluoromethane	1.021	0.901		-11.75	50
1,1,2-Trichlorotrifluoroethane	0.632	0.578		-8.54	50
1,1-Dichloroethene	0.593	0.534		-9.95	50
Acetone	0.374	0.407		8.82	50
Carbon Disulfide	1.406	0.941		-33.07	50
Methyl tert-butyl Ether	2.079	2.323		11.74	50
Methylene Chloride	0.716	0.657		-8.24	50
trans-1,2-Dichloroethene	0.596	0.545		-8.56	50
1,1-Dichloroethane	1.219	1.253	0.1	2.79	50
2-Butanone	0.543	0.635		16.94	50
Carbon Tetrachloride	0.544	0.515		-5.33	50
cis-1,2-Dichloroethene	0.718	0.723		0.7	50
Chloroform	1.271	1.332		4.8	50
1,1,1-Trichloroethane	1.101	1.145		4	50
Methylcyclohexane	0.623	0.554		-11.07	50
Benzene	1.417	1.402		-1.06	50
1,2-Dichloroethane	0.612	0.615		0.49	50
Trichloroethene	0.341	0.334		-2.05	50
1,2-Dichloropropane	0.352	0.382		8.52	50
Bromodichloromethane	0.547	0.572		4.57	50
4-Methyl-2-Pentanone	0.605	0.712		17.69	50
Toluene	0.869	0.878		1.04	50
t-1,3-Dichloropropene	0.487	0.523		7.39	50
cis-1,3-Dichloropropene	0.538	0.572		6.32	50
1,1,2-Trichloroethane	0.343	0.374		9.04	50
2-Hexanone	0.448	0.545		21.65	50
Dibromochloromethane	0.376	0.410		9.04	50
Tetrachloroethene	0.354	0.323		-8.76	50
Chlorobenzene	1.094	1.090	0.3	-0.37	50
Ethyl Benzene	1.929	1.971		2.18	50
m/p-Xylenes	0.706	0.715		1.27	50
o-Xylene	0.688	0.736		6.98	50
Styrene	1.127	1.222		8.43	50
Bromoform	0.281	0.305	0.1	8.54	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG No.: Q2186

Instrument ID: MSVOA_X Calibration Date/Time: 06/03/2025 20:25

Lab File ID: VX046486.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	4.049		4.01	50
1,1,2,2-Tetrachloroethane	1.364	1.472	0.3	7.92	50
1,3-Dichlorobenzene	1.649	1.715		4	50
1,4-Dichlorobenzene	1.684	1.700		0.95	50
1,2-Dichlorobenzene	1.655	1.774		7.19	50
1,2-Dichloroethane-d4	0.932	0.900		-3.43	50
Dibromofluoromethane	0.360	0.356		-1.11	50
Toluene-d8	1.246	1.176		-5.62	50
4-Bromofluorobenzene	0.478	0.493		3.14	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	80122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	143610	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72135	48.292	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.580%		
35) Dibromofluoromethane	5.379	113	51124	49.436	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.880%		
50) Toluene-d8	8.647	98	168819	47.165	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.340%		
62) 4-Bromofluorobenzene	11.079	95	70871	51.619	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	103.240%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	49352	40.244	ug/l	97
3) Chloromethane	1.307	50	47624	40.046	ug/l	97
4) Vinyl Chloride	1.374	62	46111	41.662	ug/l	99
5) Bromomethane	1.593	94	18265	35.580	ug/l	97
6) Chloroethane	1.666	64	27222	46.072	ug/l	97
7) Trichlorofluoromethane	1.874	101	72155	44.111	ug/l	99
8) Diethyl Ether	2.130	74	26541	47.664	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	46349	45.787	ug/l	97
10) Methyl Iodide	2.447	142	59982	50.077	ug/l	99
11) Tert butyl alcohol	2.977	59	69693	332.386	ug/l	99
12) 1,1-Dichloroethene	2.307	96	42762	45.010	ug/l	99
13) Acrolein	2.239	56	58249	243.938	ug/l	95
14) Allyl chloride	2.660	41	90860	50.041	ug/l	95
15) Acrylonitrile	3.062	53	166739	278.106	ug/l	98
16) Acetone	2.386	43	163107	272.337	ug/l	100
17) Carbon Disulfide	2.502	76	75405	33.478	ug/l	99
18) Methyl Acetate	2.703	43	101107	72.749	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	186100	55.872	ug/l	99
20) Methylene Chloride	2.782	84	52669	45.890	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	43658	45.695	ug/l	97
22) Diisopropyl ether	3.764	45	191730	54.666	ug/l	93
23) Vinyl Acetate	3.721	43	820777	266.071	ug/l	100
24) 1,1-Dichloroethane	3.605	63	100414	51.402	ug/l	100
25) 2-Butanone	4.556	43	254552	292.758	ug/l	100
26) 2,2-Dichloropropane	4.471	77	62527	40.894	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	57943	50.378	ug/l	98
28) Bromochloromethane	4.891	49	47479	50.493	ug/l	99
29) Tetrahydrofuran	5.007	42	156899	287.968	ug/l	99
30) Chloroform	5.087	83	106714	52.410	ug/l	97
31) Cyclohexane	5.464	56	80308	45.113	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	91763	51.989	ug/l	98
36) 1,1-Dichloropropene	5.690	75	62759	45.167	ug/l	96
37) Ethyl Acetate	4.715	43	91596	53.356	ug/l	100
38) Carbon Tetrachloride	5.672	117	74005	47.403	ug/l	97
39) Methylcyclohexane	7.379	83	79510	44.448	ug/l	99
40) Benzene	6.031	78	201308	49.463	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	54177	60.329	ug/l	98
42) 1,2-Dichloroethane	6.080	62	88249	50.240	ug/l	100
43) Isopropyl Acetate	6.342	43	155151	59.242	ug/l	99
44) Trichloroethene	7.123	130	47910	48.910	ug/l	92
45) 1,2-Dichloropropane	7.428	63	54922	54.270	ug/l	97
46) Dibromomethane	7.580	93	40591	50.854	ug/l	99
47) Bromodichloromethane	7.818	83	82121	52.237	ug/l	99
48) Methyl methacrylate	7.690	41	78532	58.715	ug/l	99
49) 1,4-Dioxane	7.665	88	30030	1182.487	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	511610	294.298	ug/l	100
52) Toluene	8.714	92	126020	50.497	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	75146	53.778	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	82115	53.169	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	53770	54.643	ug/l	98
56) Ethyl methacrylate	9.116	69	95030	60.593	ug/l	99
57) 1,3-Dichloropropane	9.305	76	92225	52.186	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	223496	279.526	ug/l	99
59) 2-Hexanone	9.427	43	391242	304.201	ug/l	100
60) Dibromochloromethane	9.519	129	58826	54.434	ug/l	100
61) 1,2-Dibromoethane	9.610	107	54971	53.748	ug/l	97
64) Tetrachloroethene	9.269	164	41027	45.626	ug/l	97
65) Chlorobenzene	10.080	112	138505	49.792	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	50287	52.942	ug/l	98
67) Ethyl Benzene	10.189	91	250449	51.077	ug/l	100
68) m/p-Xylenes	10.299	106	181843	101.398	ug/l	98
69) o-Xylene	10.640	106	93498	53.478	ug/l	99
70) Styrene	10.653	104	155343	54.240	ug/l	99
71) Bromoform	10.799	173	38761	54.353	ug/l #	98
73) Isopropylbenzene	10.957	105	247204	52.013	ug/l	99
74) N-amyl acetate	10.842	43	138341	58.905	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.207	83	89844	53.943	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	77432m	52.694	ug/l	
77) Bromobenzene	11.195	156	55881	50.644	ug/l	99
78) n-propylbenzene	11.299	91	286276	51.803	ug/l	100
79) 2-Chlorotoluene	11.360	91	176598	49.544	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	206301	51.957	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	22427	49.689	ug/l	99
82) 4-Chlorotoluene	11.451	91	202465	51.220	ug/l	99
83) tert-Butylbenzene	11.713	119	207546	51.893	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	211228	52.532	ug/l	100
85) sec-Butylbenzene	11.890	105	258780	52.697	ug/l	100
86) p-Isopropyltoluene	12.006	119	213478	52.665	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	104712	51.998	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	103759	50.452	ug/l	100
89) n-Butylbenzene	12.329	91	188888	53.124	ug/l	99
90) Hexachloroethane	12.536	117	36963	51.760	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	108286	53.586	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22293	60.419	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	64675	55.722	ug/l	100
94) Hexachlorobutadiene	13.719	225	26115	51.518	ug/l	97
95) Naphthalene	13.774	128	254087	59.688	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	66653	55.655	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046486.D
Acq On : 03 Jun 2025 20:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:35:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\  
Data File : VX046486.D  
Acq On    : 03 Jun 2025   20:25  
Operator  : JC/MD  
Sample    : VSTDCCC050  
Misc      : 5.0mL/MSVOA_X/WATER  
ALS Vial  : 28   Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:35:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:168 Resp: 8012
Ion Ratio Lower Upper
168 100
99 71.4 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#2
Dichlorodifluoromethane
Concen: 40.244 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion: 85 Resp: 49352
Ion Ratio Lower Upper
85 100
87 32.9 15.7 47.1





#3
Chloromethane
Concen: 40.046 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 47624
Ion Ratio Lower Upper
50 100
52 33.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#4
Vinyl Chloride
Concen: 41.662 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion: 62 Resp: 46111
Ion Ratio Lower Upper
62 100
64 31.8 25.2 37.8





#5
Bromomethane
Concen: 35.580 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 1826
Ion Ratio Lower Upper
94 100
96 93.5 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#6
Chloroethane
Concen: 46.072 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion: 64 Resp: 27222
Ion Ratio Lower Upper
64 100
66 31.9 24.3 36.5





#7
Trichlorofluoromethane
Concen: 44.111 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 7215
Ion Ratio Lower Upper
101 100
103 62.7 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#8
Diethyl Ether
Concen: 47.664 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 74 Resp: 26541
Ion Ratio Lower Upper
74 100
45 107.2 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.787 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 46349
Ion Ratio Lower Upper
101 100
85 47.8 38.6 58.0
151 72.2 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#10
Methyl Iodide
Concen: 50.077 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion:142 Resp: 59982
Ion Ratio Lower Upper
142 100
127 51.1 41.7 62.5
141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 332.386 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 6969

Ion Ratio Lower Upper

59 100





57 11.2 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#12

1,1-Dichloroethene

Concen: 45.010 ug/l

RT: 2.307 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 96 Resp: 42762


Ion Ratio Lower Upper

96 100

61 181.6 146.2 219.2


98 63.7 51.0 76.6





#13
Acrolein
Concen: 243.938 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 58249
Ion Ratio Lower Upper
56 100
55 74.0 56.2 84.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#14
Allyl chloride
Concen: 50.041 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 41 Resp: 90860
Ion Ratio Lower Upper
41 100
39 70.7 60.6 90.8
76 29.7 24.9 37.3





#15

Acrylonitrile

Concen: 278.106 ug/l

RT: 3.062 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	53	Resp:	166739
Ion	Ratio	Lower	Upper
53	100		
52	84.2	65.3	97.9
51	38.1	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#16

Acetone

Concen: 272.337 ug/l




RT: 2.386 min Scan# 213


Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:	43	Resp:	163107
Ion	Ratio	Lower	Upper
43	100		
58	26.5	21.2	31.8





#17

Carbon Disulfide

Concen: 33.478 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 7540

Ion Ratio Lower Upper

76 100





78 9.0 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#18

Methyl Acetate

Concen: 72.749 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 43 Resp: 101107

Ion Ratio Lower Upper

43 100

74 21.1 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 55.872 ug/l

RT: 3.117 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 73 Resp: 186100

Ion Ratio Lower Upper

73 100





57 21.7 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#20

Methylene Chloride

Concen: 45.890 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion: 84 Resp: 52669




Ion Ratio Lower Upper


84 100

49 138.3 113.9 170.9

51 42.2 33.5 50.3

86 66.6 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 45.695 ug/l

RT: 3.087 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 43658

Ion Ratio Lower Upper

96 100

61 150.7 119.5 179.3





98 67.4 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#22

Diisopropyl ether

Concen: 54.666 ug/l

RT: 3.764 min Scan# 439

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion: 45 Resp: 191730




Ion Ratio Lower Upper


45 100

43 73.9 66.6 100.0

87 24.9 19.8 29.6

59 10.3 8.6 12.8









#23
 Vinyl Acetate
 Concen: 266.071 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 43 Resp: 82077
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.5 11.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/04/2025
 Supervised By :Mahesh Dadoda 06/04/2025



#24
 1,1-Dichloroethane
 Concen: 51.402 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25

Tgt Ion: 63 Resp: 100414
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.6 10.8
 100 4.5 2.1 6.3





#25

2-Butanone

Concen: 292.758 ug/l

RT: 4.556 min Scan# 569

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 254552

Ion Ratio Lower Upper

43 100





72 23.1 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#26

2,2-Dichloropropane

Concen: 40.894 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 77 Resp: 62527

Ion Ratio Lower Upper

77 100

97 22.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 50.378 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 5794

Ion Ratio Lower Upper

96 100

61 158.4 0.0 322.8





98 65.2 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#28

Bromochloromethane

Concen: 50.493 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 49 Resp: 47479


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.8 56.2 84.2





#29

Tetrahydrofuran

Concen: 287.968 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 42 Resp: 15689

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#30

Chloroform

Concen: 52.410 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 83 Resp: 106714

Ion Ratio Lower Upper

83 100

85 63.7 49.3 73.9





#31

Cyclohexane

Concen: 45.113 ug/l

RT: 5.464 min Scan# 718

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 80308

Ion Ratio Lower Upper

56 100

69 32.1 24.4 36.6





84 82.6 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#32

1,1,1-Trichloroethane

Concen: 51.989 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 97 Resp: 91763


Ion Ratio Lower Upper

97 100

99 63.9 51.8 77.6

61 47.6 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.292 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 72135

Ion Ratio Lower Upper

65 100





67 49.8 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion:114 Resp: 143610


Ion Ratio Lower Upper

114 100

63 23.8 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.436 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	113	Resp:	51124
Ion	Ratio	Lower	Upper
113	100		
111	103.0	83.1	124.7
192	16.9	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#36

1,1-Dichloropropene

Concen: 45.167 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:	75	Resp:	62759
Ion	Ratio	Lower	Upper
75	100		
110	34.5	16.3	48.9
77	32.5	24.3	36.5





#37

Ethyl Acetate

Concen: 53.356 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 91590
Ion Ratio Lower Upper
43 100
61 13.0 10.3 15.5
70 10.0 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#38

Carbon Tetrachloride

Concen: 47.403 ug/l




RT: 5.672 min Scan# 752


Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:117 Resp: 74005
Ion Ratio Lower Upper
117 100
119 97.7 75.2 112.8
121 29.8 24.2 36.4





#39

Methylcyclohexane

Concen: 44.448 ug/l

RT: 7.379 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 83	Resp: 79510
Ion Ratio	Lower Upper
83	100
55	79.5 64.7 97.1
98	46.5 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#40

Benzene

Concen: 49.463 ug/l




RT: 6.031 min Scan# 811


Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25


Tgt Ion: 78	Resp: 201308
Ion Ratio	Lower Upper
78	100
77	23.2 19.0 28.4





#41
Methacrylonitrile
Concen: 60.329 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 5417
Ion Ratio Lower Upper
41 100
39 57.8 47.2 70.8
67 65.3 50.7 76.1
52 30.3 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#42
1,2-Dichloroethane
Concen: 50.240 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 62 Resp: 88249
Ion Ratio Lower Upper
62 100
98 7.7 0.0 15.2





#43

Isopropyl Acetate

Concen: 59.242 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046486.D


Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 43	Resp: 15515
Ion Ratio	Lower Upper
43	100
61	18.4 14.3 21.5
87	11.6 9.5 14.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#44

Trichloroethene


Concen: 48.910 ug/l

RT: 7.123 min Scan# 990



Delta R.T. 0.000 min


Lab File: VX046486.D

Acq: 03 Jun 2025 20:25



Tgt Ion:130	Resp: 47910
Ion Ratio	Lower Upper
130	100
95	110.4 0.0 204.2





#45

1,2-Dichloropropane

Concen: 54.270 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 5492

Ion Ratio Lower Upper

63 100





65 28.7 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#46

Dibromomethane

Concen: 50.854 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 93 Resp: 40591


Ion Ratio Lower Upper

93 100

95 81.4 65.6 98.4

174 84.7 68.2 102.2





#47

Bromodichloromethane

Concen: 52.237 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	83	Resp:	8212
Ion	Ratio	Lower	Upper
83	100		
85	62.0	50.5	75.7
127	8.0	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#48

Methyl methacrylate

Concen: 58.715 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:	41	Resp:	78532
Ion	Ratio	Lower	Upper
41	100		
69	73.8	58.5	87.7
39	63.9	51.7	77.5





#49

1,4-Dioxane

Concen: 1182.487 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	88	Resp:	30030
Ion	Ratio	Lower	Upper
88	100		
43	40.7	33.4	50.2
58	71.7	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#50

Toluene-d8

Concen: 47.165 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:	98	Resp:	168819
Ion	Ratio	Lower	Upper
98	100		
100	66.3	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 294.298 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 511610
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#52
Toluene
Concen: 50.497 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 92 Resp: 126020
Ion Ratio Lower Upper
92 100
91 171.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 53.778 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 75 Resp: 75140

Ion Ratio Lower Upper

75 100





77 31.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#54

cis-1,3-Dichloropropene

Concen: 53.169 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 75 Resp: 82115


Ion Ratio Lower Upper

75 100

77 31.3 25.4 38.0

39 59.5 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 54.643 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 53770

Ion Ratio Lower Upper

97 100

83 84.6 69.0 103.6

85 53.5 44.3 66.5





99 60.0 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#56

Ethyl methacrylate

Concen: 60.593 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 69 Resp: 95030


Ion Ratio Lower Upper

69 100

41 76.2 60.8 91.2

39 46.6 39.0 58.6





#57

1,3-Dichloropropane

Concen: 52.186 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 92225

Ion Ratio Lower Upper

76 100





78 32.8 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#58

2-Chloroethyl Vinyl ether

Concen: 279.526 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 63 Resp: 223496

Ion Ratio Lower Upper

63 100

106 26.6 21.5 32.3





#59

2-Hexanone

Concen: 304.201 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 39124

Ion Ratio Lower Upper

43 100





58 49.9 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#60

Dibromochloromethane

Concen: 54.434 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion:129 Resp: 58826

Ion Ratio Lower Upper

129 100


127 78.1 39.3 117.8





#61
1,2-Dibromoethane
Concen: 53.748 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 107 Resp: 5497
Ion Ratio Lower Upper
107 100
109 94.2 73.3 109.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 06/04/2025
Supervised By : Mahesh Dadoda 06/04/2025



#62
4-Bromofluorobenzene
Concen: 51.619 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion: 95 Resp: 70871
Ion Ratio Lower Upper
95 100
174 66.9 0.0 135.8
176 63.4 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 127089
Ion Ratio Lower Upper
117 100
82 57.3 50.6 76.0
119 31.2 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#64
Tetrachloroethene
Concen: 45.626 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion:164 Resp: 41027
Ion Ratio Lower Upper
164 100
166 129.2 105.0 157.6
129 104.9 83.5 125.3
131 104.5 76.5 114.7





#65
Chlorobenzene
Concen: 49.792 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:112 Resp: 13850
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 52.942 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion:131 Resp: 50287
Ion Ratio Lower Upper
131 100
133 92.9 47.3 141.9
119 66.0 31.6 95.0





#67
Ethyl Benzene
Concen: 51.077 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25



Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 250449
Ion Ratio Lower Upper
91 100
106 29.5 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#68
m/p-Xylenes
Concen: 101.398 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion:106 Resp: 181843
Ion Ratio Lower Upper
106 100
91 216.6 171.2 256.8





#69
o-Xylene
Concen: 53.478 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:106 Resp: 93498
Ion Ratio Lower Upper
106 100
91 223.7 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#70
Styrene
Concen: 54.240 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion:104 Resp: 155343
Ion Ratio Lower Upper
104 100
78 58.2 45.7 68.5
103 54.4 43.7 65.5





#71
Bromoform
Concen: 54.353 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:173 Resp: 3876
Ion Ratio Lower Upper
173 100
175 47.7 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25


Tgt Ion:152 Resp: 61048
Ion Ratio Lower Upper
152 100
115 93.4 46.9 140.7
150 174.9 0.0 351.0





#73
Isopropylbenzene
Concen: 52.013 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 247204
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#74
N-ethyl acetate
Concen: 58.905 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 43 Resp: 138341
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.1 18.7 28.1
61 22.3 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 53.943 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	83	Resp:	8984
Ion	Ratio	Lower	Upper
83	100		
131	9.7	5.0	14.9
85	63.8	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#76

1,2,3-Trichloropropane

Concen: 52.694 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:	75	Resp:	77432
Ion	Ratio	Lower	Upper
75	100		
77	40.2	20.5	61.5





#77

Bromobenzene

Concen: 50.644 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:156 Resp: 5588

Ion Ratio Lower Upper

156 100

77 187.4 92.8 278.3





158 96.7 48.8 146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#78

n-propylbenzene

Concen: 51.803 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 91 Resp: 286276

Ion Ratio Lower Upper

91 100


120 21.7 10.8 32.4





#79
2-Chlorotoluene
Concen: 49.544 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 176598
Ion Ratio Lower Upper
91 100
126 31.7 15.6 46.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025




#80
1,3,5-Trimethylbenzene
Concen: 51.957 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Tgt Ion:105 Resp: 206301
Ion Ratio Lower Upper
105 100
120 46.7 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 49.689 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 2242
Ion Ratio Lower Upper
75 100
53 126.9 100.5 150.7
89 47.6 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#82
4-Chlorotoluene
Concen: 51.220 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 91 Resp: 202465
Ion Ratio Lower Upper
91 100
126 26.9 13.3 39.8





#83

tert-Butylbenzene

Concen: 51.893 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:119 Resp: 207540

Ion Ratio Lower Upper

119 100

91 66.9 32.9 98.7





134 23.2 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#84

1,2,4-Trimethylbenzene

Concen: 52.532 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion:105 Resp: 211228

Ion Ratio Lower Upper

105 100


120 42.7 21.2 63.6





#85
 sec-Butylbenzene
 Concen: 52.697 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 258780
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.7 29.1

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/04/2025
 Supervised By :Mahesh Dadoda 06/04/2025




#86
 p-Isopropyltoluene
 Concen: 52.665 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25




Tgt Ion:119 Resp: 213478
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.5 37.5
 91 27.3 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 51.998 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 104712
Ion Ratio Lower Upper
146 100
111 43.7 22.1 66.3
148 63.0 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025




#88
1,4-Dichlorobenzene
Concen: 50.452 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25



Tgt Ion:146 Resp: 103759
Ion Ratio Lower Upper
146 100
111 42.9 21.3 63.9
148 63.8 31.9 95.5





#89

n-Butylbenzene

Concen: 53.124 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 18888

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7





134 24.4 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#90

Hexachloroethane

Concen: 51.760 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion:117 Resp: 36963

Ion Ratio Lower Upper

117 100


201 62.5 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 53.586 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 108280
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 62.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 60.419 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 75 Resp: 22293
Ion Ratio Lower Upper
75 100
155 70.8 34.9 104.8
157 91.0 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 55.722 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:180 Resp: 64675





Ion	Ratio	Lower	Upper
180	100		
182	95.5	47.9	143.8
145	34.3	17.1	51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#94

Hexachlorobutadiene

Concen: 51.518 ug/l

RT: 13.719 min Scan# 2072




Delta R.T. -0.006 min


Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:225 Resp: 26115


Ion	Ratio	Lower	Upper
225	100		
223	63.3	30.8	92.4
227	65.4	30.9	92.7





#95
Naphthalene
Concen: 59.688 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 25408
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 11.0 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#96
1,2,3-Trichlorobenzene
Concen: 55.655 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion:180 Resp: 66653
Ion Ratio Lower Upper
180 100
182 94.7 47.8 143.3
145 36.2 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.765	0.616	19.5	59	0.00
3 P	Chloromethane	0.742	0.594	19.9	63	0.00
4 C	Vinyl Chloride	0.691	0.576	16.6#	67	0.00
5 T	Bromomethane	0.320	0.228	28.7#	58	0.00
6 T	Chloroethane	0.369	0.340	7.9	74	0.00
7 T	Trichlorofluoromethane	1.021	0.901	11.8	70	0.00
8 T	Diethyl Ether	0.347	0.331	4.6	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.578	8.5	74	0.00
10 T	Methyl Iodide	0.747	0.749	-0.3	77	0.00
11 T	Tert butyl alcohol	0.131	0.174	-32.8#	111	0.00
12 CM	1,1-Dichloroethene	0.593	0.534	9.9#	73	0.00
13 T	Acrolein	0.149	0.145	2.7	79	0.00
14 T	Allyl chloride	1.133	1.134	-0.1	79	0.00
15 T	Acrylonitrile	0.374	0.416	-11.2	89	0.00
16 T	Acetone	0.374	0.407	-8.8	93	0.00
17 T	Carbon Disulfide	1.406	0.941	33.1#	53	0.00
18 T	Methyl Acetate	0.867	1.262	-45.6#	123	0.00
19 T	Methyl tert-butyl Ether	2.079	2.323	-11.7	89	0.00
20 T	Methylene Chloride	0.716	0.657	8.2	79	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.545	8.6	74	0.00
22 T	Diisopropyl ether	2.189	2.393	-9.3	87	0.00
23 T	Vinyl Acetate	1.925	2.049	-6.4	83	0.00
24 P	1,1-Dichloroethane	1.219	1.253	-2.8	82	0.00
25 T	2-Butanone	0.543	0.635	-16.9	95	0.00
26 T	2,2-Dichloropropane	0.954	0.780	18.2	67	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.723	-0.7	81	0.00
28 T	Bromochloromethane	0.587	0.593	-1.0	85	0.00
29 T	Tetrahydrofuran	0.340	0.392	-15.3	92	0.00
30 C	Chloroform	1.271	1.332	-4.8#	85	0.00
31 T	Cyclohexane	1.111	1.002	9.8	73	0.00
32 T	1,1,1-Trichloroethane	1.101	1.145	-4.0	84	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.900	3.4	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.360	0.356	1.1	86	0.00
36 T	1,1-Dichloropropene	0.484	0.437	9.7	75	0.00
37 T	Ethyl Acetate	0.598	0.638	-6.7	89	0.00
38 T	Carbon Tetrachloride	0.544	0.515	5.3	79	0.00
39 T	Methylcyclohexane	0.623	0.554	11.1	74	0.00
40 TM	Benzene	1.417	1.402	1.1	81	0.00
41 T	Methacrylonitrile	0.313	0.377	-20.4	93	0.00
42 TM	1,2-Dichloroethane	0.612	0.615	-0.5	84	0.00
43 T	Isopropyl Acetate	0.912	1.080	-18.4	96	0.00
44 TM	Trichloroethene	0.341	0.334	2.1	80	0.00
45 C	1,2-Dichloropropane	0.352	0.382	-8.5#	88	0.00
46 T	Dibromomethane	0.278	0.283	-1.8	84	0.00
47 T	Bromodichloromethane	0.547	0.572	-4.6	85	0.00
48 T	Methyl methacrylate	0.466	0.547	-17.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	98	0.00
50 S	Toluene-d8	1.246	1.176	5.6	82	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.712	-17.7	96	0.00
52 CM	Toluene	0.869	0.878	-1.0#	84	0.00
53 T	t-1,3-Dichloropropene	0.487	0.523	-7.4	85	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.572	-6.3	85	0.00
55 T	1,1,2-Trichloroethane	0.343	0.374	-9.0	90	0.00
56 T	Ethyl methacrylate	0.546	0.662	-21.2	95	0.00
57 T	1,3-Dichloropropane	0.615	0.642	-4.4	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.311	-11.9	87	0.00
59 T	2-Hexanone	0.448	0.545	-21.7	99	0.00
60 T	Dibromochloromethane	0.376	0.410	-9.0	88	0.00
61 T	1,2-Dibromoethane	0.356	0.383	-7.6	88	0.00
62 S	4-Bromofluorobenzene	0.478	0.493	-3.1	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.354	0.323	8.8	75	0.00
65 PM	Chlorobenzene	1.094	1.090	0.4	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.396	-5.9	88	0.00
67 C	Ethyl Benzene	1.929	1.971	-2.2#	85	0.00
68 T	m/p-Xylenes	0.706	0.715	-1.3	84	0.00
69 T	o-Xylene	0.688	0.736	-7.0	88	0.00
70 T	Styrene	1.127	1.222	-8.4	87	0.00
71 P	Bromoform	0.281	0.305	-8.5	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.893	4.049	-4.0	88	0.00
74 T	N-amyl acetate	1.924	2.266	-17.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.472	-7.9	99	0.00
76 T	1,2,3-Trichloropropane	1.204	1.268	-5.3	96	0.00
77 T	Bromobenzene	0.904	0.915	-1.2	89	0.00
78 T	n-propylbenzene	4.526	4.689	-3.6	87	0.00
79 T	2-Chlorotoluene	2.919	2.893	0.9	87	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.379	-3.9	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.367	0.8	86	0.00
82 T	4-Chlorotoluene	3.238	3.316	-2.4	87	0.00
83 T	tert-Butylbenzene	3.276	3.400	-3.8	89	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.460	-5.1	88	0.00
85 T	sec-Butylbenzene	4.022	4.239	-5.4	89	0.00
86 T	p-Isopropyltoluene	3.320	3.497	-5.3	88	0.00
87 T	1,3-Dichlorobenzene	1.649	1.715	-4.0	91	0.00
88 T	1,4-Dichlorobenzene	1.684	1.700	-1.0	90	0.00
89 T	n-Butylbenzene	2.912	3.094	-6.2	88	0.00
90 T	Hexachloroethane	0.585	0.605	-3.4	87	0.00
91 T	1,2-Dichlorobenzene	1.655	1.774	-7.2	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.365	-20.9	102	0.00
93 T	1,2,4-Trichlorobenzene	0.951	1.059	-11.4	97	0.00
94 T	Hexachlorobutadiene	0.415	0.428	-3.1	90	0.00
95 T	Naphthalene	3.487	4.162	-19.4	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046486.D
Acq On : 03 Jun 2025 20:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 04 01:35:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	1.092	-11.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	50.000	40.244	19.5	59	0.00
3 P	Chloromethane	50.000	40.046	19.9	63	0.00
4 C	Vinyl Chloride	50.000	41.662	16.7#	67	0.00
5 T	Bromomethane	50.000	35.580	28.8#	58	0.00
6 T	Chloroethane	50.000	46.072	7.9	74	0.00
7 T	Trichlorofluoromethane	50.000	44.111	11.8	70	0.00
8 T	Diethyl Ether	50.000	47.664	4.7	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	45.787	8.4	74	0.00
10 T	Methyl Iodide	50.000	50.077	-0.2	77	0.00
11 T	Tert butyl alcohol	250.000	332.386	-33.0#	111	0.00
12 CM	1,1-Dichloroethene	50.000	45.010	10.0#	73	0.00
13 T	Acrolein	250.000	243.938	2.4	79	0.00
14 T	Allyl chloride	50.000	50.041	-0.1	79	0.00
15 T	Acrylonitrile	250.000	278.106	-11.2	89	0.00
16 T	Acetone	250.000	272.337	-8.9	93	0.00
17 T	Carbon Disulfide	50.000	33.478	33.0#	53	0.00
18 T	Methyl Acetate	50.000	72.749	-45.5#	123	0.00
19 T	Methyl tert-butyl Ether	50.000	55.872	-11.7	89	0.00
20 T	Methylene Chloride	50.000	45.890	8.2	79	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.695	8.6	74	0.00
22 T	Diisopropyl ether	50.000	54.666	-9.3	87	0.00
23 T	Vinyl Acetate	250.000	266.071	-6.4	83	0.00
24 P	1,1-Dichloroethane	50.000	51.402	-2.8	82	0.00
25 T	2-Butanone	250.000	292.758	-17.1	95	0.00
26 T	2,2-Dichloropropane	50.000	40.894	18.2	67	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.378	-0.8	81	0.00
28 T	Bromochloromethane	50.000	50.493	-1.0	85	0.00
29 T	Tetrahydrofuran	250.000	287.968	-15.2	92	0.00
30 C	Chloroform	50.000	52.410	-4.8#	85	0.00
31 T	Cyclohexane	50.000	45.113	9.8	73	0.00
32 T	1,1,1-Trichloroethane	50.000	51.989	-4.0	84	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.292	3.4	82	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	86	0.00
35 S	Dibromofluoromethane	50.000	49.436	1.1	86	0.00
36 T	1,1-Dichloropropene	50.000	45.167	9.7	75	0.00
37 T	Ethyl Acetate	50.000	53.356	-6.7	89	0.00
38 T	Carbon Tetrachloride	50.000	47.403	5.2	79	0.00
39 T	Methylcyclohexane	50.000	44.448	11.1	74	0.00
40 TM	Benzene	50.000	49.463	1.1	81	0.00
41 T	Methacrylonitrile	50.000	60.329	-20.7	93	0.00
42 TM	1,2-Dichloroethane	50.000	50.240	-0.5	84	0.00
43 T	Isopropyl Acetate	50.000	59.242	-18.5	96	0.00
44 TM	Trichloroethene	50.000	48.910	2.2	80	0.00
45 C	1,2-Dichloropropane	50.000	54.270	-8.5#	88	0.00
46 T	Dibromomethane	50.000	50.854	-1.7	84	0.00
47 T	Bromodichloromethane	50.000	52.237	-4.5	85	0.00
48 T	Methyl methacrylate	50.000	58.715	-17.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1182.487	-18.2	98	0.00
50 S	Toluene-d8	50.000	47.165	5.7	82	0.00
51 T	4-Methyl-2-Pentanone	250.000	294.298	-17.7	96	0.00
52 CM	Toluene	50.000	50.497	-1.0#	84	0.00
53 T	t-1,3-Dichloropropene	50.000	53.778	-7.6	85	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.169	-6.3	85	0.00
55 T	1,1,2-Trichloroethane	50.000	54.643	-9.3	90	0.00
56 T	Ethyl methacrylate	50.000	60.593	-21.2	95	0.00
57 T	1,3-Dichloropropane	50.000	52.186	-4.4	88	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	279.526	-11.8	87	0.00
59 T	2-Hexanone	250.000	304.201	-21.7	99	0.00
60 T	Dibromochloromethane	50.000	54.434	-8.9	88	0.00
61 T	1,2-Dibromoethane	50.000	53.748	-7.5	88	0.00
62 S	4-Bromofluorobenzene	50.000	51.619	-3.2	90	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	87	0.00
64 T	Tetrachloroethene	50.000	45.626	8.7	75	0.00
65 PM	Chlorobenzene	50.000	49.792	0.4	86	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.942	-5.9	88	0.00
67 C	Ethyl Benzene	50.000	51.077	-2.2#	85	0.00
68 T	m/p-Xylenes	100.000	101.398	-1.4	84	0.00
69 T	o-Xylene	50.000	53.478	-7.0	88	0.00
70 T	Styrene	50.000	54.240	-8.5	87	0.00
71 P	Bromoform	50.000	54.353	-8.7	87	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	90	0.00
73 T	Isopropylbenzene	50.000	52.013	-4.0	88	0.00
74 T	N-ethyl acetate	50.000	58.905	-17.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	53.943	-7.9	99	0.00
76 T	1,2,3-Trichloropropane	50.000	52.694	-5.4	96	0.00
77 T	Bromobenzene	50.000	50.644	-1.3	89	0.00
78 T	n-propylbenzene	50.000	51.803	-3.6	87	0.00
79 T	2-Chlorotoluene	50.000	49.544	0.9	87	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.957	-3.9	87	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.689	0.6	86	0.00
82 T	4-Chlorotoluene	50.000	51.220	-2.4	87	0.00
83 T	tert-Butylbenzene	50.000	51.893	-3.8	89	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.532	-5.1	88	0.00
85 T	sec-Butylbenzene	50.000	52.697	-5.4	89	0.00
86 T	p-Isopropyltoluene	50.000	52.665	-5.3	88	0.00
87 T	1,3-Dichlorobenzene	50.000	51.998	-4.0	91	0.00
88 T	1,4-Dichlorobenzene	50.000	50.452	-0.9	90	0.00
89 T	n-Butylbenzene	50.000	53.124	-6.2	88	0.00
90 T	Hexachloroethane	50.000	51.760	-3.5	87	0.00
91 T	1,2-Dichlorobenzene	50.000	53.586	-7.2	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	60.419	-20.8	102	0.00
93 T	1,2,4-Trichlorobenzene	50.000	55.722	-11.4	97	0.00
94 T	Hexachlorobutadiene	50.000	51.518	-3.0	90	0.00
95 T	Naphthalene	50.000	59.688	-19.4	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046486.D
Acq On : 03 Jun 2025 20:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 04 01:35:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	55.655	-11.3	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046038.D
Acq On : 05 May 2025 09:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631


Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.1	8676	PASS
75	95	30	60	56.2	22052	PASS
95	95	100	100	100.0	39213	PASS
96	95	5	9	6.4	2505	PASS
173	174	0.00	2	0.7	185	PASS
174	95	50	100	68.8	26963	PASS
175	174	5	9	7.3	1980	PASS
176	174	95	101	97.0	26147	PASS
177	176	5	9	6.9	1802	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046459.D
Acq On : 03 Jun 2025 09:40
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	8520	PASS
75	95	30	60	56.8	21483	PASS
95	95	100	100	100.0	37835	PASS
96	95	5	9	6.6	2509	PASS
173	174	0.00	2	1.1	281	PASS
174	95	50	100	69.6	26341	PASS
175	174	5	9	7.3	1930	PASS
176	174	95	101	95.2	25067	PASS
177	176	5	9	6.2	1559	PASS

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0603WBL01		SDG No.:	Q2186
Lab Sample ID:	VX0603WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046462.D	1		06/03/25 11:01	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	
Client Sample ID:	VX0603WBL01			SDG No.:	Q2186
Lab Sample ID:	VX0603WBL01			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046462.D	1		06/03/25 11:01	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.5		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	49.2		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	49.8		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.4		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	73200	5.55				
540-36-3	1,4-Difluorobenzene	148000	6.757				
3114-55-4	Chlorobenzene-d5	139000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	60900	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046462.D
 Acq On : 03 Jun 2025 11:01
 Operator : JC/MD
 Sample : VX0603WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBL01

Quant Time: Jun 04 01:23:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	73151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147671	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60850	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70256	51.516	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.040%
35) Dibromofluoromethane	5.379	113	52267	49.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	8.647	98	183380	49.824	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	72606	51.428	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.860%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046462.D
Acq On : 03 Jun 2025 11:01
Operator : JC/MD
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Quant Time: Jun 04 01:23:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046462.D
Acq: 03 Jun 2025 11:01




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01


Tgt Ion:168 Resp: 73151
Ion Ratio Lower Upper
168 100
99 68.3 54.9 82.3



#33
1,2-Dichloroethane-d4
Concen: 51.516 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX046462.D
Acq: 03 Jun 2025 11:01

Tgt Ion: 65 Resp: 70256
Ion Ratio Lower Upper
65 100
67 50.0 0.0 99.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX046462.D

Acq: 03 Jun 2025 11:01

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBL01





Tgt Ion:114 Resp: 147671

Ion Ratio Lower Upper

114 100

63 24.0 0.0 49.2

88 17.1 0.0 33.6



#35

Dibromofluoromethane

Concen: 49.151 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046462.D

Acq: 03 Jun 2025 11:01




Tgt Ion:113 Resp: 52267


Ion Ratio Lower Upper

113 100

111 103.3 83.1 124.7

192 17.3 13.3 19.9









#50
Toluene-d8
Concen: 49.824 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046462.D
Acq: 03 Jun 2025 11:01




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01


Tgt Ion: 98 Resp: 183380
Ion Ratio Lower Upper
98 100
100 65.2 53.5 80.3



#62
4-Bromofluorobenzene
Concen: 51.428 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046462.D
Acq: 03 Jun 2025 11:01

Tgt Ion: 95 Resp: 72606
Ion Ratio Lower Upper
95 100
174 69.0 0.0 135.8
176 65.4 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046462.D
Acq: 03 Jun 2025 11:01




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Tgt Ion:117 Resp: 138701
Ion Ratio Lower Upper
117 100
82 61.1 50.6 76.0
119 32.0 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046462.D
Acq: 03 Jun 2025 11:01

Tgt Ion:152 Resp: 60850
Ion Ratio Lower Upper
152 100
115 64.3 46.9 140.7
150 155.7 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046462.D
Acq On : 03 Jun 2025 11:01
Operator : JC/MD
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046462.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.581	69	81	85	rBV3	4437	11357	2.21%	0.414%
2	5.379	692	704	722	rBV3	59541	178016	34.65%	6.489%
3	5.544	722	731	744	rVV2	82987	242365	47.17%	8.834%
4	5.952	787	798	814	rBV	67962	183834	35.78%	6.701%
5	6.757	921	930	944	rBV	153265	372864	72.57%	13.591%
6	8.647	1233	1240	1253	rBV	323391	513828	100.00%	18.729%
7	10.049	1465	1470	1487	rBV	329679	467752	91.03%	17.050%
8	11.079	1634	1639	1650	rBV	277033	356878	69.45%	13.008%
9	12.018	1788	1793	1808	rBV	337692	416570	81.07%	15.184%




Sum of corrected areas: 2743464

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046462.D
Acq On : 03 Jun 2025 11:01
Operator : JC/MD
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046462.D
Acq On : 03 Jun 2025 11:01
Operator : JC/MD
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046462.D
Acq On : 03 Jun 2025 11:01
Operator : JC/MD
Sample : VX0603WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:			
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:			
Client Sample ID:	VX0603WBS01			SDG No.:	Q2186		
Lab Sample ID:	VX0603WBS01			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046463.D	1		06/03/25 11:24	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	17.0		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	16.6		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	16.9		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.1		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.9		0.23	0.75	1.00	ug/L
67-64-1	Acetone	100		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	13.3		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.3		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	18.4		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.0		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.5		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.4		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	20.9		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.1		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.7		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.5		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.2		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.9		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.0		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.9		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.2		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.9		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	
Client Sample ID:	VX0603WBS01			SDG No.:	Q2186
Lab Sample ID:	VX0603WBS01			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046463.D	1		06/03/25 11:24	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	20.5		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.8		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.2		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.4		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	40.4		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	21.0		0.12	0.50	1.00	ug/L
100-42-5	Styrene	21.2		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.9		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	21.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.1		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.6		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	21.3		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.8		81 - 118		98%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	47.8		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.6		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	89900	5.544				
540-36-3	1,4-Difluorobenzene	159000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	65300	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046463.D
 Acq On : 03 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0603WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:23:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	89858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	159016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81765	48.808	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.620%		
35) Dibromofluoromethane	5.379	113	57691	50.381	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.760%		
50) Toluene-d8	8.647	98	189408	47.791	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.580%		
62) 4-Bromofluorobenzene	11.079	95	76982	50.637	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	101.280%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	22784	16.566	ug/l	98
3) Chloromethane	1.301	50	22687	17.010	ug/l	98
4) Vinyl Chloride	1.374	62	20626	16.617	ug/l	100
5) Bromomethane	1.599	94	9717	16.878	ug/l	98
6) Chloroethane	1.672	64	12637	19.070	ug/l	91
7) Trichlorofluoromethane	1.880	101	34563	18.840	ug/l	98
8) Diethyl Ether	2.130	74	11617	18.602	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	21267	18.733	ug/l	97
10) Methyl Iodide	2.447	142	23408	17.425	ug/l	99
11) Tert butyl alcohol	2.971	59	26285	111.778	ug/l	100
12) 1,1-Dichloroethene	2.313	96	19119	17.944	ug/l	99
13) Acrolein	2.233	56	31641	118.151	ug/l	99
14) Allyl chloride	2.654	41	40050	19.668	ug/l	98
15) Acrylonitrile	3.062	53	70799	105.292	ug/l	98
16) Acetone	2.380	43	69671	103.724	ug/l	99
17) Carbon Disulfide	2.508	76	33630	13.313	ug/l	97
18) Methyl Acetate	2.703	43	42289	27.131	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	75981	20.340	ug/l	99
20) Methylene Chloride	2.782	84	23624	18.353	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	19266	17.980	ug/l	95
22) Diisopropyl ether	3.757	45	82535	20.983	ug/l	95
23) Vinyl Acetate	3.721	43	335531	96.984	ug/l	100
24) 1,1-Dichloroethane	3.605	63	44898	20.493	ug/l	99
25) 2-Butanone	4.556	43	104022	106.672	ug/l	96
26) 2,2-Dichloropropane	4.471	77	33066	19.283	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	26341	20.421	ug/l	98
28) Bromochloromethane	4.891	49	22740	21.563	ug/l	97
29) Tetrahydrofuran	5.001	42	64707	105.894	ug/l	99
30) Chloroform	5.086	83	47703	20.890	ug/l	94
31) Cyclohexane	5.464	56	35430	17.746	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	39775	20.093	ug/l	99
36) 1,1-Dichloropropene	5.684	75	28514	18.533	ug/l	99
37) Ethyl Acetate	4.715	43	37358	19.653	ug/l	99
38) Carbon Tetrachloride	5.666	117	31976	18.497	ug/l	96
39) Methylcyclohexane	7.373	83	35078	17.710	ug/l	98
40) Benzene	6.037	78	87833	19.490	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046463.D
 Acq On : 03 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0603WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:23:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	21713	21.836	ug/l	97
42) 1,2-Dichloroethane	6.080	62	39263	20.187	ug/l	100
43) Isopropyl Acetate	6.342	43	59297	20.448	ug/l	99
44) Trichloroethene	7.123	130	20550	18.946	ug/l	98
45) 1,2-Dichloropropane	7.427	63	23439	20.917	ug/l	100
46) Dibromomethane	7.580	93	17506	19.808	ug/l	100
47) Bromodichloromethane	7.818	83	34890	20.043	ug/l	94
48) Methyl methacrylate	7.690	41	31312	21.143	ug/l	99
49) 1,4-Dioxane	7.659	88	12369	439.865	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	210362	109.285	ug/l	98
52) Toluene	8.714	92	54891	19.864	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	29748	19.227	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	34079	19.928	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	23413	21.488	ug/l	96
56) Ethyl methacrylate	9.116	69	37717	21.719	ug/l	98
57) 1,3-Dichloropropane	9.305	76	39888	20.384	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	94629	106.886	ug/l	100
59) 2-Hexanone	9.427	43	158797	111.507	ug/l	98
60) Dibromochloromethane	9.519	129	24550	20.516	ug/l	97
61) 1,2-Dibromoethane	9.610	107	22898	20.220	ug/l	95
64) Tetrachloroethene	9.269	164	19186	19.782	ug/l	96
65) Chlorobenzene	10.079	112	60633	20.209	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21133	20.627	ug/l	98
67) Ethyl Benzene	10.189	91	108068	20.434	ug/l	100
68) m/p-Xylenes	10.299	106	78193	40.424	ug/l	98
69) o-Xylene	10.640	106	39576	20.987	ug/l	98
70) Styrene	10.652	104	65394	21.169	ug/l	99
71) Bromoform	10.799	173	15282	19.868	ug/l #	99
73) Isopropylbenzene	10.957	105	106972	21.051	ug/l	99
74) N-amyl acetate	10.841	43	53285	21.221	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	38144	21.421	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	32360m	20.597	ug/l	
77) Bromobenzene	11.195	156	23864	20.228	ug/l	100
78) n-propylbenzene	11.299	91	120311	20.363	ug/l	100
79) 2-Chlorotoluene	11.360	91	76316	20.025	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	89305	21.037	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8590	17.801	ug/l	86
82) 4-Chlorotoluene	11.451	91	87167	20.625	ug/l	99
83) tert-Butylbenzene	11.713	119	89128	20.843	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	89380	20.791	ug/l	100
85) sec-Butylbenzene	11.890	105	110457	21.038	ug/l	99
86) p-Isopropyltoluene	12.006	119	90318	20.840	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	43379	20.148	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	43071	19.588	ug/l	97
89) n-Butylbenzene	12.329	91	78349	20.610	ug/l	99
90) Hexachloroethane	12.536	117	15040	19.698	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	45968	21.276	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8285	21.002	ug/l	96
93) 1,2,4-Trichlorobenzene	13.585	180	25402	20.470	ug/l	100
94) Hexachlorobutadiene	13.719	225	11133	20.542	ug/l	98
95) Naphthalene	13.774	128	94703	20.808	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	26254	20.504	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046463.D
Acq On : 03 Jun 2025 11:24
Operator : JC/MD
Sample : VX0603WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:23:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046463.D
 Acq On : 03 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0603WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01

Manual Integrations APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:23:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion:168 Resp: 8985
Ion Ratio Lower Upper
168 100
99 69.4 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#2
Dichlorodifluoromethane
Concen: 16.566 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 85 Resp: 22784
Ion Ratio Lower Upper
85 100
87 30.3 15.7 47.1









#3
Chloromethane
Concen: 17.010 ug/l
RT: 1.301 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion: 50 Resp: 2268
Ion Ratio Lower Upper
50 100
52 32.7 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#4
Vinyl Chloride
Concen: 16.617 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion: 62 Resp: 20626
Ion Ratio Lower Upper
62 100
64 31.6 25.2 37.8





#5
Bromomethane
Concen: 16.878 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion: 94 Resp: 971
Ion Ratio Lower Upper
94 100
96 92.9 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#6
Chloroethane
Concen: 19.070 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion: 64 Resp: 12637
Ion Ratio Lower Upper
64 100
66 35.3 24.3 36.5





#7
Trichlorofluoromethane
Concen: 18.840 ug/l
RT: 1.880 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion:101 Resp: 3456
Ion Ratio Lower Upper
101 100
103 65.6 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#8
Diethyl Ether
Concen: 18.602 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 74 Resp: 11617
Ion Ratio Lower Upper
74 100
45 109.2 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.733 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion:101 Resp: 2126
Ion Ratio Lower Upper
101 100
85 47.1 38.6 58.0
151 72.0 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#10
Methyl Iodide
Concen: 17.425 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:142 Resp: 23408
Ion Ratio Lower Upper
142 100
127 53.2 41.7 62.5
141 14.8 11.5 17.3





#11

Tert butyl alcohol

Concen: 111.778 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 59 Resp: 2628

Ion Ratio Lower Upper

59 100





57 10.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#12

1,1-Dichloroethene

Concen: 17.944 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 96 Resp: 19119


Ion Ratio Lower Upper

96 100

61 181.4 146.2 219.2

98 63.3 51.0 76.6





#13

Acrolein

Concen: 118.151 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 56 Resp: 3164

Ion Ratio Lower Upper

56 100





55 71.0 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#14

Allyl chloride

Concen: 19.668 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 41 Resp: 40050


Ion Ratio Lower Upper

41 100

39 73.8 60.6 90.8

76 30.3 24.9 37.3





#15

Acrylonitrile

Concen: 105.292 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 53 Resp: 70799

Ion Ratio Lower Upper

53 100

52 84.0 65.3 97.9





51 38.2 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#16

Acetone

Concen: 103.724 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 43 Resp: 69671

Ion Ratio Lower Upper

43 100

58 25.9 21.2 31.8









#17
Carbon Disulfide
Concen: 13.313 ug/l
RT: 2.508 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion: 76 Resp: 33630
Ion Ratio Lower Upper
76 100
78 9.9 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#18
Methyl Acetate
Concen: 27.131 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 43 Resp: 42289
Ion Ratio Lower Upper
43 100
74 21.0 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 20.340 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion: 73 Resp: 7598
Ion Ratio Lower Upper
73 100
57 21.5 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#20
Methylene Chloride
Concen: 18.353 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 84 Resp: 23624
Ion Ratio Lower Upper
84 100
49 136.2 113.9 170.9
51 43.7 33.5 50.3
86 65.8 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 17.980 ug/l

RT: 3.087 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 96 Resp: 1926

Ion Ratio Lower Upper

96 100

61 154.9 119.5 179.3





98 66.9 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#22

Diisopropyl ether

Concen: 20.983 ug/l

RT: 3.757 min Scan# 438

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Tgt Ion: 45 Resp: 82535




Ion Ratio Lower Upper


45 100

43 89.6 66.6 100.0

87 24.4 19.8 29.6

59 11.2 8.6 12.8









#23
Vinyl Acetate
Concen: 96.984 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion: 43 Resp: 33553
Ion Ratio Lower Upper
43 100
86 9.2 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#24
1,1-Dichloroethane
Concen: 20.493 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 63 Resp: 44898
Ion Ratio Lower Upper
63 100
98 6.5 3.6 10.8
100 4.3 2.1 6.3





#25

2-Butanone

Concen: 106.672 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 43 Resp: 10402

Ion Ratio Lower Upper

43 100





72 24.7 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#26

2,2-Dichloropropane

Concen: 19.283 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 77 Resp: 33066

Ion Ratio Lower Upper

77 100

97 21.5 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 20.421 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 96 Resp: 2634

Ion Ratio Lower Upper

96 100

61 158.3 0.0 322.8





98 63.1 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#28

Bromochloromethane

Concen: 21.563 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 49 Resp: 22740


Ion Ratio Lower Upper

49 100

129 1.5 0.0 4.0

130 68.1 56.2 84.2





#29

Tetrahydrofuran

Concen: 105.894 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion:	42	Resp:	6470
Ion	Ratio	Lower	Upper
42	100		
72	41.6	33.7	50.5
71	38.0	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#30

Chloroform

Concen: 20.890 ug/l




RT: 5.086 min Scan# 656


Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24


Tgt Ion:	83	Resp:	47703
Ion	Ratio	Lower	Upper
83	100		
85	66.1	49.3	73.9





#31
Cyclohexane
Concen: 17.746 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion: 56 Resp: 35430
Ion Ratio Lower Upper
56 100
69 28.4 24.4 36.6
84 80.0 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#32
1,1,1-Trichloroethane
Concen: 20.093 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 97 Resp: 39775
Ion Ratio Lower Upper
97 100
99 64.3 51.8 77.6
61 49.0 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.808 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 65 Resp: 8176

Ion Ratio Lower Upper

65 100





67 49.1 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion:114 Resp: 159016


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 18.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 50.381 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion:113 Resp: 5769

Ion Ratio Lower Upper

113 100

111 102.8 83.1 124.7





192 16.8 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#36

1,1-Dichloropropene

Concen: 18.533 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 75 Resp: 28514


Ion Ratio Lower Upper

75 100

110 31.8 16.3 48.9

77 29.9 24.3 36.5





#37

Ethyl Acetate

Concen: 19.653 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 43 Resp: 37358

Ion Ratio Lower Upper

43 100

61 12.8 10.3 15.5





70 9.6 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#38

Carbon Tetrachloride

Concen: 18.497 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion:117 Resp: 31976


Ion Ratio Lower Upper

117 100

119 89.5 75.2 112.8

121 28.5 24.2 36.4





#39

Methylcyclohexane

Concen: 17.710 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 83 Resp: 35078

Ion Ratio Lower Upper

83 100

55 78.7 64.7 97.1





98 47.1 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#40

Benzene

Concen: 19.490 ug/l

RT: 6.037 min Scan# 812

Delta R.T. 0.006 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 78 Resp: 87833

Ion Ratio Lower Upper

78 100


77 24.8 19.0 28.4





#41
Methacrylonitrile
Concen: 21.836 ug/l
RT: 4.916 min Scan# 611
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion: 41 Resp: 2171
Ion Ratio Lower Upper
41 100
39 60.2 47.2 70.8
67 66.9 50.7 76.1
52 33.3 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#42
1,2-Dichloroethane
Concen: 20.187 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 62 Resp: 39263
Ion Ratio Lower Upper
62 100
98 7.7 0.0 15.2





#43

Isopropyl Acetate

Concen: 20.448 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion:	43	Resp:	5929
Ion	Ratio	Lower	Upper
43	100		
61	17.8	14.3	21.5
87	11.0	9.5	14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#44

Trichloroethene

Concen: 18.946 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Tgt Ion:	130	Resp:	20550
Ion	Ratio	Lower	Upper
130	100		
95	104.3	0.0	204.2





#45

1,2-Dichloropropane

Concen: 20.917 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 63 Resp: 23439

Ion Ratio Lower Upper

63 100





65 29.9 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#46

Dibromomethane

Concen: 19.808 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 93 Resp: 17506


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 84.7 68.2 102.2





#47

Bromodichloromethane

Concen: 20.043 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion:	83	Resp:	34890
Ion	Ratio	Lower	Upper
83	100		
85	67.7	50.5	75.7
127	8.9	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#48

Methyl methacrylate

Concen: 21.143 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Tgt Ion:	41	Resp:	31312
Ion	Ratio	Lower	Upper
41	100		
69	71.3	58.5	87.7
39	64.5	51.7	77.5





#49

1,4-Dioxane

Concen: 439.865 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 88 Resp: 12369

Ion Ratio Lower Upper

88 100

43 42.9 33.4 50.2





58 78.1 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#50

Toluene-d8

Concen: 47.791 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 98 Resp: 189408

Ion Ratio Lower Upper

98 100

100 64.9 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 109.285 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion: 43 Resp: 210362
Ion Ratio Lower Upper
43 100
58 35.2 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#52
Toluene
Concen: 19.864 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 92 Resp: 54891
Ion Ratio Lower Upper
92 100
91 171.0 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.227 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 75 Resp: 29748

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#54

cis-1,3-Dichloropropene

Concen: 19.928 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 75 Resp: 34079


Ion Ratio Lower Upper

75 100

77 32.2 25.4 38.0

39 61.1 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 21.488 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 97 Resp: 2341

Ion Ratio Lower Upper

97 100

83 84.0 69.0 103.6

85 49.3 44.3 66.5





99 59.2 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#56

Ethyl methacrylate

Concen: 21.719 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 69 Resp: 37717


Ion Ratio Lower Upper

69 100

41 77.2 60.8 91.2

39 47.0 39.0 58.6





#57

1,3-Dichloropropane

Concen: 20.384 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 76 Resp: 3988

Ion Ratio Lower Upper

76 100





78 33.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#58

2-Chloroethyl Vinyl ether

Concen: 106.886 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 63 Resp: 94629

Ion Ratio Lower Upper

63 100

106 27.0 21.5 32.3





#59

2-Hexanone

Concen: 111.507 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 43 Resp: 15879

Ion Ratio Lower Upper

43 100





58 48.1 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#60

Dibromochloromethane

Concen: 20.516 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion:129 Resp: 24550

Ion Ratio Lower Upper

129 100


127 75.9 39.3 117.8





#61
1,2-Dibromoethane
Concen: 20.220 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion:107 Resp: 22898
Ion Ratio Lower Upper
107 100
109 96.1 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#62
4-Bromofluorobenzene
Concen: 50.637 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion: 95 Resp: 76982
Ion Ratio Lower Upper
95 100
174 66.7 0.0 135.8
176 64.5 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion:117 Resp: 137079
Ion Ratio Lower Upper
117 100
82 61.3 50.6 76.0
119 32.4 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#64
Tetrachloroethene
Concen: 19.782 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:164 Resp: 19186
Ion Ratio Lower Upper
164 100
166 127.0 105.0 157.6
129 98.9 83.5 125.3
131 99.7 76.5 114.7









#65
Chlorobenzene
Concen: 20.209 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion:112 Resp: 6063
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 20.627 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion:131 Resp: 21133
Ion Ratio Lower Upper
131 100
133 96.2 47.3 141.9
119 65.6 31.6 95.0





#67
Ethyl Benzene
Concen: 20.434 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24



Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion: 91 Resp: 108068
Ion Ratio Lower Upper
91 100
106 29.2 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#68
m/p-Xylenes
Concen: 40.424 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:106 Resp: 78193
Ion Ratio Lower Upper
106 100
91 217.3 171.2 256.8









#69
o-Xylene
Concen: 20.987 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion:106 Resp: 39570
Ion Ratio Lower Upper
106 100
91 229.4 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#70
Styrene
Concen: 21.169 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion:104 Resp: 65394
Ion Ratio Lower Upper
104 100
78 58.3 45.7 68.5
103 55.4 43.7 65.5





#71
Bromoform
Concen: 19.868 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion:173 Resp: 1528
Ion Ratio Lower Upper
173 100
175 47.7 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:152 Resp: 65270
Ion Ratio Lower Upper
152 100
115 76.1 46.9 140.7
150 163.5 0.0 351.0









#73
Isopropylbenzene
Concen: 21.051 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion:105 Resp: 106972
Ion Ratio Lower Upper
105 100
120 26.0 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#74
N-ethyl acetate
Concen: 21.221 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion: 43 Resp: 53285
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 23.8 18.7 28.1
61 21.4 17.1 25.7





#75
1,1,2,2-Tetrachloroethane
Concen: 21.421 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion: 83 Resp: 38144
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 63.2 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#76
1,2,3-Trichloropropane
Concen: 20.597 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 75 Resp: 32360
Ion Ratio Lower Upper
75 100
77 41.2 20.5 61.5





#77

Bromobenzene

Concen: 20.228 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion:	156	Resp:	23864
Ion Ratio	Lower	Upper	
156	100		
77	185.5	92.8	278.3
158	96.5	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#78

n-propylbenzene

Concen: 20.363 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Tgt Ion:	91	Resp:	120311
Ion Ratio	Lower	Upper	
91	100		
120	21.4	10.8	32.4









#79
2-Chlorotoluene
Concen: 20.025 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion: 91 Resp: 76310
Ion Ratio Lower Upper
91 100
126 32.0 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#80
1,3,5-Trimethylbenzene
Concen: 21.037 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion:105 Resp: 89305
Ion Ratio Lower Upper
105 100
120 45.5 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 17.801 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion: 75 Resp: 8590
Ion Ratio Lower Upper
75 100
53 146.4 100.5 150.7
89 48.4 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#82
4-Chlorotoluene
Concen: 20.625 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 91 Resp: 87167
Ion Ratio Lower Upper
91 100
126 27.1 13.3 39.8





#83

tert-Butylbenzene

Concen: 20.843 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion:	119	Resp:	89128
Ion Ratio	Lower	Upper	
119	100		
91	65.8	32.9	98.7
134	23.1	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#84

1,2,4-Trimethylbenzene

Concen: 20.791 ug/l




RT: 11.750 min Scan# 1749


Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24


Tgt Ion:	105	Resp:	89380
Ion Ratio	Lower	Upper	
105	100		
120	42.5	21.2	63.6





#85
sec-Butylbenzene
Concen: 21.038 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion:105 Resp: 11045
Ion Ratio Lower Upper
105 100
134 18.8 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#86
p-Isopropyltoluene
Concen: 20.840 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:119 Resp: 90318
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.2 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 20.148 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion:146 Resp: 43379

Ion Ratio Lower Upper

146 100

111 45.0 22.1 66.3





148 63.3 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#88

1,4-Dichlorobenzene

Concen: 19.588 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion:146 Resp: 43071


Ion Ratio Lower Upper

146 100

111 45.5 21.3 63.9

148 64.8 31.9 95.5





#89

n-Butylbenzene

Concen: 20.610 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01

Tgt Ion: 91 Resp: 78349

Ion Ratio Lower Upper

91 100

92 52.6 26.9 80.7





134 23.2 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#90

Hexachloroethane

Concen: 19.698 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion:117 Resp: 15040

Ion Ratio Lower Upper

117 100


201 59.8 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 21.276 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion:146 Resp: 4596
Ion Ratio Lower Upper
146 100
111 46.1 22.4 67.3
148 63.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.002 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion: 75 Resp: 8285
Ion Ratio Lower Upper
75 100
155 66.2 34.9 104.8
157 91.0 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 20.470 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01



Tgt Ion:180 Resp: 2540
Ion Ratio Lower Upper
180 100
182 96.2 47.9 143.8
145 33.6 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#94
Hexachlorobutadiene
Concen: 20.542 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:225 Resp: 11133
Ion Ratio Lower Upper
225 100
223 61.2 30.8 92.4
227 64.8 30.9 92.7





#95

Naphthalene

Concen: 20.808 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion:	128	Resp:	9470
Ion Ratio	Lower	Upper	
128	100		
127	13.0	10.4	15.6
129	10.7	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#96

1,2,3-Trichlorobenzene

Concen: 20.504 ug/l




RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Tgt Ion:	180	Resp:	26254
Ion Ratio	Lower	Upper	
180	100		
182	93.0	47.8	143.3
145	36.3	18.1	54.3



Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0603WBSD01		SDG No.:	Q2186
Lab Sample ID:	VX0603WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046464.D	1		06/03/25 11:50	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	16.4		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	16.0		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	16.8		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	17.8		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.1		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.1		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	110		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	13.3		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.5		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	18.3		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.0		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.1		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.6		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.1		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	20.7		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.1		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.4		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.9		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.9		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.7		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.7		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.5		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	22.2		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	VX0603WBSD01	SDG No.:	Q2186
Lab Sample ID:	VX0603WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046464.D	1		06/03/25 11:50	VX060325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	21.0		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.4		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.4		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	40.2		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	21.4		0.12	0.50	1.00	ug/L
100-42-5	Styrene	21.2		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	20.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	21.0		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.3		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.2		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.2		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	21.2		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.4		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	48.2		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	86000	5.544				
540-36-3	1,4-Difluorobenzene	151000	6.757				
3114-55-4	Chlorobenzene-d5	130000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	62400	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046464.D
 Acq On : 03 Jun 2025 11:50
 Operator : JC/MD
 Sample : VX0603WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:24:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	86032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80850	50.408	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.820%			
35) Dibromofluoromethane	5.379	113	56126	51.631	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.260%			
50) Toluene-d8	8.647	98	181255	48.175	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.340%			
62) 4-Bromofluorobenzene	11.079	95	72876	50.495	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.000%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	21879	16.616	ug/l	98
3) Chloromethane	1.307	50	20951	16.407	ug/l	99
4) Vinyl Chloride	1.374	62	18990	15.979	ug/l	100
5) Bromomethane	1.599	94	9271	16.819	ug/l	97
6) Chloroethane	1.672	64	11276	17.773	ug/l	100
7) Trichlorofluoromethane	1.880	101	31828	18.121	ug/l	100
8) Diethyl Ether	2.136	74	11245	18.807	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.325	101	20804	19.140	ug/l	99
10) Methyl Iodide	2.447	142	22438	17.446	ug/l	99
11) Tert butyl alcohol	2.971	59	26245	116.572	ug/l	100
12) 1,1-Dichloroethene	2.313	96	18188	17.829	ug/l	92
13) Acrolein	2.233	56	28734	112.067	ug/l	97
14) Allyl chloride	2.660	41	38609	19.803	ug/l	96
15) Acrylonitrile	3.062	53	71487	111.043	ug/l	99
16) Acetone	2.380	43	68723	106.863	ug/l	99
17) Carbon Disulfide	2.508	76	32208	13.317	ug/l	99
18) Methyl Acetate	2.703	43	41923	28.093	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	76793	21.472	ug/l	100
20) Methylene Chloride	2.782	84	22550	18.298	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	18461	17.995	ug/l	100
22) Diisopropyl ether	3.757	45	81514	21.645	ug/l	98
23) Vinyl Acetate	3.721	43	334608	101.019	ug/l	99
24) 1,1-Dichloroethane	3.605	63	42085	20.064	ug/l	97
25) 2-Butanone	4.556	43	104645	112.084	ug/l	100
26) 2,2-Dichloropropane	4.465	77	31623	19.261	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	24764	20.052	ug/l	99
28) Bromochloromethane	4.897	49	21879	21.669	ug/l	97
29) Tetrahydrofuran	5.001	42	63927	109.270	ug/l	99
30) Chloroform	5.093	83	45359	20.747	ug/l	97
31) Cyclohexane	5.458	56	33045	17.288	ug/l	93
32) 1,1,1-Trichloroethane	5.373	97	36112	19.054	ug/l	97
36) 1,1-Dichloropropene	5.684	75	26077	17.854	ug/l	97
37) Ethyl Acetate	4.715	43	37561	20.815	ug/l	100
38) Carbon Tetrachloride	5.672	117	30493	18.581	ug/l	99
39) Methylcyclohexane	7.379	83	32831	17.460	ug/l	97
40) Benzene	6.031	78	82785	19.351	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046464.D
 Acq On : 03 Jun 2025 11:50
 Operator : JC/MD
 Sample : VX0603WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/04/2025
 Supervised By : Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:24:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	22327	23.652	ug/l	99
42) 1,2-Dichloroethane	6.080	62	38662	20.939	ug/l	98
43) Isopropyl Acetate	6.336	43	61515	22.345	ug/l	98
44) Trichloroethene	7.123	130	19772	19.202	ug/l	91
45) 1,2-Dichloropropane	7.428	63	22490	21.141	ug/l	97
46) Dibromomethane	7.580	93	17488	20.843	ug/l	98
47) Bromodichloromethane	7.818	83	34508	20.882	ug/l	99
48) Methyl methacrylate	7.690	41	30780	21.893	ug/l	99
49) 1,4-Dioxane	7.659	88	12806	479.715	ug/l	96
51) 4-Methyl-2-Pentanone	8.568	43	210042	114.943	ug/l	98
52) Toluene	8.714	92	51638	19.684	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	28911	19.683	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	33321	20.525	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	22943	22.181	ug/l	98
56) Ethyl methacrylate	9.116	69	37370	22.668	ug/l	98
57) 1,3-Dichloropropane	9.305	76	39110	21.053	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	93351	111.071	ug/l	99
59) 2-Hexanone	9.427	43	155374	114.927	ug/l	99
60) Dibromochloromethane	9.519	129	23830	20.978	ug/l	98
61) 1,2-Dibromoethane	9.604	107	22450	20.882	ug/l	99
64) Tetrachloroethene	9.269	164	17909	19.400	ug/l	92
65) Chlorobenzene	10.073	112	57186	20.025	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	20660	21.186	ug/l	99
67) Ethyl Benzene	10.189	91	102509	20.363	ug/l	98
68) m/p-Xylenes	10.299	106	74026	40.207	ug/l	97
69) o-Xylene	10.640	106	38408	21.398	ug/l	100
70) Styrene	10.653	104	62236	21.167	ug/l	99
71) Bromoform	10.799	173	15063	20.574	ug/l #	100
73) Isopropylbenzene	10.957	105	101859	20.979	ug/l	100
74) N-amyl acetate	10.842	43	51555	21.488	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	36254	21.308	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	31973m	21.299	ug/l	
77) Bromobenzene	11.195	156	22557	20.011	ug/l	99
78) n-propylbenzene	11.299	91	115006	20.371	ug/l	100
79) 2-Chlorotoluene	11.360	91	73129	20.083	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	85390	21.051	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8927	19.361	ug/l	94
82) 4-Chlorotoluene	11.451	91	83070	20.571	ug/l	99
83) tert-Butylbenzene	11.713	119	84344	20.643	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	84267	20.514	ug/l	100
85) sec-Butylbenzene	11.890	105	106367	21.203	ug/l	100
86) p-Isopropyltoluene	12.006	119	86575	20.907	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	41494	20.170	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	42409	20.186	ug/l	98
89) n-Butylbenzene	12.329	91	75206	20.705	ug/l	99
90) Hexachloroethane	12.536	117	14400	19.739	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	43740	21.188	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8644	22.933	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	24446	20.617	ug/l	100
94) Hexachlorobutadiene	13.725	225	10733	20.726	ug/l	99
95) Naphthalene	13.774	128	93399	21.477	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	25957	21.216	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046464.D
Acq On : 03 Jun 2025 11:50
Operator : JC/MD
Sample : VX0603WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:24:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046464.D
Acq On : 03 Jun 2025 11:50
Operator : JC/MD
Sample : VX0603WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Manual Integrations

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025

Quant Time: Jun 04 01:24:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion:168 Resp: 8603
Ion Ratio Lower Upper
168 100
99 74.7 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#2
Dichlorodifluoromethane
Concen: 16.616 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 85 Resp: 21879
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1





#3
Chloromethane
Concen: 16.407 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 50 Resp: 2095
Ion Ratio Lower Upper
50 100
52 32.3 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#4
Vinyl Chloride
Concen: 15.979 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 62 Resp: 18990
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8





#5
Bromomethane
Concen: 16.819 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 94 Resp: 927
Ion Ratio Lower Upper
94 100
96 88.0 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#6
Chloroethane
Concen: 17.773 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 64 Resp: 11276
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5





#7
Trichlorofluoromethane
Concen: 18.121 ug/l
RT: 1.880 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:101 Resp: 31828
Ion Ratio Lower Upper
101 100
103 63.7 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#8
Diethyl Ether
Concen: 18.807 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 74 Resp: 11245
Ion Ratio Lower Upper
74 100
45 112.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.140 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion:101 Resp: 20804
Ion Ratio Lower Upper
101 100
85 47.2 38.6 58.0
151 70.0 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#10
Methyl Iodide
Concen: 17.446 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:142 Resp: 22438
Ion Ratio Lower Upper
142 100
127 51.8 41.7 62.5
141 13.7 11.5 17.3





#11

Tert butyl alcohol

Concen: 116.572 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 59 Resp: 2624

Ion Ratio Lower Upper

59 100





57 10.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#12

1,1-Dichloroethene

Concen: 17.829 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 96 Resp: 18188


Ion Ratio Lower Upper

96 100

61 194.8 146.2 219.2

98 68.6 51.0 76.6





#13

Acrolein

Concen: 112.067 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 56 Resp: 28734

Ion Ratio Lower Upper

56 100





55 72.4 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#14

Allyl chloride

Concen: 19.803 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 41 Resp: 38609


Ion Ratio Lower Upper

41 100

39 72.3 60.6 90.8

76 28.5 24.9 37.3





#15

Acrylonitrile

Concen: 111.043 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 53 Resp: 7148

Ion Ratio Lower Upper

53 100

52 83.2 65.3 97.9





51 37.6 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#16

Acetone

Concen: 106.863 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 43 Resp: 68723

Ion Ratio Lower Upper

43 100


58 26.1 21.2 31.8





#17
Carbon Disulfide
Concen: 13.317 ug/l
RT: 2.508 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 76 Resp: 32208
Ion Ratio Lower Upper
76 100
78 8.9 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#18
Methyl Acetate
Concen: 28.093 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 43 Resp: 41923
Ion Ratio Lower Upper
43 100
74 21.2 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 21.472 ug/l

RT: 3.111 min Scan# 311

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 73 Resp: 7679

Ion Ratio Lower Upper

73 100





57 21.9 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#20

Methylene Chloride

Concen: 18.298 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Tgt Ion: 84 Resp: 22550




Ion Ratio Lower Upper


84 100

49 143.8 113.9 170.9

51 42.1 33.5 50.3

86 68.5 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 17.995 ug/l

RT: 3.087 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 96 Resp: 1846

Ion Ratio Lower Upper

96 100

61 149.3 119.5 179.3





98 63.5 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#22

Diisopropyl ether

Concen: 21.645 ug/l

RT: 3.757 min Scan# 438

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Tgt Ion: 45 Resp: 81514




Ion Ratio Lower Upper


45 100

43 84.8 66.6 100.0

87 25.5 19.8 29.6

59 12.0 8.6 12.8









#23
Vinyl Acetate
Concen: 101.019 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion: 43 Resp: 334608
Ion Ratio Lower Upper
43 100
86 9.1 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#24
1,1-Dichloroethane
Concen: 20.064 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 63 Resp: 42085
Ion Ratio Lower Upper
63 100
98 5.8 3.6 10.8
100 4.6 2.1 6.3





#25

2-Butanone

Concen: 112.084 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 43 Resp: 10464

Ion Ratio Lower Upper

43 100





72 23.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#26

2,2-Dichloropropane

Concen: 19.261 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 77 Resp: 31623

Ion Ratio Lower Upper

77 100

97 20.4 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 20.052 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 96 Resp: 24764

Ion Ratio Lower Upper

96 100

61 159.2 0.0 322.8





98 64.4 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#28

Bromochloromethane

Concen: 21.669 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 49 Resp: 21879


Ion Ratio Lower Upper

49 100

129 1.5 0.0 4.0

130 67.6 56.2 84.2





#29

Tetrahydrofuran

Concen: 109.270 ug/l

RT: 5.001 min Scan# 64

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBSD01

Tgt Ion:	42	Resp:	6392
Ion	Ratio	Lower	Upper
42	100		
72	42.4	33.7	50.5
71	38.4	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#30

Chloroform

Concen: 20.747 ug/l




RT: 5.093 min Scan# 657


Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50


Tgt Ion:	83	Resp:	45359
Ion	Ratio	Lower	Upper
83	100		
85	64.1	49.3	73.9





#31
Cyclohexane
Concen: 17.288 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 56 Resp: 3304
Ion Ratio Lower Upper
56 100
69 28.2 24.4 36.6
84 76.0 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#32
1,1,1-Trichloroethane
Concen: 19.054 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 97 Resp: 36112
Ion Ratio Lower Upper
97 100
99 66.4 51.8 77.6
61 52.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 50.408 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 65 Resp: 80850

Ion Ratio Lower Upper

65 100





67 49.6 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion:114 Resp: 150958


Ion Ratio Lower Upper

114 100

63 22.8 0.0 49.2

88 18.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 51.631 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01





Tgt Ion:	113	Resp:	56120
Ion	Ratio	Lower	Upper
113	100		
111	101.9	83.1	124.7
192	16.8	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#36

1,1-Dichloropropene

Concen: 17.854 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Tgt Ion:	75	Resp:	26077
Ion	Ratio	Lower	Upper
75	100		
110	33.8	16.3	48.9
77	32.1	24.3	36.5





#37

Ethyl Acetate

Concen: 20.815 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 43 Resp: 3756

Ion Ratio Lower Upper

43 100

61 12.9 10.3 15.5





70 9.8 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#38

Carbon Tetrachloride

Concen: 18.581 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion:117 Resp: 30493


Ion Ratio Lower Upper

117 100

119 94.3 75.2 112.8

121 29.2 24.2 36.4





#39

Methylcyclohexane

Concen: 17.460 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 83 Resp: 3283

Ion Ratio Lower Upper

83 100

55 84.2 64.7 97.1





98 47.7 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#40

Benzene

Concen: 19.351 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 78 Resp: 82785

Ion Ratio Lower Upper

78 100


77 25.6 19.0 28.4





#41
Methacrylonitrile
Concen: 23.652 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 41 Resp: 2232
Ion Ratio Lower Upper
41 100
39 57.1 47.2 70.8
67 63.7 50.7 76.1
52 30.4 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#42
1,2-Dichloroethane
Concen: 20.939 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 62 Resp: 38662
Ion Ratio Lower Upper
62 100
98 7.0 0.0 15.2





#43

Isopropyl Acetate

Concen: 22.345 ug/l

RT: 6.336 min Scan# 861

Delta R.T. -0.000 min

Lab File: VX046464.D


Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01






Tgt Ion: 43	Resp: 6151
Ion Ratio	Lower Upper
43	100
61	17.6 14.3 21.5
87	10.7 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#44

Trichloroethene

Concen: 19.202 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Tgt Ion:130	Resp: 19772
Ion Ratio	Lower Upper
130	100
95	111.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 21.141 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 63 Resp: 22490

Ion Ratio Lower Upper

63 100





65 31.6 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#46

Dibromomethane

Concen: 20.843 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 93 Resp: 17488


Ion Ratio Lower Upper

93 100

95 79.4 65.6 98.4

174 83.5 68.2 102.2





#47

Bromodichloromethane

Concen: 20.882 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01





Tgt Ion:	83	Resp:	3450
Ion	Ratio	Lower	Upper
83	100		
85	62.4	50.5	75.7
127	8.3	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#48

Methyl methacrylate

Concen: 21.893 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Tgt Ion:	41	Resp:	30780
Ion	Ratio	Lower	Upper
41	100		
69	71.7	58.5	87.7
39	64.9	51.7	77.5





#49

1,4-Dioxane

Concen: 479.715 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 88 Resp: 12800

Ion Ratio Lower Upper

88 100

43 41.2 33.4 50.2





58 77.5 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#50

Toluene-d8

Concen: 48.175 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 98 Resp: 181255

Ion Ratio Lower Upper

98 100

100 64.6 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 114.943 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion: 43 Resp: 21004
Ion Ratio Lower Upper
43 100
58 35.2 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#52
Toluene
Concen: 19.684 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 92 Resp: 51638
Ion Ratio Lower Upper
92 100
91 173.7 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.683 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 75 Resp: 2891

Ion Ratio Lower Upper

75 100





77 31.8 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#54

cis-1,3-Dichloropropene

Concen: 20.525 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 75 Resp: 33321


Ion Ratio Lower Upper

75 100

77 33.8 25.4 38.0

39 61.4 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 22.181 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 97 Resp: 2294

Ion Ratio Lower Upper

97 100

83 82.1 69.0 103.6

85 55.0 44.3 66.5





99 59.6 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#56

Ethyl methacrylate

Concen: 22.668 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 69 Resp: 37370


Ion Ratio Lower Upper

69 100

41 77.0 60.8 91.2

39 47.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 21.053 ug/l

RT: 9.305 min Scan# 1134

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 76 Resp: 39110

Ion Ratio Lower Upper

76 100





78 32.1 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#58

2-Chloroethyl Vinyl ether

Concen: 111.071 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 63 Resp: 93351

Ion Ratio Lower Upper

63 100

106 26.5 21.5 32.3





#59

2-Hexanone

Concen: 114.927 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01

Tgt Ion: 43 Resp: 155374

Ion Ratio Lower Upper

43 100





58 49.1 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/04/2025

Supervised By :Mahesh Dadoda 06/04/2025



#60

Dibromochloromethane

Concen: 20.978 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion:129 Resp: 23830

Ion Ratio Lower Upper

129 100

127 76.8 39.3 117.8









#61
1,2-Dibromoethane
Concen: 20.882 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion:107 Resp: 22450
Ion Ratio Lower Upper
107 100
109 92.8 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#62
4-Bromofluorobenzene
Concen: 50.495 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 95 Resp: 72876
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 63.7 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:117 Resp: 130475
Ion Ratio Lower Upper
117 100
82 61.7 50.6 76.0
119 31.8 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#64
Tetrachloroethene
Concen: 19.400 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:164 Resp: 17909
Ion Ratio Lower Upper
164 100
166 139.6 105.0 157.6
129 111.2 83.5 125.3
131 105.5 76.5 114.7









#65
Chlorobenzene
Concen: 20.025 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion:112 Resp: 57180
Ion Ratio Lower Upper
112 100
114 32.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 21.186 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion:131 Resp: 20660
Ion Ratio Lower Upper
131 100
133 92.6 47.3 141.9
119 63.2 31.6 95.0





#67
Ethyl Benzene
Concen: 20.363 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 91 Resp: 102509
Ion Ratio Lower Upper
91 100
106 28.1 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#68
m/p-Xylenes
Concen: 40.207 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:106 Resp: 74026
Ion Ratio Lower Upper
106 100
91 219.2 171.2 256.8









#69
o-Xylene
Concen: 21.398 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion:106 Resp: 38408
Ion Ratio Lower Upper
106 100
91 225.9 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#70
Styrene
Concen: 21.167 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion:104 Resp: 62236
Ion Ratio Lower Upper
104 100
78 58.1 45.7 68.5
103 55.6 43.7 65.5





#71
Bromoform
Concen: 20.574 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:173 Resp: 1506
Ion Ratio Lower Upper
173 100
175 46.6 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:152 Resp: 62365
Ion Ratio Lower Upper
152 100
115 76.9 46.9 140.7
150 162.0 0.0 351.0









#73
Isopropylbenzene
Concen: 20.979 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion:105 Resp: 101855
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#74
N-ethyl acetate
Concen: 21.488 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 43 Resp: 51555
Ion Ratio Lower Upper
43 100
70 37.7 30.9 46.3
55 22.9 18.7 28.1
61 21.2 17.1 25.7





#75
1,1,2,2-Tetrachloroethane
Concen: 21.308 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 83 Resp: 36254
Ion Ratio Lower Upper
83 100
131 9.6 5.0 14.9
85 63.3 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#76
1,2,3-Trichloropropane
Concen: 21.299 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 75 Resp: 31973
Ion Ratio Lower Upper
75 100
77 39.8 20.5 61.5





#77
Bromobenzene
Concen: 20.011 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:156 Resp: 2255
Ion Ratio Lower Upper
156 100
77 188.0 92.8 278.3
158 97.3 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#78
n-propylbenzene
Concen: 20.371 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 91 Resp: 115006
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 20.083 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion: 91 Resp: 73129
Ion Ratio Lower Upper
91 100
126 30.5 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#80
1,3,5-Trimethylbenzene
Concen: 21.051 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:105 Resp: 85390
Ion Ratio Lower Upper
105 100
120 45.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 19.361 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion: 75 Resp: 892
Ion Ratio Lower Upper
75 100
53 134.1 100.5 150.7
89 48.2 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#82
4-Chlorotoluene
Concen: 20.571 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 91 Resp: 83070
Ion Ratio Lower Upper
91 100
126 26.8 13.3 39.8





#83
tert-Butylbenzene
Concen: 20.643 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:119 Resp: 84344
Ion Ratio Lower Upper
119 100
91 66.3 32.9 98.7
134 23.3 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#84
1,2,4-Trimethylbenzene
Concen: 20.514 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:105 Resp: 84267
Ion Ratio Lower Upper
105 100
120 42.7 21.2 63.6









#85
sec-Butylbenzene
Concen: 21.203 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Tgt Ion:105 Resp: 10636
Ion Ratio Lower Upper
105 100
134 19.3 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#86
p-Isopropyltoluene
Concen: 20.907 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion:119 Resp: 86575
Ion Ratio Lower Upper
119 100
134 24.9 12.5 37.5
91 27.4 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 20.170 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:146 Resp: 41494
Ion Ratio Lower Upper
146 100
111 45.2 22.1 66.3
148 65.4 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#88
1,4-Dichlorobenzene
Concen: 20.186 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion:146 Resp: 42409
Ion Ratio Lower Upper
146 100
111 43.4 21.3 63.9
148 65.0 31.9 95.5





#89
n-Butylbenzene
Concen: 20.705 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion: 91 Resp: 75200
Ion Ratio Lower Upper
91 100
92 52.9 26.9 80.7
134 23.6 11.8 35.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#90
Hexachloroethane
Concen: 19.739 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion:117 Resp: 14400
Ion Ratio Lower Upper
117 100
201 62.0 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 21.188 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:146 Resp: 43740
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 65.1 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 22.933 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50


Tgt Ion: 75 Resp: 8644
Ion Ratio Lower Upper
75 100
155 66.7 34.9 104.8
157 86.0 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 20.617 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01




Tgt Ion:180 Resp: 2444
Ion Ratio Lower Upper
180 100
182 95.5 47.9 143.8
145 34.4 17.1 51.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025




#94
Hexachlorobutadiene
Concen: 20.726 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Tgt Ion:225 Resp: 10733
Ion Ratio Lower Upper
225 100
223 63.3 30.8 92.4
227 61.8 30.9 92.7





#95
Naphthalene
Concen: 21.477 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01



Tgt Ion:128 Resp: 93395
Ion Ratio Lower Upper
128 100
127 12.6 10.4 15.6
129 10.8 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/04/2025
Supervised By :Mahesh Dadoda 06/04/2025



#96
1,2,3-Trichlorobenzene
Concen: 21.216 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:180 Resp: 25957
Ion Ratio Lower Upper
180 100
182 93.4 47.8 143.3
145 35.9 18.1 54.3



Manual Integration Report

Sequence:	VX050525	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC020	VX046041.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:13 AM	MMDadoda	5/6/2025 12:42:46 PM	Peak Integrated by Software
VSTDICCC050	VX046042.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:18 AM	MMDadoda	5/6/2025 12:42:48 PM	Peak Integrated by Software
VSTDICC100	VX046043.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:22 AM	MMDadoda	5/6/2025 12:42:50 PM	Peak Integrated by Software
VSTDICC150	VX046044.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:27 AM	MMDadoda	5/6/2025 12:42:53 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	Ethyl Acetate	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,4-Dichlorobenzene	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Bromochloromethane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Ethyl Acetate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Methyl methacrylate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICV050	VX046048.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:45 AM	MMDadoda	5/6/2025 12:41:37 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

VX050525

Instrument

MSVOA_x

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

VX060325

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX046460.D	1,2,3-Trichloropropane	JOHN	6/4/2025 9:49:57 AM	MMDadoda	6/4/2025 4:21:57 PM	Peak Integrated by Software
VX0603WBS01	VX046463.D	1,2,3-Trichloropropane	JOHN	6/4/2025 9:50:01 AM	MMDadoda	6/4/2025 4:21:58 PM	Peak Integrated by Software
VX0603WBSD0 1	VX046464.D	1,2,3-Trichloropropane	JOHN	6/4/2025 9:50:06 AM	MMDadoda	6/4/2025 4:22:00 PM	Peak Integrated by Software
VSTDCCC050	VX046486.D	1,2,3-Trichloropropane	JOHN	6/4/2025 9:51:09 AM	MMDadoda	6/4/2025 4:22:18 PM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046038.D	05 May 2025 09:37	JC/MD	Ok
2	VSTDIC001	VX046039.D	05 May 2025 10:49	JC/MD	Not Ok
3	VSTDIC005	VX046040.D	05 May 2025 11:12	JC/MD	Not Ok
4	VSTDIC020	VX046041.D	05 May 2025 11:35	JC/MD	Ok,M
5	VSTDIC050	VX046042.D	05 May 2025 11:58	JC/MD	Ok,M
6	VSTDIC100	VX046043.D	05 May 2025 12:21	JC/MD	Ok,M
7	VSTDIC150	VX046044.D	05 May 2025 12:45	JC/MD	Ok,M
8	IBLK	VX046045.D	05 May 2025 13:08	JC/MD	Ok
9	VSTDIC005	VX046046.D	05 May 2025 16:04	JC/MD	Ok,M
10	VSTDIC001	VX046047.D	05 May 2025 16:27	JC/MD	Ok,M
11	VSTDICV050	VX046048.D	05 May 2025 16:50	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX060325

Review By	John Carlone	Review On	6/4/2025 10:07:03 AM
Supervise By	Maresh Dadoda	Supervise On	6/4/2025 4:22:37 PM
SubDirectory	VX060325	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134100		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134101,VP134102		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046459.D	03 Jun 2025 09:40	JC/MD	Ok
2	VSTDCCC050	VX046460.D	03 Jun 2025 10:10	JC/MD	Ok,M
3	VX0603MBL01	VX046461.D	03 Jun 2025 10:38	JC/MD	Ok
4	VX0603WBL01	VX046462.D	03 Jun 2025 11:01	JC/MD	Ok
5	VX0603WBS01	VX046463.D	03 Jun 2025 11:24	JC/MD	Ok,M
6	VX0603WBSD01	VX046464.D	03 Jun 2025 11:50	JC/MD	Ok,M
7	Q2186-09	VX046465.D	03 Jun 2025 12:13	JC/MD	Ok
8	Q2186-01	VX046466.D	03 Jun 2025 12:36	JC/MD	Ok
9	Q2186-02	VX046467.D	03 Jun 2025 12:59	JC/MD	Ok
10	Q2186-03	VX046468.D	03 Jun 2025 13:23	JC/MD	Ok
11	Q2186-06	VX046469.D	03 Jun 2025 13:46	JC/MD	Ok
12	Q2186-07	VX046470.D	03 Jun 2025 14:09	JC/MD	Ok
13	Q2186-08	VX046471.D	03 Jun 2025 14:33	JC/MD	Ok
14	Q2159-04	VX046472.D	03 Jun 2025 14:56	JC/MD	Ok
15	Q2160-04	VX046473.D	03 Jun 2025 15:20	JC/MD	Ok,M
16	Q2160-08	VX046474.D	03 Jun 2025 15:43	JC/MD	Ok,M
17	Q2173-06	VX046475.D	03 Jun 2025 16:06	JC/MD	Ok,M
18	Q2173-12	VX046476.D	03 Jun 2025 16:30	JC/MD	Ok,M
19	Q2173-18	VX046477.D	03 Jun 2025 16:53	JC/MD	Ok
20	Q2189-01	VX046478.D	03 Jun 2025 17:17	JC/MD	ReRun
21	Q2189-02	VX046479.D	03 Jun 2025 17:40	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX060325

Review By	John Carlone	Review On	6/4/2025 10:07:03 AM
Supervise By	Mahesh Dadoda	Supervise On	6/4/2025 4:22:37 PM
SubDirectory	VX060325	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134100		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134101,VP134102		

22	Q2172-04	VX046480.D	03 Jun 2025 18:04	JC/MD	Ok
23	Q2168-04	VX046481.D	03 Jun 2025 18:28	JC/MD	Ok,M
24	Q2168-08	VX046482.D	03 Jun 2025 18:51	JC/MD	Dilution
25	Q2168-12	VX046483.D	03 Jun 2025 19:15	JC/MD	Dilution
26	Q2169-01	VX046484.D	03 Jun 2025 19:38	JC/MD	ReRun
27	Q2169-03	VX046485.D	03 Jun 2025 20:02	JC/MD	Dilution
28	VSTDCCC050	VX046486.D	03 Jun 2025 20:25	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060325

Review By	John Carlone	Review On	6/4/2025 10:07:03 AM
Supervise By	Maresh Dadoda	Supervise On	6/4/2025 4:22:37 PM
SubDirectory	VX060325	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134100
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134101,VP134102

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046459.D	03 Jun 2025 09:40		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046460.D	03 Jun 2025 10:10	pH#Lot#V12668	JC/MD	Ok,M
3	VX0603MBL01	VX0603MBL01	VX046461.D	03 Jun 2025 10:38		JC/MD	Ok
4	VX0603WBL01	VX0603WBL01	VX046462.D	03 Jun 2025 11:01		JC/MD	Ok
5	VX0603WBS01	VX0603WBS01	VX046463.D	03 Jun 2025 11:24	BSD failed high for comp. #18	JC/MD	Ok,M
6	VX0603WBSD01	VX0603WBSD01	VX046464.D	03 Jun 2025 11:50	BSD failed high for comp. #18	JC/MD	Ok,M
7	Q2186-09	VPB182-HYD-2025053	VX046465.D	03 Jun 2025 12:13	vial A pH<2	JC/MD	Ok
8	Q2186-01	BP-VPB-182-TB-20250	VX046466.D	03 Jun 2025 12:36	vial A pH<2 TB	JC/MD	Ok
9	Q2186-02	BP-VPB-182-GW-670-6	VX046467.D	03 Jun 2025 12:59	vial A pH<2	JC/MD	Ok
10	Q2186-03	BP-VPB-182-GW-680-6	VX046468.D	03 Jun 2025 13:23	vial A pH<2	JC/MD	Ok
11	Q2186-06	BP-VPB-182-GW-705-7	VX046469.D	03 Jun 2025 13:46	vial A pH<2	JC/MD	Ok
12	Q2186-07	BP-VPB-182-GW-720-7	VX046470.D	03 Jun 2025 14:09	vial A pH<2	JC/MD	Ok
13	Q2186-08	BP-VPB-182-DUP-2025	VX046471.D	03 Jun 2025 14:33	vial A pH<2	JC/MD	Ok
14	Q2159-04	TP05-MHO-WC	VX046472.D	03 Jun 2025 14:56	vial A pH#5.0	JC/MD	Ok
15	Q2160-04	TP04-MHG-WC	VX046473.D	03 Jun 2025 15:20	vial A pH#5.0	JC/MD	Ok,M
16	Q2160-08	TP05-MHH-WC	VX046474.D	03 Jun 2025 15:43	vial A pH#5.0	JC/MD	Ok,M
17	Q2173-06	OR-400-CF-402B-COM	VX046475.D	03 Jun 2025 16:06	vial A pH#5.0	JC/MD	Ok,M
18	Q2173-12	OR-400-CF-402B-COM	VX046476.D	03 Jun 2025 16:30	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060325

Review By	John Carlone	Review On	6/4/2025 10:07:03 AM
Supervise By	Maresh Dadoda	Supervise On	6/4/2025 4:22:37 PM
SubDirectory	VX060325	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134100		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134101,VP134102		

19	Q2173-18	OR-400-CF-402B-COM	VX046477.D	03 Jun 2025 16:53	vial A pH#5.0	JC/MD	Ok
20	Q2189-01	MW-7-20250602	VX046478.D	03 Jun 2025 17:17	vial A pH<2 CCAL,BS and BSD failed high	JC/MD	ReRun
21	Q2189-02	MW-8-20250602	VX046479.D	03 Jun 2025 17:40	vial A pH<2	JC/MD	Ok
22	Q2172-04	TP06-MHQ	VX046480.D	03 Jun 2025 18:04	vial A pH#5.0	JC/MD	Ok
23	Q2168-04	A3	VX046481.D	03 Jun 2025 18:28	vial A pH#5.0	JC/MD	Ok,M
24	Q2168-08	B3	VX046482.D	03 Jun 2025 18:51	vial A pH#5.0 Need 5X	JC/MD	Dilution
25	Q2168-12	C2	VX046483.D	03 Jun 2025 19:15	vial A pH#5.0 Need 10X	JC/MD	Dilution
26	Q2169-01	303-PPR-1	VX046484.D	03 Jun 2025 19:38	vial A pH<2 E flag in previous sample	JC/MD	ReRun
27	Q2169-03	303-PPR-2	VX046485.D	03 Jun 2025 20:02	vial A pH<2 Confirm concentration; Need 5X	JC/MD	Dilution
28	VSTDCCC050	VSTDCCC050EC	VX046486.D	03 Jun 2025 20:25		JC/MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q2186

Test : VOCMS Group1

Prepbatch ID :

Sequence ID/Qc Batch ID: VX060325,

Standard ID :

VP132035,VP132096,VP133174,VP133887,VP133935,VP133953,VP133991,VP134100,VP134101,VP134102,

Chemical ID :

V13391,V13457,V13460,V13465,V13466,V13706,V14290,V14432,V14435,V14503,V14504,V14525,V14526,V14613,V14614,V14620,V14626,V14630,V14631,V14632,V14633,V14711,V14717,V14718,V14721,V14749,V14750,V14811,V14812,V14843,V14921,V14944,V14945,V14946,V14947,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP132035	12/10/2024	06/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								12/12/2024

FROM 1.00000ml of V14630 + 1.00000ml of V14631 + 1.00000ml of V14632 + 1.00000ml of V14633 + 46.00000ml of V14614 = Final
Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP132096	12/12/2024	06/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								12/19/2024

FROM 1.00000ml of V13465 + 1.00000ml of V13466 + 1.50000ml of V13457 + 1.50000ml of V13460 + 20.00000ml of V14614 = Final
Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP133174	02/27/2025	08/27/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								03/04/2025

FROM 0.40000ml of V13706 + 24.60000ml of V14613 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP133887	05/12/2025	06/23/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/14/2025

FROM 0.40000ml of V14843 + 1.00000ml of V14432 + 1.00000ml of V14435 + 1.00000ml of V14503 + 1.00000ml of V14504 + 1.00000ml of V14525 + 1.00000ml of V14526 + 1.00000ml of V14711 + 1.00000ml of V14717 + 1.00000ml of V14718 + 1.00000ml of V14721 + 1.00000ml of V14749 + 1.00000ml of V14750 + 1.00000ml of V14811 + 1.00000ml of V14812 + 10.60000ml of V14921 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP133935	05/16/2025	11/12/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/21/2025

FROM 0.25000ml of V14290 + 24.75000ml of V14921 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP133953	05/19/2025	11/09/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/21/2025

FROM 0.25000ml of V13391 + 24.75000ml of V14626 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP133991	05/22/2025	06/19/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
05/24/2025								
<u>FROM</u>	1.00000ml of V14944 + 1.00000ml of V14945 + 1.00000ml of V14946 + 1.00000ml of V14947 + 21.00000ml of V14620 = Final Quantity: 25.000 ml							

[illegible]

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP134102	06/03/2025	06/04/2025	John Carlone	None	None	Mahesh Dadoda 06/05/2025
<u>FROM</u> 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP133935 + 0.01250ml of VP132035 + 0.01250ml of VP133887 + 0.01250ml of VP133991 = Final Quantity: 40.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13457

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13460

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13465

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13466

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	02/27/2026	02/27/2025 / SAM	04/12/2023 / SAM	V13706

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	05/12/2025 / SAM	08/15/2024 / SAM	V14432

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14435

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14503

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	11/12/2025	05/12/2025 / SAM	09/17/2024 / SAM	V14504

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14525

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	11/12/2025	05/12/2025 / SAM	09/18/2024 / SAM	V14526

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	08/27/2025	02/27/2025 / SAM	11/26/2024 / SAM	V14613

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	06/10/2025	12/10/2024 / SAM	11/26/2024 / SAM	V14614

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	10/25/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14620

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	11/09/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14626

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14630

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14631

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14632

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14633

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14711

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14717

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14718

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14721

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix, 500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/13/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14749

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix, 500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	11/12/2025	05/12/2025 / SAM	12/17/2024 / SAM	V14750

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	11/12/2025	05/12/2025 / SAM	01/08/2025 / SAM	V14811

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	06/30/2026	05/12/2025 / SAM	01/08/2025 / SAM	V14812

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	11/12/2025	05/12/2025 / SAM	01/21/2025 / SAM	V14843

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	11/12/2025	05/12/2025 / SAM	05/09/2025 / SAM	V14921

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14945

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14946

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	051925	06/19/2025	05/22/2025 / SAM	05/21/2025 / SAM	V14947

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis




Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KME)	Lot	DIL	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information		
														(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	CASE		
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	NA	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	NA	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	NA	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	NA	or-rat 14800mg/kg
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 76mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg
10. Methacrylonitrile	(0472)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	14002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (30mg/m3/8H)	or-rat 720mg/kg
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	NA	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	NA	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	158-59-2	NA	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	158-60-5	NA	N/A
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 1235mg/kg
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.8	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 820mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.4	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 200mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	58-23-5	2 ppm (12.6mg/m3/8H)	or-rat 933mg/kg
25. Chloroform	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.4	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 2350mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.4	74-95-3	NA	or-rat 908mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	75-34-3	100 ppm	or-rat 108mg/kg
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	NA	or-rat 725mg/kg
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	1999.8	20.4	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	108-83-4	20 ppm (8H)	or-rat 170mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 108mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (8H)	or-rat 670mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	NA	or-rat 3600mg/kg
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	NA	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10081-01-5	NA	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10081-02-6	NA	N/A
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	NA	or-rat 670mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-rat 3402mg/kg
45. Benzene	35162	050823	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	98-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-1	NA	or-rat 2699mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	NA	N/A
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	NA	or-rat 4750mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	NA	or-rat 1390mg/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	NA	or-rat 5g/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	NA	or-rat 5000mg/kg
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	NA	N/A
60. Chlorobenzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	NA	N/A
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2240mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 2290mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 3900mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 2100mg/kg
65. 1																




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methoxyvinylbenzene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2-Pentene	27.07
22	cis-1,2-Dichloroethene	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloroethene	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.84
28	1,2-Dibromomethane	33.06
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KME)	Lot	DIL	Initial Vol. (mL)	Initial Conc. (µg/mL)	Nominal Conc. (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
														(Solvent Safety Info. On Attached pg.)		
														CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102366	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 75mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 106mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10000mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-1	N/A	or-rat 2699mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40005.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	N/A	or-rat 5g/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-6	N/A	or-rat 2240mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-rat 1060mg/kg
65. 1,4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	20										




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Methoxyvinylbenzene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2-Pentene	27.07
22	cis-1,2-Dichloroethane	28.03
23	Toluene	28.73
24	Ethyl methacrylate/trans-1,2-Dichloroethane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.84
28	1,2-Dibromomethane	33.06
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.53
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.26
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324Q

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 101424

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324QExpiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTBWeight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

27

60000

50000

40000

30000

20000

10000

37

44

65 75 85

119

158 169

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp. = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT. 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1
		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm		
Skin notation	TWA 200 ppm		
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



ANAB ISO 17034 Accredited
AR-1539 Certificate Number
<https://absolutestandards.com>

CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

GHS/OSHA Compliant

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT. 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302, 332	If on skin, wash with soap and water	P305, 351, 338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1
		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Methanol	67-56-1 TWA 200 ppm		
Skin notation	TWA 200 ppm		
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : December 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix
Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix
8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dylan Murphy
Dylan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix
8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek Review of SDS per April 18, 2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : October 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water

Lot#
041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: **061925**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **5000**
NIST Test ID#: **6UTB**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): **10.0**

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM

see 05/21/25



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **051925**
Description: **Acrolein**

Solvent(s):
Water Lot# 041725Q

5.019

114944-114948

		051925
Formulated By:	Lawrence Barry	DATE
		051925
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 061925
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05170	5004.1	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.) Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 - Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
 - Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 - All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 - Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
- Rev 1.0, 2/25/2025



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax: (908) 78-8922
www.chemtech.net

Chemtech Project Number: Q218C/87
COC Number:

CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION	
COMPANY: Tetra Tech		PROJECT NAME: NWIRP Bethpage		BILL TO: SEE CONTRACT PO#	
ADDRESS: 4433 Corporation Lane Suite 300		PROJECT #: 112G08005-WE13 LOCATION: VPB-182		ADDRESS:	
CITY: Virginia Beach STATE: VA ZIP: 23462		PROJECT MANAGER: Ernie Wu		CITY: STATE: ZIP:	
ATTENTION: Ernie Wu		E-MAIL: ernie.wu@tetratech.com		ATTENTION: PHONE:	
PHONE: 757-466-4901 FAX: 757-461-4148		PHONE: 757-466-4901 FAX: 757-461-4148			

DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS										PRESERVATIVES		COMMENTS
FAX: 2 & 10 DAYS*		☐ RESULTS ONLY ☐ USEPA CLP ☐ RESULTS + QC ☐ New York State ASP "B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format		VOC(SW846-260B) 1,4 Dioxane (8270 SIM) Method 522_PRC 1,4- 1 2 3 4 5 6 7 8 9												<-- Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other
HARD COPY: 2 & 10 DAYS*																
EDD 2 & 10 DAYS*																
* TO BE APPROVED BY CHEMTECH																
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS																

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	A										<-- Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9		
1.	BP-VPB-182-TB-20250530	QA		X	5/30/25	9:00	2	2										Trip Blank
2.	BP-VPB-182-GW-670-672	AQ		X	5/30/25	10:08	3	3										
3.	BP-VPB-182-GW-680-682	AQ		X	5/30/25	12:30	3	2	1									8260B MS/MSD
4.	BP-VPB-182-GW-705-707	AQ		X	6/2/25	12:38	2	2										
5.	BP-VPB-182-GW-720-722	AQ		X	6/2/25	14:40	3	2	1									
6.	BP-VPB-182-DUP-20250602	QA		X	6/2/25	12:00	2	2										8260B Duplicate
7.	VPB182-HYD-20250530	QA		X	5/30/25	13:00	5	2	1	2								Hydrant Sample
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	1555	Conditions of bottles or coolers at receipt: q Compliant q Non Compliant q Cooler Temp 2.2°C MeOH extraction requires an additional 4oz. Jar for percent solid q Ice in Cooler?: Comments: 5 Day TAT - For VOC's see worksheet #15 of SAP 2018 for VPB program VOC list 10-DAY TAT - For 1,4 Dioxane (8270 SIM)
1.	6/2/25 1545		6-2-25	
RELINQUISHED BY	DATE/TIME	RECEIVED BY		
2.		2.		
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY		
3.	1850 6-2-25	3.		SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight CHEMTECH: ☐ Picked Up ☐ Overnight Shipment Complete ☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2186	TETR06	Order Date : 6/2/2025 4:04:00 PM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : NWIRP Bethpage 112G080	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 6/2/2025 12:00:00 AM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order : 18:50	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2186-01	BP-VPB-182-TB-20250530	Water	05/30/2025	09:00					
					VOCMS Group1		8260-Low	5 Bus. Days	
Q2186-02	BP-VPB-182-GW-670-672	Water	05/30/2025	10:08					
					VOCMS Group1		8260-Low	5 Bus. Days	
Q2186-03	BP-VPB-182-GW-680-682	Water	05/30/2025	12:30					
					VOCMS Group1		8260-Low	5 Bus. Days	
Q2186-06	BP-VPB-182-GW-705-707	Water	06/02/2025	12:38					
					VOCMS Group1		8260-Low	5 Bus. Days	
Q2186-07	BP-VPB-182-GW-720-722	Water	06/02/2025	14:40					
					VOCMS Group1		8260-Low	5 Bus. Days	
Q2186-08	BP-VPB-182-DUP-20250602	Water	06/02/2025	12:00					
					VOCMS Group1		8260-Low	5 Bus. Days	
Q2186-09	VPB182-HYD-20250530	Water	05/30/2025	13:00					
					VOCMS Group1		8260-Low	5 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2186 TETR06

Order Date : 6/2/2025 4:04:00 PM

Project Mgr :

Client Name : Tetra Tech NUS, Inc.

Project Name : NWIRP Bethpage 112G080

Report Type : Level 4

Client Contact : Ernie Wu

Receive DateTime : 6/2/2025 ~~12:00:00~~ AM
18:50

EDD Type : ADAPT

Invoice Name : Tetra Tech NUS, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Ernie Wu

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

Relinquished By : 

Date / Time : 6/3/25

Received By : 

Date / Time : 6/3/25

Storage Area : VOA Refridgerator Room

Samples received on 6/2/25 @ 1850
Samples placed in SM-REF-2 on 6/2/25

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	83671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	31909	12.820	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	25.640%#
35) Dibromofluoromethane	5.385	113	21112	12.804	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	25.600%#
50) Toluene-d8	8.646	98	73333	13.200	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	26.400%#
62) 4-Bromofluorobenzene	11.079	95	26760	13.236	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	26.480%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	23341	12.638	ug/l	96
3) Chloromethane	1.306	50	24316	13.371	ug/l	98
4) Vinyl Chloride	1.373	62	22096	13.868	ug/l	94
5) Bromomethane	1.593	94	9920	12.279	ug/l	93
6) Chloroethane	1.666	64	11832	14.414	ug/l	95
7) Trichlorofluoromethane	1.873	101	34636	14.271	ug/l	100
8) Diethyl Ether	2.129	74	11392	14.278	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.324	101	21027	14.499	ug/l	98
10) Methyl Iodide	2.446	142	25661	15.070	ug/l	99
11) Tert butyl alcohol	2.971	59	20370	68.662	ug/l	97
12) 1,1-Dichloroethene	2.312	96	18899	13.385	ug/l	98
13) Acrolein	2.233	56	26427	76.689	ug/l	95
14) Allyl chloride	2.660	41	37708	14.133	ug/l	99
15) Acrylonitrile	3.062	53	63290	71.312	ug/l	98
16) Acetone	2.379	43	60363	70.884	ug/l	99
17) Carbon Disulfide	2.501	76	43348	13.251	ug/l	100
18) Methyl Acetate	2.702	43	27234	13.397	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	68402	13.890	ug/l	100
20) Methylene Chloride	2.782	84	23050	13.362	ug/l	96
21) trans-1,2-Dichloroethene	3.086	96	19161	13.294	ug/l	96
22) Diisopropyl ether	3.763	45	74257	14.896	ug/l	95
23) Vinyl Acetate	3.720	43	322598	73.230	ug/l	100
24) 1,1-Dichloroethane	3.605	63	41257	14.082	ug/l	98
25) 2-Butanone	4.556	43	90331	73.489	ug/l	98
26) 2,2-Dichloropropane	4.470	77	30472	13.729	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23970	13.791	ug/l	98
28) Bromochloromethane	4.897	49	21028	13.497	ug/l	99
29) Tetrahydrofuran	5.007	42	56819	71.175	ug/l	100
30) Chloroform	5.086	83	43065	14.144	ug/l	95
31) Cyclohexane	5.458	56	36470	14.769	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	37003	14.092	ug/l	98
36) 1,1-Dichloropropene	5.684	75	27242	14.231	ug/l	98
37) Ethyl Acetate	4.720	43	33458	13.931	ug/l	98
38) Carbon Tetrachloride	5.671	117	31091	14.137	ug/l	98
39) Methylcyclohexane	7.372	83	35077	14.623	ug/l	99
40) Benzene	6.031	78	83898	14.140	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	18732	14.038	ug/l	95
42) 1,2-Dichloroethane	6.080	62	37203	15.165	ug/l	100
43) Isopropyl Acetate	6.342	43	53275	14.512	ug/l	99
44) Trichloroethene	7.116	130	20239	14.429	ug/l	99
45) 1,2-Dichloropropane	7.427	63	20940	14.204	ug/l	98
46) Dibromomethane	7.580	93	16781	14.439	ug/l	99
47) Bromodichloromethane	7.817	83	32762	14.545	ug/l	94
48) Methyl methacrylate	7.689	41	27374	14.521	ug/l	100
49) 1,4-Dioxane	7.659	88	10585	282.442	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	182455	75.775	ug/l	98
52) Toluene	8.713	92	52030	14.788	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27549	14.566	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	31226	14.154	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	20536	14.442	ug/l	96
56) Ethyl methacrylate	9.116	69	31787	14.425	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36341	14.417	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	79531	80.312	ug/l	100
59) 2-Hexanone	9.427	43	137054	74.997	ug/l	99
60) Dibromochloromethane	9.518	129	22267	14.387	ug/l	99
61) 1,2-Dibromoethane	9.610	107	21113	14.458	ug/l	96
64) Tetrachloroethene	9.268	164	19924	15.263	ug/l	97
65) Chlorobenzene	10.079	112	55863	14.202	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.158	131	18671	14.231	ug/l	98
67) Ethyl Benzene	10.189	91	98113	14.631	ug/l	100
68) m/p-Xylenes	10.299	106	72202	29.806	ug/l	99
69) o-Xylene	10.640	106	35178	14.408	ug/l	96
70) Styrene	10.652	104	58034	14.853	ug/l	98
71) Bromoform	10.799	173	13825	14.030	ug/l #	100
73) Isopropylbenzene	10.957	105	93013	14.096	ug/l	99
74) N-amyl acetate	10.841	43	44684	13.581	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	31830	13.560	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	34837	13.811	ug/l	89
77) Bromobenzene	11.195	156	21695	14.212	ug/l	96
78) n-propylbenzene	11.298	91	106337	14.376	ug/l	99
79) 2-Chlorotoluene	11.359	91	68534	13.748	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79261	14.468	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8104	13.371	ug/l	91
82) 4-Chlorotoluene	11.451	91	77338	14.116	ug/l	100
83) tert-Butylbenzene	11.713	119	75382	14.011	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	79232	14.480	ug/l	99
85) sec-Butylbenzene	11.890	105	95271	14.263	ug/l	100
86) p-Isopropyltoluene	12.006	119	77599	14.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	39520	13.898	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	39419	14.146	ug/l	99
89) n-Butylbenzene	12.329	91	66506	14.184	ug/l	99
90) Hexachloroethane	12.536	117	13332	13.287	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	39031	14.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7232	13.535	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	20846	13.745	ug/l	99
94) Hexachlorobutadiene	13.725	225	9524	13.426	ug/l	97
95) Naphthalene	13.774	128	77542	14.040	ug/l	99
96) 1,2,3-Trichlorobenzene	13.956	180	22301	14.023	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:168 Resp: 83671
Ion Ratio Lower Upper
168 100
99 72.7 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 12.638 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 85 Resp: 23341
Ion Ratio Lower Upper
85 100
87 33.5 15.7 47.1









#3
Chloromethane
Concen: 13.371 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 50 Resp: 24316
Ion Ratio Lower Upper
50 100
52 33.0 25.4 38.2



#4
Vinyl Chloride
Concen: 13.868 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 62 Resp: 22096
Ion Ratio Lower Upper
62 100
64 34.9 25.2 37.8









#5
Bromomethane
Concen: 12.279 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 94 Resp: 9920
Ion Ratio Lower Upper
94 100
96 98.1 72.8 109.2



#6
Chloroethane
Concen: 14.414 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 64 Resp: 11832
Ion Ratio Lower Upper
64 100
66 32.9 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 14.271 ug/l
 RT: 1.873 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion:101 Resp: 34636
 Ion Ratio Lower Upper
 101 100
 103 63.5 51.0 76.4



#8
 Diethyl Ether
 Concen: 14.278 ug/l
 RT: 2.129 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Tgt Ion: 74 Resp: 11392
 Ion Ratio Lower Upper
 74 100
 45 103.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 14.499 ug/l
RT: 2.324 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 21027
Ion Ratio Lower Upper
101 100
85 46.6 38.6 58.0
151 70.4 55.2 82.8



#10
Methyl Iodide
Concen: 15.070 ug/l
RT: 2.446 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:142 Resp: 25661
Ion Ratio Lower Upper
142 100
127 53.0 41.7 62.5
141 14.7 11.5 17.3









#11
Tert butyl alcohol
Concen: 68.662 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 59 Resp: 20370
Ion Ratio Lower Upper
59 100
57 11.7 8.6 12.8



#12
1,1-Dichloroethene
Concen: 13.385 ug/l
RT: 2.312 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 96 Resp: 18899
Ion Ratio Lower Upper
96 100
61 179.8 146.2 219.2
98 61.8 51.0 76.6





#13

Acrolein

Concen: 76.689 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 56 Resp: 26427

Ion Ratio Lower Upper

56 100

55 74.3 56.2 84.4



#14

Allyl chloride

Concen: 14.133 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 41 Resp: 37708


Ion Ratio Lower Upper

41 100

39 75.0 60.6 90.8

76 30.6 24.9 37.3









#15
Acrylonitrile
Concen: 71.312 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 53 Resp: 63290
Ion Ratio Lower Upper
53 100
52 83.5 65.3 97.9
51 38.4 29.8 44.8



#16
Acetone
Concen: 70.884 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 60363
Ion Ratio Lower Upper
43 100
58 26.0 21.2 31.8









#17
Carbon Disulfide
Concen: 13.251 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 43348
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4



#18
Methyl Acetate
Concen: 13.397 ug/l
RT: 2.702 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 27234
Ion Ratio Lower Upper
43 100
74 20.7 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 13.890 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 73 Resp: 68402
Ion Ratio Lower Upper
73 100
57 22.2 17.7 26.5



#20
Methylene Chloride
Concen: 13.362 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 84 Resp: 23050
Ion Ratio Lower Upper
84 100
49 146.0 113.9 170.9
51 43.8 33.5 50.3
86 62.2 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 13.294 ug/l
RT: 3.086 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 96 Resp: 19161
Ion Ratio Lower Upper
96 100
61 155.6 119.5 179.3
98 63.2 50.0 75.0



#22
Diisopropyl ether
Concen: 14.896 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 45 Resp: 74257
Ion Ratio Lower Upper
45 100
43 78.2 66.6 100.0
87 24.0 19.8 29.6
59 11.7 8.6 12.8









#23
 Vinyl Acetate
 Concen: 73.230 ug/l
 RT: 3.720 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 43 Resp: 322598
 Ion Ratio Lower Upper
 43 100
 86 9.3 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 14.082 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Tgt Ion: 63 Resp: 41257
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.6 10.8
 100 5.0 2.1 6.3









#25
2-Butanone
Concen: 73.489 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 90331
Ion Ratio Lower Upper
43 100
72 23.8 18.4 27.6



#26
2,2-Dichloropropane
Concen: 13.729 ug/l
RT: 4.470 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 77 Resp: 30472
Ion Ratio Lower Upper
77 100
97 20.9 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 13.791 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Tgt Ion: 96 Resp: 23970




Ion	Ratio	Lower	Upper
96	100		
61	158.3	0.0	322.8
98	63.5	0.0	129.0




#28
 Bromochloromethane
 Concen: 13.497 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Tgt Ion: 49 Resp: 21028

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.0
130	70.9	56.2	84.2





#29

Tetrahydrofuran

Concen: 71.175 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 42 Resp: 56819

Ion Ratio Lower Upper

42 100

72 42.3 33.7 50.5

71 39.2 31.4 47.2



#30

Chloroform

Concen: 14.144 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 83 Resp: 43065

Ion Ratio Lower Upper

83 100

85 65.7 49.3 73.9









#31
Cyclohexane
Concen: 14.769 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 56 Resp: 36470
Ion Ratio Lower Upper
56 100
69 28.8 24.4 36.6
84 78.5 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 14.092 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 97 Resp: 37003
Ion Ratio Lower Upper
97 100
99 62.8 51.8 77.6
61 49.5 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 12.820 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 65 Resp: 31909

Ion Ratio Lower Upper

65 100

67 49.9 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 114 Resp: 147096


Ion Ratio Lower Upper

114 100

63 25.2 0.0 49.2

88 18.0 0.0 33.6









#35
Dibromofluoromethane
Concen: 12.804 ug/l
RT: 5.385 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:113 Resp: 21112
Ion Ratio Lower Upper
113 100
111 103.0 83.1 124.7
192 17.0 13.3 19.9



#36
1,1-Dichloropropene
Concen: 14.231 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 27242
Ion Ratio Lower Upper
75 100
110 31.6 16.3 48.9
77 31.9 24.3 36.5





#37

Ethyl Acetate

Concen: 13.931 ug/l

RT: 4.720 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 43 Resp: 33458

Ion Ratio Lower Upper

43 100

61 12.6 10.3 15.5

70 10.9 7.9 11.9



#38

Carbon Tetrachloride

Concen: 14.137 ug/l

RT: 5.671 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 117 Resp: 31091


Ion Ratio Lower Upper

117 100

119 95.6 75.2 112.8

121 32.1 24.2 36.4





#39

Methylcyclohexane

Concen: 14.623 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 83 Resp: 35077

Ion Ratio Lower Upper

83 100

55 80.7 64.7 97.1

98 44.8 37.4 56.2



#40

Benzene

Concen: 14.140 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 78 Resp: 83898

Ion Ratio Lower Upper

78 100

77 22.3 19.0 28.4









#41
Methacrylonitrile
Concen: 14.038 ug/l
RT: 4.915 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:	41	Resp:	18732
Ion	Ratio	Lower	Upper
41	100		
39	62.6	47.2	70.8
67	68.3	50.7	76.1
52	32.5	24.6	37.0



#42
1,2-Dichloroethane
Concen: 15.165 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:	62	Resp:	37203
Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2









#43
Isopropyl Acetate
Concen: 14.512 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 53275
Ion Ratio Lower Upper
43 100
61 17.3 14.3 21.5
87 11.6 9.5 14.3



#44
Trichloroethene
Concen: 14.429 ug/l
RT: 7.116 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:130 Resp: 20239
Ion Ratio Lower Upper
130 100
95 102.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 14.204 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 63 Resp: 20940

Ion Ratio Lower Upper

63 100

65 29.1 24.1 36.1



#46

Dibromomethane

Concen: 14.439 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 93 Resp: 16781


Ion Ratio Lower Upper

93 100

95 81.5 65.6 98.4

174 83.1 68.2 102.2









#47
Bromodichloromethane
Concen: 14.545 ug/l
RT: 7.817 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 32762
Ion Ratio Lower Upper
83 100
85 68.4 50.5 75.7
127 8.3 6.5 9.7



#48
Methyl methacrylate
Concen: 14.521 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 41 Resp: 27374
Ion Ratio Lower Upper
41 100
69 73.5 58.5 87.7
39 64.9 51.7 77.5









#49
1,4-Dioxane
Concen: 282.442 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 88 Resp: 10585
Ion Ratio Lower Upper
88 100
43 45.5 33.4 50.2
58 73.6 58.6 88.0



#50
Toluene-d8
Concen: 13.200 ug/l
RT: 8.646 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 98 Resp: 73333
Ion Ratio Lower Upper
98 100
100 64.8 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 75.775 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 182455
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3



#52
Toluene
Concen: 14.788 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 92 Resp: 52030
Ion Ratio Lower Upper
92 100
91 172.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 14.566 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 75 Resp: 27549

Ion Ratio Lower Upper

75 100

77 31.4 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 14.154 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 75 Resp: 31226


Ion Ratio Lower Upper

75 100

77 30.8 25.4 38.0

39 64.0 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 14.442 ug/l
 RT: 9.152 min Scan# 1323
 Delta R.T. 0.006 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion:	97	Resp:	20536
Ion	Ratio	Lower	Upper
97	100		
83	87.0	69.0	103.6
85	58.7	44.3	66.5
99	66.8	47.8	71.6



#56
 Ethyl methacrylate
 Concen: 14.425 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Tgt Ion:	69	Resp:	31787
Ion	Ratio	Lower	Upper
69	100		
41	78.2	60.8	91.2
39	49.3	39.0	58.6





#57

1,3-Dichloropropane

Concen: 14.417 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 76 Resp: 36341

Ion Ratio Lower Upper

76 100

78 31.8 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 80.312 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 63 Resp: 79531

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3









#59
2-Hexanone
Concen: 74.997 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 137054
Ion Ratio Lower Upper
43 100
58 48.7 24.9 74.6



#60
Dibromochloromethane
Concen: 14.387 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 129 Resp: 22267
Ion Ratio Lower Upper
129 100
127 78.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 14.458 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:107 Resp: 21113
Ion Ratio Lower Upper
107 100
109 95.3 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 13.236 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 95 Resp: 26760
Ion Ratio Lower Upper
95 100
174 67.4 0.0 135.8
176 67.0 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:117 Resp: 127829
Ion Ratio Lower Upper
117 100
82 57.7 50.6 76.0
119 32.7 25.8 38.6



#64
Tetrachloroethene
Concen: 15.263 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:164 Resp: 19924
Ion Ratio Lower Upper
164 100
166 127.2 105.0 157.6
129 107.6 83.5 125.3
131 97.7 76.5 114.7









#65
Chlorobenzene
Concen: 14.202 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:112 Resp: 55863
Ion Ratio Lower Upper
112 100
114 32.0 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 14.231 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:131 Resp: 18671
Ion Ratio Lower Upper
131 100
133 94.5 47.3 141.9
119 66.7 31.6 95.0





#67

Ethyl Benzene

Concen: 14.631 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 91 Resp: 98113

Ion Ratio Lower Upper

91 100

106 29.3 23.4 35.2



#68

m/p-Xylenes

Concen: 29.806 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 106 Resp: 72202

Ion Ratio Lower Upper

106 100

91 216.1 171.2 256.8









#69
o-Xylene
Concen: 14.408 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:106 Resp: 35178
Ion Ratio Lower Upper
106 100
91 231.7 112.7 338.1



#70
Styrene
Concen: 14.853 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:104 Resp: 58034
Ion Ratio Lower Upper
104 100
78 58.7 45.7 68.5
103 55.6 43.7 65.5





#71

Bromoform

Concen: 14.030 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:173 Resp: 13825

Ion Ratio Lower Upper

173 100

175 46.4 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:152 Resp: 60503


Ion Ratio Lower Upper

152 100

115 76.6 46.9 140.7

150 163.6 0.0 351.0









#73
Isopropylbenzene
Concen: 14.096 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 93013
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.4



#74
N-ethyl acetate
Concen: 13.581 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 44684
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 24.1 18.7 28.1
61 22.3 17.1 25.7









#75
1,1,2-Tetrachloroethane
Concen: 13.560 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 83 Resp: 31830
Ion Ratio Lower Upper
83 100
131 10.1 5.0 14.9
85 63.3 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 13.811 ug/l
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 34837
Ion Ratio Lower Upper
75 100
77 34.4 20.5 61.5





#77

Bromobenzene

Concen: 14.212 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:156 Resp: 21695

Ion Ratio Lower Upper

156 100

77 179.9 92.8 278.3

158 93.1 48.8 146.3



#78

n-propylbenzene

Concen: 14.376 ug/l

RT: 11.298 min Scan# 1675

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 91 Resp: 106337

Ion Ratio Lower Upper

91 100

120 22.0 10.8 32.4









#79
2-Chlorotoluene
Concen: 13.748 ug/l
RT: 11.359 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 68534
Ion Ratio Lower Upper
91 100
126 31.4 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 14.468 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 105 Resp: 79261
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2









#81
 trans-1,4-Dichloro-2-butene
 Concen: 13.371 ug/l
 RT: 11.018 min Scan# 1629
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 75 Resp: 8104
 Ion Ratio Lower Upper
 75 100
 53 138.3 100.5 150.7
 89 44.6 37.0 55.6



#82
 4-Chlorotoluene
 Concen: 14.116 ug/l
 RT: 11.451 min Scan# 1700
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Tgt Ion: 91 Resp: 77338
 Ion Ratio Lower Upper
 91 100
 126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 14.011 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:119 Resp: 75382

Ion Ratio Lower Upper

119 100

91 66.8 32.9 98.7

134 23.7 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 14.480 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:105 Resp: 79232

Ion Ratio Lower Upper

105 100

120 42.0 21.2 63.6









#85
sec-Butylbenzene
Concen: 14.263 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 95271
Ion Ratio Lower Upper
105 100
134 19.3 9.7 29.1



#86
p-Isopropyltoluene
Concen: 14.387 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:119 Resp: 77599
Ion Ratio Lower Upper
119 100
134 25.9 12.5 37.5
91 28.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 13.898 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:146 Resp: 39520

Ion Ratio Lower Upper

146 100

111 42.7 22.1 66.3

148 62.9 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 14.146 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:146 Resp: 39419


Ion Ratio Lower Upper

146 100

111 42.8 21.3 63.9

148 64.2 31.9 95.5





#89

n-Butylbenzene

Concen: 14.184 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 91 Resp: 66506

Ion Ratio Lower Upper

91 100

92 52.9 26.9 80.7

134 23.1 11.8 35.3



#90

Hexachloroethane

Concen: 13.287 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 117 Resp: 13332

Ion Ratio Lower Upper

117 100

201 62.9 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 14.092 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:146 Resp: 39031
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 64.2 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 13.535 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 7232
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 13.745 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:180 Resp: 20846

Ion Ratio Lower Upper

180 100

182 96.6 47.9 143.8

145 34.1 17.1 51.2



#94

Hexachlorobutadiene

Concen: 13.426 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:225 Resp: 9524


Ion Ratio Lower Upper

225 100

223 65.6 30.8 92.4

227 61.4 30.9 92.7





#95

Naphthalene

Concen: 14.040 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:128 Resp: 77542

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6

129 11.1 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 14.023 ug/l

RT: 13.956 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:180 Resp: 22301

Ion Ratio Lower Upper

180 100

182 92.1 47.8 143.3

145 36.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	102335	36.110	ug/l	86
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 97076
Ion Ratio Lower Upper
168 100
99 68.6 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 39.120 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 85 Resp: 83826
Ion Ratio Lower Upper
85 100
87 31.4 15.7 47.1









#3
Chloromethane
Concen: 35.658 ug/l
RT: 1.307 min Scan# 36
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 75238
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2



#4
Vinyl Chloride
Concen: 37.273 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 62 Resp: 68901
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8









#5
Bromomethane
Concen: 33.765 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 31648
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2



#6
Chloroethane
Concen: 38.558 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 64 Resp: 36722
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 36.824 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:101 Resp: 103693
 Ion Ratio Lower Upper
 101 100
 103 63.7 51.0 76.4



#8
 Diethyl Ether
 Concen: 35.350 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 74 Resp: 32724
 Ion Ratio Lower Upper
 74 100
 45 109.9 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 36.997 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 62251
Ion Ratio Lower Upper
101 100
85 48.3 38.6 58.0
151 69.0 55.2 82.8



#10
Methyl Iodide
Concen: 39.622 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:142 Resp: 78277
Ion Ratio Lower Upper
142 100
127 52.1 41.7 62.5
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 181.745 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 62557
Ion Ratio Lower Upper
59 100
57 10.7 8.6 12.8



#12
1,1-Dichloroethene
Concen: 35.610 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 96 Resp: 58334
Ion Ratio Lower Upper
96 100
61 182.7 146.2 219.2
98 63.8 51.0 76.6









#13
Acrolein
Concen: 184.907 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 73927
Ion Ratio Lower Upper
56 100
55 70.3 56.2 84.4



#14
Allyl chloride
Concen: 36.971 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 41 Resp: 114445
Ion Ratio Lower Upper
41 100
39 75.7 60.6 90.8
76 31.1 24.9 37.3









#15
Acrylonitrile
Concen: 182.874 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 53 Resp: 188304
Ion Ratio Lower Upper
53 100
52 81.6 65.3 97.9
51 37.3 29.8 44.8



#16
Acetone
Concen: 177.910 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 175777
Ion Ratio Lower Upper
43 100
58 26.5 21.2 31.8









#17
Carbon Disulfide
Concen: 37.223 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 141281
Ion Ratio Lower Upper
76 100
78 8.7 7.0 10.4



#18
Methyl Acetate
Concen: 34.915 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 82347
Ion Ratio Lower Upper
43 100
74 20.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 36.695 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 209665
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5



#20
Methylene Chloride
Concen: 33.182 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 84 Resp: 66412
Ion Ratio Lower Upper
84 100
49 142.4 113.9 170.9
51 41.9 33.5 50.3
86 67.3 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 35.421 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 59234
Ion Ratio Lower Upper
96 100
61 149.4 119.5 179.3
98 62.5 50.0 75.0



#22
Diisopropyl ether
Concen: 38.241 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 45 Resp: 221172
Ion Ratio Lower Upper
45 100
43 83.3 66.6 100.0
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
 Vinyl Acetate
 Concen: 194.528 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 994236
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 36.067 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 63 Resp: 122601
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.6 10.8
 100 4.2 2.1 6.3





#25

2-Butanone

Concen: 188.783 ug/l

RT: 4.556 min Scan# 569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 43 Resp: 269224

Ion Ratio Lower Upper

43 100

72 23.0 18.4 27.6



#26

2,2-Dichloropropane

Concen: 36.059 ug/l

RT: 4.465 min Scan# 554

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 77 Resp: 92857

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 35.466 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 96 Resp: 71518

Ion Ratio Lower Upper

96 100

61 161.4 0.0 322.8

98 64.5 0.0 129.0



#28

Bromochloromethane

Concen: 31.031 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 49 Resp: 56091


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.0

130 70.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 183.679 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 170124

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5

71 39.3 31.4 47.2



#30

Chloroform

Concen: 35.626 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 83 Resp: 125850

Ion Ratio Lower Upper

83 100

85 61.6 49.3 73.9









#31
Cyclohexane
Concen: 38.205 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 109459
Ion Ratio Lower Upper
56 100
69 30.5 24.4 36.6
84 83.6 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 36.034 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 97 Resp: 109781
Ion Ratio Lower Upper
97 100
99 64.7 51.8 77.6
61 50.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 30.601 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 65 Resp: 88371

Ion Ratio Lower Upper

65 100

67 49.5 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 114 Resp: 167947


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 16.8 0.0 33.6









#35
Dibromofluoromethane
Concen: 31.633 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:113 Resp: 59550
Ion Ratio Lower Upper
113 100
111 103.9 83.1 124.7
192 16.6 13.3 19.9



#36
1,1-Dichloropropene
Concen: 38.073 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 83215
Ion Ratio Lower Upper
75 100
110 32.6 16.3 48.9
77 30.4 24.3 36.5









#37
Ethyl Acetate
Concen: 37.394 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 102537
Ion Ratio Lower Upper
43 100
61 12.9 10.3 15.5
70 9.9 7.9 11.9



#38
Carbon Tetrachloride
Concen: 37.320 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 117 Resp: 93712
Ion Ratio Lower Upper
117 100
119 94.0 75.2 112.8
121 30.3 24.2 36.4









#39
Methylcyclohexane
Concen: 39.307 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 107651
Ion Ratio Lower Upper
83 100
55 80.9 64.7 97.1
98 46.8 37.4 56.2



#40
Benzene
Concen: 36.530 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 78 Resp: 247476
Ion Ratio Lower Upper
78 100
77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 38.123 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:	41	Resp:	58082
Ion	Ratio	Lower	Upper
41	100		
39	59.0	47.2	70.8
67	63.4	50.7	76.1
52	30.8	24.6	37.0



#42
1,2-Dichloroethane
Concen: 37.605 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:	62	Resp:	105332
Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2





#43

Isopropyl Acetate

Concen: 38.598 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 43 Resp: 161787

Ion Ratio Lower Upper

43 100

61 17.9 14.3 21.5

87 11.9 9.5 14.3



#44

Trichloroethene

Concen: 37.230 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:130 Resp: 59623

Ion Ratio Lower Upper

130 100

95 102.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 37.064 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 63 Resp: 62387

Ion Ratio Lower Upper

63 100

65 30.1 24.1 36.1



#46

Dibromomethane

Concen: 36.326 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 93 Resp: 48201


Ion Ratio Lower Upper

93 100

95 82.0 65.6 98.4

174 85.2 68.2 102.2









#47
Bromodichloromethane
Concen: 37.685 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 96916
Ion Ratio Lower Upper
83 100
85 63.1 50.5 75.7
127 8.1 6.5 9.7



#48
Methyl methacrylate
Concen: 39.157 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 41 Resp: 84277
Ion Ratio Lower Upper
41 100
69 73.1 58.5 87.7
39 64.6 51.7 77.5





#49

1,4-Dioxane

Concen: 713.477 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 30529

Ion Ratio Lower Upper

88 100

43 41.8 33.4 50.2

58 73.3 58.6 88.0



#50

Toluene-d8

Concen: 32.395 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 98 Resp: 205479

Ion Ratio Lower Upper

98 100

100 66.9 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 193.653 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 532388
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3



#52
Toluene
Concen: 37.530 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 92 Resp: 150757
Ion Ratio Lower Upper
92 100
91 170.8 136.6 205.0









#53
t-1,3-Dichloropropene
Concen: 41.056 ug/l
RT: 8.976 min Scan# 1294
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 88655
Ion Ratio Lower Upper
75 100
77 31.3 25.0 37.6



#54
cis-1,3-Dichloropropene
Concen: 38.521 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 97031
Ion Ratio Lower Upper
75 100
77 31.7 25.4 38.0
39 65.3 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 36.648 ug/l
 RT: 9.147 min Scan# 1322
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:	97	Resp:	59499
Ion	Ratio	Lower	Upper
97	100		
83	86.3	69.0	103.6
85	55.4	44.3	66.5
99	59.7	47.8	71.6



#56
 Ethyl methacrylate
 Concen: 39.726 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion:	69	Resp:	99947
Ion	Ratio	Lower	Upper
69	100		
41	76.0	60.8	91.2
39	48.8	39.0	58.6





#57

1,3-Dichloropropane

Concen: 36.347 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 104606

Ion Ratio Lower Upper

76 100

78 32.9 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 227.902 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 63 Resp: 257678

Ion Ratio Lower Upper

63 100

106 26.9 21.5 32.3









#59
2-Hexanone
Concen: 190.168 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 396784
Ion Ratio Lower Upper
43 100
58 49.7 24.9 74.6



#60
Dibromochloromethane
Concen: 38.027 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 129 Resp: 67198
Ion Ratio Lower Upper
129 100
127 78.5 39.3 117.8









#61
1,2-Dibromoethane
Concen: 37.567 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 62633
Ion Ratio Lower Upper
107 100
109 91.6 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 34.229 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 95 Resp: 79012
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.7 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:117 Resp: 146257
Ion Ratio Lower Upper
117 100
82 63.3 50.6 76.0
119 32.2 25.8 38.6



#64
Tetrachloroethene
Concen: 36.726 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:164 Resp: 54853
Ion Ratio Lower Upper
164 100
166 131.3 105.0 157.6
129 104.4 83.5 125.3
131 95.6 76.5 114.7





#65

Chlorobenzene

Concen: 35.686 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:112 Resp: 160600

Ion Ratio Lower Upper

112 100

114 31.8 25.4 38.2



#66

1,1,1,2-Tetrachloroethane

Concen: 38.017 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:131 Resp: 57066


Ion Ratio Lower Upper

131 100

133 94.6 47.3 141.9

119 63.3 31.6 95.0









#67
Ethyl Benzene
Concen: 38.543 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 295712
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2



#68
m/p-Xylenes
Concen: 78.142 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 106 Resp: 216578
Ion Ratio Lower Upper
106 100
91 214.0 171.2 256.8





#69

o-Xylene

Concen: 38.064 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:106 Resp: 106335

Ion Ratio Lower Upper

106 100

91 225.4 112.7 338.1



#70

Styrene

Concen: 39.888 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:104 Resp: 178313


Ion Ratio Lower Upper

104 100

78 57.1 45.7 68.5

103 54.6 43.7 65.5





#71

Bromoform

Concen: 39.457 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:173 Resp: 44486

Ion Ratio Lower Upper

173 100

175 46.7 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:152 Resp: 67976


Ion Ratio Lower Upper

152 100

115 93.8 46.9 140.7

150 175.5 0.0 351.0









#73
Isopropylbenzene
Concen: 37.865 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:105 Resp: 280719
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.4



#74
N-ethyl acetate
Concen: 38.012 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 140509
Ion Ratio Lower Upper
43 100
70 38.6 30.9 46.3
55 23.4 18.7 28.1
61 21.4 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 34.495 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 90975
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 63.8 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 36.110 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 102335
Ion Ratio Lower Upper
75 100
77 32.3 20.5 61.5









#77
Bromobenzene
Concen: 36.763 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:156 Resp: 63053
Ion Ratio Lower Upper
156 100
77 185.5 92.8 278.3
158 97.5 48.8 146.3



#78
n-propylbenzene
Concen: 39.707 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 329981
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 36.336 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 203509
Ion Ratio Lower Upper
91 100
126 31.1 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 38.516 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 105 Resp: 237066
Ion Ratio Lower Upper
105 100
120 46.1 23.1 69.2









#81
 trans-1,4-Dichloro-2-butene
 Concen: 38.445 ug/l
 RT: 11.018 min Scan# 1629
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 75 Resp: 26179
 Ion Ratio Lower Upper
 75 100
 53 125.6 100.5 150.7
 89 46.3 37.0 55.6



#82
 4-Chlorotoluene
 Concen: 37.880 ug/l
 RT: 11.451 min Scan# 1700
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 233172
 Ion Ratio Lower Upper
 91 100
 126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 38.622 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 233468

Ion Ratio Lower Upper

119 100

91 65.8 32.9 98.7

134 22.7 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 38.946 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:105 Resp: 239433

Ion Ratio Lower Upper

105 100

120 42.4 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 38.786 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:105 Resp: 291074
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 39.879 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion:119 Resp: 241658
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.5 37.5
 91 27.6 13.8 41.4









#87
1,3-Dichlorobenzene
Concen: 36.199 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 115651
Ion Ratio Lower Upper
146 100
111 44.2 22.1 66.3
148 64.3 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 36.763 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:146 Resp: 115097
Ion Ratio Lower Upper
146 100
111 42.6 21.3 63.9
148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 40.603 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 91 Resp: 213899

Ion Ratio Lower Upper

91 100

92 53.8 26.9 80.7

134 23.5 11.8 35.3



#90

Hexachloroethane

Concen: 37.534 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 117 Resp: 42314

Ion Ratio Lower Upper

117 100

201 63.1 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 37.047 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 115286
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.3
148 63.2 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 36.496 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 21909
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 87.6 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 39.117 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:180 Resp: 66655
Ion Ratio Lower Upper
180 100
182 95.9 47.9 143.8
145 34.1 17.1 51.2



#94
Hexachlorobutadiene
Concen: 36.395 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:225 Resp: 29007
Ion Ratio Lower Upper
225 100
223 61.6 30.8 92.4
227 61.8 30.9 92.7





#95

Naphthalene

Concen: 39.575 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:128 Resp: 245568

Ion Ratio Lower Upper

128 100

127 13.0 10.4 15.6

129 10.8 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 38.765 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:180 Resp: 69262

Ion Ratio Lower Upper

180 100

182 95.5 47.8 143.3

145 36.2 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161877	62.526	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.060%#	
35) Dibromofluoromethane	5.379	113	111455	65.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 130.020%#	
50) Toluene-d8	8.647	98	387230	67.031	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 134.060%#	
62) 4-Bromofluorobenzene	11.079	95	153085	72.818	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 145.640%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	149508	77.828	ug/l	99
3) Chloromethane	1.307	50	136937	72.392	ug/l	98
4) Vinyl Chloride	1.374	62	126609	76.398	ug/l	98
5) Bromomethane	1.593	94	59246	70.507	ug/l	99
6) Chloroethane	1.660	64	57332	67.148	ug/l	99
7) Trichlorofluoromethane	1.867	101	171042	67.754	ug/l	99
8) Diethyl Ether	2.136	74	58762	70.807	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	109520	72.604	ug/l	99
10) Methyl Iodide	2.440	142	138045	77.943	ug/l	100
11) Tert butyl alcohol	2.995	59	125094	405.394	ug/l	100
12) 1,1-Dichloroethene	2.306	96	105566	71.884	ug/l	96
13) Acrolein	2.239	56	134456	375.132	ug/l	99
14) Allyl chloride	2.660	41	206519	74.419	ug/l	98
15) Acrylonitrile	3.068	53	331142	358.723	ug/l	97
16) Acetone	2.386	43	314383	354.936	ug/l	98
17) Carbon Disulfide	2.501	76	264956	77.867	ug/l	100
18) Methyl Acetate	2.703	43	147014	69.531	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	378113	73.818	ug/l	100
20) Methylene Chloride	2.782	84	120270	67.030	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	106606	71.109	ug/l	99
22) Diisopropyl ether	3.763	45	399494	77.048	ug/l	90
23) Vinyl Acetate	3.721	43	1811550	395.363	ug/l	100
24) 1,1-Dichloroethane	3.605	63	219759	72.114	ug/l	99
25) 2-Butanone	4.556	43	485448	379.704	ug/l	100
26) 2,2-Dichloropropane	4.464	77	174579	75.621	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	128454	71.056	ug/l	99
28) Bromochloromethane	4.891	49	103602	63.932	ug/l	100
29) Tetrahydrofuran	5.001	42	305757	368.234	ug/l	99
30) Chloroform	5.086	83	222253	70.180	ug/l	95
31) Cyclohexane	5.458	56	196302	76.428	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	201112	73.635	ug/l	98
36) 1,1-Dichloropropene	5.690	75	147873	74.285	ug/l	99
37) Ethyl Acetate	4.714	43	186215	74.565	ug/l	99
38) Carbon Tetrachloride	5.672	117	168946	73.875	ug/l	99
39) Methylcyclohexane	7.372	83	191941	76.952	ug/l	96
40) Benzene	6.031	78	440885	71.457	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	104863	75.574	ug/l	98
42) 1,2-Dichloroethane	6.080	62	187036	73.318	ug/l	99
43) Isopropyl Acetate	6.342	43	300491	78.714	ug/l	100
44) Trichloroethene	7.123	130	105628	72.420	ug/l	98
45) 1,2-Dichloropropane	7.427	63	112554	73.420	ug/l	100
46) Dibromomethane	7.574	93	85728	70.938	ug/l	100
47) Bromodichloromethane	7.818	83	175301	74.845	ug/l	99
48) Methyl methacrylate	7.689	41	153093	78.101	ug/l	99
49) 1,4-Dioxane	7.659	88	57313	1470.687	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	963319	384.740	ug/l	100
52) Toluene	8.714	92	270763	74.010	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	169863	86.372	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	184134	80.264	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	107264	72.543	ug/l	99
56) Ethyl methacrylate	9.116	69	188717	82.359	ug/l	99
57) 1,3-Dichloropropane	9.305	76	187649	71.590	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	463169	449.791	ug/l	100
59) 2-Hexanone	9.433	43	729602	383.945	ug/l	100
60) Dibromochloromethane	9.518	129	127012	78.918	ug/l	98
61) 1,2-Dibromoethane	9.610	107	112669	74.200	ug/l	98
64) Tetrachloroethene	9.268	164	93195	67.493	ug/l	96
65) Chlorobenzene	10.079	112	293284	70.490	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	103177	74.349	ug/l	99
67) Ethyl Benzene	10.189	91	535122	75.444	ug/l	99
68) m/p-Xylenes	10.299	106	389935	152.181	ug/l	99
69) o-Xylene	10.640	106	190833	73.890	ug/l	97
70) Styrene	10.652	104	328195	79.411	ug/l	100
71) Bromoform	10.799	173	84353	80.927	ug/l #	98
73) Isopropylbenzene	10.957	105	514528	71.083	ug/l	100
74) N-amyl acetate	10.841	43	274553	76.073	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	170481	66.207	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	196706	71.091	ug/l	86
77) Bromobenzene	11.195	156	117151	69.959	ug/l	99
78) n-propylbenzene	11.305	91	608332	74.974	ug/l	100
79) 2-Chlorotoluene	11.360	91	372392	68.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	432096	71.903	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	54398	81.820	ug/l	93
82) 4-Chlorotoluene	11.451	91	432053	71.890	ug/l	100
83) tert-Butylbenzene	11.713	119	432011	73.197	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	442728	73.758	ug/l	100
85) sec-Butylbenzene	11.890	105	543597	74.188	ug/l	100
86) p-Isopropyltoluene	12.006	119	457898	77.393	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	219753	70.449	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	217613	71.190	ug/l	99
89) n-Butylbenzene	12.329	91	416602	80.996	ug/l	100
90) Hexachloroethane	12.536	117	82528	74.978	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	216834	71.367	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43673	74.512	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	137430	82.605	ug/l	98
94) Hexachlorobutadiene	13.725	225	55494	71.314	ug/l	98
95) Naphthalene	13.774	128	489778	80.842	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	139525	79.981	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDIC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:168 Resp: 87028
Ion Ratio Lower Upper
168 100
99 72.1 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 77.828 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 85 Resp: 149508
Ion Ratio Lower Upper
85 100
87 32.0 15.7 47.1









#3
Chloromethane
Concen: 72.392 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 50 Resp: 136937
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2



#4
Vinyl Chloride
Concen: 76.398 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 126609
Ion Ratio Lower Upper
62 100
64 30.6 25.2 37.8









#5
Bromomethane
Concen: 70.507 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 94 Resp: 59246
Ion Ratio Lower Upper
94 100
96 91.7 72.8 109.2



#6
Chloroethane
Concen: 67.148 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 64 Resp: 57332
Ion Ratio Lower Upper
64 100
66 30.8 24.3 36.5









#7
Trichlorofluoromethane
Concen: 67.754 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 171042
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4



#8
Diethyl Ether
Concen: 70.807 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 74 Resp: 58762
Ion Ratio Lower Upper
74 100
45 110.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 72.604 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 109520
Ion Ratio Lower Upper
101 100
85 48.1 38.6 58.0
151 69.9 55.2 82.8



#10
Methyl Iodide
Concen: 77.943 ug/l
RT: 2.440 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:142 Resp: 138045
Ion Ratio Lower Upper
142 100
127 52.0 41.7 62.5
141 14.5 11.5 17.3









#11
Tert butyl alcohol
Concen: 405.394 ug/l
RT: 2.995 min Scan# 310
Delta R.T. 0.018 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 125094
Ion Ratio Lower Upper
59 100
57 10.9 8.6 12.8



#12
1,1-Dichloroethene
Concen: 71.884 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 96 Resp: 105566
Ion Ratio Lower Upper
96 100
61 189.9 146.2 219.2
98 63.3 51.0 76.6









#13
Acrolein
Concen: 375.132 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 134456
Ion Ratio Lower Upper
56 100
55 71.3 56.2 84.4



#14
Allyl chloride
Concen: 74.419 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 41 Resp: 206519
Ion Ratio Lower Upper
41 100
39 73.8 60.6 90.8
76 31.4 24.9 37.3





#15

Acrylonitrile

Concen: 358.723 ug/l

RT: 3.068 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 53 Resp: 331142

Ion Ratio Lower Upper

53 100

52 84.7 65.3 97.9

51 37.9 29.8 44.8



#16

Acetone

Concen: 354.936 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 43 Resp: 314383

Ion Ratio Lower Upper

43 100

58 27.6 21.2 31.8









#17
Carbon Disulfide
Concen: 77.867 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 76 Resp: 264956
Ion Ratio Lower Upper
76 100
78 8.7 7.0 10.4



#18
Methyl Acetate
Concen: 69.531 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 147014
Ion Ratio Lower Upper
43 100
74 22.6 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 73.818 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 73 Resp: 378113
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5



#20
Methylene Chloride
Concen: 67.030 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 84 Resp: 120270
Ion Ratio Lower Upper
84 100
49 136.4 113.9 170.9
51 41.1 33.5 50.3
86 64.0 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 71.109 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 106606
Ion Ratio Lower Upper
96 100
61 149.7 119.5 179.3
98 64.4 50.0 75.0



#22
Diisopropyl ether
Concen: 77.048 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 45 Resp: 399494
Ion Ratio Lower Upper
45 100
43 71.2 66.6 100.0
87 25.7 19.8 29.6
59 11.0 8.6 12.8









#23
 Vinyl Acetate
 Concen: 395.363 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 43 Resp: 1811550
 Ion Ratio Lower Upper
 43 100
 86 9.5 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 72.114 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Tgt Ion: 63 Resp: 219759
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.6 10.8
 100 4.5 2.1 6.3





#25

2-Butanone

Concen: 379.704 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 485448

Ion Ratio Lower Upper

43 100

72 23.1 18.4 27.6



#26

2,2-Dichloropropane

Concen: 75.621 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 77 Resp: 174579

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 71.056 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100

Tgt Ion: 96 Resp: 128454




Ion	Ratio	Lower	Upper
96	100		
61	160.1	0.0	322.8
98	64.9	0.0	129.0




#28
 Bromochloromethane
 Concen: 63.932 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Tgt Ion: 49 Resp: 103602

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.0
130	70.1	56.2	84.2





#29

Tetrahydrofuran

Concen: 368.234 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 42 Resp: 305757

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5

71 38.7 31.4 47.2



#30

Chloroform

Concen: 70.180 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 83 Resp: 222253

Ion Ratio Lower Upper

83 100

85 65.1 49.3 73.9









#31
Cyclohexane
Concen: 76.428 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 196302
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.3 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 73.635 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 97 Resp: 201112
Ion Ratio Lower Upper
97 100
99 63.8 51.8 77.6
61 48.7 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 62.526 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 65 Resp: 161877

Ion Ratio Lower Upper

65 100

67 49.5 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 114 Resp: 152958


Ion Ratio Lower Upper

114 100

63 24.3 0.0 49.2

88 17.1 0.0 33.6









#35
Dibromofluoromethane
Concen: 65.006 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:113 Resp: 111455
Ion Ratio Lower Upper
113 100
111 101.7 83.1 124.7
192 16.6 13.3 19.9



#36
1,1-Dichloropropene
Concen: 74.285 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 147873
Ion Ratio Lower Upper
75 100
110 33.5 16.3 48.9
77 31.1 24.3 36.5









#37
Ethyl Acetate
Concen: 74.565 ug/l
RT: 4.714 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 186215
Ion Ratio Lower Upper
43 100
61 12.4 10.3 15.5
70 9.9 7.9 11.9



#38
Carbon Tetrachloride
Concen: 73.875 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 117 Resp: 168946
Ion Ratio Lower Upper
117 100
119 95.0 75.2 112.8
121 29.6 24.2 36.4









#39
Methylcyclohexane
Concen: 76.952 ug/l
RT: 7.372 min Scan# 1031
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 83 Resp: 191941
Ion Ratio Lower Upper
83 100
55 84.2 64.7 97.1
98 49.0 37.4 56.2



#40
Benzene
Concen: 71.457 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 78 Resp: 440885
Ion Ratio Lower Upper
78 100
77 24.6 19.0 28.4









#41
Methacrylonitrile
Concen: 75.574 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 41 Resp: 104863
Ion Ratio Lower Upper
41 100
39 57.0 47.2 70.8
67 64.6 50.7 76.1
52 29.7 24.6 37.0



#42
1,2-Dichloroethane
Concen: 73.318 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 187036
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2









#43
Isopropyl Acetate
Concen: 78.714 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 300491
Ion Ratio Lower Upper
43 100
61 18.0 14.3 21.5
87 11.6 9.5 14.3



#44
Trichloroethene
Concen: 72.420 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 130 Resp: 105628
Ion Ratio Lower Upper
130 100
95 103.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 73.420 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 63 Resp: 112554

Ion Ratio Lower Upper

63 100

65 30.0 24.1 36.1



#46

Dibromomethane

Concen: 70.938 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 93 Resp: 85728


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2









#47
Bromodichloromethane
Concen: 74.845 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 175301
Ion Ratio Lower Upper
83 100
85 63.6 50.5 75.7
127 7.8 6.5 9.7



#48
Methyl methacrylate
Concen: 78.101 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 41 Resp: 153093
Ion Ratio Lower Upper
41 100
69 74.1 58.5 87.7
39 65.5 51.7 77.5





#49

1,4-Dioxane

Concen: 1470.687 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 88 Resp: 57313

Ion Ratio Lower Upper

88 100

43 40.1 33.4 50.2

58 73.3 58.6 88.0



#50

Toluene-d8

Concen: 67.031 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 98 Resp: 387230

Ion Ratio Lower Upper

98 100

100 65.2 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 384.740 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 963319
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3



#52
Toluene
Concen: 74.010 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 92 Resp: 270763
Ion Ratio Lower Upper
92 100
91 171.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 86.372 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 75 Resp: 169863

Ion Ratio Lower Upper

75 100

77 31.5 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 80.264 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 184134


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 59.6 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 72.543 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion: 97 Resp: 107264




Ion	Ratio	Lower	Upper
97	100		
83	86.8	69.0	103.6
85	56.4	44.3	66.5
99	61.8	47.8	71.6




#56
Ethyl methacrylate
Concen: 82.359 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 69 Resp: 188717

Ion	Ratio	Lower	Upper
69	100		
41	75.7	60.8	91.2
39	47.7	39.0	58.6





#57

1,3-Dichloropropane

Concen: 71.590 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 76 Resp: 187649

Ion Ratio Lower Upper

76 100

78 32.1 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 449.791 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 63 Resp: 463169

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 383.945 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 729602

Ion Ratio Lower Upper

43 100

58 49.9 24.9 74.6



#60

Dibromochloromethane

Concen: 78.918 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 129 Resp: 127012

Ion Ratio Lower Upper

129 100

127 77.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 74.200 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:107 Resp: 112669
Ion Ratio Lower Upper
107 100
109 93.9 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 72.818 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 95 Resp: 153085
Ion Ratio Lower Upper
95 100
174 67.1 0.0 135.8
176 65.3 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:117 Resp: 135214
Ion Ratio Lower Upper
117 100
82 61.7 50.6 76.0
119 30.6 25.8 38.6



#64
Tetrachloroethene
Concen: 67.493 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:164 Resp: 93195
Ion Ratio Lower Upper
164 100
166 123.4 105.0 157.6
129 101.4 83.5 125.3
131 98.0 76.5 114.7









#65
Chlorobenzene
Concen: 70.490 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:112 Resp: 293284
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 74.349 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:131 Resp: 103177
Ion Ratio Lower Upper
131 100
133 95.3 47.3 141.9
119 64.5 31.6 95.0









#67
Ethyl Benzene
Concen: 75.444 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 535122
Ion Ratio Lower Upper
91 100
106 29.6 23.4 35.2



#68
m/p-Xylenes
Concen: 152.181 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:106 Resp: 389935
Ion Ratio Lower Upper
106 100
91 215.8 171.2 256.8









#69
o-Xylene
Concen: 73.890 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:106 Resp: 190833
Ion Ratio Lower Upper
106 100
91 230.3 112.7 338.1



#70
Styrene
Concen: 79.411 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:104 Resp: 328195
Ion Ratio Lower Upper
104 100
78 57.4 45.7 68.5
103 54.8 43.7 65.5





#71

Bromoform

Concen: 80.927 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:173 Resp: 84353

Ion Ratio Lower Upper

173 100

175 48.3 23.4 70.0

254 0.0 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:152 Resp: 66369


Ion Ratio Lower Upper

152 100

115 120.4 46.9 140.7

150 190.4 0.0 351.0









#73
Isopropylbenzene
Concen: 71.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:105 Resp: 514528
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4



#74
N-ethyl acetate
Concen: 76.073 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 274553
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.5 18.7 28.1
61 22.2 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 66.207 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 170481
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 64.9 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 71.091 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 196706
Ion Ratio Lower Upper
75 100
77 32.2 20.5 61.5









#77
Bromobenzene
Concen: 69.959 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:156 Resp: 117151
Ion Ratio Lower Upper
156 100
77 187.4 92.8 278.3
158 96.1 48.8 146.3



#78
n-propylbenzene
Concen: 74.974 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 91 Resp: 608332
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 68.100 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 91 Resp: 372392

Ion Ratio Lower Upper

91 100

126 30.9 15.6 46.7



#80

1,3,5-Trimethylbenzene

Concen: 71.903 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 105 Resp: 432096

Ion Ratio Lower Upper

105 100

120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 81.820 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 54398
Ion Ratio Lower Upper
75 100
53 115.8 100.5 150.7
89 47.0 37.0 55.6



#82
4-Chlorotoluene
Concen: 71.890 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 91 Resp: 432053
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 73.197 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:119 Resp: 432011

Ion Ratio Lower Upper

119 100

91 64.4 32.9 98.7

134 22.8 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 73.758 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 442728

Ion Ratio Lower Upper

105 100

120 42.4 21.2 63.6









#85
sec-Butylbenzene
Concen: 74.188 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:105 Resp: 543597
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1



#86
p-Isopropyltoluene
Concen: 77.393 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:119 Resp: 457898
Ion Ratio Lower Upper
119 100
134 24.6 12.5 37.5
91 26.8 13.8 41.4









#87
1,3-Dichlorobenzene
Concen: 70.449 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion:146 Resp: 219753
Ion Ratio Lower Upper
146 100
111 44.1 22.1 66.3
148 63.5 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 71.190 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:146 Resp: 217613
Ion Ratio Lower Upper
146 100
111 43.4 21.3 63.9
148 63.2 31.9 95.5





#89

n-Butylbenzene

Concen: 80.996 ug/l

RT: 12.329 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 91 Resp: 416602

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7

134 24.0 11.8 35.3



#90

Hexachloroethane

Concen: 74.978 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 117 Resp: 82528

Ion Ratio Lower Upper

117 100

201 63.3 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 71.367 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:146 Resp: 216834
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 64.0 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 74.512 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 43673
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 90.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 82.605 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:180 Resp: 137430

Ion Ratio Lower Upper

180 100

182 93.7 47.9 143.8

145 34.6 17.1 51.2



#94

Hexachlorobutadiene

Concen: 71.314 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:225 Resp: 55494


Ion Ratio Lower Upper

225 100

223 62.3 30.8 92.4

227 64.9 30.9 92.7









#95
Naphthalene
Concen: 80.842 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion:128 Resp: 489778
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 10.9 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 79.981 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:180 Resp: 139525
Ion Ratio Lower Upper
180 100
182 93.7 47.8 143.3
145 35.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60345	50.000	ug/l	# 0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	231952	93.983	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 187.960%#	
35) Dibromofluoromethane	5.385	113	160168	98.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 197.120%#	
50) Toluene-d8	8.647	98	554390	101.252	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 202.500%#	
62) 4-Bromofluorobenzene	11.079	95	217536	109.174	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 218.340%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	217793	118.929	ug/l	98
3) Chloromethane	1.307	50	196843	109.161	ug/l	99
4) Vinyl Chloride	1.374	62	187924	118.953	ug/l	100
5) Bromomethane	1.593	94	83008	103.625	ug/l	95
6) Chloroethane	1.660	64	78806	96.822	ug/l	98
7) Trichlorofluoromethane	1.868	101	245108	101.851	ug/l	97
8) Diethyl Ether	2.136	74	88441	111.791	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.313	101	161404	112.243	ug/l	99
10) Methyl Iodide	2.441	142	189843	112.441	ug/l	100
11) Tert butyl alcohol	3.002	59	181238	616.119	ug/l	100
12) 1,1-Dichloroethene	2.307	96	155451	111.039	ug/l	99
13) Acrolein	2.239	56	203257	594.872	ug/l	98
14) Allyl chloride	2.654	41	297776	112.561	ug/l	98
15) Acrylonitrile	3.069	53	484854	550.973	ug/l	98
16) Acetone	2.392	43	460820	545.754	ug/l	98
17) Carbon Disulfide	2.502	76	397353	122.499	ug/l	100
18) Methyl Acetate	2.709	43	217885	108.099	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	557210	114.112	ug/l	100
20) Methylene Chloride	2.782	84	172035	100.579	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	154694	108.241	ug/l	96
22) Diisopropyl ether	3.764	45	577610	116.858	ug/l	90
23) Vinyl Acetate	3.721	43	2655751	608.005	ug/l	100
24) 1,1-Dichloroethane	3.605	63	320083	110.182	ug/l	99
25) 2-Butanone	4.562	43	708463	581.292	ug/l	98
26) 2,2-Dichloropropane	4.471	77	258698	117.548	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	187835	108.994	ug/l	99
28) Bromochloromethane	4.898	49	146923	95.108	ug/l	100
29) Tetrahydrofuran	5.007	42	450456	569.081	ug/l	99
30) Chloroform	5.087	83	323659	107.208	ug/l	95
31) Cyclohexane	5.458	56	286312	116.934	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	295702	113.572	ug/l	98
36) 1,1-Dichloropropene	5.684	75	219731	116.462	ug/l	99
37) Ethyl Acetate	4.715	43	274312	115.889	ug/l	99
38) Carbon Tetrachloride	5.666	117	250916	115.760	ug/l	99
39) Methylcyclohexane	7.373	83	286311	121.107	ug/l	97
40) Benzene	6.031	78	642178	109.813	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	151424	115.139	ug/l	98
42) 1,2-Dichloroethane	6.080	62	271992	112.492	ug/l	99
43) Isopropyl Acetate	6.342	43	447871	123.780	ug/l	100
44) Trichloroethene	7.117	130	157641	114.032	ug/l	96
45) 1,2-Dichloropropane	7.428	63	164505	113.217	ug/l	99
46) Dibromomethane	7.574	93	125846	109.869	ug/l	100
47) Bromodichloromethane	7.818	83	258326	116.365	ug/l	98
48) Methyl methacrylate	7.690	41	231040	124.356	ug/l	99
49) 1,4-Dioxane	7.665	88	82800	2241.695	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	1371275	577.831	ug/l	100
52) Toluene	8.714	92	393323	113.430	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	257233	138.001	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	271054	124.658	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	154932	110.551	ug/l	97
56) Ethyl methacrylate	9.116	69	277921	127.968	ug/l	99
57) 1,3-Dichloropropane	9.305	76	272814	109.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	680608	697.344	ug/l	100
59) 2-Hexanone	9.433	43	1028597	571.094	ug/l	99
60) Dibromochloromethane	9.519	129	187497	122.915	ug/l	98
61) 1,2-Dibromoethane	9.604	107	165634	115.087	ug/l	97
64) Tetrachloroethene	9.269	164	132488	100.918	ug/l	96
65) Chlorobenzene	10.080	112	429656	108.615	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	152173	115.333	ug/l	99
67) Ethyl Benzene	10.189	91	785121	116.421	ug/l	99
68) m/p-Xylenes	10.299	106	570913	234.349	ug/l	100
69) o-Xylene	10.640	106	279906	113.990	ug/l	99
70) Styrene	10.653	104	474254	120.694	ug/l	99
71) Bromoform	10.799	173	125986	127.128	ug/l #	99
73) Isopropylbenzene	10.957	105	752337	114.312	ug/l	99
74) N-amyl acetate	10.842	43	396877	120.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	243532	104.018	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	284571	113.113	ug/l	85
77) Bromobenzene	11.195	156	167934	110.297	ug/l	99
78) n-propylbenzene	11.305	91	881218	119.448	ug/l	100
79) 2-Chlorotoluene	11.360	91	534560	107.514	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	616477	112.825	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	81372	134.609	ug/l	93
82) 4-Chlorotoluene	11.451	91	611736	111.948	ug/l	100
83) tert-Butylbenzene	11.713	119	617473	115.065	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	623456	114.236	ug/l	99
85) sec-Butylbenzene	11.890	105	786293	118.023	ug/l	99
86) p-Isopropyltoluene	12.006	119	651531	121.113	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	313100	110.394	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	311752	112.167	ug/l	100
89) n-Butylbenzene	12.329	91	605783	129.533	ug/l	100
90) Hexachloroethane	12.536	117	123188	123.091	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	308039	111.506	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64438	120.915	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	203255	134.366	ug/l	98
94) Hexachlorobutadiene	13.719	225	80558	113.857	ug/l	99
95) Naphthalene	13.774	128	721236	130.929	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	200379	126.331	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:168 Resp: 82963
Ion Ratio Lower Upper
168 100
99 67.6 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 118.929 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 85 Resp: 217793
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1









#3
Chloromethane
Concen: 109.161 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 50 Resp: 196843
Ion Ratio Lower Upper
50 100
52 32.5 25.4 38.2



#4
Vinyl Chloride
Concen: 118.953 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 187924
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8









#5
Bromomethane
Concen: 103.625 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 94 Resp: 83008
Ion Ratio Lower Upper
94 100
96 95.8 72.8 109.2



#6
Chloroethane
Concen: 96.822 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 64 Resp: 78806
Ion Ratio Lower Upper
64 100
66 31.3 24.3 36.5









#7
Trichlorofluoromethane
Concen: 101.851 ug/l
RT: 1.868 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 245108
Ion Ratio Lower Upper
101 100
103 66.4 51.0 76.4



#8
Diethyl Ether
Concen: 111.791 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 74 Resp: 88441
Ion Ratio Lower Upper
74 100
45 104.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 112.243 ug/l
RT: 2.313 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 161404
Ion Ratio Lower Upper
101 100
85 46.8 38.6 58.0
151 69.8 55.2 82.8



#10
Methyl Iodide
Concen: 112.441 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:142 Resp: 189843
Ion Ratio Lower Upper
142 100
127 52.4 41.7 62.5
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 616.119 ug/l
RT: 3.002 min Scan# 31
Delta R.T. 0.024 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 59 Resp: 181238
Ion Ratio Lower Upper
59 100
57 10.8 8.6 12.8



#12
1,1-Dichloroethene
Concen: 111.039 ug/l
RT: 2.307 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 96 Resp: 155451
Ion Ratio Lower Upper
96 100
61 184.5 146.2 219.2
98 64.3 51.0 76.6









#13
Acrolein
Concen: 594.872 ug/l
RT: 2.239 min Scan# 188
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 203257
Ion Ratio Lower Upper
56 100
55 72.1 56.2 84.4



#14
Allyl chloride
Concen: 112.561 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 41 Resp: 297776
Ion Ratio Lower Upper
41 100
39 73.7 60.6 90.8
76 31.8 24.9 37.3









#15
Acrylonitrile
Concen: 550.973 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 53 Resp: 484854
Ion Ratio Lower Upper
53 100
52 84.2 65.3 97.9
51 37.5 29.8 44.8



#16
Acetone
Concen: 545.754 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 460820
Ion Ratio Lower Upper
43 100
58 27.3 21.2 31.8









#17
Carbon Disulfide
Concen: 122.499 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 76 Resp: 397353
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4



#18
Methyl Acetate
Concen: 108.099 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 217885
Ion Ratio Lower Upper
43 100
74 21.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 114.112 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 73 Resp: 557210
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5



#20
Methylene Chloride
Concen: 100.579 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 84 Resp: 172035
Ion Ratio Lower Upper
84 100
49 136.0 113.9 170.9
51 41.5 33.5 50.3
86 63.7 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 108.241 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 154694
Ion Ratio Lower Upper
96 100
61 154.6 119.5 179.3
98 64.3 50.0 75.0



#22
Diisopropyl ether
Concen: 116.858 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 45 Resp: 577610
Ion Ratio Lower Upper
45 100
43 70.3 66.6 100.0
87 24.7 19.8 29.6
59 10.6 8.6 12.8









#23
 Vinyl Acetate
 Concen: 608.005 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 43 Resp: 2655751
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 110.182 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 63 Resp: 320083
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.6 10.8
 100 4.7 2.1 6.3









#25
2-Butanone
Concen: 581.292 ug/l
RT: 4.562 min Scan# 51
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 708463
Ion Ratio Lower Upper
43 100
72 24.0 18.4 27.6



#26
2,2-Dichloropropane
Concen: 117.548 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 77 Resp: 258698
Ion Ratio Lower Upper
77 100
97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 108.994 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 96 Resp: 187835

Ion Ratio Lower Upper

96 100

61 159.8 0.0 322.8

98 64.8 0.0 129.0



#28

Bromochloromethane

Concen: 95.108 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 49 Resp: 146923


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 569.081 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 42 Resp: 450456

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5

71 38.7 31.4 47.2



#30

Chloroform

Concen: 107.208 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 83 Resp: 323659

Ion Ratio Lower Upper

83 100

85 65.1 49.3 73.9









#31
Cyclohexane
Concen: 116.934 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 286312
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.8 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 113.572 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 97 Resp: 295702
Ion Ratio Lower Upper
97 100
99 64.3 51.8 77.6
61 48.0 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 93.983 ug/l
 RT: 5.952 min Scan# 798
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 65 Resp: 231952
 Ion Ratio Lower Upper
 65 100
 67 49.4 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 114 Resp: 144975
 Ion Ratio Lower Upper
 114 100
 63 23.9 0.0 49.2
 88 18.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 98.562 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:113 Resp: 160168

Ion Ratio Lower Upper

113 100

111 103.0 83.1 124.7

192 16.9 13.3 19.9



#36

1,1-Dichloropropene

Concen: 116.462 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 219731


Ion Ratio Lower Upper

75 100

110 32.9 16.3 48.9

77 30.8 24.3 36.5









#37
Ethyl Acetate
Concen: 115.889 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 274312
Ion Ratio Lower Upper
43 100
61 12.5 10.3 15.5
70 10.2 7.9 11.9



#38
Carbon Tetrachloride
Concen: 115.760 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 117 Resp: 250916
Ion Ratio Lower Upper
117 100
119 94.3 75.2 112.8
121 29.7 24.2 36.4





#39

Methylcyclohexane

Concen: 121.107 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 83 Resp: 286311

Ion Ratio Lower Upper

83 100

55 83.8 64.7 97.1

98 47.8 37.4 56.2



#40

Benzene

Concen: 109.813 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 78 Resp: 642178

Ion Ratio Lower Upper

78 100

77 23.4 19.0 28.4









#41
Methacrylonitrile
Concen: 115.139 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 41 Resp: 151424




Ion	Ratio	Lower	Upper
41	100		
39	60.0	47.2	70.8
67	65.9	50.7	76.1
52	30.9	24.6	37.0




#42
1,2-Dichloroethane
Concen: 112.492 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 271992

Ion	Ratio	Lower	Upper
62	100		
98	7.4	0.0	15.2









#43
Isopropyl Acetate
Concen: 123.780 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 447871
Ion Ratio Lower Upper
43 100
61 18.1 14.3 21.5
87 11.7 9.5 14.3



#44
Trichloroethene
Concen: 114.032 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:130 Resp: 157641
Ion Ratio Lower Upper
130 100
95 105.9 0.0 204.2









#45
1,2-Dichloropropane
Concen: 113.217 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 63 Resp: 164505
Ion Ratio Lower Upper
63 100
65 30.8 24.1 36.1



#46
Dibromomethane
Concen: 109.869 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 93 Resp: 125846
Ion Ratio Lower Upper
93 100
95 82.4 65.6 98.4
174 85.0 68.2 102.2









#47
 Bromodichloromethane
 Concen: 116.365 ug/l
 RT: 7.818 min Scan# 1104
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 83 Resp: 258326
 Ion Ratio Lower Upper
 83 100
 85 64.8 50.5 75.7
 127 8.0 6.5 9.7



#48
 Methyl methacrylate
 Concen: 124.356 ug/l
 RT: 7.690 min Scan# 1083
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 41 Resp: 231040
 Ion Ratio Lower Upper
 41 100
 69 73.0 58.5 87.7
 39 63.8 51.7 77.5





#49

1,4-Dioxane

Concen: 2241.695 ug/l

RT: 7.665 min Scan# 1079

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 88 Resp: 82800

Ion Ratio Lower Upper

88 100

43 40.0 33.4 50.2

58 74.8 58.6 88.0



#50

Toluene-d8

Concen: 101.252 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 98 Resp: 554390

Ion Ratio Lower Upper

98 100

100 65.4 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 577.831 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1371275
Ion Ratio Lower Upper
43 100
58 36.2 28.9 43.3



#52
Toluene
Concen: 113.430 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 92 Resp: 393323
Ion Ratio Lower Upper
92 100
91 171.0 136.6 205.0









#53
t-1,3-Dichloropropene
Concen: 138.001 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 257233
Ion Ratio Lower Upper
75 100
77 31.7 25.0 37.6



#54
cis-1,3-Dichloropropene
Concen: 124.658 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 271054
Ion Ratio Lower Upper
75 100
77 31.4 25.4 38.0
39 60.0 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 110.551 ug/l
 RT: 9.147 min Scan# 1322
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150

Tgt Ion: 97 Resp: 154932




Ion	Ratio	Lower	Upper
97	100		
83	88.5	69.0	103.6
85	56.0	44.3	66.5
99	64.5	47.8	71.6




#56
 Ethyl methacrylate
 Concen: 127.968 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 69 Resp: 277921

Ion	Ratio	Lower	Upper
69	100		
41	75.4	60.8	91.2
39	47.2	39.0	58.6





#57

1,3-Dichloropropane

Concen: 109.813 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 272814

Ion Ratio Lower Upper

76 100

78 32.3 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 697.344 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 63 Resp: 680608

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3









#59
2-Hexanone
Concen: 571.094 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1028597
Ion Ratio Lower Upper
43 100
58 50.2 24.9 74.6



#60
Dibromochloromethane
Concen: 122.915 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 129 Resp: 187497
Ion Ratio Lower Upper
129 100
127 77.1 39.3 117.8









#61
1,2-Dibromoethane
Concen: 115.087 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:107 Resp: 165634
Ion Ratio Lower Upper
107 100
109 94.6 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 109.174 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 95 Resp: 217536
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 64.4 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 128557
Ion Ratio Lower Upper
117 100
82 61.4 50.6 76.0
119 31.9 25.8 38.6



#64
Tetrachloroethene
Concen: 100.918 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:164 Resp: 132488
Ion Ratio Lower Upper
164 100
166 123.3 105.0 157.6
129 101.6 83.5 125.3
131 97.3 76.5 114.7









#65
Chlorobenzene
Concen: 108.615 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:112 Resp: 429656
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 115.333 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:131 Resp: 152173
Ion Ratio Lower Upper
131 100
133 95.7 47.3 141.9
119 64.0 31.6 95.0





#67

Ethyl Benzene

Concen: 116.421 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 91 Resp: 785121

Ion Ratio Lower Upper

91 100

106 29.9 23.4 35.2



#68

m/p-Xylenes

Concen: 234.349 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:106 Resp: 570913

Ion Ratio Lower Upper

106 100

91 214.2 171.2 256.8









#69
o-Xylene
Concen: 113.990 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:106 Resp: 279906
Ion Ratio Lower Upper
106 100
91 226.9 112.7 338.1



#70
Styrene
Concen: 120.694 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:104 Resp: 474254
Ion Ratio Lower Upper
104 100
78 57.2 45.7 68.5
103 55.5 43.7 65.5





#71

Bromoform

Concen: 127.128 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:173 Resp: 125986

Ion Ratio Lower Upper

173 100

175 47.3 23.4 70.0

254 0.1 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:152 Resp: 60345


Ion Ratio Lower Upper

152 100

115 150.7 46.9 140.7#

150 211.6 0.0 351.0





#73

Isopropylbenzene

Concen: 114.312 ug/l

RT: 10.957 min Scan# 1619

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion:105 Resp: 752337

Ion Ratio Lower Upper

105 100

120 25.3 12.8 38.4



#74

N-ethyl acetate

Concen: 120.944 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 396877




Ion Ratio Lower Upper


43 100

70 38.5 30.9 46.3

55 22.5 18.7 28.1

61 22.5 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 104.018 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 243532
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 64.4 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 113.113 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 284571
Ion Ratio Lower Upper
75 100
77 31.8 20.5 61.5









#77
Bromobenzene
Concen: 110.297 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:156 Resp: 167934
Ion Ratio Lower Upper
156 100
77 187.1 92.8 278.3
158 97.3 48.8 146.3



#78
n-propylbenzene
Concen: 119.448 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 881218
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 107.514 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 91 Resp: 534560
Ion Ratio Lower Upper
91 100
126 31.0 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 112.825 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 105 Resp: 616477
Ion Ratio Lower Upper
105 100
120 46.0 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 134.609 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 81372
Ion Ratio Lower Upper
75 100
53 114.9 100.5 150.7
89 46.5 37.0 55.6



#82
4-Chlorotoluene
Concen: 111.948 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 611736
Ion Ratio Lower Upper
91 100
126 26.7 13.3 39.8





#83

tert-Butylbenzene

Concen: 115.065 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:119 Resp: 617473

Ion Ratio Lower Upper

119 100

91 65.6 32.9 98.7

134 22.6 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 114.236 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 623456

Ion Ratio Lower Upper

105 100

120 43.1 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 118.023 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 786293
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 121.113 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion:119 Resp: 651531
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.5 37.5
 91 27.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 110.394 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:146 Resp: 313100

Ion Ratio Lower Upper

146 100

111 44.5 22.1 66.3

148 64.0 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 112.167 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:146 Resp: 311752


Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.9

148 63.9 31.9 95.5





#89

n-Butylbenzene

Concen: 129.533 ug/l

RT: 12.329 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150





Tgt Ion: 91 Resp: 605783

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7

134 24.0 11.8 35.3



#90

Hexachloroethane

Concen: 123.091 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 117 Resp: 123188

Ion Ratio Lower Upper

117 100

201 63.2 31.6 94.7









#91
 1,2-Dichlorobenzene
 Concen: 111.506 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:146 Resp: 308039
 Ion Ratio Lower Upper
 146 100
 111 46.3 22.4 67.3
 148 63.6 31.6 94.8



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 120.915 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 64438
 Ion Ratio Lower Upper
 75 100
 155 69.6 34.9 104.8
 157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 134.366 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:180 Resp: 203255

Ion Ratio Lower Upper

180 100

182 94.1 47.9 143.8

145 34.5 17.1 51.2



#94

Hexachlorobutadiene

Concen: 113.857 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:225 Resp: 80558


Ion Ratio Lower Upper

225 100

223 62.5 30.8 92.4

227 63.1 30.9 92.7





#95

Naphthalene

Concen: 130.929 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:128 Resp: 721236

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6

129 11.0 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 126.331 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:180 Resp: 200379

Ion Ratio Lower Upper

180 100

182 96.5 47.8 143.3

145 36.7 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	168484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	9067	3.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	6.280%#
35) Dibromofluoromethane	5.373	113	5969	3.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	6.320%#
50) Toluene-d8	8.647	98	20566	3.232	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	6.460%#
62) 4-Bromofluorobenzene	11.079	95	7822	3.378	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	6.760%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	6195	2.894	ug/l	95
3) Chloromethane	1.307	50	6587	3.125	ug/l	99
4) Vinyl Chloride	1.374	62	5999	3.249	ug/l	96
5) Bromomethane	1.593	94	2962	3.164	ug/l	88
6) Chloroethane	1.666	64	3567	3.750	ug/l	94
7) Trichlorofluoromethane	1.874	101	9599	3.413	ug/l	88
8) Diethyl Ether	2.130	74	3020	3.266	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	5911	3.517	ug/l	99
10) Methyl Iodide	2.441	142	5900	2.990	ug/l	99
11) Tert butyl alcohol	2.965	59	5539	16.111	ug/l	98
12) 1,1-Dichloroethene	2.306	96	5496	3.359	ug/l	97
13) Acrolein	2.233	56	5673	14.206	ug/l	95
14) Allyl chloride	2.654	41	10262	3.319	ug/l	98
15) Acrylonitrile	3.062	53	16746	16.282	ug/l	97
16) Acetone	2.380	43	19773	20.036	ug/l	98
17) Carbon Disulfide	2.501	76	11068	2.919	ug/l	96
18) Methyl Acetate	2.703	43	7909	3.357	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	18497	3.241	ug/l	97
20) Methylene Chloride	2.782	84	6680	3.341	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	5402	3.234	ug/l	100
22) Diisopropyl ether	3.751	45	18657	3.230	ug/l #	65
23) Vinyl Acetate	3.715	43	82328	16.127	ug/l	99
24) 1,1-Dichloroethane	3.605	63	11193	3.297	ug/l	99
25) 2-Butanone	4.562	43	26125	18.340	ug/l	98
26) 2,2-Dichloropropane	4.465	77	8245	3.205	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	6222	3.089	ug/l	95
28) Bromochloromethane	4.891	49	5360	2.969	ug/l #	100
29) Tetrahydrofuran	5.001	42	15431	16.680	ug/l	98
30) Chloroform	5.080	83	11624	3.294	ug/l	89
31) Cyclohexane	5.458	56	10269	3.588	ug/l	96
32) 1,1,1-Trichloroethane	5.367	97	9827	3.229	ug/l	97
36) 1,1-Dichloropropene	5.678	75	7787	3.551	ug/l	98
37) Ethyl Acetate	4.721	43	9021	3.279	ug/l	98
38) Carbon Tetrachloride	5.659	117	8505	3.376	ug/l	97
39) Methylcyclohexane	7.373	83	9882	3.597	ug/l	93
40) Benzene	6.031	78	22528	3.315	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	4851	3.174	ug/l #	65
42) 1,2-Dichloroethane	6.080	62	10011	3.563	ug/l	100
43) Isopropyl Acetate	6.348	43	13914	3.309	ug/l	96
44) Trichloroethene	7.123	130	5313	3.307	ug/l	94
45) 1,2-Dichloropropane	7.427	63	5451	3.228	ug/l #	89
46) Dibromomethane	7.580	93	4417	3.318	ug/l	98
47) Bromodichloromethane	7.818	83	8394	3.254	ug/l	98
48) Methyl methacrylate	7.696	41	7171	3.321	ug/l	97
49) 1,4-Dioxane	7.659	88	2907	67.721	ug/l	96
51) 4-Methyl-2-Pentanone	8.567	43	46794	16.967	ug/l	98
52) Toluene	8.714	92	14126	3.505	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	6834	3.155	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	7908	3.129	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	5685	3.490	ug/l	98
56) Ethyl methacrylate	9.116	69	8555	3.390	ug/l	98
57) 1,3-Dichloropropane	9.305	76	10127	3.508	ug/l	94
58) 2-Chloroethyl Vinyl ether	8.238	63	20809	18.346	ug/l	99
59) 2-Hexanone	9.427	43	34853	16.651	ug/l	97
60) Dibromochloromethane	9.518	129	5500	3.102	ug/l	98
61) 1,2-Dibromoethane	9.604	107	5614	3.356	ug/l	99
64) Tetrachloroethene	9.269	164	4752	3.162	ug/l	96
65) Chlorobenzene	10.079	112	15391	3.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	5024	3.326	ug/l	99
67) Ethyl Benzene	10.195	91	26732	3.463	ug/l	99
68) m/p-Xylenes	10.299	106	19957	7.156	ug/l	98
69) o-Xylene	10.640	106	9398	3.343	ug/l	95
70) Styrene	10.652	104	14900	3.312	ug/l	99
71) Bromoform	10.799	173	3470	3.059	ug/l #	99
73) Isopropylbenzene	10.957	105	24201	3.266	ug/l	98
74) N-amyl acetate	10.841	43	11224	3.038	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	9170	3.479	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	8548	3.018	ug/l	95
77) Bromobenzene	11.195	156	5857	3.417	ug/l	96
78) n-propylbenzene	11.299	91	28436	3.424	ug/l	100
79) 2-Chlorotoluene	11.360	91	18672	3.336	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	20742	3.372	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	1828	2.686	ug/l	99
82) 4-Chlorotoluene	11.451	91	19966	3.245	ug/l	98
83) tert-Butylbenzene	11.713	119	21049	3.484	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	20613	3.355	ug/l	97
85) sec-Butylbenzene	11.890	105	25590	3.412	ug/l	99
86) p-Isopropyltoluene	12.006	119	20955	3.460	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10585	3.315	ug/l	97
88) 1,4-Dichlorobenzene	12.036	146	10908	3.486	ug/l	90
89) n-Butylbenzene	12.329	91	18004	3.419	ug/l	94
90) Hexachloroethane	12.536	117	3553	3.153	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	10712	3.444	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1687	2.812	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	5720	3.359	ug/l	96
94) Hexachlorobutadiene	13.719	225	2816	3.535	ug/l	96
95) Naphthalene	13.774	128	19902	3.209	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	5745	3.217	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On : 05 May 2025 16:04
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion:168 Resp: 96964
Ion Ratio Lower Upper
168 100
99 74.0 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 2.894 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 85 Resp: 6195
Ion Ratio Lower Upper
85 100
87 28.7 15.7 47.1









#3
Chloromethane
Concen: 3.125 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 50 Resp: 6587
Ion Ratio Lower Upper
50 100
52 32.2 25.4 38.2



#4
Vinyl Chloride
Concen: 3.249 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 62 Resp: 5999
Ion Ratio Lower Upper
62 100
64 29.1 25.2 37.8









#5
 Bromomethane
 Concen: 3.164 ug/l
 RT: 1.593 min Scan# 83
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 94 Resp: 2962
 Ion Ratio Lower Upper
 94 100
 96 102.2 72.8 109.2



#6
 Chloroethane
 Concen: 3.750 ug/l
 RT: 1.666 min Scan# 95
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 64 Resp: 3567
 Ion Ratio Lower Upper
 64 100
 66 33.7 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 3.413 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:101 Resp: 9599
 Ion Ratio Lower Upper
 101 100
 103 54.0 51.0 76.4



#8
 Diethyl Ether
 Concen: 3.266 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 74 Resp: 3020
 Ion Ratio Lower Upper
 74 100
 45 115.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 3.517 ug/l
RT: 2.325 min Scan# 20
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 5911
Ion Ratio Lower Upper
101 100
85 46.3 38.6 58.0
151 68.8 55.2 82.8



#10
Methyl Iodide
Concen: 2.990 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:142 Resp: 5900
Ion Ratio Lower Upper
142 100
127 52.8 41.7 62.5
141 14.3 11.5 17.3









#11
Tert butyl alcohol
Concen: 16.111 ug/l
RT: 2.965 min Scan# 30
Delta R.T. -0.012 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 59 Resp: 5539
Ion Ratio Lower Upper
59 100
57 10.0 8.6 12.8



#12
1,1-Dichloroethene
Concen: 3.359 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 96 Resp: 5496
Ion Ratio Lower Upper
96 100
61 178.7 146.2 219.2
98 59.5 51.0 76.6





#13

Acrolein

Concen: 14.206 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 56 Resp: 5673

Ion Ratio Lower Upper

56 100

55 74.5 56.2 84.4



#14

Allyl chloride

Concen: 3.319 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 41 Resp: 10262


Ion Ratio Lower Upper

41 100

39 74.1 60.6 90.8

76 29.1 24.9 37.3









#15
Acrylonitrile
Concen: 16.282 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion:	53	Resp:	16746
Ion	Ratio	Lower	Upper
53	100		
52	84.3	65.3	97.9
51	38.5	29.8	44.8



#16
Acetone
Concen: 20.036 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:	43	Resp:	19773
Ion	Ratio	Lower	Upper
43	100		
58	27.7	21.2	31.8





#17

Carbon Disulfide

Concen: 2.919 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 76 Resp: 11068

Ion Ratio Lower Upper

76 100

78 10.2 7.0 10.4



#18

Methyl Acetate

Concen: 3.357 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 43 Resp: 7909

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 3.241 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 18497
Ion Ratio Lower Upper
73 100
57 20.5 17.7 26.5



#20
Methylene Chloride
Concen: 3.341 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 84 Resp: 6680
Ion Ratio Lower Upper
84 100
49 137.0 113.9 170.9
51 43.8 33.5 50.3
86 67.6 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 3.234 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 96 Resp: 5402
Ion Ratio Lower Upper
96 100
61 149.9 119.5 179.3
98 62.9 50.0 75.0



#22
Diisopropyl ether
Concen: 3.230 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 45 Resp: 18657
Ion Ratio Lower Upper
45 100
43 128.0 66.6 100.0#
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
 Vinyl Acetate
 Concen: 16.127 ug/l
 RT: 3.715 min Scan# 411
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 82328
 Ion Ratio Lower Upper
 43 100
 86 9.1 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 3.297 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 63 Resp: 11193
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.6 10.8
 100 4.7 2.1 6.3





#25

2-Butanone

Concen: 18.340 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 43 Resp: 26125

Ion Ratio Lower Upper

43 100

72 21.8 18.4 27.6



#26

2,2-Dichloropropane

Concen: 3.205 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 77 Resp: 8245

Ion Ratio Lower Upper

77 100

97 23.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 3.089 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 96 Resp: 6222

Ion Ratio Lower Upper

96 100

61 168.5 0.0 322.8

98 67.6 0.0 129.0



#28

Bromochloromethane

Concen: 2.969 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 49 Resp: 5360


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 16.680 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 42 Resp: 15431

Ion Ratio Lower Upper

42 100

72 42.2 33.7 50.5

71 37.4 31.4 47.2



#30

Chloroform

Concen: 3.294 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 83 Resp: 11624

Ion Ratio Lower Upper

83 100

85 69.9 49.3 73.9









#31
Cyclohexane
Concen: 3.588 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 56 Resp: 10269
Ion Ratio Lower Upper
56 100
69 31.6 24.4 36.6
84 79.1 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 3.229 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 97 Resp: 9827
Ion Ratio Lower Upper
97 100
99 62.6 51.8 77.6
61 51.8 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 3.143 ug/l
 RT: 5.946 min Scan# 798
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 65 Resp: 9067
 Ion Ratio Lower Upper
 65 100
 67 50.1 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.751 min Scan# 929
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 114 Resp: 168484
 Ion Ratio Lower Upper
 114 100
 63 24.5 0.0 49.2
 88 17.7 0.0 33.6









#35
Dibromofluoromethane
Concen: 3.161 ug/l
RT: 5.373 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:113 Resp: 5969
Ion Ratio Lower Upper
113 100
111 104.4 83.1 124.7
192 17.5 13.3 19.9



#36
1,1-Dichloropropene
Concen: 3.551 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 75 Resp: 7787
Ion Ratio Lower Upper
75 100
110 30.6 16.3 48.9
77 30.5 24.3 36.5





#37

Ethyl Acetate

Concen: 3.279 ug/l

RT: 4.721 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 9021

Ion Ratio Lower Upper

43 100

61 11.9 10.3 15.5

70 9.2 7.9 11.9



#38

Carbon Tetrachloride

Concen: 3.376 ug/l

RT: 5.659 min Scan# 750

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 117 Resp: 8505


Ion Ratio Lower Upper

117 100

119 90.6 75.2 112.8

121 29.5 24.2 36.4









#39
Methylcyclohexane
Concen: 3.597 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 83	Resp: 9882
Ion Ratio	Lower Upper
83	100
55	73.3 64.7 97.1
98	50.3 37.4 56.2



#40
Benzene
Concen: 3.315 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 78	Resp: 22528
Ion Ratio	Lower Upper
78	100
77	23.9 19.0 28.4









#41
Methacrylonitrile
Concen: 3.174 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion:	41	Resp:	4851
Ion	Ratio	Lower	Upper
41	100		
39	59.7	47.2	70.8
67	0.0	50.7	76.1#
52	28.8	24.6	37.0



#42
1,2-Dichloroethane
Concen: 3.563 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:	62	Resp:	10011
Ion	Ratio	Lower	Upper
62	100		
98	7.8	0.0	15.2





#43

Isopropyl Acetate

Concen: 3.309 ug/l

RT: 6.348 min Scan# 863

Delta R.T. 0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 13914

Ion Ratio Lower Upper

43 100

61 16.2 14.3 21.5

87 10.6 9.5 14.3



#44

Trichloroethene

Concen: 3.307 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 130 Resp: 5313

Ion Ratio Lower Upper

130 100

95 108.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 3.228 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 63 Resp: 5451

Ion Ratio Lower Upper

63 100

65 36.2 24.1 36.1#



#46

Dibromomethane

Concen: 3.318 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 93 Resp: 4417


Ion Ratio Lower Upper

93 100

95 80.1 65.6 98.4

174 87.6 68.2 102.2









#47
Bromodichloromethane
Concen: 3.254 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 83	Resp:	8394
Ion Ratio	Lower	Upper
83	100	
85	61.9	50.5 75.7
127	9.0	6.5 9.7



#48
Methyl methacrylate
Concen: 3.321 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 41	Resp:	7171
Ion Ratio	Lower	Upper
41	100	
69	68.4	58.5 87.7
39	65.1	51.7 77.5





#49

1,4-Dioxane

Concen: 67.721 ug/l

RT: 7.659 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 88 Resp: 2907

Ion Ratio Lower Upper

88 100

43 45.3 33.4 50.2

58 71.0 58.6 88.0



#50

Toluene-d8

Concen: 3.232 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 98 Resp: 20566

Ion Ratio Lower Upper

98 100

100 66.1 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 16.967 ug/l
RT: 8.567 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 46794
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3



#52
Toluene
Concen: 3.505 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 92 Resp: 14126
Ion Ratio Lower Upper
92 100
91 167.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 3.155 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 75 Resp: 6834

Ion Ratio Lower Upper

75 100

77 31.1 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 3.129 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 7908


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 55.3 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 3.490 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:	97	Resp:	5685
Ion	Ratio	Lower	Upper
97	100		
83	88.0	69.0	103.6
85	54.4	44.3	66.5
99	62.2	47.8	71.6



#56
Ethyl methacrylate
Concen: 3.390 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:	69	Resp:	8555
Ion	Ratio	Lower	Upper
69	100		
41	73.7	60.8	91.2
39	49.3	39.0	58.6





#57

1,3-Dichloropropane

Concen: 3.508 ug/l

RT: 9.305 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 76 Resp: 10127

Ion Ratio Lower Upper

76 100

78 29.5 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 18.346 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 63 Resp: 20809

Ion Ratio Lower Upper

63 100

106 26.4 21.5 32.3





#59

2-Hexanone

Concen: 16.651 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 43 Resp: 34853

Ion Ratio Lower Upper

43 100

58 47.5 24.9 74.6



#60

Dibromochloromethane

Concen: 3.102 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 129 Resp: 5500

Ion Ratio Lower Upper

129 100

127 80.4 39.3 117.8









#61
1,2-Dibromoethane
Concen: 3.356 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:107 Resp: 5614
Ion Ratio Lower Upper
107 100
109 91.0 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 3.378 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 95 Resp: 7822
Ion Ratio Lower Upper
95 100
174 65.0 0.0 135.8
176 66.6 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 147167
Ion Ratio Lower Upper
117 100
82 61.1 50.6 76.0
119 32.1 25.8 38.6



#64
Tetrachloroethene
Concen: 3.162 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:164 Resp: 4752
Ion Ratio Lower Upper
164 100
166 127.8 105.0 157.6
129 105.8 83.5 125.3
131 104.1 76.5 114.7









#65
Chlorobenzene
Concen: 3.399 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:112 Resp: 15391
Ion Ratio Lower Upper
112 100
114 32.6 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 3.326 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:131 Resp: 5024
Ion Ratio Lower Upper
131 100
133 95.0 47.3 141.9
119 64.8 31.6 95.0









#67
Ethyl Benzene
Concen: 3.463 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 91 Resp: 26732
Ion Ratio Lower Upper
91 100
106 29.9 23.4 35.2



#68
m/p-Xylenes
Concen: 7.156 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 106 Resp: 19957
Ion Ratio Lower Upper
106 100
91 210.1 171.2 256.8









#69
o-Xylene
Concen: 3.343 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 9398
Ion Ratio Lower Upper
106 100
91 233.6 112.7 338.1



#70
Styrene
Concen: 3.312 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:104 Resp: 14900
Ion Ratio Lower Upper
104 100
78 57.9 45.7 68.5
103 55.6 43.7 65.5





#71

Bromoform

Concen: 3.059 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion:173 Resp: 3470

Ion Ratio Lower Upper

173 100

175 45.8 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:152 Resp: 67939


Ion Ratio Lower Upper

152 100

115 66.8 46.9 140.7

150 157.5 0.0 351.0









#73
Isopropylbenzene
Concen: 3.266 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:105 Resp: 24201
Ion Ratio Lower Upper
105 100
120 26.4 12.8 38.4



#74
N-ethyl acetate
Concen: 3.038 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 43 Resp: 11224
Ion Ratio Lower Upper
43 100
70 35.4 30.9 46.3
55 25.0 18.7 28.1
61 19.6 17.1 25.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 3.479 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 83 Resp: 9170
 Ion Ratio Lower Upper
 83 100
 131 8.3 5.0 14.9
 85 61.7 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 3.018 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 75 Resp: 8548
 Ion Ratio Lower Upper
 75 100
 77 37.7 20.5 61.5









#77
Bromobenzene
Concen: 3.417 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:156 Resp: 5857
Ion Ratio Lower Upper
156 100
77 177.0 92.8 278.3
158 97.7 48.8 146.3



#78
n-propylbenzene
Concen: 3.424 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 28436
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 3.336 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 18672
Ion Ratio Lower Upper
91 100
126 29.7 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 3.372 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 105 Resp: 20742
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2









#81
 trans-1,4-Dichloro-2-butene
 Concen: 2.686 ug/l
 RT: 11.018 min Scan# 1629
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 75 Resp: 1828
 Ion Ratio Lower Upper
 75 100
 53 125.8 100.5 150.7
 89 47.9 37.0 55.6



#82
 4-Chlorotoluene
 Concen: 3.245 ug/l
 RT: 11.451 min Scan# 1700
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 19966
 Ion Ratio Lower Upper
 91 100
 126 27.3 13.3 39.8





#83

tert-Butylbenzene

Concen: 3.484 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion:119 Resp: 21049

Ion Ratio Lower Upper

119 100

91 67.0 32.9 98.7

134 23.8 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 3.355 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20613

Ion Ratio Lower Upper

105 100

120 44.5 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 3.412 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 25590
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 3.460 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion:119 Resp: 20955
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.5 37.5
 91 28.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 3.315 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:146 Resp: 10585

Ion Ratio Lower Upper

146 100

111 46.2 22.1 66.3

148 62.4 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 3.486 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:146 Resp: 10908


Ion Ratio Lower Upper

146 100

111 52.1 21.3 63.9

148 69.4 31.9 95.5





#89

n-Butylbenzene

Concen: 3.419 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 91 Resp: 18004

Ion Ratio Lower Upper

91 100

92 49.5 26.9 80.7

134 21.2 11.8 35.3



#90

Hexachloroethane

Concen: 3.153 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 117 Resp: 3553

Ion Ratio Lower Upper

117 100

201 58.5 31.6 94.7





#91

1,2-Dichlorobenzene

Concen: 3.444 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:146 Resp: 10712

Ion Ratio Lower Upper

146 100

111 44.7 22.4 67.3

148 64.2 31.6 94.8



#92

1,2-Dibromo-3-Chloropropane

Concen: 2.812 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 1687


Ion Ratio Lower Upper

75 100

155 71.7 34.9 104.8

157 87.0 43.8 131.4









#93
 1,2,4-Trichlorobenzene
 Concen: 3.359 ug/l
 RT: 13.585 min Scan# 2050
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:180 Resp: 5720
 Ion Ratio Lower Upper
 180 100
 182 91.6 47.9 143.8
 145 35.7 17.1 51.2



#94
 Hexachlorobutadiene
 Concen: 3.535 ug/l
 RT: 13.719 min Scan# 2072
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion:225 Resp: 2816
 Ion Ratio Lower Upper
 225 100
 223 65.1 30.8 92.4
 227 58.8 30.9 92.7





#95

Naphthalene

Concen: 3.209 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion:128 Resp: 19902

Ion Ratio Lower Upper

128 100

127 13.0 10.4 15.6

129 11.5 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 3.217 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:180 Resp: 5745

Ion Ratio Lower Upper

180 100

182 99.1 47.8 143.3

145 34.8 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	2188	0.799	ug/l	0.02
Spiked Amount	50.000	Range	74 - 125	Recovery	=	1.600%#
35) Dibromofluoromethane	5.391	113	1350	0.721	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	1.440%#
50) Toluene-d8	8.647	98	3964	0.628	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	1.260%#
62) 4-Bromofluorobenzene	11.085	95	1250	0.545	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	1.080%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
5) Bromomethane	1.593	94	770	0.866	ug/l	88
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
10) Methyl Iodide	2.441	142	1297	0.692	ug/l #	87
11) Tert butyl alcohol	2.965	59	1176	3.604	ug/l #	78
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
13) Acrolein	2.233	56	1374	3.625	ug/l	83
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	682	0.398	ug/l #	63
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
31) Cyclohexane	5.458	56	2072	0.763	ug/l	88
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1447	0.531	ug/l #	72
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1033	0.483	ug/l #	71
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1858	0.754	ug/l	77
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.024	75	252	0.425	ug/l #	71
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	11.969	146	1914	0.703	ug/l	98
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046047.D
Acq On : 05 May 2025 16:27
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Quant Time: May 06 06:13:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:13:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :




MSVOA_X

ClientSampleId :

VSTDIC001







Tgt Ion:168 Resp: 92040
Ion Ratio Lower Upper
168 100
99 71.5 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 0.596 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 85 Resp: 1211
Ion Ratio Lower Upper
85 100
87 32.4 15.7 47.1









#3
Chloromethane
Concen: 0.639 ug/l
RT: 1.307 min Scan# 36
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 50 Resp: 1278
Ion Ratio Lower Upper
50 100
52 30.3 25.4 38.2



#4
Vinyl Chloride
Concen: 0.707 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 62 Resp: 1239
Ion Ratio Lower Upper
62 100
64 41.7 25.2 37.8#









#5
 Bromomethane
 Concen: 0.866 ug/l
 RT: 1.593 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 94 Resp: 770
 Ion Ratio Lower Upper
 94 100
 96 79.9 72.8 109.2



#6
 Chloroethane
 Concen: 0.951 ug/l
 RT: 1.678 min Scan# 97
 Delta R.T. 0.012 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 64 Resp: 859
 Ion Ratio Lower Upper
 64 100
 66 35.2 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 0.734 ug/l
 RT: 1.880 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 1959
 Ion Ratio Lower Upper
 101 100
 103 67.8 51.0 76.4



#8
 Diethyl Ether
 Concen: 0.845 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 74 Resp: 742
 Ion Ratio Lower Upper
 74 100
 45 123.6 54.9 164.8









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.731 ug/l
 RT: 2.319 min Scan# 202
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 1166
 Ion Ratio Lower Upper
 101 100
 85 47.3 38.6 58.0
 151 73.1 55.2 82.8



#10
 Methyl Iodide
 Concen: 0.692 ug/l
 RT: 2.441 min Scan# 222
 Delta R.T. -0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion:142 Resp: 1297
 Ion Ratio Lower Upper
 142 100
 127 45.5 41.7 62.5
 141 3.9 11.5 17.3#









#11
Tert butyl alcohol
Concen: 3.604 ug/l
RT: 2.965 min Scan# 308
Delta R.T. -0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 59 Resp: 1176
Ion Ratio Lower Upper
59 100
57 2.4 8.6 12.8#



#12
1,1-Dichloroethene
Concen: 0.704 ug/l
RT: 2.313 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 96 Resp: 1093
Ion Ratio Lower Upper
96 100
61 199.6 146.2 219.2
98 49.6 51.0 76.6#





#13

Acrolein

Concen: 3.625 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 56 Resp: 1374

Ion Ratio Lower Upper

56 100

55 83.9 56.2 84.4



#14

Allyl chloride

Concen: 0.660 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 41 Resp: 1936


Ion Ratio Lower Upper

41 100

39 85.5 60.6 90.8

76 26.4 24.9 37.3









#15
Acrylonitrile
Concen: 3.424 ug/l
RT: 3.075 min Scan# 31
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Tgt Ion: 53 Resp: 3343
Ion Ratio Lower Upper
53 100
52 82.9 65.3 97.9
51 34.4 29.8 44.8



#16
Acetone
Concen: 3.736 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 3500
Ion Ratio Lower Upper
43 100
58 28.1 21.2 31.8





#17

Carbon Disulfide

Concen: 0.728 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 76 Resp: 2619

Ion Ratio Lower Upper

76 100

78 8.0 7.0 10.4



#18

Methyl Acetate

Concen: 0.828 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 1852

Ion Ratio Lower Upper

43 100

74 22.0 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 0.662 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 73 Resp: 3588
Ion Ratio Lower Upper
73 100
57 20.3 17.7 26.5



#20
Methylene Chloride
Concen: 0.828 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 84 Resp: 1571
Ion Ratio Lower Upper
84 100
49 140.0 113.9 170.9
51 51.3 33.5 50.3#
86 85.2 53.8 80.8#









#21
trans-1,2-Dichloroethene
Concen: 0.701 ug/l
RT: 3.081 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 96 Resp: 1111
Ion Ratio Lower Upper
96 100
61 119.0 119.5 179.3#
98 30.1 50.0 75.0#



#22
Diisopropyl ether
Concen: 0.703 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 45 Resp: 3857
Ion Ratio Lower Upper
45 100
43 176.2 66.6 100.0#
87 28.9 19.8 29.6
59 9.1 8.6 12.8









#23
 Vinyl Acetate
 Concen: 3.154 ug/l
 RT: 3.727 min Scan# 413
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 43 Resp: 15283
 Ion Ratio Lower Upper
 43 100
 86 10.8 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 0.637 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 63 Resp: 2054
 Ion Ratio Lower Upper
 63 100
 98 6.3 3.6 10.8
 100 5.8 2.1 6.3









#25
2-Butanone
Concen: 3.371 ug/l
RT: 4.593 min Scan# 51
Delta R.T. 0.037 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 4558
Ion Ratio Lower Upper
43 100
72 19.2 18.4 27.6



#26
2,2-Dichloropropane
Concen: 0.728 ug/l
RT: 4.458 min Scan# 553
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 77 Resp: 1777
Ion Ratio Lower Upper
77 100
97 13.7 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 0.693 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 96	Resp: 1324		
Ion Ratio	Lower	Upper	
96	100		
61	145.5	0.0	322.8
98	63.1	0.0	129.0



#28
 Bromochloromethane
 Concen: 0.398 ug/l
 RT: 4.910 min Scan# 627
 Delta R.T. 0.012 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 49	Resp: 682		
Ion Ratio	Lower	Upper	
49 100			
129 0.0	0.0	4.0	
130 101.6	56.2	84.2#	









#29
Tetrahydrofuran
Concen: 3.338 ug/l
RT: 5.019 min Scan# 642
Delta R.T. 0.018 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 42 Resp: 2931
Ion Ratio Lower Upper
42 100
72 41.9 33.7 50.5
71 37.7 31.4 47.2



#30
Chloroform
Concen: 0.695 ug/l
RT: 5.093 min Scan# 657
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 83 Resp: 2328
Ion Ratio Lower Upper
83 100
85 60.4 49.3 73.9








#31
Cyclohexane
Concen: 0.763 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001







Tgt Ion:	56	Resp:	2072
Ion	Ratio	Lower	Upper
56	100		
69	25.1	24.4	36.6
84	95.5	66.9	100.3



#32
1,1,1-Trichloroethane
Concen: 0.647 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:	97	Resp:	1869
Ion	Ratio	Lower	Upper
97	100		
99	0.0	51.8	77.6#
61	55.0	40.1	60.1





#33

1,2-Dichloroethane-d4

Concen: 0.799 ug/l

RT: 5.970 min Scan# 801

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 65 Resp: 2188

Ion Ratio Lower Upper

65 100

67 42.3 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 114 Resp: 167022


Ion Ratio Lower Upper

114 100

63 24.2 0.0 49.2

88 17.1 0.0 33.6









#35
 Dibromofluoromethane
 Concen: 0.721 ug/l
 RT: 5.391 min Scan# 704
 Delta R.T. 0.012 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:113 Resp: 1350
 Ion Ratio Lower Upper
 113 100
 111 117.0 83.1 124.7
 192 12.8 13.3 19.9#



#36
 1,1-Dichloropropene
 Concen: 0.758 ug/l
 RT: 5.684 min Scan# 754
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 75 Resp: 1648
 Ion Ratio Lower Upper
 75 100
 110 28.2 16.3 48.9
 77 15.8 24.3 36.5#









#37
Ethyl Acetate
Concen: 0.531 ug/l
RT: 4.733 min Scan# 598
Delta R.T. 0.018 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Tgt Ion: 43 Resp: 1447
Ion Ratio Lower Upper
43 100
61 0.0 10.3 15.5#
70 1.4 7.9 11.9#



#38
Carbon Tetrachloride
Concen: 0.724 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 117 Resp: 1807
Ion Ratio Lower Upper
117 100
119 67.0 75.2 112.8#
121 28.5 24.2 36.4









#39
Methylcyclohexane
Concen: 0.770 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Tgt Ion: 83 Resp: 2096
Ion Ratio Lower Upper
83 100
55 80.1 64.7 97.1
98 41.9 37.4 56.2



#40
Benzene
Concen: 0.668 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 78 Resp: 4502
Ion Ratio Lower Upper
78 100
77 29.1 19.0 28.4#









#41
Methacrylonitrile
Concen: 0.513 ug/l
RT: 4.946 min Scan# 61
Delta R.T. 0.030 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:	41	Resp:	778
Ion	Ratio	Lower	Upper
41	100		
39	51.7	47.2	70.8
67	0.0	50.7	76.1#
52	13.6	24.6	37.0#



#42
1,2-Dichloroethane
Concen: 0.694 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:	62	Resp:	1934
Ion	Ratio	Lower	Upper
62	100		
98	1.1	0.0	15.2





#43

Isopropyl Acetate

Concen: 0.612 ug/l

RT: 6.354 min Scan# 864

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 43 Resp: 2553

Ion Ratio Lower Upper

43 100

61 10.7 14.3 21.5#

87 3.2 9.5 14.3#



#44

Trichloroethene

Concen: 0.680 ug/l

RT: 7.135 min Scan# 992

Delta R.T. 0.012 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:130 Resp: 1083

Ion Ratio Lower Upper

130 100

95 117.3 0.0 204.2





#45

1,2-Dichloropropane

Concen: 0.633 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 63 Resp: 1059

Ion Ratio Lower Upper

63 100

65 36.1 24.1 36.1



#46

Dibromomethane

Concen: 0.667 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 93 Resp: 880


Ion Ratio Lower Upper

93 100

95 84.5 65.6 98.4

174 88.5 68.2 102.2









#47
Bromodichloromethane
Concen: 0.633 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:	83	Resp:	1620
Ion	Ratio	Lower	Upper
83	100		
85	57.3	50.5	75.7
127	0.0	6.5	9.7#



#48
Methyl methacrylate
Concen: 0.483 ug/l
RT: 7.714 min Scan# 1087
Delta R.T. 0.024 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:	41	Resp:	1033
Ion	Ratio	Lower	Upper
41	100		
69	84.4	58.5	87.7
39	101.4	51.7	77.5#









#49
1,4-Dioxane
Concen: 11.703 ug/l
RT: 7.677 min Scan# 1081
Delta R.T. 0.018 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 88 Resp: 498
Ion Ratio Lower Upper
88 100
43 44.4 33.4 50.2
58 30.3 58.6 88.0#



#50
Toluene-d8
Concen: 0.628 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 98 Resp: 3964
Ion Ratio Lower Upper
98 100
100 73.5 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 3.430 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 9378
Ion Ratio Lower Upper
43 100
58 34.4 28.9 43.3



#52
Toluene
Concen: 0.672 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 92 Resp: 2684
Ion Ratio Lower Upper
92 100
91 165.7 136.6 205.0









#53
 t-1,3-Dichloropropene
 Concen: 0.576 ug/l
 RT: 8.988 min Scan# 1196
 Delta R.T. 0.012 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 75 Resp: 1238
 Ion Ratio Lower Upper
 75 100
 77 42.7 25.0 37.6#



#54
 cis-1,3-Dichloropropene
 Concen: 0.564 ug/l
 RT: 8.379 min Scan# 1196
 Delta R.T. 0.018 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 75 Resp: 1414
 Ion Ratio Lower Upper
 75 100
 77 44.2 25.4 38.0#
 39 59.0 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 0.637 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 97	Resp: 1029
Ion Ratio	Lower Upper
97 100	
83 74.8	69.0 103.6
85 59.5	44.3 66.5
99 63.0	47.8 71.6



#56
 Ethyl methacrylate
 Concen: 0.504 ug/l
 RT: 9.128 min Scan# 1319
 Delta R.T. 0.012 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 69	Resp: 1260		
Ion Ratio	Lower	Upper	
69 100			
41 86.6	60.8	91.2	
39 78.0	39.0	58.6	#





#57

1,3-Dichloropropane

Concen: 0.711 ug/l

RT: 9.311 min Scan# 1134

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 76 Resp: 2036

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 3.420 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 63 Resp: 3846

Ion Ratio Lower Upper

63 100

106 24.0 21.5 32.3









#59
2-Hexanone
Concen: 3.096 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 6424
Ion Ratio Lower Upper
43 100
58 41.6 24.9 74.6



#60
Dibromochloromethane
Concen: 0.582 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 129 Resp: 1023
Ion Ratio Lower Upper
129 100
127 72.5 39.3 117.8





#61

1,2-Dibromoethane

Concen: 0.650 ug/l

RT: 9.610 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:107 Resp: 1077

Ion Ratio Lower Upper

107 100

109 94.0 73.3 109.9



#62

4-Bromofluorobenzene

Concen: 0.545 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 95 Resp: 1250


Ion Ratio Lower Upper

95 100

174 72.2 0.0 135.8

176 68.9 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:117 Resp: 140108
Ion Ratio Lower Upper
117 100
82 61.0 50.6 76.0
119 32.1 25.8 38.6



#64
Tetrachloroethene
Concen: 0.679 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:164 Resp: 972
Ion Ratio Lower Upper
164 100
166 119.1 105.0 157.6
129 103.4 83.5 125.3
131 95.4 76.5 114.7









#65
Chlorobenzene
Concen: 0.735 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:112 Resp: 3170
Ion Ratio Lower Upper
112 100
114 30.6 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 0.720 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:131 Resp: 1035
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 0.0 31.6 95.0#





#67

Ethyl Benzene

Concen: 0.687 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 91 Resp: 5052

Ion Ratio Lower Upper

91 100

106 32.5 23.4 35.2



#68

m/p-Xylenes

Concen: 1.367 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 106 Resp: 3630

Ion Ratio Lower Upper

106 100

91 209.8 171.2 256.8









#69
o-Xylene
Concen: 0.672 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:106 Resp: 1799
Ion Ratio Lower Upper
106 100
91 214.9 112.7 338.1



#70
Styrene
Concen: 0.622 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:104 Resp: 2664
Ion Ratio Lower Upper
104 100
78 54.3 45.7 68.5
103 53.9 43.7 65.5





#71

Bromoform

Concen: 0.608 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:173 Resp: 657

Ion Ratio Lower Upper

173 100

175 46.6 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:152 Resp: 59123


Ion Ratio Lower Upper

152 100

115 66.6 46.9 140.7

150 156.3 0.0 351.0









#73
Isopropylbenzene
Concen: 0.695 ug/l
RT: 10.963 min Scan# 1619
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 4480
Ion Ratio Lower Upper
105 100
120 23.7 12.8 38.4



#74
N-ethyl acetate
Concen: 0.631 ug/l
RT: 10.854 min Scan# 1602
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 2028
Ion Ratio Lower Upper
43 100
70 31.5 30.9 46.3
55 24.5 18.7 28.1
61 15.4 17.1 25.7#





#75

1,1,2,2-Tetrachloroethane

Concen: 0.800 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 1835

Ion Ratio Lower Upper

83 100

131 7.5 5.0 14.9

85 58.7 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 0.754 ug/l

RT: 11.244 min Scan# 1666

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 75 Resp: 1858

Ion Ratio Lower Upper

75 100

77 26.5 20.5 61.5









#77
Bromobenzene
Concen: 0.734 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:156 Resp: 1095
Ion Ratio Lower Upper
156 100
77 179.5 92.8 278.3
158 93.5 48.8 146.3



#78
n-propylbenzene
Concen: 0.699 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 5052
Ion Ratio Lower Upper
91 100
120 20.1 10.8 32.4









#79
2-Chlorotoluene
Concen: 0.773 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 3765
Ion Ratio Lower Upper
91 100
126 28.0 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 0.671 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 105 Resp: 3590
Ion Ratio Lower Upper
105 100
120 43.4 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 0.425 ug/l
RT: 11.024 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 252
Ion Ratio Lower Upper
75 100
53 164.7 100.5 150.7#
89 37.3 37.0 55.6



#82
4-Chlorotoluene
Concen: 0.713 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 3815
Ion Ratio Lower Upper
91 100
126 24.8 13.3 39.8









#83
 tert-Butylbenzene
 Concen: 0.751 ug/l
 RT: 11.713 min Scan# 1743
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:	119	Resp:	3951
Ion	Ratio	Lower	Upper
119	100		
91	63.9	32.9	98.7
134	25.2	11.4	34.1



#84
 1,2,4-Trimethylbenzene
 Concen: 0.697 ug/l
 RT: 11.756 min Scan# 1750
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion:	105	Resp:	3725
Ion	Ratio	Lower	Upper
105	100		
120	40.3	21.2	63.6









#85
 sec-Butylbenzene
 Concen: 0.672 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:105 Resp: 4385
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 0.679 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion:119 Resp: 3577
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.5 37.5
 91 33.1 13.8 41.4









#87
1,3-Dichlorobenzene
Concen: 0.689 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:146	Resp:	1914	
Ion	Ratio	Lower	Upper
146	100		
111	43.9	22.1	66.3
148	65.3	32.1	96.5



#88
1,4-Dichlorobenzene
Concen: 0.703 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.067 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:146	Resp:	1914	
Ion	Ratio	Lower	Upper
146	100		
111	43.9	21.3	63.9
148	65.3	31.9	95.5





#89

n-Butylbenzene

Concen: 0.631 ug/l

RT: 12.335 min Scan# 1844

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 91 Resp: 2889

Ion Ratio Lower Upper

91 100

92 48.1 26.9 80.7

134 20.1 11.8 35.3



#90

Hexachloroethane

Concen: 0.616 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 117 Resp: 604

Ion Ratio Lower Upper

117 100

201 64.2 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 0.747 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:146 Resp: 2022
Ion Ratio Lower Upper
146 100
111 48.8 22.4 67.3
148 59.4 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.586 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 75 Resp: 306
Ion Ratio Lower Upper
75 100
155 49.3 34.9 104.8
157 78.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 0.688 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:180 Resp: 1019

Ion Ratio Lower Upper

180 100

182 100.6 47.9 143.8

145 29.5 17.1 51.2



#94

Hexachlorobutadiene

Concen: 0.671 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:225 Resp: 465


Ion Ratio Lower Upper

225 100

223 69.7 30.8 92.4

227 70.5 30.9 92.7





#95

Naphthalene

Concen: 0.767 ug/l

RT: 13.780 min Scan# 2081

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:128 Resp: 4138

Ion Ratio Lower Upper

128 100

127 12.4 10.4 15.6

129 8.9 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 0.716 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:180 Resp: 1113

Ion Ratio Lower Upper

180 100

182 92.5 47.8 143.3

145 32.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83733	48.509	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.020%
35) Dibromofluoromethane	5.379	113	58192	49.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.647	98	198575	48.984	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.079	95	75568	48.596	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	78783	55.594	ug/l	99
3) Chloromethane	1.307	50	72275	52.592	ug/l	100
4) Vinyl Chloride	1.374	62	65994	51.598	ug/l	98
5) Bromomethane	1.593	94	30033	50.627	ug/l	97
6) Chloroethane	1.666	64	35690	52.271	ug/l	94
7) Trichlorofluoromethane	1.874	101	99352	52.560	ug/l	99
8) Diethyl Ether	2.130	74	31875	49.536	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59451	50.822	ug/l	99
10) Methyl Iodide	2.447	142	73940	53.419	ug/l	99
11) Tert butyl alcohol	2.977	59	61065	252.025	ug/l	99
12) 1,1-Dichloroethene	2.313	96	56097	51.097	ug/l	98
13) Acrolein	2.233	56	64157	232.505	ug/l	99
14) Allyl chloride	2.660	41	108757	51.833	ug/l	100
15) Acrylonitrile	3.063	53	181808	262.412	ug/l	99
16) Acetone	2.380	43	169923	245.518	ug/l	98
17) Carbon Disulfide	2.502	76	132695	50.981	ug/l	99
18) Methyl Acetate	2.703	43	80251	49.968	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	195442	50.777	ug/l	99
20) Methylene Chloride	2.782	84	63369	47.779	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	55770	50.513	ug/l	99
22) Diisopropyl ether	3.757	45	209946	51.800	ug/l	97
23) Vinyl Acetate	3.715	43	946746	265.585	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114733	50.825	ug/l	98
25) 2-Butanone	4.556	43	258339	257.110	ug/l	97
26) 2,2-Dichloropropane	4.471	77	88229	49.934	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	67727	50.956	ug/l	100
28) Bromochloromethane	4.891	49	55390	50.975	ug/l	98
29) Tetrahydrofuran	5.001	42	163675	259.958	ug/l	99
30) Chloroform	5.093	83	120782	51.332	ug/l	97
31) Cyclohexane	5.458	56	103158	50.147	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	104243	51.108	ug/l	100
36) 1,1-Dichloropropene	5.690	75	77931	49.521	ug/l	99
37) Ethyl Acetate	4.715	43	98604	50.714	ug/l	100
38) Carbon Tetrachloride	5.666	117	88100	49.825	ug/l	97
39) Methylcyclohexane	7.373	83	100808	49.757	ug/l	98
40) Benzene	6.031	78	233977	50.759	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	56017	55.076	ug/l	100
42) 1,2-Dichloroethane	6.080	62	101185	50.861	ug/l	98
43) Isopropyl Acetate	6.336	43	154818	52.195	ug/l	99
44) Trichloroethene	7.123	130	55491	50.018	ug/l	100
45) 1,2-Dichloropropane	7.428	63	59582	51.983	ug/l	98
46) Dibromomethane	7.580	93	45181	49.978	ug/l	99
47) Bromodichloromethane	7.818	83	92388	51.889	ug/l	100
48) Methyl methacrylate	7.690	41	82144	54.226	ug/l	98
49) 1,4-Dioxane	7.659	88	30477	1059.599	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	510206	259.133	ug/l	100
52) Toluene	8.714	92	144574	51.150	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	83751	52.920	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	92517	52.892	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	56713	50.887	ug/l	96
56) Ethyl methacrylate	9.116	69	95772	53.918	ug/l	99
57) 1,3-Dichloropropane	9.305	76	99663	49.793	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.238	63	240681	265.780	ug/l	99
59) 2-Hexanone	9.427	43	390701	268.218	ug/l	100
60) Dibromochloromethane	9.519	129	63738	52.075	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59134	51.050	ug/l	98
64) Tetrachloroethene	9.269	164	48034	46.795	ug/l	98
65) Chlorobenzene	10.073	112	154640	48.699	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	53735	49.557	ug/l	100
67) Ethyl Benzene	10.189	91	281239	50.245	ug/l	100
68) m/p-Xylenes	10.299	106	206813	101.022	ug/l	99
69) o-Xylene	10.640	106	101469	50.841	ug/l	99
70) Styrene	10.653	104	169703	51.907	ug/l	99
71) Bromoform	10.799	173	40364	49.583	ug/l #	97
73) Isopropylbenzene	10.957	105	269484	50.733	ug/l	100
74) N-amyl acetate	10.842	43	137559	52.407	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	89027	47.827	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	99886	60.821	ug/l	86
77) Bromobenzene	11.195	156	59871	48.549	ug/l	98
78) n-propylbenzene	11.299	91	312285	50.562	ug/l	100
79) 2-Chlorotoluene	11.360	91	195630	49.107	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	228538	51.500	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	24857	49.276	ug/l	98
82) 4-Chlorotoluene	11.451	91	223811	50.661	ug/l	100
83) tert-Butylbenzene	11.713	119	225783	50.511	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	231371	51.485	ug/l	100
85) sec-Butylbenzene	11.890	105	280578	51.122	ug/l	99
86) p-Isopropyltoluene	12.006	119	231942	51.198	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	113146	50.272	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	111173	48.368	ug/l	99
89) n-Butylbenzene	12.329	91	205858	51.804	ug/l	100
90) Hexachloroethane	12.536	117	39823	49.895	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111482	49.361	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21220	51.458	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	64421	49.662	ug/l	100
94) Hexachlorobutadiene	13.719	225	27425	48.408	ug/l	98
95) Naphthalene	13.774	128	239805	50.404	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	64950	48.525	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:168 Resp: 92588
Ion Ratio Lower Upper
168 100
99 74.5 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 55.594 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 85 Resp: 78783
Ion Ratio Lower Upper
85 100
87 32.2 15.7 47.1





#3

Chloromethane

Concen: 52.592 ug/l

RT: 1.307 min Scan# 30

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 50 Resp: 72275

Ion Ratio Lower Upper

50 100

52 31.8 25.4 38.2



#4

Vinyl Chloride

Concen: 51.598 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 62 Resp: 65994

Ion Ratio Lower Upper

62 100

64 32.3 25.2 37.8









#5
 Bromomethane
 Concen: 50.627 ug/l
 RT: 1.593 min Scan# 81
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion: 94 Resp: 30033
 Ion Ratio Lower Upper
 94 100
 96 93.5 72.8 109.2



#6
 Chloroethane
 Concen: 52.271 ug/l
 RT: 1.666 min Scan# 95
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 64 Resp: 35690
 Ion Ratio Lower Upper
 64 100
 66 33.7 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 52.560 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion:101 Resp: 99352
 Ion Ratio Lower Upper
 101 100
 103 62.6 51.0 76.4



#8
 Diethyl Ether
 Concen: 49.536 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 74 Resp: 31875
 Ion Ratio Lower Upper
 74 100
 45 107.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.822 ug/l
RT: 2.319 min Scan# 20
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:101 Resp: 59451
Ion Ratio Lower Upper
101 100
85 47.8 38.6 58.0
151 69.4 55.2 82.8



#10
Methyl Iodide
Concen: 53.419 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:142 Resp: 73940
Ion Ratio Lower Upper
142 100
127 52.9 41.7 62.5
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 252.025 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 59 Resp: 61065
Ion Ratio Lower Upper
59 100
57 11.1 8.6 12.8



#12
1,1-Dichloroethene
Concen: 51.097 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 96 Resp: 56097
Ion Ratio Lower Upper
96 100
61 179.7 146.2 219.2
98 62.1 51.0 76.6





#13

Acrolein

Concen: 232.505 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 56 Resp: 64157

Ion Ratio Lower Upper

56 100

55 70.8 56.2 84.4



#14

Allyl chloride

Concen: 51.833 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 41 Resp: 108757


Ion Ratio Lower Upper

41 100

39 75.6 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 262.412 ug/l

RT: 3.063 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion: 53 Resp: 181808

Ion Ratio Lower Upper

53 100

52 82.9 65.3 97.9

51 36.5 29.8 44.8



#16

Acetone

Concen: 245.518 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 169923

Ion Ratio Lower Upper

43 100

58 27.3 21.2 31.8





#17

Carbon Disulfide

Concen: 50.981 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 76 Resp: 132695

Ion Ratio Lower Upper

76 100

78 9.1 7.0 10.4



#18

Methyl Acetate

Concen: 49.968 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 80251

Ion Ratio Lower Upper

43 100

74 22.1 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 50.777 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 73 Resp: 195442
Ion Ratio Lower Upper
73 100
57 21.8 17.7 26.5



#20
Methylene Chloride
Concen: 47.779 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 84 Resp: 63369
Ion Ratio Lower Upper
84 100
49 142.7 113.9 170.9
51 43.3 33.5 50.3
86 66.8 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 50.513 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 96 Resp: 55770
Ion Ratio Lower Upper
96 100
61 149.3 119.5 179.3
98 64.6 50.0 75.0



#22
Diisopropyl ether
Concen: 51.800 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 45 Resp: 209946
Ion Ratio Lower Upper
45 100
43 86.9 66.6 100.0
87 24.8 19.8 29.6
59 10.8 8.6 12.8









#23
 Vinyl Acetate
 Concen: 265.585 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion: 43 Resp: 946746
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 50.825 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 63 Resp: 114733
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.6 10.8
 100 4.8 2.1 6.3









#25
2-Butanone
Concen: 257.110 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 258339
Ion Ratio Lower Upper
43 100
72 24.3 18.4 27.6



#26
2,2-Dichloropropane
Concen: 49.934 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 77 Resp: 88229
Ion Ratio Lower Upper
77 100
97 21.0 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 50.956 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Tgt Ion: 96 Resp: 67727




Ion	Ratio	Lower	Upper
96	100		
61	160.7	0.0	322.8
98	64.4	0.0	129.0




#28
 Bromochloromethane
 Concen: 50.975 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 49 Resp: 55390

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.0
130	68.7	56.2	84.2





#29

Tetrahydrofuran

Concen: 259.958 ug/l

RT: 5.001 min Scan# 64

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion: 42 Resp: 163675

Ion Ratio Lower Upper

42 100

72 42.0 33.7 50.5

71 38.5 31.4 47.2



#30

Chloroform

Concen: 51.332 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 83 Resp: 120782

Ion Ratio Lower Upper

83 100

85 63.6 49.3 73.9









#31
Cyclohexane
Concen: 50.147 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 56 Resp: 103158
Ion Ratio Lower Upper
56 100
69 27.8 24.4 36.6
84 80.5 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 51.108 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 97 Resp: 104243
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.6
61 50.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.509 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 65 Resp: 83733

Ion Ratio Lower Upper

65 100

67 50.0 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 114 Resp: 162651


Ion Ratio Lower Upper

114 100

63 23.1 0.0 49.2

88 17.2 0.0 33.6









#35
Dibromofluoromethane
Concen: 49.683 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:113 Resp: 58192
Ion Ratio Lower Upper
113 100
111 102.1 83.1 124.7
192 16.2 13.3 19.9



#36
1,1-Dichloropropene
Concen: 49.521 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 77931
Ion Ratio Lower Upper
75 100
110 32.7 16.3 48.9
77 31.1 24.3 36.5









#37
Ethyl Acetate
Concen: 50.714 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 98604
Ion Ratio Lower Upper
43 100
61 12.6 10.3 15.5
70 10.0 7.9 11.9



#38
Carbon Tetrachloride
Concen: 49.825 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 117 Resp: 88100
Ion Ratio Lower Upper
117 100
119 90.7 75.2 112.8
121 29.4 24.2 36.4









#39
Methylcyclohexane
Concen: 49.757 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 83 Resp: 100808
Ion Ratio Lower Upper
83 100
55 82.7 64.7 97.1
98 48.3 37.4 56.2



#40
Benzene
Concen: 50.759 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 78 Resp: 233977
Ion Ratio Lower Upper
78 100
77 24.0 19.0 28.4









#41
Methacrylonitrile
Concen: 55.076 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:	41	Resp:	56017
Ion	Ratio	Lower	Upper
41	100		
39	58.7	47.2	70.8
67	63.7	50.7	76.1
52	30.5	24.6	37.0



#42
1,2-Dichloroethane
Concen: 50.861 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:	62	Resp:	101185
Ion	Ratio	Lower	Upper
62	100		
98	7.1	0.0	15.2









#43
Isopropyl Acetate
Concen: 52.195 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 154818
Ion Ratio Lower Upper
43 100
61 17.9 14.3 21.5
87 11.4 9.5 14.3



#44
Trichloroethene
Concen: 50.018 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 130 Resp: 55491
Ion Ratio Lower Upper
130 100
95 102.6 0.0 204.2









#45
1,2-Dichloropropane
Concen: 51.983 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 63 Resp: 59582
Ion Ratio Lower Upper
63 100
65 31.4 24.1 36.1



#46
Dibromomethane
Concen: 49.978 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 93 Resp: 45181
Ion Ratio Lower Upper
93 100
95 83.0 65.6 98.4
174 84.8 68.2 102.2









#47
Bromodichloromethane
Concen: 51.889 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 83 Resp: 92388
Ion Ratio Lower Upper
83 100
85 63.0 50.5 75.7
127 8.8 6.5 9.7



#48
Methyl methacrylate
Concen: 54.226 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 41 Resp: 82144
Ion Ratio Lower Upper
41 100
69 70.4 58.5 87.7
39 64.2 51.7 77.5





#49

1,4-Dioxane

Concen: 1059.599 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion: 88 Resp: 30477

Ion Ratio Lower Upper

88 100

43 39.9 33.4 50.2

58 76.2 58.6 88.0



#50

Toluene-d8

Concen: 48.984 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 98 Resp: 198575

Ion Ratio Lower Upper

98 100

100 65.6 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 259.133 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 510206
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3



#52
Toluene
Concen: 51.150 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 92 Resp: 144574
Ion Ratio Lower Upper
92 100
91 171.4 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 52.920 ug/l

RT: 8.976 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVX050525

Tgt Ion: 75 Resp: 83751

Ion Ratio Lower Upper

75 100

77 31.3 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 52.892 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 75 Resp: 92517


Ion Ratio Lower Upper

75 100

77 31.5 25.4 38.0

39 63.2 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 50.887 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 97 Resp: 56713





Ion Ratio Lower Upper

97 100

83 89.8 69.0 103.6

85 58.4 44.3 66.5

99 63.9 47.8 71.6



#56

Ethyl methacrylate

Concen: 53.918 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 69 Resp: 95772


Ion Ratio Lower Upper

69 100

41 76.1 60.8 91.2

39 47.4 39.0 58.6





#57

1,3-Dichloropropane

Concen: 49.793 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 76 Resp: 99663

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 265.780 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 63 Resp: 240681

Ion Ratio Lower Upper

63 100

106 26.6 21.5 32.3









#59
2-Hexanone
Concen: 268.218 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 390701
Ion Ratio Lower Upper
43 100
58 49.6 24.9 74.6



#60
Dibromochloromethane
Concen: 52.075 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 129 Resp: 63738
Ion Ratio Lower Upper
129 100
127 77.9 39.3 117.8









#61
1,2-Dibromoethane
Concen: 51.050 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:107 Resp: 59134
Ion Ratio Lower Upper
107 100
109 93.8 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 48.596 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 95 Resp: 75568
Ion Ratio Lower Upper
95 100
174 68.2 0.0 135.8
176 66.5 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:117 Resp: 145078
Ion Ratio Lower Upper
117 100
82 62.1 50.6 76.0
119 32.9 25.8 38.6



#64
Tetrachloroethene
Concen: 46.795 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:164 Resp: 48034
Ion Ratio Lower Upper
164 100
166 129.9 105.0 157.6
129 105.3 83.5 125.3
131 98.8 76.5 114.7









#65
Chlorobenzene
Concen: 48.699 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:112 Resp: 154640
Ion Ratio Lower Upper
112 100
114 31.1 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 49.557 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:131 Resp: 53735
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 63.7 31.6 95.0





#67

Ethyl Benzene

Concen: 50.245 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 91 Resp: 281239

Ion Ratio Lower Upper

91 100

106 29.1 23.4 35.2



#68

m/p-Xylenes

Concen: 101.022 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:106 Resp: 206813

Ion Ratio Lower Upper

106 100

91 215.6 171.2 256.8









#69
o-Xylene
Concen: 50.841 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:106 Resp: 101469
Ion Ratio Lower Upper
106 100
91 227.7 112.7 338.1



#70
Styrene
Concen: 51.907 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:104 Resp: 169703
Ion Ratio Lower Upper
104 100
78 57.6 45.7 68.5
103 55.5 43.7 65.5





#71

Bromoform

Concen: 49.583 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:173 Resp: 40364

Ion Ratio Lower Upper

173 100

175 48.8 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:152 Resp: 68229


Ion Ratio Lower Upper

152 100

115 93.3 46.9 140.7

150 175.3 0.0 351.0









#73
Isopropylbenzene
Concen: 50.733 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:105 Resp: 269484
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4



#74
N-ethyl acetate
Concen: 52.407 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 43 Resp: 137559
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 22.7 18.7 28.1
61 21.8 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 47.827 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 83 Resp: 89027
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 65.5 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 60.821 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 99886
Ion Ratio Lower Upper
75 100
77 32.5 20.5 61.5





#77

Bromobenzene

Concen: 48.549 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:156 Resp: 59871

Ion Ratio Lower Upper

156 100

77 189.0 92.8 278.3

158 98.9 48.8 146.3



#78

n-propylbenzene

Concen: 50.562 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 91 Resp: 312285

Ion Ratio Lower Upper

91 100

120 21.4 10.8 32.4





#79

2-Chlorotoluene

Concen: 49.107 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 91 Resp: 195630

Ion Ratio Lower Upper

91 100

126 30.8 15.6 46.7



#80

1,3,5-Trimethylbenzene

Concen: 51.500 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 105 Resp: 228538

Ion Ratio Lower Upper

105 100

120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 49.276 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 75 Resp: 24857
Ion Ratio Lower Upper
75 100
53 122.3 100.5 150.7
89 46.2 37.0 55.6



#82
4-Chlorotoluene
Concen: 50.661 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 91 Resp: 223811
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 50.511 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:119 Resp: 225783

Ion Ratio Lower Upper

119 100

91 64.9 32.9 98.7

134 22.6 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 51.485 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 231371

Ion Ratio Lower Upper

105 100

120 42.3 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 51.122 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion:105 Resp: 280578
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 51.198 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion:119 Resp: 231942
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.5 37.5
 91 27.5 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 50.272 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:146 Resp: 113146

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3

148 63.5 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 48.368 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:146 Resp: 111173


Ion Ratio Lower Upper

146 100

111 43.9 21.3 63.9

148 63.3 31.9 95.5









#89
n-Butylbenzene
Concen: 51.804 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 91 Resp: 205858
Ion Ratio Lower Upper
91 100
92 53.7 26.9 80.7
134 23.7 11.8 35.3



#90
Hexachloroethane
Concen: 49.895 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 117 Resp: 39823
Ion Ratio Lower Upper
117 100
201 62.4 31.6 94.7









#91
 1,2-Dichlorobenzene
 Concen: 49.361 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion:146 Resp: 111482
 Ion Ratio Lower Upper
 146 100
 111 45.8 22.4 67.3
 148 64.2 31.6 94.8



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 51.458 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 21220
 Ion Ratio Lower Upper
 75 100
 155 68.6 34.9 104.8
 157 86.7 43.8 131.4









#93
 1,2,4-Trichlorobenzene
 Concen: 49.662 ug/l
 RT: 13.585 min Scan# 2050
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion:180 Resp: 64421
 Ion Ratio Lower Upper
 180 100
 182 95.5 47.9 143.8
 145 34.4 17.1 51.2



#94
 Hexachlorobutadiene
 Concen: 48.408 ug/l
 RT: 13.719 min Scan# 2072
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion:225 Resp: 27425
 Ion Ratio Lower Upper
 225 100
 223 63.1 30.8 92.4
 227 63.7 30.9 92.7









#95
Naphthalene
Concen: 50.404 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:128 Resp: 239805
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.7 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 48.525 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:180 Resp: 64950
Ion Ratio Lower Upper
180 100
182 97.0 47.8 143.3
145 37.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	97221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	166757	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	142113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69603	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	83472	46.053	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.100%
35) Dibromofluoromethane	5.379	113	58765	48.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	8.641	98	191895	46.171	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.340%
62) 4-Bromofluorobenzene	11.079	95	76697	48.108	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.220%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	71992	48.381	ug/l	96
3) Chloromethane	1.301	50	71578	49.603	ug/l	100
4) Vinyl Chloride	1.374	62	65400	48.697	ug/l	100
5) Bromomethane	1.593	94	28400	45.593	ug/l	98
6) Chloroethane	1.666	64	36039	50.267	ug/l	99
7) Trichlorofluoromethane	1.874	101	99535	50.148	ug/l	100
8) Diethyl Ether	2.130	74	33963	50.265	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	62523	50.901	ug/l	99
10) Methyl Iodide	2.447	142	74693	51.391	ug/l	99
11) Tert butyl alcohol	2.971	59	60143	236.391	ug/l	100
12) 1,1-Dichloroethene	2.313	96	57843	50.176	ug/l	96
13) Acrolein	2.233	56	48798	168.416	ug/l	98
14) Allyl chloride	2.654	41	116985	53.097	ug/l	98
15) Acrylonitrile	3.056	53	186943	256.965	ug/l	99
16) Acetone	2.380	43	215550	296.602	ug/l	99
17) Carbon Disulfide	2.502	76	123668	45.249	ug/l	99
18) Methyl Acetate	2.703	43	106195	62.971	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	212708	52.629	ug/l	100
20) Methylene Chloride	2.782	84	67564	48.515	ug/l	94
21) trans-1,2-Dichloroethene	3.087	96	58500	50.461	ug/l	99
22) Diisopropyl ether	3.758	45	227076	53.357	ug/l	94
23) Vinyl Acetate	3.715	43	941542	251.538	ug/l	99
24) 1,1-Dichloroethane	3.599	63	123119	51.940	ug/l	99
25) 2-Butanone	4.550	43	278637	264.096	ug/l	99
26) 2,2-Dichloropropane	4.465	77	96338	51.925	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	72243	51.764	ug/l	99
28) Bromochloromethane	4.885	49	54439	47.712	ug/l	99
29) Tetrahydrofuran	4.995	42	166146	251.308	ug/l	100
30) Chloroform	5.080	83	126054	51.020	ug/l	95
31) Cyclohexane	5.458	56	106868	49.475	ug/l	95
32) 1,1,1-Trichloroethane	5.373	97	109132	50.955	ug/l	99
36) 1,1-Dichloropropene	5.678	75	81293	50.385	ug/l	98
37) Ethyl Acetate	4.709	43	98903	49.615	ug/l	100
38) Carbon Tetrachloride	5.666	117	91791	50.634	ug/l	99
39) Methylcyclohexane	7.373	83	106672	51.355	ug/l	98
40) Benzene	6.025	78	244281	51.690	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046460.D
 Acq On : 03 Jun 2025 10:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Jun 04 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	58928	56.511	ug/l	99
42) 1,2-Dichloroethane	6.080	62	105576	51.762	ug/l	100
43) Isopropyl Acetate	6.330	43	160957	52.928	ug/l	100
44) Trichloroethene	7.117	130	58531	51.459	ug/l	94
45) 1,2-Dichloropropane	7.422	63	63306	53.872	ug/l	98
46) Dibromomethane	7.574	93	47187	50.912	ug/l	99
47) Bromodichloromethane	7.812	83	97879	53.619	ug/l	97
48) Methyl methacrylate	7.690	41	83841	53.984	ug/l	99
49) 1,4-Dioxane	7.653	88	27329	926.756	ug/l	98
51) 4-Methyl-2-Pentanone	8.568	43	526838	260.992	ug/l	99
52) Toluene	8.714	92	149333	51.533	ug/l	98
53) t-1,3-Dichloropropene	8.970	75	87793	54.108	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	97218	54.211	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	59472	52.048	ug/l	98
56) Ethyl methacrylate	9.110	69	99973	54.897	ug/l	98
57) 1,3-Dichloropropane	9.305	76	106135	51.721	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	241312	259.915	ug/l	99
59) 2-Hexanone	9.427	43	398970	267.150	ug/l	100
60) Dibromochloromethane	9.519	129	67274	53.611	ug/l	100
61) 1,2-Dibromoethane	9.604	107	61399	51.700	ug/l	98
64) Tetrachloroethene	9.269	164	53715	53.421	ug/l	97
65) Chlorobenzene	10.073	112	160933	51.739	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	57122	53.780	ug/l	99
67) Ethyl Benzene	10.189	91	297648	54.286	ug/l	100
68) m/p-Xylenes	10.299	106	214306	106.866	ug/l	99
69) o-Xylene	10.640	106	106129	54.285	ug/l	97
70) Styrene	10.653	104	178562	55.756	ug/l	99
71) Bromoform	10.799	173	42440	53.221	ug/l #	97
73) Isopropylbenzene	10.957	105	284040	52.417	ug/l	99
74) N-amyl acetate	10.842	43	141380	52.800	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	92189	48.548	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	104375	62.299	ug/l	85
77) Bromobenzene	11.195	156	63843	50.748	ug/l	99
78) n-propylbenzene	11.299	91	332789	52.818	ug/l	100
79) 2-Chlorotoluene	11.360	91	201843	49.667	ug/l	99
80) 1,3,5-Trimethylbenzene	11.445	105	238098	52.595	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.012	75	26522	51.539	ug/l	99
82) 4-Chlorotoluene	11.451	91	234790	52.097	ug/l	100
83) tert-Butylbenzene	11.713	119	235173	51.573	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	240382	52.434	ug/l	100
85) sec-Butylbenzene	11.884	105	298046	53.233	ug/l	100
86) p-Isopropyltoluene	12.006	119	249289	53.941	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	118632	51.669	ug/l	100
88) 1,4-Dichlorobenzene	12.037	146	118528	50.549	ug/l	99
89) n-Butylbenzene	12.329	91	227595	56.143	ug/l	99
90) Hexachloroethane	12.536	117	42405	52.082	ug/l	100
91) 1,2-Dichlorobenzene	12.329	146	116846	50.715	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21419	50.916	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	70545	53.309	ug/l	100
94) Hexachlorobutadiene	13.719	225	31033	53.695	ug/l	99
95) Naphthalene	13.774	128	248647	51.231	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	71443	52.322	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046460.D
Acq On : 03 Jun 2025 10:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Quant Time: Jun 04 01:21:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 04 01:21:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 97221
Ion Ratio Lower Upper
168 100
99 70.0 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 48.381 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 85 Resp: 71992
Ion Ratio Lower Upper
85 100
87 33.7 15.7 47.1









#3
Chloromethane
Concen: 49.603 ug/l
RT: 1.301 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 71578
Ion Ratio Lower Upper
50 100
52 31.7 25.4 38.2



#4
Vinyl Chloride
Concen: 48.697 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 62 Resp: 65400
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8









#5
Bromomethane
Concen: 45.593 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 28400
Ion Ratio Lower Upper
94 100
96 92.6 72.8 109.2



#6
Chloroethane
Concen: 50.267 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 64 Resp: 36039
Ion Ratio Lower Upper
64 100
66 30.9 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 50.148 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 99535
 Ion Ratio Lower Upper
 101 100
 103 63.7 51.0 76.4



#8
 Diethyl Ether
 Concen: 50.265 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Tgt Ion: 74 Resp: 33963
 Ion Ratio Lower Upper
 74 100
 45 106.2 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.901 ug/l
RT: 2.319 min Scan# 20
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 62523
Ion Ratio Lower Upper
101 100
85 48.2 38.6 58.0
151 70.4 55.2 82.8



#10
Methyl Iodide
Concen: 51.391 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:142 Resp: 74693
Ion Ratio Lower Upper
142 100
127 52.5 41.7 62.5
141 14.8 11.5 17.3









#11
Tert butyl alcohol
Concen: 236.391 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 60143
Ion Ratio Lower Upper
59 100
57 10.8 8.6 12.8



#12
1,1-Dichloroethene
Concen: 50.176 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 96 Resp: 57843
Ion Ratio Lower Upper
96 100
61 187.2 146.2 219.2
98 68.1 51.0 76.6





#13

Acrolein

Concen: 168.416 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 56 Resp: 48798

Ion Ratio Lower Upper

56 100

55 71.6 56.2 84.4



#14

Allyl chloride

Concen: 53.097 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 41 Resp: 116985


Ion Ratio Lower Upper

41 100

39 73.9 60.6 90.8

76 30.2 24.9 37.3





#15

Acrylonitrile

Concen: 256.965 ug/l

RT: 3.056 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 53 Resp: 186943

Ion Ratio Lower Upper

53 100

52 82.9 65.3 97.9

51 37.4 29.8 44.8



#16

Acetone

Concen: 296.602 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 43 Resp: 215550

Ion Ratio Lower Upper

43 100

58 26.8 21.2 31.8









#17
Carbon Disulfide
Concen: 45.249 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 123668
Ion Ratio Lower Upper
76 100
78 8.9 7.0 10.4



#18
Methyl Acetate
Concen: 62.971 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 43 Resp: 106195
Ion Ratio Lower Upper
43 100
74 22.1 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 52.629 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 212708
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5



#20
Methylene Chloride
Concen: 48.515 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 84 Resp: 67564
Ion Ratio Lower Upper
84 100
49 135.4 113.9 170.9
51 40.7 33.5 50.3
86 61.3 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 50.461 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 58500
Ion Ratio Lower Upper
96 100
61 148.7 119.5 179.3
98 60.7 50.0 75.0



#22
Diisopropyl ether
Concen: 53.357 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 45 Resp: 227076
Ion Ratio Lower Upper
45 100
43 76.3 66.6 100.0
87 24.2 19.8 29.6
59 10.4 8.6 12.8









#23
 Vinyl Acetate
 Concen: 251.538 ug/l
 RT: 3.715 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 941542
 Ion Ratio Lower Upper
 43 100
 86 9.1 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 51.940 ug/l
 RT: 3.599 min Scan# 412
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Tgt Ion: 63 Resp: 123119
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.6 10.8
 100 4.3 2.1 6.3









#25
2-Butanone
Concen: 264.096 ug/l
RT: 4.550 min Scan# 569
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 278637
Ion Ratio Lower Upper
43 100
72 23.5 18.4 27.6



#26
2,2-Dichloropropane
Concen: 51.925 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 77 Resp: 96338
Ion Ratio Lower Upper
77 100
97 21.4 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 51.764 ug/l
 RT: 4.477 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tgt Ion: 96 Resp: 72243




Ion	Ratio	Lower	Upper
96	100		
61	160.5	0.0	322.8
98	65.6	0.0	129.0




#28
 Bromochloromethane
 Concen: 47.712 ug/l
 RT: 4.885 min Scan# 623
 Delta R.T. -0.012 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Tgt Ion: 49 Resp: 54439

Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	4.0
130	69.7	56.2	84.2





#29

Tetrahydrofuran

Concen: 251.308 ug/l

RT: 4.995 min Scan# 641

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 166146

Ion Ratio Lower Upper

42 100

72 42.0 33.7 50.5

71 38.9 31.4 47.2



#30

Chloroform

Concen: 51.020 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 83 Resp: 126054

Ion Ratio Lower Upper

83 100

85 65.3 49.3 73.9









#31
Cyclohexane
Concen: 49.475 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 106868
Ion Ratio Lower Upper
56 100
69 28.8 24.4 36.6
84 78.6 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 50.955 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 97 Resp: 109132
Ion Ratio Lower Upper
97 100
99 63.5 51.8 77.6
61 50.5 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 46.053 ug/l
 RT: 5.946 min Scan# 798
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 65 Resp: 83472
 Ion Ratio Lower Upper
 65 100
 67 48.5 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.751 min Scan# 929
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Tgt Ion: 114 Resp: 166757
 Ion Ratio Lower Upper
 114 100
 63 23.8 0.0 49.2
 88 17.7 0.0 33.6









#35
Dibromofluoromethane
Concen: 48.937 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 58765
Ion Ratio Lower Upper
113 100
111 102.9 83.1 124.7
192 16.9 13.3 19.9



#36
1,1-Dichloropropene
Concen: 50.385 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 75 Resp: 81293
Ion Ratio Lower Upper
75 100
110 33.5 16.3 48.9
77 31.6 24.3 36.5





#37

Ethyl Acetate

Concen: 49.615 ug/l

RT: 4.709 min Scan# 594

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 98903

Ion Ratio Lower Upper

43 100

61 12.8 10.3 15.5

70 9.8 7.9 11.9



#38

Carbon Tetrachloride

Concen: 50.634 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 117 Resp: 91791


Ion Ratio Lower Upper

117 100

119 92.5 75.2 112.8

121 29.8 24.2 36.4









#39
Methylcyclohexane
Concen: 51.355 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 106672
Ion Ratio Lower Upper
83 100
55 82.0 64.7 97.1
98 48.6 37.4 56.2



#40
Benzene
Concen: 51.690 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 78 Resp: 244281
Ion Ratio Lower Upper
78 100
77 24.2 19.0 28.4









#41
Methacrylonitrile
Concen: 56.511 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	58928
Ion	Ratio	Lower	Upper
41	100		
39	58.5	47.2	70.8
67	63.9	50.7	76.1
52	30.2	24.6	37.0



#42
1,2-Dichloroethane
Concen: 51.762 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:	62	Resp:	105576
Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2









#43
Isopropyl Acetate
Concen: 52.928 ug/l
RT: 6.330 min Scan# 861
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 160957
Ion Ratio Lower Upper
43 100
61 18.0 14.3 21.5
87 11.7 9.5 14.3



#44
Trichloroethene
Concen: 51.459 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:130 Resp: 58531
Ion Ratio Lower Upper
130 100
95 108.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 53.872 ug/l

RT: 7.422 min Scan# 1039

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 63306

Ion Ratio Lower Upper

63 100

65 29.2 24.1 36.1



#46

Dibromomethane

Concen: 50.912 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 93 Resp: 47187


Ion Ratio Lower Upper

93 100

95 83.2 65.6 98.4

174 85.1 68.2 102.2









#47
Bromodichloromethane
Concen: 53.619 ug/l
RT: 7.812 min Scan# 1103
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 97879
Ion Ratio Lower Upper
83 100
85 60.6 50.5 75.7
127 7.3 6.5 9.7



#48
Methyl methacrylate
Concen: 53.984 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 41 Resp: 83841
Ion Ratio Lower Upper
41 100
69 72.3 58.5 87.7
39 64.7 51.7 77.5









#49
1,4-Dioxane
Concen: 926.756 ug/l
RT: 7.653 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 88 Resp: 27329
Ion Ratio Lower Upper
88 100
43 42.5 33.4 50.2
58 75.3 58.6 88.0



#50
Toluene-d8
Concen: 46.171 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 98 Resp: 191895
Ion Ratio Lower Upper
98 100
100 64.3 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 260.992 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 526838
Ion Ratio Lower Upper
43 100
58 35.7 28.9 43.3



#52
Toluene
Concen: 51.533 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 92 Resp: 149333
Ion Ratio Lower Upper
92 100
91 172.9 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 54.108 ug/l

RT: 8.970 min Scan# 1193

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 75 Resp: 87793

Ion Ratio Lower Upper

75 100

77 32.1 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 54.211 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion: 75 Resp: 97218


Ion Ratio Lower Upper

75 100

77 32.0 25.4 38.0

39 62.9 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 52.048 ug/l
RT: 9.147 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	97	Resp:	59472
Ion	Ratio	Lower	Upper
97	100		
83	85.8	69.0	103.6
85	55.4	44.3	66.5
99	63.0	47.8	71.6



#56
Ethyl methacrylate
Concen: 54.897 ug/l
RT: 9.110 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:	69	Resp:	99973
Ion	Ratio	Lower	Upper
69	100		
41	77.8	60.8	91.2
39	47.9	39.0	58.6





#57

1,3-Dichloropropane

Concen: 51.721 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 106135

Ion Ratio Lower Upper

76 100

78 32.5 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 259.915 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 63 Resp: 241312

Ion Ratio Lower Upper

63 100

106 26.4 21.5 32.3









#59
2-Hexanone
Concen: 267.150 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 398970
Ion Ratio Lower Upper
43 100
58 49.6 24.9 74.6



#60
Dibromochloromethane
Concen: 53.611 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 129 Resp: 67274
Ion Ratio Lower Upper
129 100
127 78.8 39.3 117.8









#61
1,2-Dibromoethane
Concen: 51.700 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 61399
Ion Ratio Lower Upper
107 100
109 93.9 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 48.108 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 95 Resp: 76697
Ion Ratio Lower Upper
95 100
174 67.0 0.0 135.8
176 65.1 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 142113
Ion Ratio Lower Upper
117 100
82 62.9 50.6 76.0
119 31.7 25.8 38.6



#64
Tetrachloroethene
Concen: 53.421 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:164 Resp: 53715
Ion Ratio Lower Upper
164 100
166 131.0 105.0 157.6
129 105.9 83.5 125.3
131 103.2 76.5 114.7









#65
Chlorobenzene
Concen: 51.739 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 160933
Ion Ratio Lower Upper
112 100
114 31.2 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 53.780 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:131 Resp: 57122
Ion Ratio Lower Upper
131 100
133 93.1 47.3 141.9
119 63.0 31.6 95.0





#67

Ethyl Benzene

Concen: 54.286 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 297648

Ion Ratio Lower Upper

91 100

106 29.3 23.4 35.2



#68

m/p-Xylenes

Concen: 106.866 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 106 Resp: 214306

Ion Ratio Lower Upper

106 100

91 216.3 171.2 256.8









#69
o-Xylene
Concen: 54.285 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 106129
Ion Ratio Lower Upper
106 100
91 229.8 112.7 338.1



#70
Styrene
Concen: 55.756 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:104 Resp: 178562
Ion Ratio Lower Upper
104 100
78 57.7 45.7 68.5
103 54.8 43.7 65.5









#71
Bromoform
Concen: 53.221 ug/l
RT: 10.799 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 42440
Ion Ratio Lower Upper
173 100
175 48.7 23.4 70.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:152 Resp: 69603
Ion Ratio Lower Upper
152 100
115 93.8 46.9 140.7
150 172.1 0.0 351.0









#73
Isopropylbenzene
Concen: 52.417 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 284040
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4



#74
N-ethyl acetate
Concen: 52.800 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 43 Resp: 141380
Ion Ratio Lower Upper
43 100
70 37.7 30.9 46.3
55 22.6 18.7 28.1
61 22.0 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 48.548 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 92189
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 64.2 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 62.299 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 75 Resp: 104375
Ion Ratio Lower Upper
75 100
77 31.8 20.5 61.5









#77
Bromobenzene
Concen: 50.748 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 63843
Ion Ratio Lower Upper
156 100
77 186.5 92.8 278.3
158 96.5 48.8 146.3



#78
n-propylbenzene
Concen: 52.818 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 91 Resp: 332789
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 49.667 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 201843
Ion Ratio Lower Upper
91 100
126 31.6 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 52.595 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 105 Resp: 238098
Ion Ratio Lower Upper
105 100
120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 51.539 ug/l
RT: 11.012 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 26522
Ion Ratio Lower Upper
75 100
53 123.8 100.5 150.7
89 45.9 37.0 55.6



#82
4-Chlorotoluene
Concen: 52.097 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion: 91 Resp: 234790
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 51.573 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:119 Resp: 235173

Ion Ratio Lower Upper

119 100

91 65.9 32.9 98.7

134 22.7 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 52.434 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion:105 Resp: 240382

Ion Ratio Lower Upper

105 100

120 42.1 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 53.233 ug/l
 RT: 11.884 min Scan# 1771
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 298046
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 53.941 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Tgt Ion:119 Resp: 249289
 Ion Ratio Lower Upper
 119 100
 134 25.3 12.5 37.5
 91 27.7 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 51.669 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 118632

Ion Ratio Lower Upper

146 100

111 44.2 22.1 66.3

148 64.4 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 50.549 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion:146 Resp: 118528


Ion Ratio Lower Upper

146 100

111 43.5 21.3 63.9

148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 56.143 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 91 Resp: 227595

Ion Ratio Lower Upper

91 100

92 53.1 26.9 80.7

134 23.9 11.8 35.3



#90

Hexachloroethane

Concen: 52.082 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046460.D




Acq: 03 Jun 2025 10:10


Tgt Ion: 117 Resp: 42405

Ion Ratio Lower Upper

117 100

201 63.3 31.6 94.7









#91
 1,2-Dichlorobenzene
 Concen: 50.715 ug/l
 RT: 12.329 min Scan# 1844
 Delta R.T. -0.006 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:146 Resp: 116846
 Ion Ratio Lower Upper
 146 100
 111 46.7 22.4 67.3
 148 63.5 31.6 94.8



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 50.916 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. 0.000 min
 Lab File: VX046460.D
 Acq: 03 Jun 2025 10:10

Tgt Ion: 75 Resp: 21419
 Ion Ratio Lower Upper
 75 100
 155 68.3 34.9 104.8
 157 89.7 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 53.309 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:180 Resp: 70545

Ion Ratio Lower Upper

180 100

182 96.0 47.9 143.8

145 34.0 17.1 51.2



#94

Hexachlorobutadiene

Concen: 53.695 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX046460.D

Acq: 03 Jun 2025 10:10




Tgt Ion:225 Resp: 31033


Ion Ratio Lower Upper

225 100

223 61.2 30.8 92.4

227 61.3 30.9 92.7









#95
Naphthalene
Concen: 51.231 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 248647
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 11.1 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 52.322 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046460.D
Acq: 03 Jun 2025 10:10

Tgt Ion:180 Resp: 71443
Ion Ratio Lower Upper
180 100
182 95.1 47.8 143.3
145 36.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046463.D
 Acq On : 03 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0603WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01

Quant Time: Jun 04 01:23:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	89858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	159016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81765	48.808	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.620%
35) Dibromofluoromethane	5.379	113	57691	50.381	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.760%
50) Toluene-d8	8.647	98	189408	47.791	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.580%
62) 4-Bromofluorobenzene	11.079	95	76982	50.637	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.280%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	22784	16.566	ug/l	98
3) Chloromethane	1.301	50	22687	17.010	ug/l	98
4) Vinyl Chloride	1.374	62	20626	16.617	ug/l	100
5) Bromomethane	1.599	94	9717	16.878	ug/l	98
6) Chloroethane	1.672	64	12637	19.070	ug/l	91
7) Trichlorofluoromethane	1.880	101	34563	18.840	ug/l	98
8) Diethyl Ether	2.130	74	11617	18.602	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	21267	18.733	ug/l	97
10) Methyl Iodide	2.447	142	23408	17.425	ug/l	99
11) Tert butyl alcohol	2.971	59	26285	111.778	ug/l	100
12) 1,1-Dichloroethene	2.313	96	19119	17.944	ug/l	99
13) Acrolein	2.233	56	31641	118.151	ug/l	99
14) Allyl chloride	2.654	41	40050	19.668	ug/l	98
15) Acrylonitrile	3.062	53	70799	105.292	ug/l	98
16) Acetone	2.380	43	69671	103.724	ug/l	99
17) Carbon Disulfide	2.508	76	33630	13.313	ug/l	97
18) Methyl Acetate	2.703	43	42289	27.131	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	75981	20.340	ug/l	99
20) Methylene Chloride	2.782	84	23624	18.353	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	19266	17.980	ug/l	95
22) Diisopropyl ether	3.757	45	82535	20.983	ug/l	95
23) Vinyl Acetate	3.721	43	335531	96.984	ug/l	100
24) 1,1-Dichloroethane	3.605	63	44898	20.493	ug/l	99
25) 2-Butanone	4.556	43	104022	106.672	ug/l	96
26) 2,2-Dichloropropane	4.471	77	33066	19.283	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	26341	20.421	ug/l	98
28) Bromochloromethane	4.891	49	22740	21.563	ug/l	97
29) Tetrahydrofuran	5.001	42	64707	105.894	ug/l	99
30) Chloroform	5.086	83	47703	20.890	ug/l	94
31) Cyclohexane	5.464	56	35430	17.746	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	39775	20.093	ug/l	99
36) 1,1-Dichloropropene	5.684	75	28514	18.533	ug/l	99
37) Ethyl Acetate	4.715	43	37358	19.653	ug/l	99
38) Carbon Tetrachloride	5.666	117	31976	18.497	ug/l	96
39) Methylcyclohexane	7.373	83	35078	17.710	ug/l	98
40) Benzene	6.037	78	87833	19.490	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046463.D
 Acq On : 03 Jun 2025 11:24
 Operator : JC/MD
 Sample : VX0603WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01

Quant Time: Jun 04 01:23:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	21713	21.836	ug/l	97
42) 1,2-Dichloroethane	6.080	62	39263	20.187	ug/l	100
43) Isopropyl Acetate	6.342	43	59297	20.448	ug/l	99
44) Trichloroethene	7.123	130	20550	18.946	ug/l	98
45) 1,2-Dichloropropane	7.427	63	23439	20.917	ug/l	100
46) Dibromomethane	7.580	93	17506	19.808	ug/l	100
47) Bromodichloromethane	7.818	83	34890	20.043	ug/l	94
48) Methyl methacrylate	7.690	41	31312	21.143	ug/l	99
49) 1,4-Dioxane	7.659	88	12369	439.865	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	210362	109.285	ug/l	98
52) Toluene	8.714	92	54891	19.864	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	29748	19.227	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	34079	19.928	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	23413	21.488	ug/l	96
56) Ethyl methacrylate	9.116	69	37717	21.719	ug/l	98
57) 1,3-Dichloropropane	9.305	76	39888	20.384	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	94629	106.886	ug/l	100
59) 2-Hexanone	9.427	43	158797	111.507	ug/l	98
60) Dibromochloromethane	9.519	129	24550	20.516	ug/l	97
61) 1,2-Dibromoethane	9.610	107	22898	20.220	ug/l	95
64) Tetrachloroethene	9.269	164	19186	19.782	ug/l	96
65) Chlorobenzene	10.079	112	60633	20.209	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21133	20.627	ug/l	98
67) Ethyl Benzene	10.189	91	108068	20.434	ug/l	100
68) m/p-Xylenes	10.299	106	78193	40.424	ug/l	98
69) o-Xylene	10.640	106	39576	20.987	ug/l	98
70) Styrene	10.652	104	65394	21.169	ug/l	99
71) Bromoform	10.799	173	15282	19.868	ug/l #	99
73) Isopropylbenzene	10.957	105	106972	21.051	ug/l	99
74) N-amyl acetate	10.841	43	53285	21.221	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	38144	21.421	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	40431	25.735	ug/l	87
77) Bromobenzene	11.195	156	23864	20.228	ug/l	100
78) n-propylbenzene	11.299	91	120311	20.363	ug/l	100
79) 2-Chlorotoluene	11.360	91	76316	20.025	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	89305	21.037	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8590	17.801	ug/l	86
82) 4-Chlorotoluene	11.451	91	87167	20.625	ug/l	99
83) tert-Butylbenzene	11.713	119	89128	20.843	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	89380	20.791	ug/l	100
85) sec-Butylbenzene	11.890	105	110457	21.038	ug/l	99
86) p-Isopropyltoluene	12.006	119	90318	20.840	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	43379	20.148	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	43071	19.588	ug/l	97
89) n-Butylbenzene	12.329	91	78349	20.610	ug/l	99
90) Hexachloroethane	12.536	117	15040	19.698	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	45968	21.276	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8285	21.002	ug/l	96
93) 1,2,4-Trichlorobenzene	13.585	180	25402	20.470	ug/l	100
94) Hexachlorobutadiene	13.719	225	11133	20.542	ug/l	98
95) Naphthalene	13.774	128	94703	20.808	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	26254	20.504	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046463.D
Acq On : 03 Jun 2025 11:24
Operator : JC/MD
Sample : VX0603WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Quant Time: Jun 04 01:23:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046463.D
Acq On : 03 Jun 2025 11:24
Operator : JC/MD
Sample : VX0603WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Quant Time: Jun 04 01:23:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:168 Resp: 89858
Ion Ratio Lower Upper
168 100
99 69.4 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 16.566 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 85 Resp: 22784
Ion Ratio Lower Upper
85 100
87 30.3 15.7 47.1









#3
Chloromethane
Concen: 17.010 ug/l
RT: 1.301 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 50 Resp: 22687
Ion Ratio Lower Upper
50 100
52 32.7 25.4 38.2



#4
Vinyl Chloride
Concen: 16.617 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 62 Resp: 20626
Ion Ratio Lower Upper
62 100
64 31.6 25.2 37.8









#5
Bromomethane
Concen: 16.878 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 94 Resp: 9717
Ion Ratio Lower Upper
94 100
96 92.9 72.8 109.2



#6
Chloroethane
Concen: 19.070 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 64 Resp: 12637
Ion Ratio Lower Upper
64 100
66 35.3 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 18.840 ug/l
 RT: 1.880 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01


Tgt Ion:101 Resp: 34563
 Ion Ratio Lower Upper
 101 100
 103 65.6 51.0 76.4



#8
 Diethyl Ether
 Concen: 18.602 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24

Tgt Ion: 74 Resp: 11617
 Ion Ratio Lower Upper
 74 100
 45 109.2 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.733 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:101 Resp: 21267
Ion Ratio Lower Upper
101 100
85 47.1 38.6 58.0
151 72.0 55.2 82.8



#10
Methyl Iodide
Concen: 17.425 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:142 Resp: 23408
Ion Ratio Lower Upper
142 100
127 53.2 41.7 62.5
141 14.8 11.5 17.3









#11
Tert butyl alcohol
Concen: 111.778 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 59 Resp: 26285
Ion Ratio Lower Upper
59 100
57 10.8 8.6 12.8



#12
1,1-Dichloroethene
Concen: 17.944 ug/l
RT: 2.313 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 96 Resp: 19119
Ion Ratio Lower Upper
96 100
61 181.4 146.2 219.2
98 63.3 51.0 76.6





#13

Acrolein

Concen: 118.151 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion: 56 Resp: 31641

Ion Ratio Lower Upper

56 100

55 71.0 56.2 84.4



#14

Allyl chloride

Concen: 19.668 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 41 Resp: 40050


Ion Ratio Lower Upper

41 100

39 73.8 60.6 90.8

76 30.3 24.9 37.3









#15
Acrylonitrile
Concen: 105.292 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 53 Resp: 70799
Ion Ratio Lower Upper
53 100
52 84.0 65.3 97.9
51 38.2 29.8 44.8



#16
Acetone
Concen: 103.724 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 43 Resp: 69671
Ion Ratio Lower Upper
43 100
58 25.9 21.2 31.8





#17

Carbon Disulfide

Concen: 13.313 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion: 76 Resp: 33630

Ion Ratio Lower Upper

76 100

78 9.9 7.0 10.4



#18

Methyl Acetate

Concen: 27.131 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 43 Resp: 42289

Ion Ratio Lower Upper

43 100

74 21.0 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 20.340 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 73 Resp: 75981
Ion Ratio Lower Upper
73 100
57 21.5 17.7 26.5



#20
Methylene Chloride
Concen: 18.353 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 84 Resp: 23624
Ion Ratio Lower Upper
84 100
49 136.2 113.9 170.9
51 43.7 33.5 50.3
86 65.8 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 17.980 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 96 Resp: 19266
Ion Ratio Lower Upper
96 100
61 154.9 119.5 179.3
98 66.9 50.0 75.0



#22
Diisopropyl ether
Concen: 20.983 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 45 Resp: 82535
Ion Ratio Lower Upper
45 100
43 89.6 66.6 100.0
87 24.4 19.8 29.6
59 11.2 8.6 12.8









#23
 Vinyl Acetate
 Concen: 96.984 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01


Tgt Ion: 43 Resp: 335531
 Ion Ratio Lower Upper
 43 100
 86 9.2 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 20.493 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24

Tgt Ion: 63 Resp: 44898
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.6 10.8
 100 4.3 2.1 6.3









#25
2-Butanone
Concen: 106.672 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 43 Resp: 104022
Ion Ratio Lower Upper
43 100
72 24.7 18.4 27.6



#26
2,2-Dichloropropane
Concen: 19.283 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 77 Resp: 33066
Ion Ratio Lower Upper
77 100
97 21.5 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 20.421 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01

Tgt Ion: 96 Resp: 26341




Ion	Ratio	Lower	Upper
96	100		
61	158.3	0.0	322.8
98	63.1	0.0	129.0




#28
 Bromochloromethane
 Concen: 21.563 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24

Tgt Ion: 49 Resp: 22740

Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	4.0
130	68.1	56.2	84.2









#29
Tetrahydrofuran
Concen: 105.894 ug/l
RT: 5.001 min Scan# 642
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:	42	Resp:	64707
Ion	Ratio	Lower	Upper
42	100		
72	41.6	33.7	50.5
71	38.0	31.4	47.2



#30
Chloroform
Concen: 20.890 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24


Tgt Ion:	83	Resp:	47703
Ion	Ratio	Lower	Upper
83	100		
85	66.1	49.3	73.9








#31
Cyclohexane
Concen: 17.746 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01







Tgt Ion:	56	Resp:	35430
Ion	Ratio	Lower	Upper
56	100		
69	28.4	24.4	36.6
84	80.0	66.9	100.3



#32
1,1,1-Trichloroethane
Concen: 20.093 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:	97	Resp:	39775
Ion	Ratio	Lower	Upper
97	100		
99	64.3	51.8	77.6
61	49.0	40.1	60.1





#33

1,2-Dichloroethane-d4

Concen: 48.808 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion: 65 Resp: 81765

Ion Ratio Lower Upper

65 100

67 49.1 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 114 Resp: 159016


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 18.3 0.0 33.6









#35
Dibromofluoromethane
Concen: 50.381 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:113 Resp: 57691
Ion Ratio Lower Upper
113 100
111 102.8 83.1 124.7
192 16.8 13.3 19.9



#36
1,1-Dichloropropene
Concen: 18.533 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 75 Resp: 28514
Ion Ratio Lower Upper
75 100
110 31.8 16.3 48.9
77 29.9 24.3 36.5









#37
Ethyl Acetate
Concen: 19.653 ug/l
RT: 4.715 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:	43	Resp:	37358
Ion	Ratio	Lower	Upper
43	100		
61	12.8	10.3	15.5
70	9.6	7.9	11.9



#38
Carbon Tetrachloride
Concen: 18.497 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:	117	Resp:	31976
Ion	Ratio	Lower	Upper
117	100		
119	89.5	75.2	112.8
121	28.5	24.2	36.4





#39

Methylcyclohexane

Concen: 17.710 ug/l

RT: 7.373 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion: 83 Resp: 35078

Ion Ratio Lower Upper

83 100

55 78.7 64.7 97.1

98 47.1 37.4 56.2



#40

Benzene

Concen: 19.490 ug/l

RT: 6.037 min Scan# 812

Delta R.T. 0.006 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 78 Resp: 87833

Ion Ratio Lower Upper

78 100

77 24.8 19.0 28.4









#41
Methacrylonitrile
Concen: 21.836 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:	41	Resp:	21713
Ion	Ratio	Lower	Upper
41	100		
39	60.2	47.2	70.8
67	66.9	50.7	76.1
52	33.3	24.6	37.0



#42
1,2-Dichloroethane
Concen: 20.187 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:	62	Resp:	39263
Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	15.2









#43
Isopropyl Acetate
Concen: 20.448 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 43 Resp: 59297
Ion Ratio Lower Upper
43 100
61 17.8 14.3 21.5
87 11.0 9.5 14.3



#44
Trichloroethene
Concen: 18.946 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:130 Resp: 20550
Ion Ratio Lower Upper
130 100
95 104.3 0.0 204.2









#45
1,2-Dichloropropane
Concen: 20.917 ug/l
RT: 7.427 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 63 Resp: 23439
Ion Ratio Lower Upper
63 100
65 29.9 24.1 36.1



#46
Dibromomethane
Concen: 19.808 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 93 Resp: 17506
Ion Ratio Lower Upper
93 100
95 82.4 65.6 98.4
174 84.7 68.2 102.2









#47
Bromodichloromethane
Concen: 20.043 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 83 Resp: 34890
Ion Ratio Lower Upper
83 100
85 67.7 50.5 75.7
127 8.9 6.5 9.7



#48
Methyl methacrylate
Concen: 21.143 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 41 Resp: 31312
Ion Ratio Lower Upper
41 100
69 71.3 58.5 87.7
39 64.5 51.7 77.5





#49

1,4-Dioxane

Concen: 439.865 ug/l

RT: 7.659 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion: 88 Resp: 12369

Ion Ratio Lower Upper

88 100

43 42.9 33.4 50.2

58 78.1 58.6 88.0



#50

Toluene-d8

Concen: 47.791 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 98 Resp: 189408

Ion Ratio Lower Upper

98 100

100 64.9 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 109.285 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 43 Resp: 210362
Ion Ratio Lower Upper
43 100
58 35.2 28.9 43.3



#52
Toluene
Concen: 19.864 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 92 Resp: 54891
Ion Ratio Lower Upper
92 100
91 171.0 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.227 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion: 75 Resp: 29748

Ion Ratio Lower Upper

75 100

77 31.3 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 19.928 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion: 75 Resp: 34079


Ion Ratio Lower Upper

75 100

77 32.2 25.4 38.0

39 61.1 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 21.488 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 97	Resp: 23413
Ion Ratio	Lower Upper
97 100	
83 84.0	69.0 103.6
85 49.3	44.3 66.5
99 59.2	47.8 71.6



#56
Ethyl methacrylate
Concen: 21.719 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 69	Resp: 37717
Ion Ratio	Lower Upper
69 100	
41 77.2	60.8 91.2
39 47.0	39.0 58.6





#57

1,3-Dichloropropane

Concen: 20.384 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion: 76 Resp: 39888

Ion Ratio Lower Upper

76 100

78 33.3 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 106.886 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 63 Resp: 94629

Ion Ratio Lower Upper

63 100

106 27.0 21.5 32.3









#59
2-Hexanone
Concen: 111.507 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 43 Resp: 158797
Ion Ratio Lower Upper
43 100
58 48.1 24.9 74.6



#60
Dibromochloromethane
Concen: 20.516 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 129 Resp: 24550
Ion Ratio Lower Upper
129 100
127 75.9 39.3 117.8









#61
1,2-Dibromoethane
Concen: 20.220 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:107 Resp: 22898
Ion Ratio Lower Upper
107 100
109 96.1 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 50.637 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 95 Resp: 76982
Ion Ratio Lower Upper
95 100
174 66.7 0.0 135.8
176 64.5 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:117 Resp: 137079
Ion Ratio Lower Upper
117 100
82 61.3 50.6 76.0
119 32.4 25.8 38.6



#64
Tetrachloroethene
Concen: 19.782 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:164 Resp: 19186
Ion Ratio Lower Upper
164 100
166 127.0 105.0 157.6
129 98.9 83.5 125.3
131 99.7 76.5 114.7





#65

Chlorobenzene

Concen: 20.209 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion:112 Resp: 60633

Ion Ratio Lower Upper

112 100

114 31.5 25.4 38.2



#66

1,1,1,2-Tetrachloroethane

Concen: 20.627 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion:131 Resp: 21133


Ion Ratio Lower Upper

131 100

133 96.2 47.3 141.9

119 65.6 31.6 95.0





#67

Ethyl Benzene

Concen: 20.434 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion: 91 Resp: 108068

Ion Ratio Lower Upper

91 100

106 29.2 23.4 35.2



#68

m/p-Xylenes

Concen: 40.424 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion:106 Resp: 78193

Ion Ratio Lower Upper

106 100

91 217.3 171.2 256.8









#69
o-Xylene
Concen: 20.987 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:106 Resp: 39576
Ion Ratio Lower Upper
106 100
91 229.4 112.7 338.1



#70
Styrene
Concen: 21.169 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:104 Resp: 65394
Ion Ratio Lower Upper
104 100
78 58.3 45.7 68.5
103 55.4 43.7 65.5









#71
Bromoform
Concen: 19.868 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:173 Resp: 15282
Ion Ratio Lower Upper
173 100
175 47.7 23.4 70.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:152 Resp: 65270
Ion Ratio Lower Upper
152 100
115 76.1 46.9 140.7
150 163.5 0.0 351.0









#73
Isopropylbenzene
Concen: 21.051 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:105 Resp: 106972
Ion Ratio Lower Upper
105 100
120 26.0 12.8 38.4



#74
N-ethyl acetate
Concen: 21.221 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 43 Resp: 53285
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 23.8 18.7 28.1
61 21.4 17.1 25.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 21.421 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBS01


Tgt Ion: 83 Resp: 38144
 Ion Ratio Lower Upper
 83 100
 131 9.7 5.0 14.9
 85 63.2 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 25.735 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046463.D
 Acq: 03 Jun 2025 11:24

Tgt Ion: 75 Resp: 40431
 Ion Ratio Lower Upper
 75 100
 77 32.9 20.5 61.5









#77
Bromobenzene
Concen: 20.228 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:156 Resp: 23864
Ion Ratio Lower Upper
156 100
77 185.5 92.8 278.3
158 96.5 48.8 146.3



#78
n-propylbenzene
Concen: 20.363 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 91 Resp: 120311
Ion Ratio Lower Upper
91 100
120 21.4 10.8 32.4





#79

2-Chlorotoluene

Concen: 20.025 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBS01

Tgt Ion: 91 Resp: 76316

Ion Ratio Lower Upper

91 100

126 32.0 15.6 46.7



#80

1,3,5-Trimethylbenzene

Concen: 21.037 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046463.D




Acq: 03 Jun 2025 11:24


Tgt Ion: 105 Resp: 89305

Ion Ratio Lower Upper

105 100

120 45.5 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 17.801 ug/l
RT: 11.018 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 75 Resp: 8590
Ion Ratio Lower Upper
75 100
53 146.4 100.5 150.7
89 48.4 37.0 55.6



#82
4-Chlorotoluene
Concen: 20.625 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 91 Resp: 87167
Ion Ratio Lower Upper
91 100
126 27.1 13.3 39.8









#83
tert-Butylbenzene
Concen: 20.843 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:119 Resp: 89128
Ion Ratio Lower Upper
119 100
91 65.8 32.9 98.7
134 23.1 11.4 34.1



#84
1,2,4-Trimethylbenzene
Concen: 20.791 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:105 Resp: 89380
Ion Ratio Lower Upper
105 100
120 42.5 21.2 63.6









#85
sec-Butylbenzene
Concen: 21.038 ug/l
RT: 11.890 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:105 Resp: 110457
Ion Ratio Lower Upper
105 100
134 18.8 9.7 29.1



#86
p-Isopropyltoluene
Concen: 20.840 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:119 Resp: 90318
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.2 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 20.148 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBS01





Tgt Ion:146 Resp: 43379

Ion Ratio Lower Upper

146 100

111 45.0 22.1 66.3

148 63.3 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 19.588 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046463.D

Acq: 03 Jun 2025 11:24




Tgt Ion:146 Resp: 43071


Ion Ratio Lower Upper

146 100

111 45.5 21.3 63.9

148 64.8 31.9 95.5









#89
n-Butylbenzene
Concen: 20.610 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion: 91 Resp: 78349
Ion Ratio Lower Upper
91 100
92 52.6 26.9 80.7
134 23.2 11.8 35.3



#90
Hexachloroethane
Concen: 19.698 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 117 Resp: 15040
Ion Ratio Lower Upper
117 100
201 59.8 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 21.276 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:146 Resp: 45968
Ion Ratio Lower Upper
146 100
111 46.1 22.4 67.3
148 63.6 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.002 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion: 75 Resp: 8285
Ion Ratio Lower Upper
75 100
155 66.2 34.9 104.8
157 91.0 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 20.470 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01


Tgt Ion:180 Resp: 25402
Ion Ratio Lower Upper
180 100
182 96.2 47.9 143.8
145 33.6 17.1 51.2



#94
Hexachlorobutadiene
Concen: 20.542 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:225 Resp: 11133
Ion Ratio Lower Upper
225 100
223 61.2 30.8 92.4
227 64.8 30.9 92.7









#95
Naphthalene
Concen: 20.808 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBS01

Tgt Ion:	128	Resp:	94703
Ion	Ratio	Lower	Upper
128	100		
127	13.0	10.4	15.6
129	10.7	8.6	13.0



#96
1,2,3-Trichlorobenzene
Concen: 20.504 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046463.D
Acq: 03 Jun 2025 11:24

Tgt Ion:	180	Resp:	26254
Ion	Ratio	Lower	Upper
180	100		
182	93.0	47.8	143.3
145	36.3	18.1	54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046464.D
 Acq On : 03 Jun 2025 11:50
 Operator : JC/MD
 Sample : VX0603WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01

Quant Time: Jun 04 01:24:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	86032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62365	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80850	50.408	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.820%	
35) Dibromofluoromethane	5.379	113	56126	51.631	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.260%	
50) Toluene-d8	8.647	98	181255	48.175	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.340%	
62) 4-Bromofluorobenzene	11.079	95	72876	50.495	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.000%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	21879	16.616	ug/l	98
3) Chloromethane	1.307	50	20951	16.407	ug/l	99
4) Vinyl Chloride	1.374	62	18990	15.979	ug/l	100
5) Bromomethane	1.599	94	9271	16.819	ug/l	97
6) Chloroethane	1.672	64	11276	17.773	ug/l	100
7) Trichlorofluoromethane	1.880	101	31828	18.121	ug/l	100
8) Diethyl Ether	2.136	74	11245	18.807	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.325	101	20804	19.140	ug/l	99
10) Methyl Iodide	2.447	142	22438	17.446	ug/l	99
11) Tert butyl alcohol	2.971	59	26245	116.572	ug/l	100
12) 1,1-Dichloroethene	2.313	96	18188	17.829	ug/l	92
13) Acrolein	2.233	56	28734	112.067	ug/l	97
14) Allyl chloride	2.660	41	38609	19.803	ug/l	96
15) Acrylonitrile	3.062	53	71487	111.043	ug/l	99
16) Acetone	2.380	43	68723	106.863	ug/l	99
17) Carbon Disulfide	2.508	76	32208	13.317	ug/l	99
18) Methyl Acetate	2.703	43	41923	28.093	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	76793	21.472	ug/l	100
20) Methylene Chloride	2.782	84	22550	18.298	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	18461	17.995	ug/l	100
22) Diisopropyl ether	3.757	45	81514	21.645	ug/l	98
23) Vinyl Acetate	3.721	43	334608	101.019	ug/l	99
24) 1,1-Dichloroethane	3.605	63	42085	20.064	ug/l	97
25) 2-Butanone	4.556	43	104645	112.084	ug/l	100
26) 2,2-Dichloropropane	4.465	77	31623	19.261	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	24764	20.052	ug/l	99
28) Bromochloromethane	4.897	49	21879	21.669	ug/l	97
29) Tetrahydrofuran	5.001	42	63927	109.270	ug/l	99
30) Chloroform	5.093	83	45359	20.747	ug/l	97
31) Cyclohexane	5.458	56	33045	17.288	ug/l	93
32) 1,1,1-Trichloroethane	5.373	97	36112	19.054	ug/l	97
36) 1,1-Dichloropropene	5.684	75	26077	17.854	ug/l	97
37) Ethyl Acetate	4.715	43	37561	20.815	ug/l	100
38) Carbon Tetrachloride	5.672	117	30493	18.581	ug/l	99
39) Methylcyclohexane	7.379	83	32831	17.460	ug/l	97
40) Benzene	6.031	78	82785	19.351	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046464.D
 Acq On : 03 Jun 2025 11:50
 Operator : JC/MD
 Sample : VX0603WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01

Quant Time: Jun 04 01:24:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	22327	23.652	ug/l	99
42) 1,2-Dichloroethane	6.080	62	38662	20.939	ug/l	98
43) Isopropyl Acetate	6.336	43	61515	22.345	ug/l	98
44) Trichloroethene	7.123	130	19772	19.202	ug/l	91
45) 1,2-Dichloropropane	7.428	63	22490	21.141	ug/l	97
46) Dibromomethane	7.580	93	17488	20.843	ug/l	98
47) Bromodichloromethane	7.818	83	34508	20.882	ug/l	99
48) Methyl methacrylate	7.690	41	30780	21.893	ug/l	99
49) 1,4-Dioxane	7.659	88	12806	479.715	ug/l	96
51) 4-Methyl-2-Pentanone	8.568	43	210042	114.943	ug/l	98
52) Toluene	8.714	92	51638	19.684	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	28911	19.683	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	33321	20.525	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	22943	22.181	ug/l	98
56) Ethyl methacrylate	9.116	69	37370	22.668	ug/l	98
57) 1,3-Dichloropropane	9.305	76	39110	21.053	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	93351	111.071	ug/l	99
59) 2-Hexanone	9.427	43	155374	114.927	ug/l	99
60) Dibromochloromethane	9.519	129	23830	20.978	ug/l	98
61) 1,2-Dibromoethane	9.604	107	22450	20.882	ug/l	99
64) Tetrachloroethene	9.269	164	17909	19.400	ug/l	92
65) Chlorobenzene	10.073	112	57186	20.025	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	20660	21.186	ug/l	99
67) Ethyl Benzene	10.189	91	102509	20.363	ug/l	98
68) m/p-Xylenes	10.299	106	74026	40.207	ug/l	97
69) o-Xylene	10.640	106	38408	21.398	ug/l	100
70) Styrene	10.653	104	62236	21.167	ug/l	99
71) Bromoform	10.799	173	15063	20.574	ug/l #	100
73) Isopropylbenzene	10.957	105	101859	20.979	ug/l	100
74) N-amyl acetate	10.842	43	51555	21.488	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	36254	21.308	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	37965	25.291	ug/l	88
77) Bromobenzene	11.195	156	22557	20.011	ug/l	99
78) n-propylbenzene	11.299	91	115006	20.371	ug/l	100
79) 2-Chlorotoluene	11.360	91	73129	20.083	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	85390	21.051	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8927	19.361	ug/l	94
82) 4-Chlorotoluene	11.451	91	83070	20.571	ug/l	99
83) tert-Butylbenzene	11.713	119	84344	20.643	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	84267	20.514	ug/l	100
85) sec-Butylbenzene	11.890	105	106367	21.203	ug/l	100
86) p-Isopropyltoluene	12.006	119	86575	20.907	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	41494	20.170	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	42409	20.186	ug/l	98
89) n-Butylbenzene	12.329	91	75206	20.705	ug/l	99
90) Hexachloroethane	12.536	117	14400	19.739	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	43740	21.188	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8644	22.933	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	24446	20.617	ug/l	100
94) Hexachlorobutadiene	13.725	225	10733	20.726	ug/l	99
95) Naphthalene	13.774	128	93399	21.477	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	25957	21.216	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046464.D
Acq On : 03 Jun 2025 11:50
Operator : JC/MD
Sample : VX0603WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Quant Time: Jun 04 01:24:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046464.D
Acq On : 03 Jun 2025 11:50
Operator : JC/MD
Sample : VX0603WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01

Quant Time: Jun 04 01:24:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:168 Resp: 86032
Ion Ratio Lower Upper
168 100
99 74.7 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 16.616 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 85 Resp: 21879
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1









#3
Chloromethane
Concen: 16.407 ug/l
RT: 1.307 min Scan# 36
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 50 Resp: 20951
Ion Ratio Lower Upper
50 100
52 32.3 25.4 38.2



#4
Vinyl Chloride
Concen: 15.979 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 62 Resp: 18990
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8









#5
Bromomethane
Concen: 16.819 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 94 Resp: 9271
Ion Ratio Lower Upper
94 100
96 88.0 72.8 109.2



#6
Chloroethane
Concen: 17.773 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 64 Resp: 11276
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 18.121 ug/l
 RT: 1.880 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01


Tgt Ion:101 Resp: 31828
 Ion Ratio Lower Upper
 101 100
 103 63.7 51.0 76.4



#8
 Diethyl Ether
 Concen: 18.807 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50

Tgt Ion: 74 Resp: 11245
 Ion Ratio Lower Upper
 74 100
 45 112.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.140 ug/l
RT: 2.325 min Scan# 20
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:101 Resp: 20804
Ion Ratio Lower Upper
101 100
85 47.2 38.6 58.0
151 70.0 55.2 82.8



#10
Methyl Iodide
Concen: 17.446 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:142 Resp: 22438
Ion Ratio Lower Upper
142 100
127 51.8 41.7 62.5
141 13.7 11.5 17.3









#11
Tert butyl alcohol
Concen: 116.572 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 59 Resp: 26245
Ion Ratio Lower Upper
59 100
57 10.7 8.6 12.8



#12
1,1-Dichloroethene
Concen: 17.829 ug/l
RT: 2.313 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 96 Resp: 18188
Ion Ratio Lower Upper
96 100
61 194.8 146.2 219.2
98 68.6 51.0 76.6









#13
Acrolein
Concen: 112.067 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 56 Resp: 28734
Ion Ratio Lower Upper
56 100
55 72.4 56.2 84.4



#14
Allyl chloride
Concen: 19.803 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 41 Resp: 38609
Ion Ratio Lower Upper
41 100
39 72.3 60.6 90.8
76 28.5 24.9 37.3









#15
Acrylonitrile
Concen: 111.043 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:	53	Resp:	71487
Ion	Ratio	Lower	Upper
53	100		
52	83.2	65.3	97.9
51	37.6	29.8	44.8



#16
Acetone
Concen: 106.863 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:	43	Resp:	68723
Ion	Ratio	Lower	Upper
43	100		
58	26.1	21.2	31.8





#17

Carbon Disulfide

Concen: 13.317 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBSD01

Tgt Ion: 76 Resp: 32208

Ion Ratio Lower Upper

76 100

78 8.9 7.0 10.4



#18

Methyl Acetate

Concen: 28.093 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 43 Resp: 41923

Ion Ratio Lower Upper

43 100

74 21.2 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 21.472 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 73 Resp: 76793
Ion Ratio Lower Upper
73 100
57 21.9 17.7 26.5



#20
Methylene Chloride
Concen: 18.298 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 84 Resp: 22550
Ion Ratio Lower Upper
84 100
49 143.8 113.9 170.9
51 42.1 33.5 50.3
86 68.5 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 17.995 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 96 Resp: 18461
Ion Ratio Lower Upper
96 100
61 149.3 119.5 179.3
98 63.5 50.0 75.0



#22
Diisopropyl ether
Concen: 21.645 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 45 Resp: 81514
Ion Ratio Lower Upper
45 100
43 84.8 66.6 100.0
87 25.5 19.8 29.6
59 12.0 8.6 12.8









#23
 Vinyl Acetate
 Concen: 101.019 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01


Tgt Ion: 43 Resp: 334608
 Ion Ratio Lower Upper
 43 100
 86 9.1 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 20.064 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50

Tgt Ion: 63 Resp: 42085
 Ion Ratio Lower Upper
 63 100
 98 5.8 3.6 10.8
 100 4.6 2.1 6.3









#25
2-Butanone
Concen: 112.084 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 43 Resp: 104645
Ion Ratio Lower Upper
43 100
72 23.0 18.4 27.6



#26
2,2-Dichloropropane
Concen: 19.261 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 77 Resp: 31623
Ion Ratio Lower Upper
77 100
97 20.4 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 20.052 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01

Tgt Ion: 96 Resp: 24764




Ion	Ratio	Lower	Upper
96	100		
61	159.2	0.0	322.8
98	64.4	0.0	129.0




#28
 Bromochloromethane
 Concen: 21.669 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50

Tgt Ion: 49 Resp: 21879

Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	4.0
130	67.6	56.2	84.2









#29
Tetrahydrofuran
Concen: 109.270 ug/l
RT: 5.001 min Scan# 64
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 42 Resp: 63927
Ion Ratio Lower Upper
42 100
72 42.4 33.7 50.5
71 38.4 31.4 47.2



#30
Chloroform
Concen: 20.747 ug/l
RT: 5.093 min Scan# 657
Delta R.T. 0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 83 Resp: 45359
Ion Ratio Lower Upper
83 100
85 64.1 49.3 73.9









#31
Cyclohexane
Concen: 17.288 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 56 Resp: 33045
Ion Ratio Lower Upper
56 100
69 28.2 24.4 36.6
84 76.0 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 19.054 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 97 Resp: 36112
Ion Ratio Lower Upper
97 100
99 66.4 51.8 77.6
61 52.1 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 50.408 ug/l
 RT: 5.952 min Scan# 798
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01


Tgt Ion: 65 Resp: 80850
 Ion Ratio Lower Upper
 65 100
 67 49.6 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50

Tgt Ion: 114 Resp: 150958
 Ion Ratio Lower Upper
 114 100
 63 22.8 0.0 49.2
 88 18.1 0.0 33.6









#35
Dibromofluoromethane
Concen: 51.631 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:113 Resp: 56126
Ion Ratio Lower Upper
113 100
111 101.9 83.1 124.7
192 16.8 13.3 19.9



#36
1,1-Dichloropropene
Concen: 17.854 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 75 Resp: 26077
Ion Ratio Lower Upper
75 100
110 33.8 16.3 48.9
77 32.1 24.3 36.5









#37
Ethyl Acetate
Concen: 20.815 ug/l
RT: 4.715 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:	43	Resp:	37561
Ion	Ratio	Lower	Upper
43	100		
61	12.9	10.3	15.5
70	9.8	7.9	11.9



#38
Carbon Tetrachloride
Concen: 18.581 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:	117	Resp:	30493
Ion	Ratio	Lower	Upper
117	100		
119	94.3	75.2	112.8
121	29.2	24.2	36.4





#39

Methylcyclohexane

Concen: 17.460 ug/l

RT: 7.379 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01





Tgt Ion: 83 Resp: 32831

Ion Ratio Lower Upper

83 100

55 84.2 64.7 97.1

98 47.7 37.4 56.2



#40

Benzene

Concen: 19.351 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 78 Resp: 82785

Ion Ratio Lower Upper

78 100

77 25.6 19.0 28.4









#41
Methacrylonitrile
Concen: 23.652 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:	41	Resp:	22327
Ion	Ratio	Lower	Upper
41	100		
39	57.1	47.2	70.8
67	63.7	50.7	76.1
52	30.4	24.6	37.0



#42
1,2-Dichloroethane
Concen: 20.939 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:	62	Resp:	38662
Ion	Ratio	Lower	Upper
62	100		
98	7.0	0.0	15.2









#43
Isopropyl Acetate
Concen: 22.345 ug/l
RT: 6.336 min Scan# 861
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 43 Resp: 61515
Ion Ratio Lower Upper
43 100
61 17.6 14.3 21.5
87 10.7 9.5 14.3



#44
Trichloroethene
Concen: 19.202 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:130 Resp: 19772
Ion Ratio Lower Upper
130 100
95 111.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 21.141 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBSD01

Tgt Ion: 63 Resp: 22490

Ion Ratio Lower Upper

63 100

65 31.6 24.1 36.1



#46

Dibromomethane

Concen: 20.843 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 93 Resp: 17488


Ion Ratio Lower Upper

93 100

95 79.4 65.6 98.4

174 83.5 68.2 102.2









#47
Bromodichloromethane
Concen: 20.882 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 83 Resp: 34508
Ion Ratio Lower Upper
83 100
85 62.4 50.5 75.7
127 8.3 6.5 9.7



#48
Methyl methacrylate
Concen: 21.893 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 41 Resp: 30780
Ion Ratio Lower Upper
41 100
69 71.7 58.5 87.7
39 64.9 51.7 77.5









#49
1,4-Dioxane
Concen: 479.715 ug/l
RT: 7.659 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 88 Resp: 12806
Ion Ratio Lower Upper
88 100
43 41.2 33.4 50.2
58 77.5 58.6 88.0



#50
Toluene-d8
Concen: 48.175 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 98 Resp: 181255
Ion Ratio Lower Upper
98 100
100 64.6 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 114.943 ug/l
RT: 8.568 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 43 Resp: 210042
Ion Ratio Lower Upper
43 100
58 35.2 28.9 43.3



#52
Toluene
Concen: 19.684 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 92 Resp: 51638
Ion Ratio Lower Upper
92 100
91 173.7 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.683 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBSD01

Tgt Ion: 75 Resp: 28911

Ion Ratio Lower Upper

75 100

77 31.8 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 20.525 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion: 75 Resp: 33321


Ion Ratio Lower Upper

75 100

77 33.8 25.4 38.0

39 61.4 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 22.181 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:	97	Resp:	22943
Ion	Ratio	Lower	Upper
97	100		
83	82.1	69.0	103.6
85	55.0	44.3	66.5
99	59.6	47.8	71.6



#56
Ethyl methacrylate
Concen: 22.668 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:	69	Resp:	37370
Ion	Ratio	Lower	Upper
69	100		
41	77.0	60.8	91.2
39	47.3	39.0	58.6





#57

1,3-Dichloropropane

Concen: 21.053 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBSD01

Tgt Ion: 76 Resp: 39110

Ion Ratio Lower Upper

76 100

78 32.1 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 111.071 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 63 Resp: 93351

Ion Ratio Lower Upper

63 100

106 26.5 21.5 32.3









#59
2-Hexanone
Concen: 114.927 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 43 Resp: 155374
Ion Ratio Lower Upper
43 100
58 49.1 24.9 74.6



#60
Dibromochloromethane
Concen: 20.978 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 129 Resp: 23830
Ion Ratio Lower Upper
129 100
127 76.8 39.3 117.8









#61
1,2-Dibromoethane
Concen: 20.882 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:107 Resp: 22450
Ion Ratio Lower Upper
107 100
109 92.8 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 50.495 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 95 Resp: 72876
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 63.7 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:117 Resp: 130475
Ion Ratio Lower Upper
117 100
82 61.7 50.6 76.0
119 31.8 25.8 38.6



#64
Tetrachloroethene
Concen: 19.400 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:164 Resp: 17909
Ion Ratio Lower Upper
164 100
166 139.6 105.0 157.6
129 111.2 83.5 125.3
131 105.5 76.5 114.7









#65
Chlorobenzene
Concen: 20.025 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:112 Resp: 57186
Ion Ratio Lower Upper
112 100
114 32.5 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 21.186 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:131 Resp: 20660
Ion Ratio Lower Upper
131 100
133 92.6 47.3 141.9
119 63.2 31.6 95.0





#67

Ethyl Benzene

Concen: 20.363 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :





VX0603WBSD01

Tgt Ion: 91 Resp: 102509

Ion Ratio Lower Upper

91 100

106 28.1 23.4 35.2



#68

m/p-Xylenes

Concen: 40.207 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion: 106 Resp: 74026

Ion Ratio Lower Upper

106 100

91 219.2 171.2 256.8









#69
o-Xylene
Concen: 21.398 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:106 Resp: 38408
Ion Ratio Lower Upper
106 100
91 225.9 112.7 338.1



#70
Styrene
Concen: 21.167 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:104 Resp: 62236
Ion Ratio Lower Upper
104 100
78 58.1 45.7 68.5
103 55.6 43.7 65.5









#71
Bromoform
Concen: 20.574 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:173 Resp: 15063
Ion Ratio Lower Upper
173 100
175 46.6 23.4 70.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion:152 Resp: 62365
Ion Ratio Lower Upper
152 100
115 76.9 46.9 140.7
150 162.0 0.0 351.0









#73
Isopropylbenzene
Concen: 20.979 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:105 Resp: 101859
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4



#74
N-ethyl acetate
Concen: 21.488 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 43 Resp: 51555
Ion Ratio Lower Upper
43 100
70 37.7 30.9 46.3
55 22.9 18.7 28.1
61 21.2 17.1 25.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 21.308 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01


Tgt Ion: 83 Resp: 36254
 Ion Ratio Lower Upper
 83 100
 131 9.6 5.0 14.9
 85 63.3 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 25.291 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50

Tgt Ion: 75 Resp: 37965
 Ion Ratio Lower Upper
 75 100
 77 33.5 20.5 61.5









#77
Bromobenzene
Concen: 20.011 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:156 Resp: 22557
Ion Ratio Lower Upper
156 100
77 188.0 92.8 278.3
158 97.3 48.8 146.3



#78
n-propylbenzene
Concen: 20.371 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 91 Resp: 115006
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 20.083 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 91 Resp: 73129
Ion Ratio Lower Upper
91 100
126 30.5 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 21.051 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 105 Resp: 85390
Ion Ratio Lower Upper
105 100
120 45.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 19.361 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 75 Resp: 8927
Ion Ratio Lower Upper
75 100
53 134.1 100.5 150.7
89 48.2 37.0 55.6



#82
4-Chlorotoluene
Concen: 20.571 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 91 Resp: 83070
Ion Ratio Lower Upper
91 100
126 26.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 20.643 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01





Tgt Ion:119 Resp: 84344

Ion Ratio Lower Upper

119 100

91 66.3 32.9 98.7

134 23.3 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 20.514 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046464.D




Acq: 03 Jun 2025 11:50


Tgt Ion:105 Resp: 84267

Ion Ratio Lower Upper

105 100

120 42.7 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 21.203 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0603WBSD01


Tgt Ion:105 Resp: 106367
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 20.907 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046464.D
 Acq: 03 Jun 2025 11:50

Tgt Ion:119 Resp: 86575
 Ion Ratio Lower Upper
 119 100
 134 24.9 12.5 37.5
 91 27.4 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 20.170 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01





Tgt Ion:146 Resp: 41494

Ion Ratio Lower Upper

146 100

111 45.2 22.1 66.3

148 65.4 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 20.186 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion:146 Resp: 42409


Ion Ratio Lower Upper

146 100

111 43.4 21.3 63.9

148 65.0 31.9 95.5









#89
n-Butylbenzene
Concen: 20.705 ug/l
RT: 12.329 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion: 91 Resp: 75206
Ion Ratio Lower Upper
91 100
92 52.9 26.9 80.7
134 23.6 11.8 35.3



#90
Hexachloroethane
Concen: 19.739 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 117 Resp: 14400
Ion Ratio Lower Upper
117 100
201 62.0 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 21.188 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50




Instrument :
MSVOA_X
ClientSampleId :
VX0603WBSD01


Tgt Ion:146 Resp: 43740
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 65.1 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 22.933 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046464.D
Acq: 03 Jun 2025 11:50

Tgt Ion: 75 Resp: 8644
Ion Ratio Lower Upper
75 100
155 66.7 34.9 104.8
157 86.0 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 20.617 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01





Tgt Ion:180 Resp: 24446

Ion Ratio Lower Upper

180 100

182 95.5 47.9 143.8

145 34.4 17.1 51.2



#94

Hexachlorobutadiene

Concen: 20.726 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion:225 Resp: 10733


Ion Ratio Lower Upper

225 100

223 63.3 30.8 92.4

227 61.8 30.9 92.7





#95

Naphthalene

Concen: 21.477 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50

Instrument :

MSVOA_X

ClientSampleId :

VX0603WBSD01





Tgt Ion:128 Resp: 93399

Ion Ratio Lower Upper

128 100

127 12.6 10.4 15.6

129 10.8 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 21.216 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046464.D

Acq: 03 Jun 2025 11:50




Tgt Ion:180 Resp: 25957

Ion Ratio Lower Upper

180 100

182 93.4 47.8 143.3

145 35.9 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	80122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	143610	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61048	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	72135	48.292	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.580%
35) Dibromofluoromethane	5.379	113	51124	49.436	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	8.647	98	168819	47.165	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.340%
62) 4-Bromofluorobenzene	11.079	95	70871	51.619	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.240%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	49352	40.244	ug/l	97
3) Chloromethane	1.307	50	47624	40.046	ug/l	97
4) Vinyl Chloride	1.374	62	46111	41.662	ug/l	99
5) Bromomethane	1.593	94	18265	35.580	ug/l	97
6) Chloroethane	1.666	64	27222	46.072	ug/l	97
7) Trichlorofluoromethane	1.874	101	72155	44.111	ug/l	99
8) Diethyl Ether	2.130	74	26541	47.664	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	46349	45.787	ug/l	97
10) Methyl Iodide	2.447	142	59982	50.077	ug/l	99
11) Tert butyl alcohol	2.977	59	69693	332.386	ug/l	99
12) 1,1-Dichloroethene	2.307	96	42762	45.010	ug/l	99
13) Acrolein	2.239	56	58249	243.938	ug/l	95
14) Allyl chloride	2.660	41	90860	50.041	ug/l	95
15) Acrylonitrile	3.062	53	166739	278.106	ug/l	98
16) Acetone	2.386	43	163107	272.337	ug/l	100
17) Carbon Disulfide	2.502	76	75405	33.478	ug/l	99
18) Methyl Acetate	2.703	43	101107	72.749	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	186100	55.872	ug/l	99
20) Methylene Chloride	2.782	84	52669	45.890	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	43658	45.695	ug/l	97
22) Diisopropyl ether	3.764	45	191730	54.666	ug/l	93
23) Vinyl Acetate	3.721	43	820777	266.071	ug/l	100
24) 1,1-Dichloroethane	3.605	63	100414	51.402	ug/l	100
25) 2-Butanone	4.556	43	254552	292.758	ug/l	100
26) 2,2-Dichloropropane	4.471	77	62527	40.894	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	57943	50.378	ug/l	98
28) Bromochloromethane	4.891	49	47479	50.493	ug/l	99
29) Tetrahydrofuran	5.007	42	156899	287.968	ug/l	99
30) Chloroform	5.087	83	106714	52.410	ug/l	97
31) Cyclohexane	5.464	56	80308	45.113	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	91763	51.989	ug/l	98
36) 1,1-Dichloropropene	5.690	75	62759	45.167	ug/l	96
37) Ethyl Acetate	4.715	43	91596	53.356	ug/l	100
38) Carbon Tetrachloride	5.672	117	74005	47.403	ug/l	97
39) Methylcyclohexane	7.379	83	79510	44.448	ug/l	99
40) Benzene	6.031	78	201308	49.463	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
 Data File : VX046486.D
 Acq On : 03 Jun 2025 20:25
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Quant Time: Jun 04 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	54177	60.329	ug/l	98
42) 1,2-Dichloroethane	6.080	62	88249	50.240	ug/l	100
43) Isopropyl Acetate	6.342	43	155151	59.242	ug/l	99
44) Trichloroethene	7.123	130	47910	48.910	ug/l	92
45) 1,2-Dichloropropane	7.428	63	54922	54.270	ug/l	97
46) Dibromomethane	7.580	93	40591	50.854	ug/l	99
47) Bromodichloromethane	7.818	83	82121	52.237	ug/l	99
48) Methyl methacrylate	7.690	41	78532	58.715	ug/l	99
49) 1,4-Dioxane	7.665	88	30030	1182.487	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	511610	294.298	ug/l	100
52) Toluene	8.714	92	126020	50.497	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	75146	53.778	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	82115	53.169	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	53770	54.643	ug/l	98
56) Ethyl methacrylate	9.116	69	95030	60.593	ug/l	99
57) 1,3-Dichloropropane	9.305	76	92225	52.186	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	223496	279.526	ug/l	99
59) 2-Hexanone	9.427	43	391242	304.201	ug/l	100
60) Dibromochloromethane	9.519	129	58826	54.434	ug/l	100
61) 1,2-Dibromoethane	9.610	107	54971	53.748	ug/l	97
64) Tetrachloroethene	9.269	164	41027	45.626	ug/l	97
65) Chlorobenzene	10.080	112	138505	49.792	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	50287	52.942	ug/l	98
67) Ethyl Benzene	10.189	91	250449	51.077	ug/l	100
68) m/p-Xylenes	10.299	106	181843	101.398	ug/l	98
69) o-Xylene	10.640	106	93498	53.478	ug/l	99
70) Styrene	10.653	104	155343	54.240	ug/l	99
71) Bromoform	10.799	173	38761	54.353	ug/l #	98
73) Isopropylbenzene	10.957	105	247204	52.013	ug/l	99
74) N-amyl acetate	10.842	43	138341	58.905	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.207	83	89844	53.943	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	97018	66.023	ug/l	86
77) Bromobenzene	11.195	156	55881	50.644	ug/l	99
78) n-propylbenzene	11.299	91	286276	51.803	ug/l	100
79) 2-Chlorotoluene	11.360	91	176598	49.544	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	206301	51.957	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	22427	49.689	ug/l	99
82) 4-Chlorotoluene	11.451	91	202465	51.220	ug/l	99
83) tert-Butylbenzene	11.713	119	207546	51.893	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	211228	52.532	ug/l	100
85) sec-Butylbenzene	11.890	105	258780	52.697	ug/l	100
86) p-Isopropyltoluene	12.006	119	213478	52.665	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	104712	51.998	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	103759	50.452	ug/l	100
89) n-Butylbenzene	12.329	91	188888	53.124	ug/l	99
90) Hexachloroethane	12.536	117	36963	51.760	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	108286	53.586	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22293	60.419	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	64675	55.722	ug/l	100
94) Hexachlorobutadiene	13.719	225	26115	51.518	ug/l	97
95) Naphthalene	13.774	128	254087	59.688	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	66653	55.655	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060325\
Data File : VX046486.D
Acq On : 03 Jun 2025 20:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Quant Time: Jun 04 01:35:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 04 01:35:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 80122
Ion Ratio Lower Upper
168 100
99 71.4 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 40.244 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 85 Resp: 49352
Ion Ratio Lower Upper
85 100
87 32.9 15.7 47.1









#3
Chloromethane
Concen: 40.046 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 47624
Ion Ratio Lower Upper
50 100
52 33.6 25.4 38.2



#4
Vinyl Chloride
Concen: 41.662 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 62 Resp: 46111
Ion Ratio Lower Upper
62 100
64 31.8 25.2 37.8









#5
Bromomethane
Concen: 35.580 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 18265
Ion Ratio Lower Upper
94 100
96 93.5 72.8 109.2



#6
Chloroethane
Concen: 46.072 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 64 Resp: 27222
Ion Ratio Lower Upper
64 100
66 31.9 24.3 36.5









#7
Trichlorofluoromethane
Concen: 44.111 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 72155
Ion Ratio Lower Upper
101 100
103 62.7 51.0 76.4



#8
Diethyl Ether
Concen: 47.664 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 74 Resp: 26541
Ion Ratio Lower Upper
74 100
45 107.2 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.787 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 46349
Ion Ratio Lower Upper
101 100
85 47.8 38.6 58.0
151 72.2 55.2 82.8



#10
Methyl Iodide
Concen: 50.077 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion:142 Resp: 59982
Ion Ratio Lower Upper
142 100
127 51.1 41.7 62.5
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 332.386 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 69693
Ion Ratio Lower Upper
59 100
57 11.2 8.6 12.8



#12
1,1-Dichloroethene
Concen: 45.010 ug/l
RT: 2.307 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 96 Resp: 42762
Ion Ratio Lower Upper
96 100
61 181.6 146.2 219.2
98 63.7 51.0 76.6









#13
Acrolein
Concen: 243.938 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 58249
Ion Ratio Lower Upper
56 100
55 74.0 56.2 84.4



#14
Allyl chloride
Concen: 50.041 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 41 Resp: 90860
Ion Ratio Lower Upper
41 100
39 70.7 60.6 90.8
76 29.7 24.9 37.3









#15
Acrylonitrile
Concen: 278.106 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 166739
Ion Ratio Lower Upper
53 100
52 84.2 65.3 97.9
51 38.1 29.8 44.8



#16
Acetone
Concen: 272.337 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 43 Resp: 163107
Ion Ratio Lower Upper
43 100
58 26.5 21.2 31.8









#17
Carbon Disulfide
Concen: 33.478 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 75405
Ion Ratio Lower Upper
76 100
78 9.0 7.0 10.4



#18
Methyl Acetate
Concen: 72.749 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 43 Resp: 101107
Ion Ratio Lower Upper
43 100
74 21.1 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 55.872 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 186100
Ion Ratio Lower Upper
73 100
57 21.7 17.7 26.5



#20
Methylene Chloride
Concen: 45.890 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 84 Resp: 52669
Ion Ratio Lower Upper
84 100
49 138.3 113.9 170.9
51 42.2 33.5 50.3
86 66.6 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 45.695 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 43658
Ion Ratio Lower Upper
96 100
61 150.7 119.5 179.3
98 67.4 50.0 75.0



#22
Diisopropyl ether
Concen: 54.666 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 45 Resp: 191730
Ion Ratio Lower Upper
45 100
43 73.9 66.6 100.0
87 24.9 19.8 29.6
59 10.3 8.6 12.8









#23
 Vinyl Acetate
 Concen: 266.071 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 820777
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 51.402 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25

Tgt Ion: 63 Resp: 100414
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.6 10.8
 100 4.5 2.1 6.3





#25

2-Butanone

Concen: 292.758 ug/l

RT: 4.556 min Scan# 5

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 254552

Ion Ratio Lower Upper

43 100

72 23.1 18.4 27.6



#26

2,2-Dichloropropane

Concen: 40.894 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 77 Resp: 62527

Ion Ratio Lower Upper

77 100

97 22.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 50.378 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 96 Resp: 57943

Ion Ratio Lower Upper

96 100

61 158.4 0.0 322.8

98 65.2 0.0 129.0



#28

Bromochloromethane

Concen: 50.493 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 49 Resp: 47479


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.8 56.2 84.2





#29

Tetrahydrofuran

Concen: 287.968 ug/l

RT: 5.007 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 156899

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5

71 38.7 31.4 47.2



#30

Chloroform

Concen: 52.410 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 83 Resp: 106714

Ion Ratio Lower Upper

83 100

85 63.7 49.3 73.9









#31
Cyclohexane
Concen: 45.113 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 80308
Ion Ratio Lower Upper
56 100
69 32.1 24.4 36.6
84 82.6 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 51.989 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 97 Resp: 91763
Ion Ratio Lower Upper
97 100
99 63.9 51.8 77.6
61 47.6 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.292 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 72135

Ion Ratio Lower Upper

65 100

67 49.8 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion: 114 Resp: 143610


Ion Ratio Lower Upper

114 100

63 23.8 0.0 49.2

88 17.1 0.0 33.6









#35
Dibromofluoromethane
Concen: 49.436 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 51124
Ion Ratio Lower Upper
113 100
111 103.0 83.1 124.7
192 16.9 13.3 19.9



#36
1,1-Dichloropropene
Concen: 45.167 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 75 Resp: 62759
Ion Ratio Lower Upper
75 100
110 34.5 16.3 48.9
77 32.5 24.3 36.5









#37
Ethyl Acetate
Concen: 53.356 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 91596
Ion Ratio Lower Upper
43 100
61 13.0 10.3 15.5
70 10.0 7.9 11.9



#38
Carbon Tetrachloride
Concen: 47.403 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 117 Resp: 74005
Ion Ratio Lower Upper
117 100
119 97.7 75.2 112.8
121 29.8 24.2 36.4









#39
Methylcyclohexane
Concen: 44.448 ug/l
RT: 7.379 min Scan# 1031
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 79510
Ion Ratio Lower Upper
83 100
55 79.5 64.7 97.1
98 46.5 37.4 56.2



#40
Benzene
Concen: 49.463 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 78 Resp: 201308
Ion Ratio Lower Upper
78 100
77 23.2 19.0 28.4









#41
Methacrylonitrile
Concen: 60.329 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 41	Resp: 54177
Ion Ratio	Lower Upper
41	100
39	57.8 47.2 70.8
67	65.3 50.7 76.1
52	30.3 24.6 37.0



#42
1,2-Dichloroethane
Concen: 50.240 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 62	Resp: 88249
Ion Ratio	Lower Upper
62	100
98	7.7 0.0 15.2









#43
Isopropyl Acetate
Concen: 59.242 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 155151
Ion Ratio Lower Upper
43 100
61 18.4 14.3 21.5
87 11.6 9.5 14.3



#44
Trichloroethene
Concen: 48.910 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion:130 Resp: 47910
Ion Ratio Lower Upper
130 100
95 110.4 0.0 204.2









#45
1,2-Dichloropropane
Concen: 54.270 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 63 Resp: 54922
Ion Ratio Lower Upper
63 100
65 28.7 24.1 36.1



#46
Dibromomethane
Concen: 50.854 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 93 Resp: 40591
Ion Ratio Lower Upper
93 100
95 81.4 65.6 98.4
174 84.7 68.2 102.2









#47
Bromodichloromethane
Concen: 52.237 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 82121
Ion Ratio Lower Upper
83 100
85 62.0 50.5 75.7
127 8.0 6.5 9.7



#48
Methyl methacrylate
Concen: 58.715 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 41 Resp: 78532
Ion Ratio Lower Upper
41 100
69 73.8 58.5 87.7
39 63.9 51.7 77.5









#49
1,4-Dioxane
Concen: 1182.487 ug/l
RT: 7.665 min Scan# 110
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 88 Resp: 30030
Ion Ratio Lower Upper
88 100
43 40.7 33.4 50.2
58 71.7 58.6 88.0



#50
Toluene-d8
Concen: 47.165 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 98 Resp: 168819
Ion Ratio Lower Upper
98 100
100 66.3 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 294.298 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 511610
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3



#52
Toluene
Concen: 50.497 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 92 Resp: 126020
Ion Ratio Lower Upper
92 100
91 171.2 136.6 205.0









#53
 t-1,3-Dichloropropene
 Concen: 53.778 ug/l
 RT: 8.976 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 75 Resp: 75146
 Ion Ratio Lower Upper
 75 100
 77 31.1 25.0 37.6



#54
 cis-1,3-Dichloropropene
 Concen: 53.169 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.006 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25

Tgt Ion: 75 Resp: 82115
 Ion Ratio Lower Upper
 75 100
 77 31.3 25.4 38.0
 39 59.5 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 54.643 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	97	Resp:	53770
Ion	Ratio	Lower	Upper
97	100		
83	84.6	69.0	103.6
85	53.5	44.3	66.5
99	60.0	47.8	71.6



#56
Ethyl methacrylate
Concen: 60.593 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion:	69	Resp:	95030
Ion	Ratio	Lower	Upper
69	100		
41	76.2	60.8	91.2
39	46.6	39.0	58.6









#57
1,3-Dichloropropane
Concen: 52.186 ug/l
RT: 9.305 min Scan# 11348
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 92225
Ion Ratio Lower Upper
76 100
78 32.8 26.3 39.5



#58
2-Chloroethyl Vinyl ether
Concen: 279.526 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 63 Resp: 223496
Ion Ratio Lower Upper
63 100
106 26.6 21.5 32.3









#59
2-Hexanone
Concen: 304.201 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 391242
Ion Ratio Lower Upper
43 100
58 49.9 24.9 74.6



#60
Dibromochloromethane
Concen: 54.434 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 129 Resp: 58826
Ion Ratio Lower Upper
129 100
127 78.1 39.3 117.8









#61
1,2-Dibromoethane
Concen: 53.748 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 54971
Ion Ratio Lower Upper
107 100
109 94.2 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 51.619 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 95 Resp: 70871
Ion Ratio Lower Upper
95 100
174 66.9 0.0 135.8
176 63.4 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:117 Resp: 127089

Ion Ratio Lower Upper

117 100

82 57.3 50.6 76.0

119 31.2 25.8 38.6



#64

Tetrachloroethene

Concen: 45.626 ug/l

RT: 9.269 min Scan# 1342

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:164 Resp: 41027




Ion Ratio Lower Upper


164 100

166 129.2 105.0 157.6

129 104.9 83.5 125.3

131 104.5 76.5 114.7









#65
Chlorobenzene
Concen: 49.792 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 138505
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 52.942 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion:131 Resp: 50287
Ion Ratio Lower Upper
131 100
133 92.9 47.3 141.9
119 66.0 31.6 95.0





#67

Ethyl Benzene

Concen: 51.077 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 250449

Ion Ratio Lower Upper

91 100

106 29.5 23.4 35.2



#68

m/p-Xylenes

Concen: 101.398 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 106 Resp: 181843

Ion Ratio Lower Upper

106 100

91 216.6 171.2 256.8





#69

o-Xylene

Concen: 53.478 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:106 Resp: 93498

Ion Ratio Lower Upper

106 100

91 223.7 112.7 338.1



#70

Styrene

Concen: 54.240 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion:104 Resp: 155343


Ion Ratio Lower Upper

104 100

78 58.2 45.7 68.5

103 54.4 43.7 65.5





#71

Bromoform

Concen: 54.353 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:173 Resp: 38761

Ion Ratio Lower Upper

173 100

175 47.7 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion:152 Resp: 61048


Ion Ratio Lower Upper

152 100

115 93.4 46.9 140.7

150 174.9 0.0 351.0









#73
Isopropylbenzene
Concen: 52.013 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 247204
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4



#74
N-ethyl acetate
Concen: 58.905 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 43 Resp: 138341
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.1 18.7 28.1
61 22.3 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 53.943 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 89844

Ion Ratio Lower Upper

83 100

131 9.7 5.0 14.9

85 63.8 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 66.023 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 75 Resp: 97018

Ion Ratio Lower Upper

75 100

77 32.1 20.5 61.5









#77
Bromobenzene
Concen: 50.644 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 55881
Ion Ratio Lower Upper
156 100
77 187.4 92.8 278.3
158 96.7 48.8 146.3



#78
n-propylbenzene
Concen: 51.803 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 91 Resp: 286276
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 49.544 ug/l
RT: 11.360 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 176598
Ion Ratio Lower Upper
91 100
126 31.7 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 51.957 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 105 Resp: 206301
Ion Ratio Lower Upper
105 100
120 46.7 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 49.689 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 22427
Ion Ratio Lower Upper
75 100
53 126.9 100.5 150.7
89 47.6 37.0 55.6



#82
4-Chlorotoluene
Concen: 51.220 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 91 Resp: 202465
Ion Ratio Lower Upper
91 100
126 26.9 13.3 39.8









#83
tert-Butylbenzene
Concen: 51.893 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:119 Resp: 207546
Ion Ratio Lower Upper
119 100
91 66.9 32.9 98.7
134 23.2 11.4 34.1



#84
1,2,4-Trimethylbenzene
Concen: 52.532 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion:105 Resp: 211228
Ion Ratio Lower Upper
105 100
120 42.7 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 52.697 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 258780
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 52.665 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046486.D
 Acq: 03 Jun 2025 20:25

Tgt Ion:119 Resp: 213478
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.5 37.5
 91 27.3 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 51.998 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 104712

Ion Ratio Lower Upper

146 100

111 43.7 22.1 66.3

148 63.0 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 50.452 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion:146 Resp: 103759


Ion Ratio Lower Upper

146 100

111 42.9 21.3 63.9

148 63.8 31.9 95.5





#89

n-Butylbenzene

Concen: 53.124 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 188888

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7

134 24.4 11.8 35.3



#90

Hexachloroethane

Concen: 51.760 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046486.D




Acq: 03 Jun 2025 20:25


Tgt Ion: 117 Resp: 36963

Ion Ratio Lower Upper

117 100

201 62.5 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 53.586 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 108286
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 62.6 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 60.419 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046486.D
Acq: 03 Jun 2025 20:25

Tgt Ion: 75 Resp: 22293
Ion Ratio Lower Upper
75 100
155 70.8 34.9 104.8
157 91.0 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 55.722 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 64675

Ion Ratio Lower Upper

180 100

182 95.5 47.9 143.8

145 34.3 17.1 51.2



#94

Hexachlorobutadiene

Concen: 51.518 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25




Tgt Ion:225 Resp: 26115


Ion Ratio Lower Upper

225 100

223 63.3 30.8 92.4

227 65.4 30.9 92.7





#95

Naphthalene

Concen: 59.688 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 254087

Ion Ratio Lower Upper

128 100

127 12.8 10.4 15.6

129 11.0 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 55.655 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046486.D

Acq: 03 Jun 2025 20:25

Tgt Ion:180 Resp: 66653

Ion Ratio Lower Upper

180 100

182 94.7 47.8 143.3

145 36.2 18.1 54.3

