

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG NO.: Q2186
 Lab File ID: VX046460.D Date Analyzed: 06/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:10
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	69603	12.018				
UPPER LIMIT	139206	12.518				
LOWER LIMIT	34801.5	11.518				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250530	59352	12.02				
BP-VPB-182-GW-670-672	57166	12.02				
BP-VPB-182-GW-680-682	57376	12.02				
BP-VPB-182-GW-705-707	56111	12.02				
BP-VPB-182-GW-720-722	50694	12.02				
BP-VPB-182-DUP-20250602	55355	12.02				
VPB182-HYD-20250530	61107	12.02				
VX0603WBL01	60850	12.02				
VX0603WBS01	65270	12.02				
VX0603WBSD01	62365	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.