

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2186 SAS No.: Q2186 SDG NO.: Q2186
Lab File ID: VX046459.D BFB Injection Date: 06/03/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:40
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	56.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1.1) 1
174	50.0 - 100.0% of mass 95	69.6
175	5.0 - 9.0% of mass 174	5.1 (7.3) 1
176	95.0 - 101.0% of mass 174	66.3 (95.2) 1
177	5.0 - 9.0% of mass 176	4.1 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046460.D	06/03/2025	10:10
VX0603WBL01	VX0603WBL01	VX046462.D	06/03/2025	11:01
VX0603WBS01	VX0603WBS01	VX046463.D	06/03/2025	11:24
VX0603WBSD01	VX0603WBSD01	VX046464.D	06/03/2025	11:50
VPB182-HYD-20250530	Q2186-09	VX046465.D	06/03/2025	12:13
BP-VPB-182-TB-20250530	Q2186-01	VX046466.D	06/03/2025	12:36
BP-VPB-182-GW-670-672	Q2186-02	VX046467.D	06/03/2025	12:59
BP-VPB-182-GW-680-682	Q2186-03	VX046468.D	06/03/2025	13:23
BP-VPB-182-GW-705-707	Q2186-06	VX046469.D	06/03/2025	13:46
BP-VPB-182-GW-720-722	Q2186-07	VX046470.D	06/03/2025	14:09
BP-VPB-182-DUP-20250602	Q2186-08	VX046471.D	06/03/2025	14:33
VSTDCCC050EC	VSTDCCC050	VX046486.D	06/03/2025	20:25