

## Initial and Continuing Calibration Blank Summary

**Client:** PARSONS Engineering of New York, Inc.

**SDG No.:** Q2189

**Project:** Con Ed Non MGP – Atlantic Ave 453957.600024.05

**RunNo.:** LB135991

| Analyte                | Units | Result   | Acceptance Limits | Conc Qual | MDL   | RDL | Analysis Date |
|------------------------|-------|----------|-------------------|-----------|-------|-----|---------------|
| <b>Sample ID: ICB1</b> |       |          |                   |           |       |     |               |
| Bromide                | mg/L  | < 1.0000 | 1.0000            | U         | 0.37  | 2   | 05/22/2025    |
| Chloride               | mg/L  | < 0.3000 | 0.3000            | U         | 0.19  | 0.6 | 05/22/2025    |
| Fluoride               | mg/L  | < 0.2000 | 0.2000            | U         | 0.11  | 0.4 | 05/22/2025    |
| Nitrite                | mg/L  | < 0.3000 | 0.3000            | U         | 0.074 | 0.6 | 05/22/2025    |
| Nitrate                | mg/L  | < 0.2500 | 0.2500            | U         | 0.095 | 0.5 | 05/22/2025    |
| Sulfate                | mg/L  | < 1.5000 | 1.5000            | U         | 0.46  | 3   | 05/22/2025    |
| Orthophosphate as P    | mg/L  | < 0.5000 | 0.5000            | U         | 0.34  | 1   | 05/22/2025    |
| <b>Sample ID: CCB1</b> |       |          |                   |           |       |     |               |
| Bromide                | mg/L  | < 1.0000 | 1.0000            | U         | 0.37  | 2   | 06/03/2025    |
| Chloride               | mg/L  | < 0.3000 | 0.3000            | U         | 0.19  | 0.6 | 06/03/2025    |
| Fluoride               | mg/L  | < 0.2000 | 0.2000            | U         | 0.11  | 0.4 | 06/03/2025    |
| Nitrite                | mg/L  | < 0.3000 | 0.3000            | U         | 0.074 | 0.6 | 06/03/2025    |
| Nitrate                | mg/L  | < 0.2500 | 0.2500            | U         | 0.095 | 0.5 | 06/03/2025    |
| Sulfate                | mg/L  | < 1.5000 | 1.5000            | U         | 0.46  | 3   | 06/03/2025    |
| Orthophosphate as P    | mg/L  | < 0.5000 | 0.5000            | U         | 0.34  | 1   | 06/03/2025    |
| <b>Sample ID: CCB2</b> |       |          |                   |           |       |     |               |
| Bromide                | mg/L  | < 1.0000 | 1.0000            | U         | 0.37  | 2   | 06/03/2025    |
| Chloride               | mg/L  | < 0.3000 | 0.3000            | U         | 0.19  | 0.6 | 06/03/2025    |
| Fluoride               | mg/L  | < 0.2000 | 0.2000            | U         | 0.11  | 0.4 | 06/03/2025    |
| Nitrite                | mg/L  | < 0.3000 | 0.3000            | U         | 0.074 | 0.6 | 06/03/2025    |
| Nitrate                | mg/L  | < 0.2500 | 0.2500            | U         | 0.095 | 0.5 | 06/03/2025    |
| Sulfate                | mg/L  | < 1.5000 | 1.5000            | U         | 0.46  | 3   | 06/03/2025    |
| Orthophosphate as P    | mg/L  | < 0.5000 | 0.5000            | U         | 0.34  | 1   | 06/03/2025    |
| <b>Sample ID: CCB3</b> |       |          |                   |           |       |     |               |
| Bromide                | mg/L  | < 1.0000 | 1.0000            | U         | 0.37  | 2   | 06/04/2025    |
| Chloride               | mg/L  | < 0.3000 | 0.3000            | U         | 0.19  | 0.6 | 06/04/2025    |
| Fluoride               | mg/L  | < 0.2000 | 0.2000            | U         | 0.11  | 0.4 | 06/04/2025    |
| Nitrite                | mg/L  | < 0.3000 | 0.3000            | U         | 0.074 | 0.6 | 06/04/2025    |
| Nitrate                | mg/L  | < 0.2500 | 0.2500            | U         | 0.095 | 0.5 | 06/04/2025    |
| Sulfate                | mg/L  | < 1.5000 | 1.5000            | U         | 0.46  | 3   | 06/04/2025    |
| Orthophosphate as P    | mg/L  | < 0.5000 | 0.5000            | U         | 0.34  | 1   | 06/04/2025    |
| <b>Sample ID: CCB4</b> |       |          |                   |           |       |     |               |
| Bromide                | mg/L  | < 1.0000 | 1.0000            | U         | 0.37  | 2   | 06/04/2025    |
| Chloride               | mg/L  | < 0.3000 | 0.3000            | U         | 0.19  | 0.6 | 06/04/2025    |
| Fluoride               | mg/L  | < 0.2000 | 0.2000            | U         | 0.11  | 0.4 | 06/04/2025    |
| Nitrite                | mg/L  | < 0.3000 | 0.3000            | U         | 0.074 | 0.6 | 06/04/2025    |
| Nitrate                | mg/L  | < 0.2500 | 0.2500            | U         | 0.095 | 0.5 | 06/04/2025    |
| Sulfate                | mg/L  | < 1.5000 | 1.5000            | U         | 0.46  | 3   | 06/04/2025    |
| Orthophosphate as P    | mg/L  | < 0.5000 | 0.5000            | U         | 0.34  | 1   | 06/04/2025    |