

Matrix Spike Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2189
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID:	Q2189-01
Client ID:	MW-7-20250602MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.5		0.43	J	10	1	101		06/03/2025
Chloride	mg/L	80-120	445	OR	462	OR	3	1	-567	*	06/03/2025
Fluoride	mg/L	80-120	2.10		0.11	U	2	1	105		06/03/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		06/03/2025
Nitrate	mg/L	80-120	11.6	OR	9.20	OR	2.5	1	96		06/03/2025
Sulfate	mg/L	80-120	78.3	OR	64.2	OR	15	1	94		06/03/2025
Orthophosphate as P	mg/L	80-120	5.20		0.34	U	5	1	104		06/03/2025

Matrix Spike Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2189
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID:	Q2189-01
Client ID:	MW-7-20250602MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.5		0.43	J	10	1	101		06/03/2025
Chloride	mg/L	80-120	444	OR	462	OR	3	1	-600	*	06/03/2025
Fluoride	mg/L	80-120	2.10		0.11	U	2	1	105		06/03/2025
Nitrite	mg/L	80-120	3.00		0.074	U	3	1	100		06/03/2025
Nitrate	mg/L	80-120	11.6	OR	9.20	OR	2.5	1	96		06/03/2025
Sulfate	mg/L	80-120	78.3	OR	64.2	OR	15	1	94		06/03/2025
Orthophosphate as P	mg/L	80-120	5.20		0.34	U	5	1	104		06/03/2025