

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2189
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB135991

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135991BSW							
Bromide	mg/L	10	10.5		105	1	90-110	06/03/2025
Chloride	mg/L	3	3.10		103	1	90-110	06/03/2025
Fluoride	mg/L	2	2.10		105	1	90-110	06/03/2025
Nitrite	mg/L	3	3.10		103	1	90-110	06/03/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	06/03/2025
Sulfate	mg/L	15	15.5		103	1	90-110	06/03/2025
Orthophosphate as P	mg/L	5	5.30		106	1	90-110	06/03/2025

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2189
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB135991

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135991BSW2							
Bromide	mg/L	10	10.7		107	1	90-110	06/04/2025
Chloride	mg/L	3	3.20		107	1	90-110	06/04/2025
Fluoride	mg/L	2	2.20		110	1	90-110	06/04/2025
Nitrite	mg/L	3	3.20		107	1	90-110	06/04/2025
Nitrate	mg/L	2.5	2.70		108	1	90-110	06/04/2025
Sulfate	mg/L	15	15.8		105	1	90-110	06/04/2025
Orthophosphate as P	mg/L	5	5.40		108	1	90-110	06/04/2025

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2189
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB136008

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136008BS							
TDS	mg/L	100	95.0		95	1	90-110	06/04/2025