

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2189

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VX046463.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0603WBS01	Dichlorodifluoromethane	20	16.6	ug/L	83			69	116	
	Chloromethane	20	17.0	ug/L	85			65	116	
	Vinyl chloride	20	16.6	ug/L	83			65	117	
	Bromomethane	20	16.9	ug/L	85			58	125	
	Chloroethane	20	19.1	ug/L	96			56	128	
	Trichlorofluoromethane	20	18.8	ug/L	94			73	115	
	1,1,2-Trichlorotrifluoroethane	20	18.7	ug/L	94			80	112	
	1,1-Dichloroethene	20	17.9	ug/L	90			74	110	
	Acetone	100	100	ug/L	100			60	125	
	Carbon disulfide	20	13.3	ug/L	67			64	112	
	Methyl tert-butyl Ether	20	20.3	ug/L	102			78	114	
	Methyl Acetate	20	27.1	ug/L	136		*	67	125	
	Methylene Chloride	20	18.4	ug/L	92			72	114	
	trans-1,2-Dichloroethene	20	18.0	ug/L	90			75	108	
	1,1-Dichloroethane	20	20.5	ug/L	103			78	112	
	Cyclohexane	20	17.7	ug/L	89			75	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	18.5	ug/L	93			77	113	
	cis-1,2-Dichloroethene	20	20.4	ug/L	102			77	110	
	Bromochloromethane	20	21.6	ug/L	108			70	124	
	Chloroform	20	20.9	ug/L	104			79	113	
	1,1,1-Trichloroethane	20	20.1	ug/L	101			80	108	
	Methylcyclohexane	20	17.7	ug/L	89			72	115	
	Benzene	20	19.5	ug/L	98			82	109	
	1,2-Dichloroethane	20	20.2	ug/L	101			80	115	
	Trichloroethene	20	18.9	ug/L	95			77	113	
	1,2-Dichloropropane	20	20.9	ug/L	104			83	111	
	Bromodichloromethane	20	20.0	ug/L	100			83	110	
	4-Methyl-2-Pentanone	100	110	ug/L	110			74	118	
	Toluene	20	19.9	ug/L	100			82	110	
	t-1,3-Dichloropropene	20	19.2	ug/L	96			79	110	
	cis-1,3-Dichloropropene	20	19.9	ug/L	100			82	110	
	1,1,2-Trichloroethane	20	21.5	ug/L	108			83	112	
	2-Hexanone	100	110	ug/L	110			73	117	
	Dibromochloromethane	20	20.5	ug/L	103			82	110	
	1,2-Dibromoethane	20	20.2	ug/L	101			81	110	
	Tetrachloroethene	20	19.8	ug/L	99			67	123	
	Chlorobenzene	20	20.2	ug/L	101			82	109	
	Ethyl Benzene	20	20.4	ug/L	102			83	109	
	m/p-Xylenes	40	40.4	ug/L	101			82	110	
	o-Xylene	20	21.0	ug/L	105			83	109	
	Styrene	20	21.2	ug/L	106			80	111	
	Bromoform	20	19.9	ug/L	100			79	109	
	Isopropylbenzene	20	21.1	ug/L	106			83	112	
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107			76	118	
	1,3-Dichlorobenzene	20	20.1	ug/L	101			82	108	
	1,4-Dichlorobenzene	20	19.6	ug/L	98			82	107	
	1,2-Dichlorobenzene	20	21.3	ug/L	106			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2189

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VX046463.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0603WBS01	1,2-Dibromo-3-Chloropropane	20	21.0	ug/L	105			68	112	
	1,2,4-Trichlorobenzene	20	20.5	ug/L	103			75	113	
	1,2,3-Trichlorobenzene	20	20.5	ug/L	103			76	114	