

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

RunNo.: LB135992

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.500	0.50	100	90-110	06/03/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.496	0.5	99	90-110	06/03/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.502	0.5	100	90-110	06/03/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

RunNo.: LB136007

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.494	0.50	99	90-110	06/04/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.497	0.50	99	90-110	06/04/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.496	0.50	99	90-110	06/04/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

RunNo.: LB136039

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	06/06/2025
Sample ID: CCV2 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV4 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV5 Ammonia as N	mg/L	1	1	100	90-110	06/06/2025