

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2191
Project: Pre Treatment Plant 2025	Sample ID: Q2191-01
Client ID: EFFLUENTDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.16		0.16		1	1.86		06/03/2025
Phosphorus, Total	mg/L	+/-20	0.24		0.24		1	2.52		06/04/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2191
Project: Pre Treatment Plant 2025	Sample ID: Q2191-01
Client ID: EFFLUENTMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Orthophosphate as P	mg/L	+/-20	0.64		0.64		1	0.47		06/03/2025
Phosphorus, Total	mg/L	+/-20	0.65		0.66		1	0.46		06/04/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2191
Project: Pre Treatment Plant 2025	Sample ID: Q2191-01
Client ID: EFFLUENTMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	90.8		90.9		1	0.11		06/09/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2191
Project: Pre Treatment Plant 2025	Sample ID: Q2197-05
Client ID: DSN003DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
BOD5	mg/L	+/-20	8.84		8.64		1	2.29		06/04/2025
TSS	mg/L	+/-5	7.90		8.00		1	1.26		06/05/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2191
Project: Pre Treatment Plant 2025	Sample ID: Q2203-01
Client ID: 001-WILLETS-PT-BLVD(JUNE)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	5.50	OR	5.40	OR	1	2		06/06/2025
Ammonia as N	mg/L	+/-20	5.30	D	5.30	D	5	0		06/06/2025

Duplicate Sample Summary

Client: Holland Manufacturing Co.	SDG No.: Q2191
Project: Pre Treatment Plant 2025	Sample ID: Q2203-01
Client ID: 001-WILLETS-PT-BLVD(JUNE)MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	6.70	OR	6.50	OR	1	3		06/06/2025

Duplicate Sample Summary

Client:	Holland Manufacturing Co.	SDG No.:	Q2191
Project:	Pre Treatment Plant 2025	Sample ID:	Q2243-02
Client ID:	WATER-TREATMENT-DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	23.9		24.3		1	1.66		06/09/2025