

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

Run No.: LB135992

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135992BS							
Orthophosphate as P	mg/L	0.5	0.51		102	1	90-110	06/03/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

Run No.: LB136002

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136002BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	06/04/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

Run No.: LB136013

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136013BS							
TSS	mg/L	550	532		97	1	90-110	06/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

Run No.: LB136056

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136056BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	06/09/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

Run No.: LB136007

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168270BS							
Phosphorus, Total	mg/L	0.50	0.49		98	1	90-110	06/04/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q2191

Project: Pre Treatment Plant 2025

Run No.: LB136039

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168326BS							
Ammonia as N	mg/L	1	1.10		110	1	90-110	06/06/2025