

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2197

Project: Linden Generating Station - LRSA

RunNo.: LB136064

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 COD	mg/L	47.915	50	96	95-105	06/09/2025
Sample ID: CCV2 COD	mg/L	48.931	50	98	95-105	06/09/2025
Sample ID: CCV3 COD	mg/L	49.946	50	100	95-105	06/09/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2197

Project: Linden Generating Station - LRSA

RunNo.: LB136084

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.02	7	100	90-110	06/03/2025
Sample ID:	CCV1						
Field pH		pH	7.01	7	100	90-110	06/03/2025
Sample ID:	CCV2						
Field pH		pH	7.03	7	100	90-110	06/03/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2197

Project: Linden Generating Station - LRSA

RunNo.: LB136084

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------