

## Matrix Spike Summary

<b>Client:</b>	Parkway Generation Operating LLC	<b>SDG No.:</b>	Q2197
<b>Project:</b>	Linden Generating Station - LRSA	<b>Sample ID:</b>	Q2197-01
<b>Client ID:</b>	DSN002MS	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	79.4		36.7		50.0	1	85		06/09/2025

## Matrix Spike Summary

<b>Client:</b>	Parkway Generation Operating LLC	<b>SDG No.:</b>	Q2197
<b>Project:</b>	Linden Generating Station - LRSA	<b>Sample ID:</b>	Q2197-01
<b>Client ID:</b>	DSN002MSD	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	80.4		36.7		50.0	1	87		06/09/2025