

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2197

Project: Linden Generating Station - LRSA

Run No.: LB136002

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136002BS							
BOD5	mg/L	198	178		90	1	84.6-115.4	06/04/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2197

Project: Linden Generating Station - LRSA

Run No.: LB136013

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136013BS							
TSS	mg/L	550	532		97	1	90-110	06/05/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2197

Project: Linden Generating Station - LRSA

Run No.: LB136064

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136064BS							
COD	mg/L	50	52.0		104	1	90-110	06/09/2025