

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136006

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/04/2025	11:30
2	CAL2	1	Water	NA	NA	20.2	7.00	06/04/2025	11:31
3	CAL3	1	Water	NA	NA	20.2	10.02	06/04/2025	11:33
4	ICV	1	Water	NA	NA	20.2	7.02	06/04/2025	11:35
5	CCV1	1	Water	NA	NA	20.2	2.01	06/04/2025	11:37
6	Q2198-01	1	Solid	20.05	20	22.1	6.61	06/04/2025	11:55
7	Q2198-01DUP	1	Solid	20.03	20	22.2	6.62	06/04/2025	11:56
8	Q2198-03	1	Solid	20.01	20	22.1	7.87	06/04/2025	12:00
9	CCV2	1	Water	NA	NA	20.2	12.02	06/04/2025	12:19

WORKLIST(Hardcopy Internal Chain)

136006

WorkList Name : PH S Q2198 WorkList ID : 189920 Department : Wet-Chemistry Date : 06-04-2025 09:52:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2198-01	B-202-SB02	Solid	pH	Cool 4 deg C	PORT06	N22	05/31/2025	9045D
Q2198-03	B-207-SB01	Solid	pH	Cool 4 deg C	PORT06	N22	06/01/2025	9045D

Date/Time 06/04/25 11:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/04/25 17:13

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]