

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136042

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/06/2025	12:15
2	CAL2	1	Water	NA	NA	20.2	7.00	06/06/2025	12:16
3	CAL3	1	Water	NA	NA	20.3	10.02	06/06/2025	12:17
4	ICV	1	Water	NA	NA	20.2	7.01	06/06/2025	12:20
5	CCV1	1	Water	NA	NA	20.2	2.01	06/06/2025	12:25
6	Q2198-05	1	Water	NA	NA	20.6	6.42	06/06/2025	12:35
7	Q2198-05DUP	1	Water	NA	NA	20.7	6.43	06/06/2025	12:37
8	Q2221-02	1	Water	NA	NA	21.1	5.50	06/06/2025	12:44
9	Q2221-03	1	Water	NA	NA	20.9	7.90	06/06/2025	12:55
10	Q2245-01	1	Water	NA	NA	21.2	5.22	06/06/2025	13:10
11	Q2245-02	1	Water	NA	NA	21.3	4.55	06/06/2025	13:15
12	CCV2	1	Water	NA	NA	20.2	12.02	06/06/2025	13:20

WORKLIST(Hardcopy Internal Chain)

2136042

WorkList Name : ph q2245

WorkList ID : 190000

Department : Wet-Chemistry

Date : 06-06-2025 11:56:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2198-05	B-202-GW01	Water	pH	Cool 4 deg C	PORT06	N22	05/31/2025	9040C
Q2221-02	34900	Water	pH	Cool 4 deg C	PSEG03	N41	06/04/2025	9040C
Q2221-03	34867	Water	pH	Cool 4 deg C	PSEG03	N41	06/04/2025	9040C
Q2245-01	AU-25-0090	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2025	9040C
Q2245-02	AU-25-0091	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2025	9040C

Date/Time 06/06/25 12:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/06/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]