



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture
Project: Amtrak Sawtooth Bridges 2025

SDG No.: Q2198
RunNo.: LB136000

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.508	0.5	102	90-110	06/04/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	06/04/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	06/04/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB136005

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	06/04/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	06/04/2025
Sample ID: Corrosivity	CCV2	pH	12.01	12.00	100	90-110	06/04/2025
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	06/04/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB136006

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.02	7	100	90-110	06/04/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	06/04/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	06/04/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB136021

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.1	0.099	101	85-115	06/05/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB136024

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	06/05/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	06/05/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	06/05/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	06/05/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	06/05/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB136042

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	06/06/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	06/06/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	06/06/2025