

Laboratory Control Sample Summary

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

Run No.: LB136000

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136000BS							
Hexavalent Chromium	mg/L	0.5	0.51		102	1	90-111	06/04/2025

Laboratory Control Sample Summary

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Project: Amtrak Sawtooth Bridges 2025

Run No.: LB136024

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168265BS							
Hexavalent Chromium	mg/Kg	20	19.8		99	1	84-110	06/05/2025