

LB136001


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

54631



A= 0.0000e+000

B= 1.8159e-004

C= -4.0693e-002

Rho= 0.9997523

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.011	-0.011	164	0.000	164					-
0.2	0.200	0.193	-0.007	1289	0.0 %	1289					-4
2.5	2.500	2.435	-0.065	13631	0.0 %	13631					-3
5.0	5.000	5.079	0.079	28193	0.0 %	28193					2
7.5	7.500	7.624	0.124	42208	0.0 %	42208					2
10.0	10.000	9.880	-0.120	54631	0.0 %	54631					-1

LB136001 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Jun 2025 14:32:34

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	164	-	0.0000	7471B	PPB	04 Jun 2025 14:45:44	S	Std
0.2 - 1	S0-2	1289	-	0.2000	7471B	PPB	04 Jun 2025 14:48:01	S	Std
2.5 - 1	S2-5	13631	-	2.5000	7471B	PPB	04 Jun 2025 14:50:20	S	Std
5.0 - 1	S5	28193	-	5.0000	7471B	PPB	04 Jun 2025 14:52:37	S	Std
7.5 - 1	S7-5	42208	-	7.5000	7471B	PPB	04 Jun 2025 14:54:56	S	Std
10.0 - 1	S10	54631	-	10.0000	7471B	PPB	04 Jun 2025 14:57:15	S	Std
ICV27 - 1	ICV27	23321	4.1942	-	7471B	PPB	04 Jun 2025 15:04:33	U	SMPL
ICB27 - 1	ICB27	-269	-0.0895	-	7471B	PPB	04 Jun 2025 15:06:49	U	SMPL
CCV01 - 1	CCV01	27728	4.9945	-	7471B	PPB	04 Jun 2025 15:09:08	U	SMPL
CCB01 - 1	CCB01	668	0.0806	-	7471B	PPB	04 Jun 2025 15:14:02	U	SMPL
CRA - 1	CRA	1300	0.1954	-	7471B	PPB	04 Jun 2025 15:21:19	U	SMPL
HighStd - 1	HighStd	51830	9.3713	-	7471B	PPB	04 Jun 2025 15:23:34	U	SMPL
ChkStd - 1	ChkStd	37287	6.7304	-	7471B	PPB	04 Jun 2025 15:28:34	U	SMPL
PB168290BL - 1	PBS	176	-0.0087	-	7471B	PPB	04 Jun 2025 15:33:40	U	SMPL
PB168290BS - 1	LCSS	19835	3.5612	-	7471B	PPB	04 Jun 2025 15:38:36	U	SMPL
Q2182-01 - 1	OR-03-06022025	1051	0.1502	-	7471B	PPB	04 Jun 2025 15:40:54	U	SMPL
Q2194-01 - 1	COMP-12	976	0.1365	-	7471B	PPB	04 Jun 2025 15:43:14	U	SMPL
Q2194-03 - 1	COMP-13	790	0.1028	-	7471B	PPB	04 Jun 2025 15:45:32	U	SMPL
Q2195-01 - 1	OK-01-060325	934	0.1289	-	7471B	PPB	04 Jun 2025 15:47:50	U	SMPL
Q2195-01DUP - 1	OK-01-060325DUP	563	0.0615	-	7471B	PPB	04 Jun 2025 15:50:06	U	SMPL
CCV02 - 1	CCV02	28539	5.1418	-	7471B	PPB	04 Jun 2025 15:59:30	U	SMPL
CCB02 - 1	CCB02	98	-0.0229	-	7471B	PPB	04 Jun 2025 16:01:45	U	SMPL
Q2195-01MS - 1	OK-01-060325MS	18282	3.2792	-	7471B	PPB	04 Jun 2025 16:04:06	U	SMPL
Q2195-01MSD - 1	OK-01-060325MSD	19012	3.4118	-	7471B	PPB	04 Jun 2025 16:06:22	U	SMPL
Q2198-01 - 1	B-202-SB02	4028	0.6908	-	7471B	PPB	04 Jun 2025 16:08:40	U	SMPL
Q2198-03 - 1	B-207-SB01	8968	1.5878	-	7471B	PPB	04 Jun 2025 16:10:59	U	SMPL
Q2199-01 - 1	ETGI-343	10321	1.8335	-	7471B	PPB	04 Jun 2025 16:13:17	U	SMPL
Q2199-03 - 1	VNJ-231	2128	0.3457	-	7471B	PPB	04 Jun 2025 16:15:36	U	SMPL
Q2199-05 - 1	72-11978	4120	0.7075	-	7471B	PPB	04 Jun 2025 16:17:55	U	SMPL
Q2195-01LX5 - 1		46	-0.0323	-	7471B	PPB	04 Jun 2025 16:20:13	U	SMPL
Q2195-01A - 1		24774	4.4581	-	7471B	PPB	04 Jun 2025 16:25:07	U	SMPL
CCV03 - 1	CCV03	28956	5.2175	-	7471B	PPB	04 Jun 2025 16:27:23	U	SMPL
CCB03 - 1	CCB03	-260	-0.0879	-	7471B	PPB	04 Jun 2025 16:29:41	U	SMPL