

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168290BL		SOLID		Batch Number:	PB168290		Prep Date:	06/04/2025	
	Mercury	0.0070	<0.013	U	0.013	CV	06/04/2025	15:33	LB136001
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168317BL		WATER		Batch Number:	PB168317		Prep Date:	06/05/2025	
	Mercury	0.076	<0.2	U	0.20	CV	06/06/2025	11:16	LB136036

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168276BL	WATER			Batch Number:	PB168276		Prep Date:	06/04/2025	
	Aluminum	5.67	<25	U	50.0	P	06/04/2025	20:46	LB136011
	Antimony	3.38	<12.5	U	25.0	P	06/04/2025	20:46	LB136011
	Arsenic	2.56	<5	U	10.0	P	06/04/2025	20:46	LB136011
	Barium	7.28	<25	U	50.0	P	06/04/2025	20:46	LB136011
	Beryllium	0.28	<1.5	U	3.00	P	06/04/2025	20:46	LB136011
	Cadmium	0.25	<1.5	U	3.00	P	06/04/2025	20:46	LB136011
	Calcium	117	<500	U	1000	P	06/04/2025	20:46	LB136011
	Chromium	1.06	<2.5	U	5.00	P	06/04/2025	20:46	LB136011
	Cobalt	1.13	<7.5	U	15.0	P	06/04/2025	20:46	LB136011
	Copper	2.30	<5	U	10.0	P	06/04/2025	20:46	LB136011
	Iron	11.7	<25	U	50.0	P	06/04/2025	20:46	LB136011
	Lead	1.15	<3	U	6.00	P	06/04/2025	20:46	LB136011
	Magnesium	122	<500	U	1000	P	06/04/2025	20:46	LB136011
	Manganese	2.97	<5	U	10.0	P	06/04/2025	20:46	LB136011
	Nickel	1.53	<10	U	20.0	P	06/04/2025	20:46	LB136011
	Potassium	459	<500	U	1000	P	06/04/2025	20:46	LB136011
	Selenium	4.82	<5	U	10.0	P	06/04/2025	20:46	LB136011
	Silver	0.81	<2.5	U	5.00	P	06/04/2025	20:46	LB136011
	Sodium	434	<500	U	1000	P	06/04/2025	20:46	LB136011
	Thallium	2.19	<10	U	20.0	P	06/04/2025	20:46	LB136011
	Vanadium	3.13	<10	U	20.0	P	06/04/2025	20:46	LB136011
	Zinc	8.33	<10	U	20.0	P	06/04/2025	20:46	LB136011
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168279BL	SOLID			Batch Number:	PB168279		Prep Date:	06/04/2025	
	Aluminum	0.84	<2.5	U	5.00	P	06/06/2025	14:13	LB136052
	Antimony	0.22	<1.25	U	2.50	P	06/06/2025	14:13	LB136052
	Arsenic	0.19	<0.5	U	1.00	P	06/06/2025	14:13	LB136052
	Barium	0.73	<2.5	U	5.00	P	06/06/2025	14:13	LB136052
	Beryllium	0.025	<0.15	U	0.30	P	06/06/2025	14:13	LB136052
	Cadmium	0.024	<0.15	U	0.30	P	06/06/2025	14:13	LB136052
	Calcium	11.1	<50	U	100	P	06/06/2025	14:13	LB136052
	Chromium	0.047	<0.25	U	0.50	P	06/06/2025	14:13	LB136052
	Cobalt	0.10	<0.75	U	1.50	P	06/06/2025	14:13	LB136052
	Copper	0.22	<0.5	U	1.00	P	06/06/2025	14:13	LB136052
	Iron	3.99	<2.5	U	5.00	P	06/06/2025	14:13	LB136052
	Lead	0.13	<0.3	U	0.60	P	06/06/2025	14:13	LB136052
	Magnesium	12.0	<50	U	100	P	06/06/2025	14:13	LB136052
	Manganese	0.14	<0.5	U	1.00	P	06/06/2025	14:13	LB136052
	Nickel	0.13	<1	U	2.00	P	06/06/2025	14:13	LB136052

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Instrument: P4

Potassium	27.7	<50	U	100	P	06/06/2025	14:13	LB136052
Selenium	0.26	<0.5	U	1.00	P	06/06/2025	14:13	LB136052
Silver	0.12	<0.25	U	0.50	P	06/06/2025	14:13	LB136052
Sodium	17.8	<50	U	100	P	06/06/2025	14:13	LB136052
Thallium	0.23	<1	U	2.00	P	06/06/2025	14:13	LB136052
Vanadium	0.25	<1	U	2.00	P	06/06/2025	14:13	LB136052
Zinc	0.23	<1	U	2.00	P	06/06/2025	14:13	LB136052