



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q2198
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2198 **SAS No.:** Q2198
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	107	70 - 130	CV	06/06/2025	10:12	LB136036
CRI01	Arsenic	21.3	20.0	107	65 - 135	P	06/06/2025	13:37	LB136052
	Barium	91.8	100	92	65 - 135	P	06/06/2025	13:37	LB136052
	Cadmium	5.92	6.0	99	65 - 135	P	06/06/2025	13:37	LB136052
	Chromium	9.75	10.0	98	65 - 135	P	06/06/2025	13:37	LB136052
	Lead	13.8	12.0	115	65 - 135	P	06/06/2025	13:37	LB136052
	Selenium	19.7	20.0	98	65 - 135	P	06/06/2025	13:37	LB136052
	Silver	10.1	10.0	101	65 - 135	P	06/06/2025	13:37	LB136052
CRI01	Arsenic	19.0	20.0	95	65 - 135	P	06/09/2025	12:44	LB136071
	Barium	91.7	100	92	65 - 135	P	06/09/2025	12:44	LB136071
	Cadmium	5.74	6.0	96	65 - 135	P	06/09/2025	12:44	LB136071
	Chromium	9.13	10.0	91	65 - 135	P	06/09/2025	12:44	LB136071
	Lead	12.3	12.0	103	65 - 135	P	06/09/2025	12:44	LB136071
	Selenium	18.5	20.0	93	65 - 135	P	06/09/2025	12:44	LB136071
	Silver	9.51	10.0	95	65 - 135	P	06/09/2025	12:44	LB136071