

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168271TB		WATER		Batch Number:	PB168316		Prep Date:	06/05/2025	
	Mercury	0.76	<2	U	2.00	CV	06/06/2025	12:11	LB136036
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168316BL		WATER		Batch Number:	PB168316		Prep Date:	06/05/2025	
	Mercury	0.076	<0.2	U	0.20	CV	06/06/2025	10:19	LB136036

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168271TB		WATER		Batch Number:	PB168307		Prep Date:	06/05/2025	
	Arsenic	25.6	<50	U	100	P	06/06/2025	15:10	LB136052
	Barium	72.8	<250	U	500	P	06/06/2025	15:10	LB136052
	Cadmium	2.50	<15	U	30.0	P	06/06/2025	15:10	LB136052
	Chromium	10.6	<25	U	50.0	P	06/06/2025	15:10	LB136052
	Lead	11.5	<30	U	60.0	P	06/06/2025	15:10	LB136052
	Selenium	48.2	<50	U	100	P	06/06/2025	15:10	LB136052
	Silver	8.10	<25	U	50.0	P	06/06/2025	15:10	LB136052
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168307BL		WATER		Batch Number:	PB168307		Prep Date:	06/05/2025	
	Arsenic	25.6	<50	U	100	P	06/06/2025	14:39	LB136052
	Barium	72.8	<250	U	500	P	06/06/2025	14:39	LB136052
	Cadmium	2.50	<15	U	30.0	P	06/06/2025	14:39	LB136052
	Chromium	10.6	<25	U	50.0	P	06/06/2025	14:39	LB136052
	Lead	11.5	<30	U	60.0	P	06/06/2025	14:39	LB136052
	Selenium	48.2	<50	U	100	P	06/06/2025	14:39	LB136052
	Silver	8.10	<25	U	50.0	P	06/06/2025	14:39	LB136052