

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Portal Partners Tri-Venture</u>	SDG No.: <u>Q2198</u>
Contract: <u>PORT06</u>	Lab Code: <u>CHEM</u>
ICS Source: <u>EPA</u>	Case No.: <u>Q2198</u>
	SAS No.: <u>Q2198</u>
	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.57			-20	20	06/06/2025	13:42	LB136052
	Barium	1.92	6.0	32	-94	106	06/06/2025	13:42	LB136052
	Cadmium	2.20	1.0	220	-5	7	06/06/2025	13:42	LB136052
	Chromium	45.0	52.0	86	42	62	06/06/2025	13:42	LB136052
	Lead	2.11			-12	12	06/06/2025	13:42	LB136052
	Selenium	0.16			-20	20	06/06/2025	13:42	LB136052
	Silver	0.046			-10	10	06/06/2025	13:42	LB136052
ICSAB01	Arsenic	106	104	102	88.4	120	06/06/2025	13:46	LB136052
	Barium	477	537	89	437	637	06/06/2025	13:46	LB136052
	Cadmium	970	972	100	826	1120	06/06/2025	13:46	LB136052
	Chromium	545	542	101	460	624	06/06/2025	13:46	LB136052
	Lead	47.1	49.0	96	37	61	06/06/2025	13:46	LB136052
	Selenium	42.7	46.0	93	26	66	06/06/2025	13:46	LB136052
	Silver	213	201	106	170	232	06/06/2025	13:46	LB136052
ICSA01	Arsenic	1.84			-20	20	06/09/2025	12:48	LB136071
	Barium	3.38	6.0	56	-94	106	06/09/2025	12:48	LB136071
	Cadmium	0.56	1.0	56	-5	7	06/09/2025	12:48	LB136071
	Chromium	44.9	52.0	86	42	62	06/09/2025	12:48	LB136071
	Lead	-4.92			-12	12	06/09/2025	12:48	LB136071
	Selenium	-6.19			-20	20	06/09/2025	12:48	LB136071
	Silver	-0.39			-10	10	06/09/2025	12:48	LB136071
ICSAB01	Arsenic	103	104	99	88.4	120	06/09/2025	12:54	LB136071
	Barium	491	537	91	437	637	06/09/2025	12:54	LB136071
	Cadmium	962	972	99	826	1120	06/09/2025	12:54	LB136071
	Chromium	541	542	100	460	624	06/09/2025	12:54	LB136071
	Lead	40.2	49.0	82	37	61	06/09/2025	12:54	LB136071
	Selenium	31.4	46.0	68	26	66	06/09/2025	12:54	LB136071
	Silver	213	201	106	170	232	06/09/2025	12:54	LB136071