

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** Q2198
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2198 **SAS No.:** Q2198
Matrix: Water **Sample ID:** Q2208-36 **Client ID:** BU-703-COMP-09DUP
Percent Solids for Sample: NA **Duplicate ID** Q2208-36DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1050		872		19		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	14.6	J	14.4	J	1		P
Lead	ug/L	20	14.2	J	60.0	U	200.0		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **Level:** LOW **SDG No.:** Q2198
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2198 **SAS No.:** Q2198
Matrix: Water **Sample ID:** Q2208-36MS **Client ID:** BU-703-COMP-09MSD
Percent Solids for Sample: NA **Duplicate ID** Q2208-36MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3850		3830		1		P
Barium	ug/L	20	2090		1980		5		P
Cadmium	ug/L	20	981		951		3		P
Chromium	ug/L	20	1950		1940		1		P
Lead	ug/L	20	4660		4560		2		P
Mercury	ug/L	20	41.2		41.0		0		CV
Selenium	ug/L	20	9560		9570		0		P
Silver	ug/L	20	331		347		5		P