

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** Q2198
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q2198 **SAS No.:** Q2198

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168307BS							
Arsenic	ug/L	4000	4100		102	80 - 120	P
Barium	ug/L	1000	970		97	80 - 120	P
Cadmium	ug/L	1000	1020		102	80 - 120	P
Chromium	ug/L	2000	2090		104	80 - 120	P
Lead	ug/L	5000	4990		100	80 - 120	P
Selenium	ug/L	10000	10400		104	80 - 120	P
Silver	ug/L	380	376		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2198

Contract: PORT06

Lab Code: CHEM

Case No.: Q2198

SAS No.: Q2198

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168316BS Mercury	ug/L	4.0	3.75		94	80 - 120	CV