

Continuing Calibration Report for SequenceID : FC060925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069133.D

Aliphatic C9-C12	6438999.000	60.000	3.299	6.602	107316.650	105717.491	-1.513
Aliphatic C12-C16	4050508.000	40.000	6.603	10.008	101262.700	101985.435	0.709
Aliphatic C16-C21	5664814.000	60.000	10.009	13.379	94413.567	97865.081	3.527
Aliphatic C21-C28	7292919.000	80.000	13.380	17.047	91161.488	93803.915	2.817
Aliphatic C28-C40	11227163.000	120.000	17.048	22.037	93559.692	94439.035	0.931
Aliphatic EPH	34674403.000	360.000	3.299	22.037	96317.786	97587.136	1.301