

Continuing Calibration Report for SequenceID : FC060925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069140.D

Aliphatic C9-C12	6898571.000	60.000	3.299	6.602	114976.183	105717.491	-8.758
Aliphatic C12-C16	4221397.000	40.000	6.603	10.008	105534.925	101985.435	-3.480
Aliphatic C16-C21	5932054.000	60.000	10.009	13.379	98867.567	97865.081	-1.024
Aliphatic C21-C28	8098980.000	80.000	13.380	17.047	101237.250	93803.915	-7.924
Aliphatic C28-C40	12168142.000	120.000	17.048	22.037	101401.183	94439.035	-7.372
Aliphatic EPH	37319144.000	360.000	3.299	22.037	103664.289	97587.136	-6.227