

Continuing Calibration Report for SequenceID : FC061125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069156.D

Aliphatic C9-C12	6783670.000	60.000	3.303	6.605	113061.167	105717.491	-6.947
Aliphatic C12-C16	4187470.000	40.000	6.606	10.010	104686.750	101985.435	-2.649
Aliphatic C16-C21	5888827.000	60.000	10.011	13.380	98147.117	97865.081	-0.288
Aliphatic C21-C28	8136653.000	80.000	13.381	17.048	101708.163	93803.915	-8.426
Aliphatic C28-C40	12161407.000	120.000	17.049	22.033	101345.058	94439.035	-7.313
Aliphatic EPH	37158027.000	360.000	3.303	22.033	103216.742	97587.136	-5.769