

Continuing Calibration Report for SequenceID : FC061125AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC069168.D

|                   |              |         |        |        |            |            |       |
|-------------------|--------------|---------|--------|--------|------------|------------|-------|
| Aliphatic C9-C12  | 6110551.000  | 60.000  | 3.303  | 6.605  | 101842.517 | 105717.491 | 3.665 |
| Aliphatic C12-C16 | 3798337.000  | 40.000  | 6.606  | 10.010 | 94958.425  | 101985.435 | 6.890 |
| Aliphatic C16-C21 | 5309111.000  | 60.000  | 10.011 | 13.380 | 88485.183  | 97865.081  | 9.585 |
| Aliphatic C21-C28 | 7214648.000  | 80.000  | 13.381 | 17.048 | 90183.100  | 93803.915  | 3.860 |
| Aliphatic C28-C40 | 10985455.000 | 120.000 | 17.049 | 22.033 | 91545.458  | 94439.035  | 3.064 |
| Aliphatic EPH     | 33418102.000 | 360.000 | 3.303  | 22.033 | 92828.061  | 97587.136  | 4.877 |